

RICH NOT WORKING CHANNELS MAY 30, 2001

Column	Channel	Status	New/Old
3	28	Hot	NEW
4	31	No signal	OLD
6	15	No signal	OLD
9	9	Weak signal	OLD
9	30	Weak signal	NEW
12	19	Hot	NEW
13	26	No signal. Pot DC	OLD
14	9	No signal	NEW
15	7	Bad positive weak sign	NEW
16	22	Weak signal	OLD
22	6	No signal	NEW
22	22	Potential DC	OLD
23	15	No signal	OLD
31	16	Hot	NEW
32	2	Weak signal	NEW
32	20	No signal	OLD
36	8	Bad double pulse	OLD
37	13	Hot	OLD
37	32	No signal	OLD
38	18	HV on signal !!!	NEW
43	25	Bad signal	NEW
44	30	No signal	OLD
46	14	Bad double pulse	NEW
48	17	No signal	OLD
49	29	No signal	OLD
50	19	No signal	NEW
53	26	Hot	NEW
56	29	Weak signal	NEW
56	30	Bad signal	OLD
58	24	HV problem/Tube disc	NEW
62	16	HV problem/Tube disc	NEW
62	21	Bad signal	NEW
64	12	No signal	OLD
64	24	No signal	OLD
66	29	Bad double pulse	OLD
67	8	No signal/HV probl?	NEW
67	19	No signal/HV probl?	NEW
69	9	Weak signal	NEW
69	19	No signal	OLD
70	29	Bad double pulse	NEW
70	30	No signal	OLD
70	31	No signal	OLD
70	32	No signal	OLD
73	22	Hot	NEW
76	20	Weak signal	NEW
78	3	No signal. DC pot	NEW
78	17	No signal. DC pot	NEW
80	5	HV problem/Tube disc	NEW
83	12	No signal. DC pot	NEW
84	2	No signal/HV problem?	NEW
84	8	No signal/HV problem?	NEW
84	24	No signal/HV problem?	NEW
85	16	Hot	NEW
88	3	Weak signal	NEW
88	10	Weak signal	NEW
88	18	No signal	NEW
88	22	No signal	NEW
89	1	Weak signal	NEW

TOTAL NO SIGNAL: 16 OLD + 9 NEW = 25  
 TOTAL BAD SIGNAL: 5 OLD + 12 NEW = 17  
 TOTAL HOT: 1 OLD + 6 NEW = 7  
 TOTAL HV PROBLEMS: 9 NEW = 9  
**TOTAL: 22 OLD + 36 NEW = 58**

RICH NOT WORKING CHANNELS MAY 22, 2001

Column	Channel	Status	New/Old	
3	28	Hot	NEW	
4	31	No signal	OLD	
6	15	No signal	OLD	
9	9	Weak signal		OLD
9	30	Weak signal	NEW	
12	19	Hot	NEW	
13	26	No signal. Pot DC	OLD	
14	9	No signal	NEW	
15	7	Bad positive weak sign	NEW	
16	22	Weak signal		OLD
22	6	No signal	NEW	
22	22	Potential DC		OLD
23	15	No signal	OLD	
31	16	Hot	NEW	
32	2	Weak signal	NEW	
32	20	No signal	OLD	
36	8	Bad double pulse	OLD	
37	13	Hot		OLD
37	32	No signal	OLD	
38	18	HV on signal !!!	NEW	
43	25	Bad signal	NEW	
44	30	No signal	OLD	
46	14	Bad double pulse	NEW	
48	17	No signal	OLD	
49	29	No signal	OLD	
50	19	No signal	NEW	
53	26	Hot	NEW	
56	29	Weak signal	NEW	
56	30	Bad signal	OLD	
58	1-32	HV problem	NEW	
62	1-32	HV problem	NEW	
64	12	No signal	OLD	
64	24	No signal	OLD	
66	29	Bad double pulse	OLD	
67	1-32	HV problem	NEW	
69	9	Weak signal	NEW	
69	19	No signal	OLD	
70	29	Bad double pulse	NEW	
70	30	No signal	OLD	
70	31	No signal	OLD	
70	32	No signal	OLD	
73	22	Hot	NEW	
76	20	Weak signal	NEW	
78	3	No signal. DC pot	NEW	
78	17	No signal. DC pot	NEW	
80	1-32	HV problem	NEW	
83	12	No signal. DC pot	NEW	
84	1-32	HV problem	NEW	
85	16	Hot	NEW	
88	1-32	HV problem	NEW	
89	1	Weak signal	NEW	

TOTAL NO SIGNAL: 16 OLD + 7 NEW = 23  
 TOTAL BAD SIGNAL: 5 OLD + 9 NEW = 14  
 TOTAL HOT: 1 OLD + 6 NEW = 7  
 TOTAL HV PROBLEMS: 6 COL + 1 CH = 7

TOTAL: 22 OLD + 29 NEW = 51