

<b>U11A SAFETY CHECKLIST</b>			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
<b>Rev: B</b>	<b>Effective: 5/24/2007</b>	<b>Page 1 of 1</b>	Number: LS-SCL- 0080
Reviewed by: Thomas McDonald		Reviewed by:	<b>Approved by: J. Aloï</b>

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

**A. FRONT-END VACUUM**

With mask closed, or no beam in the machine, pressure must be  $2.0 \times 10^{-9}$  or less.

**B. BEAMLINE ISOLATION VALVES**

All beamline isolation valves must be unsealed and in the open position.

**C. NUMBERED CHECKPOINTS**

The following checkpoints start at the Front End. The purpose of the list is to ensure that all shielding and exclusion zones are in place and secured. Refer to U4B Beamline Diagram.

- 1. Bremsstrahlung Shielding/Ring shielding in place as per photo.
- 2. Bremsstrahlung Shied (BS #2) in place and secure as per photo.
- 3. Bremsstrahlung Shied (BS #3) in place and secure as per photo.

OPCO / Beamline Rep.: \_\_\_\_\_ Date: \_\_\_\_\_