

<b>U3C SAFETY CHECKLIST</b>			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
<b>Rev: B</b>	<b>Effective: 5/24/2007</b>	<b>Page 1 of 1</b>	Number: LS-SCL- 0068
Reviewed by: Thomas McDonald		Reviewed by:	<b>Approved by: J. Aloï</b>

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

**A. Front End Vacuum**

With mask and shutter closed or no beam in the machine, pressure must be 6 nTorr or less

**B. Beamline Isolation Valves**

All beamline U3A, U3B and U3C isolation valves must be in the open position or the photon masks closed.

**C. Numbered Checkpoints**

*Checkpoints 1 through 9 are checked from the U2 side of beamline.*

- 1. Bremsstrahlung Shielding (BS #1)/Front end shielding in place as per photo.
- 2. Glass View ports on mirror chamber are covered (5 on U2 side of tank, 2 on top)
- 3. Bremsstrahlung Shielding (BS #2A, 2B, 2C) in place and secure.
- 4. Exclusion Zone (EZ #1), blue cage, in place and secure.
- 5. Bremsstrahlung Shielding (BS #3A) in place and secure.
- 6. Exclusion Zone (EZ #2A) in place and secure. Two sections of aluminum plates between EZ #1 and U3A Valve 2A.
- 7. Bremsstrahlung Shielding (BS #4A) in place and secure inside EZ #2A.
- 8. Bremsstrahlung Shielding (BS #5A) in place and secure inside EZ #2A.
- 9. Bremsstrahlung Shielding (BS #3B) in place and secure.

Note: Bremsstrahlung Shielding (BS # 6A) is located inside the U3A chamber.

*Checkpoints 10 and 11 are checked on the U4 side.*

- 10. All beamline visible light glass view ports covered.
- 11. Bremsstrahlung Shielding (BS #3C) in place and secured.

OPCO / Beamline Rep.: \_\_\_\_\_ Date: \_\_\_\_\_