

<b>U4IR SAFETY CHECKLIST</b>			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
<b>Rev: B</b>	<b>Effective: 5/24/2007</b>	<b>Page 1 of 1</b>	Number: LS-SCL- 0071
Reviewed by: Thomas McDonald		Reviewed by:	<b>Approved by: J. Aloï</b>

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

### SAFETY - VACUUM

- 1. The M5 vacuum is  $2 \times 10^{-9}$  Torr or less.

### SAFETY – VISIBLE LIGHT HAZARD

- 2. Any viewport capable of transmitting the visible part of the beam must be covered. The cover must be labeled “Warning-Intense Light Hazard, DO NOT STARE INTO THE BEAM”

### SAFETY – SHIELDING AND RADIATION

- 3. No additional passive shielding is required since the beam is extracted vertically behind the existing shielding.
- 4. No one should remain on the platform during injection.

OPCO / Beamline Rep.: \_\_\_\_\_ Date: \_\_\_\_\_