

X16C SAFETY CHECKLIST			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
Rev: B	Effective: 5/24/2007	Page 1 of 1	Number: LS-SCL-0036
Reviewed by: Thomas McDonald		Reviewed by:	Approved by: J. Aloï

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

1. Valid Padlock Index and log, no padlocks open

Checkpoints 2-4 are viewable from the X15 side of beamline

2. Lead Exclusion Zone (EZ # 1), lead enclosure just downstream of shield wall, in place and secured.
3. Bremsstrahlung Shield (BS #1) in place and secured.
4. Exclusion Zone (EZ #2) in place and secured.

Checkpoints 5-12 viewable from the X17 side of beamline

5. Bremsstrahlung Shield (BS # 2) in place and secured.
6. Lead Exclusion Zone (EZ #3) in place and secured.
7. Lead scatter shielding on beamline components as per photo.
8. View ports (3) covered with leaded glass; labeled 8A, 8B, 8C.
9. Exclusion Zone (EZ #4) in place and secured.
10. Bremsstrahlung Shield (BS # 3) in place and secured.
11. Bremsstrahlung Shield (BS # 4) around photon shutter in place and secured.
12. Bremsstrahlung Shield (BS # 5) located downstream of the X16C hutch, in place and secured.

OPCO / Beamline Staff: _____ Date: _____