

<b>X24C SAFETY CHECKLIST</b>			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
<b>Rev: B</b>	<b>Effective: 5/24/07</b>	<b>Page 1 of 1</b>	Number: LS-SCL-0055
Reviewed by: Thomas McDonald		Reviewed by:	<b>Approved by: J. Aloï</b>

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

1. Not used (no padlocks)

## START ON X24A SIDE

- 2. Bremsstrahlung shield (BS #1) located outside of shield wall banded as per photo.
- 3. Exclusion Zone (EZ #1) located upstream of isolation valve V1.
- 4. Bremsstrahlung shield (BS #2) located upstream of V1 as per photo.
- 5. Scatter Shield down stream of Valve 1 to mirror Tank (M1).

## ON X24C SIDE

- 6. Exclusion Zone (EZ #2) located downstream of first mirror tank.
- 7. Bremsstrahlung shield (BS # 3) in place and banded as per photo. Located downstream of valve 2
- 8. Viewport shielded and secured on Phosphor screen flipper cross-section.
- 9. 2 viewports upstream on monochromator, 3 viewports downstream on monochromator shielded and secured.

OPCO / Beamline Staff: \_\_\_\_\_ Date: \_\_\_\_\_