

X28B SAFETY CHECKLIST			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
Rev: C	Effective: 5/24/07	Page 1 of 1	Number: LS-SCL-0062
Reviewed by: Thomas McDonald		Reviewed by:	Approved by: J. Aloï

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

Note: The upstream section of this beamline has been posted as a Radiation area. This part of the beamline includes Bremsstrahlung shielding (BS #1).

- 1. Valid Padlock Index and log, no padlocks open
- 2. BS #2 in place and banded as per photo
- 3. Lead around Beam pipe banded
- 4. BS #3 in place and banded as per photo
- 5. Bellows covered with lead and banded
- 6. BS #4 in place and banded as per photo
- 7. Bellows covered with lead and banded
- 8. Scatter shields in place and banded as per photo
- 9. Scatter shields secure on X28B beamline components as per photo
- 10. BS #5 in place and banded as per photo
- 11. Hutch closure panel lined with lead in place and secured (Inside hutch downstream wall)
- 12. Hutch closure panel lined with lead in place and secured (Inside hutch downstream wall)

OPCO / Beamline Staff: _____ Date: _____