

X4A SAFETY CHECKLIST			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
Rev: B	Effective: 5/24/07	Page 1 of 1	Number: LS-SCL-0010
Reviewed by: Thomas McDonald		Reviewed by:	Approved by: J. Aloï

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

- 1. Valid Padlock Index and log, no padlocks open.

FOR CHECK POINTS 2 THRU 30 GO ON THE X3 SIDE OF BEAMLIN

- 2. Bremsstrahlung shield (BS #1) in place and banded. See photo #1
- 3. Exclusion Zone (EZ #1) located after the Bremsstrahlung shield (BS #1). See photo #1
- 4. Exclusion Zone (EZ #2) starting after the Vat valve #1A.
- 5. Bremsstrahlung shield (BS # 2) in place and banded. See photo #2
- 6. Exclusion Zone (EZ #3) located after the Bremsstrahlung shield (BS #2). See photo #2
- 7. Vacuum Bellows #1 wrapped in lead. See photo #2
- 8. Bremsstrahlung shield (BS # 3) in place and banded. See photo #3
- 9. Exclusion Zone (EZ #4) located after the Bremsstrahlung shield (BS #3). See photo #3

**10 AND 11 ARE VISIBLE FROM THE WINDOW ON THE X4C HUTCH
LOOK UNDER CHIPMUNK**

- 10. Bremsstrahlung shield (BS # 4) in place and banded.
- 11. Exclusion Zone (EZ #5) located after the Bremsstrahlung shield (BS #4).
- 12. Exclusion Zone (EZ #6) preceding the Bremsstrahlung shield #6. See photo #4
- 13. Vacuum Bellows #3 wrapped in lead. See photo #4
- 14. Viewport #1 must see "lead glass" label on the housing. See photo #4
- 15. Water cooled Beryllium window #1 wrapped in lead. See photo #4
- 16. Bremsstrahlung shield (BS # 6) in place and banded. See photo #5
- 17. Lead Exclusion Zone (EZ #7). See photo #5
- 18. Exclusion Zone (EZ #8).
- 19. Vacuum Bellows #4 wrapped in lead. See photo #6
- 20. Water cooled step-slit wrapped in lead. Viewport 2 & 3 covered in leaded glass. See photo #6
- 21. Bremsstrahlung shield (BS # 7) in place and banded. See photo #6
- 22. Exclusion Zone (EZ #9). See photo #6
- 23. Viewport 4, 5, and 6 on the Kohzu monochromator must have a "lead glass" label on the housing. See photo #7
- 24. Viewport #7 must see "lead glass" label on the housing. See photo #8
- 25. Shielding flange at the end of the kohzu monochromator must be lead wrapped. See photo #8
- 26. Bremsstrahlung shield (BS # 8) in place and banded. See photo #8
- 27. Vacuum Bellows #5 wrapped in lead. See photo #8
- 28. Viewport #8, 9 on the mirror tank must have a "lead glass" label on the housing. See photo #9
- 29. Viewport #10, 11 must see "lead glass" label on the housing. See photo #10
- 30. Water cooled Beryllium window #2 wrapped in lead. See photo #10

NOW MOVE ON THE X5 SIDE, INTO THE 4C HUTCH

- 31. Bremsstrahlung shield (BS # 5) in place and banded.
- 32. Exclusion Zone (EZ #10).

NOW OUT OF THE X4C HUTCH AND MOVE TO THE KOHZU MONOCHROMATOR

- 33. Two lead covers in the back of the kohzu monochromator. See photo #11
- 34. Viewport #12 must see "lead glass" label on the housing. See photo #11

OPCO / Beamline Staff: _____ Date: _____