

ARCS PROCEDURE: Author: TWPO	MANUS MONTHLY ROUNDS CHECKLIST	PRO(ARCS)-010.010 31 March 2006 Page 1 of 3
-------------------------------------	---	---

SMET

See Procedure PRO(SMET)-001

#	Question	Yes	No
1.	Remove the T/RH probe from the aspirated shield and replace the membrane filter; is the replacement completed successfully?		
2.	Clean the ORG lens gently using the optics brush (see PRO(SMET)-001); was this completed successfully and is the lens clean?		

SKYRAD – PSP (unshaded)

#	Question	Yes	No
1.	Following PRO(PSP)-001, check the PSP level bubbles; is the PSP level? <i>If NO, follow the procedure to level the instrument.</i>		
2.	Following PRO(PSP)-002, change the PSP desiccant. Are the desiccant replacement steps completed successfully? Note any comments in the SDL Daily Rounds comments field.		

SKYRAD – PSP (B/W, shaded)

#	Question	Yes	No
1.	Following PRO(PSP)-001, check the PSP level bubbles; is the PSP level? <i>If NO, follow the procedure to level the instrument.</i>		
2.	Following PRO(PSP)-002, change the PSP desiccant. Are the desiccant replacement steps completed successfully? Note any comments in the SDL Daily Rounds comments field.		

SKYRAD – PIR (shaded #1)

#	Question	Yes	No
1.	Following PRO(PIR)-001, check the PIR level bubbles; are the PIRs level? <i>If NO, follow the procedure to level the instrument.</i>		
2.	Following PRO(PIR)-002, change the PIR desiccant. Are the desiccant replacement steps completed successfully? Note any comments in the SDL daily rounds comments field.		

ARCS PROCEDURE: Author: TWPO	MANUS MONTHLY ROUNDS CHECKLIST	PRO(ARCS)-010.010 31 March 2006 Page 2 of 3
-------------------------------------	---	---

SKYRAD – PIR (shaded #2)

#	Question	Yes	No
1.	Following PRO(PIR)-001, check the PIR level bubbles; are the PIRs level? <i>If NO, follow the procedure to level the instrument.</i>		
2.	Following PRO(PIR)-002, change the PIR desiccant. Are the desiccant replacement steps completed successfully? Note any comments in the SDL daily rounds comments field.		

GNDRAD – PSP (down-looking)

#	Question	Yes	No
1.	Following PRO(PSP)-001, check the PSP level bubbles; is the PSP level? <i>If NO, follow the procedure to level the instrument.</i>		
2.	Following PRO(PSP)-002, change the PSP desiccant. Are the desiccant replacement steps completed successfully? Note any comments in the SDL Daily Rounds comments field.		

GNDRAD – PIR (down-looking)

#	Question	Yes	No
1.	Following PRO(PIR)-001, check the PIR level bubbles; is the PIR level? <i>If NO, follow the procedure to level the instrument.</i>		
2.	Following PRO(PIR)-002, change the PIR desiccant. Are the desiccant replacement steps completed successfully? Note any comments in the SDL daily rounds comments field.		

STAND-ALONE INSTRUMENT – Total Sky Imager (TSI)

#	Question	Yes	No
1.	Has the TSI mirror been cleaned and waxed? Clean and then wax the TSI mirror; Was this completed successfully? See PRO(TSI)-004.		

HYDROGEN GENERATOR

#	Question	Yes	No
1.	Did you perform Hydrogen Generator monthly maintenance check? See PRO(BBSS)-007, <i>BBSS: Electrolyzer Operations and Maintenance.</i>		

ARCS PROCEDURE: Author: TWPO	MANUS MONTHLY ROUNDS CHECKLIST	PRO(ARCS)-010.010 31 March 2006 Page 3 of 3
-------------------------------------	---	---

STAND-ALONE INSTRUMENT (outside I-Van) – Millimeter Cloud Radar (MMCR) and Micropulse Lidar (MPL)

#	Question	Yes	No
1.	Climb on top of the I-Van. Is the MPL window clean and not leaking? <i>If NO, see PRO(MPL)-002.</i>		
2.	Is there corrosion or damage on the external components of the MMCR? Inspect the MMCR external components; Are they free of any damage or corrosion?		