



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Honeywell FM&T Metrology
2000 East 95th Street
P.O. Box 419159
Kansas City, MO 64141-6159
Mr. Roger N. Burton
Phone: 816-997-5431 Fax: 816-997-3856
E-mail: rburton@kcp.com

CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

NVLAP Code: 20/A01 ANSI/NCSL Z540-1-1994; Part 1 Compliant

DIMENSIONAL

NVLAP Code: 20/D01
Angle Blocks

| <i>Range</i> | <i>Best Uncertainty (±) ^{note 1}</i> | <i>Remarks</i> |
|--|---|----------------------------------|
| to 45° | 1.3 arc seconds | Comparison Method |
| Autocollimators | | |
| ≤ 30 arc seconds | 0.10 arc seconds | Small Angle Generator and Mirror |
| >30 to 1200 arc sec | (0.2 arc seconds +0.2 % of angle) | |
| Index Table/Polygons | | |
| 0° to 360° (in 10° or 30° increments) | 0.5 arc seconds | 3 Stack Method |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

NVLAP Code: 20/D03

Gage Blocks

| Range | Best Uncertainty (\pm) ^{notes 1, 4} | Remarks |
|--------------------|--|----------------|
| < 0.04 in | 2.7 μin | Comparison |
| 0.04 in to 4 in | (2.6 + 0.3 L) μin ^{note 2} | Comparison |
| > 4 in to 20 in | (6.7 + 0.3 L) μin ^{note 2} | Comparison |
| < 1 mm | 0.07 μm | Comparison |
| 1 mm to 100 mm | (0.07 + 0.3 L) μm ^{note 3} | Comparison |
| > 100 mm to 500 mm | (0.13 + 0.3 L) μm ^{note 3} | Comparison |

NVLAP Code: 20/D04

Laser Frequency/Wavelength

| Laser Type | Best Uncertainty (\pm) ^{note 1} | Remarks |
|----------------------|--|----------------|
| HeNe (632.991370 nm) | 0.027 ppm | Comparison |

NVLAP Code: 20/D05

Length
Line Scales (Chrome on Glass)

| Range | Best Uncertainty (\pm) ^{note 1} | Remarks |
|-------------------|--|---|
| 0 in to 16 in | (15 + 0.5 L) μin | Vision System with Laser Interferometer |
| Step Gages | | |
| 0 in to 40 in | (20 + 1.0 L) μin ^{note 2} | Shelton CMM |
| 1-D Ball Plates | | |
| to 48 in | (20 + 1.0 L) μin ^{note 2} | CMM Single – Axis Method |
| Squares | | |
| to 24 in by 36 in | 20 μin | CMM, Self Closing Method |

Straight Edges

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

to 48 in

5 μ m

CMM, Reversal Method

NVLAP Code: 20/D07
Thread Measuring Wires

Range

All 29° and 60° Wires

Best Uncertainty (\pm) *note 1*

8.0 μ m

Remarks

Direct Measurement

NVLAP Code: 20/D08
Optical Reference Planes
Optical Flats, Mirrors

Range

0 in to 12 in
0 to 10 in

Best Uncertainty (\pm) *note 1*

1.2 μ m
2.7 μ m

Remarks

3 Flat Method
Comparison to Master / Interferometer

NVLAP Code: 20/D09
Roundness

Range

to 18 in Diameter

Best Uncertainty (\pm) *note 1*

2 μ m

Remarks

Roundness Machine

NVLAP Code: 20/D11
Spherical Diameter

Range

1/16 in to 1
(1 to 25 mm)

Best Uncertainty (\pm) *note 1*

8.3 μ m

Remarks

Comparison to Master

NVLAP Code: 20/D12
Surface Plates

Range

to 8 ft Diagonal

Best Uncertainty (\pm) *note 1*

(30 μ m + 2 μ m/ft²)

Remarks

Moody and Least Squares
Method with Autocollimator

NVLAP Code: 20/D14
Plain Plug Gages

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

| <i>Range</i> | <i>Best Uncertainty (±) ^{note 1}</i> | <i>Remarks</i> |
|--------------|---|----------------------|
| 0 in to 1 in | 6.2 μin | Comparison to Master |

Plain Ring Gages

| | | |
|--------------|--------|------------|
| 0 in to 1 in | 10 μin | Comparison |
|--------------|--------|------------|

NVLAP Code: 20/D15
2-D Ball Plates

| <i>Range</i> | <i>Best Uncertainty (±) ^{note 1}</i> | <i>Remarks</i> |
|---------------|---|---------------------------|
| 36 in x 36 in | (20 + 1.0 L) μin ^{note 2} | CMM Single – Axial Method |

NVLAP Code: 20/D16
Coordinate Measuring Machines

| <i>Range</i> | <i>Best Uncertainty (±) ^{notes 1, 2}</i> | <i>Remarks</i> |
|---------------------|---|------------------------|
| To 56 in | Axial (60 + 3 L) μin | Step Gage / Ball Plate |
| Volumetric Diagonal | Planar (65 + 3 L) μin | |
| | Spatial (70 +3 L) μin | |

ELECTROMAGNETICS – DC/LOW FREQUENCY

NVLAP Code: 20/E05
DC Resistance

| <i>Range</i> | <i>Best Uncertainty (±) in ppm ^{note 1}</i> | <i>Remarks</i> |
|-------------------|--|-------------------|
| 0.001 Ω | 3.6 | Comparison Method |
| 0.01 Ω | 3.6 | Comparison Method |
| 0.1 Ω | 3.6 | Comparison Method |
| 1 Ω | 1.1 | Comparison Method |
| 10 kΩ | 1.9 | Comparison Method |
| 1 Ω to 100 Ω | 3.6 | Comparison Method |
| 100 Ω to 1 MΩ | 3.6 | Comparison Method |
| 10 ⁷ Ω | 13.4 | Comparison Method |
| 10 ⁸ Ω | 16.7 | Comparison Method |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

NVLAP Code: 20/E06

DC Voltage

| Range | Best Uncertainty (\pm) in ppm^{note 1} | Remarks |
|--------------|--|-------------------|
| 1.018 V | 0.41 | Comparison Method |
| 10.00 V | 0.45 | Comparison Method |
| 1.000 V | 3.1 | Comparison Method |
| 10.00 V | 3.1 | Comparison Method |
| 100.0 V | 3.1 | Comparison Method |
| 1000.0 V | 3.1 | Comparison Method |
| 10.0 kV | 235 | Comparison Method |
| 100.0 kV | 220 | Comparison Method |

NVLAP Code: 20/E09

LF AC Voltage

| Range | Best Uncertainty (\pm) in ppm + μV^{note 1} | | | | | | | |
|--------------|---|-----------------|-------------------|---------------------|----------------------|-----------------------|-----------------------|--------------------|
| | Frequency in Hz | | | | | | | |
| | 10 to 20 | 20 to 40 | 40 to 20 k | 20 k to 50 k | 50 k to 100 k | 100 k to 300 k | 300 k to 500 k | 500 k to 1M |
| 2.2 mV | 1700 + 1.3 | 740 + 1.3 | 420 + 1.3 | 810 + 2.0 | 1200 + 2.5 | 2300 + 4.0 | 2400 + 8.0 | 3500 + 8.0 |
| 7 mV | 850 + 1.3 | 370 + 1.3 | 210 + 1.3 | 400 + 2.0 | 600 + 2.5 | 1200 + 4.0 | 1300 + 8.0 | 2300 + 8.0 |
| 22 mV | 290 + 1.3 | 190 + 1.3 | 110 + 1.3 | 210 + 2.0 | 310 + 2.5 | 810 + 4.0 | 890 + 8.0 | 1700 + 8.0 |
| 70 mV | 240 + 1.5 | 120 + 1.5 | 100 + 1.5 | 130 + 2.0 | 260 + 2.5 | 510 + 4.0 | 670 + 8.0 | 1100 + 8.0 |
| 220 mV | 210 + 1.5 | 100 + 1.5 | 100 + 1.5 | 100 + 2.0 | 160 + 2.5 | 250 + 4.0 | 380 + 8.0 | 1000 + 8.0 |
| 700 mV | 210 + 1.5 | 100 + 1.5 | 100 + 1.5 | 100 + 2.0 | 150 + 2.5 | 200 + 4.0 | 350 + 8.0 | 960 + 8.0 |
| 2.2 V | 200 | 100 | 100 | 100 | 150 | 200 | 300 | 1200 |
| 7 V | 200 | 100 | 100 | 100 | 150 | 200 | 400 | 1200 |
| 22 V | 200 | 100 | 100 | 100 | 150 | 200 | 400 | 1200 |
| 70 V | 200 | 100 | 100 | 100 | 150 | 200 | 410 | 1200 |
| 220 V | 200 | 100 | 100 | 100 | 150 | 400 | 500 | |
| 700 V | 200 | 100 | 100 | 200 | 500 | | | |
| 1000 V | 200 | 100 | 100 | 200 | 800 | | | |

NVLAP Code: 20/E10

LF Capacitance

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

| <i>Range</i> | <i>Best Uncertainty (±) ^{note 1}</i> | <i>Remarks</i> |
|------------------|---|----------------------------------|
| 0.01 pF to 10 pF | 100 ppm | AH2700 @ 1kHz Direct Measurement |
| 10 pF to 1000 pF | 10 ppm | AH2700 @ 1kHz Direct Measurement |

NVLAP Code: 20/E11
LF Inductance

Best Uncertainty (±) % of Reading ^{note 1}

| <i>Range</i> | <i>100 Hz</i> | <i>1 kHz</i> | <i>10 kHz</i> | <i>Remarks</i> |
|--------------|---------------|--------------|---------------|-------------------|
| 100 μH | 0.19 | 0.18 | 0.19 | Comparison Method |
| 200 μH | | 0.09 | 0.11 | Comparison Method |
| 500 μH | | 0.09 | 0.08 | Comparison Method |
| 1 mH | 0.05 | 0.04 | 0.09 | Comparison Method |
| 5 mH | | 0.034 | 0.082 | Comparison Method |
| 10 mH | 0.035 | 0.035 | 0.08 | Comparison Method |
| 20 mH | | 0.04 | 0.1 | Comparison Method |
| 50 mH | | 0.034 | 0.079 | Comparison Method |
| 100 mH | 0.033 | 0.035 | 0.098 | Comparison Method |
| 200 mH | 0.035 | 0.045 | | Comparison Method |
| 500 mH | 0.046 | 0.044 | | Comparison Method |
| 1 H | 0.03 | 0.08 | | Comparison Method |
| 2 H | | 0.074 | | Comparison Method |
| 5 H | | 0.158 | | Comparison Method |
| 10 H | 0.033 | 0.34 | | Comparison Method |

NVLAP Code: 20/E18
Resistance Dividers

| <i>Range</i> | <i>Best Uncertainty (±) in ppm ^{note 1}</i> | <i>Remarks</i> |
|---------------|--|-------------------|
| 0 V to 1200 V | 2 | Ratio Measurement |
| 0 kV to 10 kV | 44 | Ratio Measurement |

TIME AND FREQUENCY

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

NVLAP Code: 20/F01
Frequency Dissemination

| | Range | Best Uncertainty (\pm)^{note 1} | Remarks |
|---------------------|------------------|---|----------------|
| Frequency | 0.1 MHz | 1×10^{-12} | |
| Frequency | 1.0 MHz | 1×10^{-12} | |
| Frequency | 10.0 MHz | 1×10^{-12} | |
| Frequency Counters | | | |
| Internal Oscillator | (1 to 10) MHz | 1.01×10^{-12} | |
| Frequency | 1.0 Hz to 50 GHz | (0.001 to 0.012) ppm + 1 LSD | |
| Time | 10 ns to 10 s | (0.001 to 0.012) ppm + 1 LSD | |
| Generator Frequency | | | |
| Internal Oscillator | 10 MHz | 1.01×10^{-12} | |
| Frequency | 1 Hz to 50 GHz | (0.001 to 0.012) ppm | |

NVLAP Code: 20/F02
Stopwatches and Timers

| | Range | Best Uncertainty (\pm)^{note 1} | Remarks |
|------|--------------|---|----------------|
| Time | 1 day | 0.053 s/day + 1 LSD | |

MECHANICAL

NVLAP Code: 20/M03
Air Speed

| Range in fpm | Best Uncertainty (\pm) % of Reading^{note 1} | Remarks |
|---------------------|--|----------------|
| 50 to 9500 | 6 | |

NVLAP Code: 20/M06
Force

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

| <i>Range</i> | <i>Best Uncertainty (±) % ^{note 1,5}</i> | <i>Remarks</i> |
|----------------------------|---|------------------------------------|
| 5 lbf thru 500 lbf | 0.005 | Deadweight |
| > 500 lbf thru 100 000 lbf | 0.025 | Proving Ring, Uncertainty of Range |

NVLAP Code: 20/M08
Mass

| <i>Range</i> | <i>Best Uncertainty (±) in mg ^{note 1,6}</i> | <i>Remarks</i> |
|--------------|---|----------------|
| 5 kg | 2.3 | Echelon II |
| 3 kg | 1.6 | Echelon II |
| 2 kg | 1.1 | Echelon II |
| 1 kg | 0.19 | Echelon II |
| 500 g | 0.14 | Echelon II |
| 300 g | 0.086 | Echelon II |
| 200 g | 0.062 | Echelon II |
| 100 g | 0.047 | Echelon II |
| 50 g | 0.024 | Echelon II |
| 30 g | 0.014 | Echelon II |
| 20 g | 0.0095 | Echelon II |
| 10 g | 0.0051 | Echelon II |
| 5 g | 0.0026 | Echelon II |
| 3 g | 0.0016 | Echelon II |
| 2 g | 0.0011 | Echelon II |
| 1 g | 0.00071 | Echelon II |
| 500 mg | 0.00044 | Echelon II |
| 300 mg | 0.00033 | Echelon II |
| 200 mg | 0.00030 | Echelon II |
| 100 mg | 0.00032 | Echelon II |
| 50 mg | 0.00018 | Echelon II |
| 30 mg | 0.00012 | Echelon II |
| 20 mg | 0.00010 | Echelon II |
| 10 mg | 0.00010 | Echelon II |
| 5 mg | 0.00020 | Echelon II |
| 3 mg | 0.00018 | Echelon II |
| 2 mg | 0.00018 | Echelon II |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

1 mg 0.00022 Echelon II
NVLAP Code: 20/M11
 Vibration/Acceleration

| <i>Range</i> | <i>Best Uncertainty (±) in % ^{note 1}</i> |
|---|--|
| 0.5 g @ 5 Hz | 2.5 |
| 1 g @ 10 Hz thru 50 Hz | 1.8 |
| 2 g, 5 g @ 10 Hz | 2.5 |
| 5 g @ 100 Hz | 1.8 |
| 10 g @ 30 Hz thru < 100 Hz | 2.5 |
| 10 g @ 100 Hz thru 2000 Hz | 1.8 |
| 10 g @ > 2000 Hz thru 10 000 Hz | 2.5 |
| Temperature Coefficient | 3.0 |
| 10 g, 100 Hz thru 4000 Hz -65°C thru 125°C | |

Shock
 100 thru 10 000 g at the test duration 3.0
 @ 0.1 thru 50 ms. Over the range at the customer's
 requested duration 3.7

NVLAP Code: 20/M14
 Tachometers

| <i>Range</i> | <i>Best Uncertainty (±) ^{note 1}</i> | <i>Remarks</i> |
|-----------------------|---|----------------|
| RPM | | |
| 1.5 rpm to 10 000 rpm | 0.052 to 10.2 rpm | |

RF MICROWAVE

NVLAP Code: 20/R01
 Coaxial Air Line Standards
 Air Lines (Air-Dielectric)

| <i>Connector Type</i> | <i>Quantity</i> | <i>Range</i> | <i>Frequency (Hz)</i> | <i>Best Uncertainty (±) ^{note 1}</i> |
|-----------------------|-----------------|--------------|-----------------------|---|
|-----------------------|-----------------|--------------|-----------------------|---|

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

| | | | | |
|---|-------------------|---|-----------------|-------------------------|
| GR-900 | Impedance | 50 Ω | 50 M to 8.5 G | (0.028 to 0.050) Ω |
| APC-7 | Impedance | 50 Ω | 50 M to 18.0 G | (0.045 to 0.082) Ω |
| N | Impedance | 50 Ω | 50 M to 18.0 G | (0.047 to 0.210) Ω |
| APC-3.5 | Impedance | 50 Ω | 50 M to 26.5 G | (0.069 to 0.176) Ω |
| 2.4 mm | Impedance | 50 Ω | 150 M to 50.0 G | (0.093 to 0.222) Ω |
| GR-900 | Electrical Length | (3 to 30) cm | 50 M to 8.5 G | (0.0017 to 0.012) cm |
| APC-7 | Electrical Length | (3 to 30) cm | 50 M to 18.0 G | (0.0018 to 0.042) cm |
| N | Electrical Length | (3 to 30) cm | 50 M to 18.0 G | (0.0018 to 0.042) cm |
| APC-3.5 | Electrical Length | (5 to 15) cm | 50 M to 26.5 G | (0.0018 to 0.042) cm |
| 2.4 mm | Electrical Length | (3 to 5) cm | 150 M to 50 G | (0.0018 to 0.012) cm |
| GR-900, 7 mm, N, 3.5 mm, & 2.4 mm | Diameter | (0.10 to 1.5) cm | | (0.00033 to 0.00045) cm |
| GR-900, 7 mm, N, 3.5 mm, & 2.4 mm | Physical Length | (3 to 30) cm (5 to 15) cm (3 to 5) cm | | (0.0010 to 0.0026) cm |

NVLAP Code: 20/R02

Coaxial/Waveguide Terminations

Reflection Coefficient (Scattering Parameter S_{ii}) on a Vector Automatic Network Analyzer

| Connector Type | Quantity | Range | Frequency in Hz | Best Uncertainty (\pm) ^{note 1} |
|----------------|----------------------|--|-----------------|--|
| GR-900 | $ S_{ii} $ | 0 to 1 | 300 k to 8.5 G | 0.004 to 0.009 |
| APC-7 | $ S_{ii} $ | 0 to 1 | 300 k to 18.0 G | 0.003 to 0.018 |
| N | $ S_{ii} $ | 0 to 1 | 300 k to 18.0 G | 0.007 to 0.030 |
| APC-3.5 | $ S_{ii} $ | 0 to 1 | 50 M to 26.5 G | 0.008 to 0.045 |
| 2.4 mm | $ S_{ii} $ | 0 to 1 | 500 M to 50 G | 0.015 to 0.065 |
| GR-900 | $\text{Arg}(S_{ii})$ | -180° to 180° , $0 < S_{ii} < 1$ | 300 k to 8.5 G | (0.52 to 180°) |
| APC-7 | $\text{Arg}(S_{ii})$ | -180° to 180° , $0 < S_{ii} < 1$ | 300 k to 18.0 G | (0.5 to 180°) |
| N | $\text{Arg}(S_{ii})$ | -180° to 180° , $0 < S_{ii} < 1$ | 300 k to 18.0 G | (1.0 to 180°) |
| APC-3.5 | $\text{Arg}(S_{ii})$ | -180° to 180° , $0 < S_{ii} < 1$ | 50 M to 26.5 G | (1.25 to 180°) |
| 2.4 mm | $\text{Arg}(S_{ii})$ | -180° to 180° , $0 < S_{ii} < 1$ | 500 M to 50 G | (2.0 to 180°) |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

NVLAP Code: 20/R12
RF/Microwave Bolometer Units
Thermistor Mounts

| Connector Type | Quantity | Range in dBm | Frequency Range | Best Uncertainty (\pm) ^{note 1} |
|----------------|--------------------|----------------|-------------------|--|
| N | Calibration Factor | (- 10 to + 10) | 1 MHz to 18.0 GHz | (1.25 to 3.0) % |

NVLAP Code: 20/R13
RF/Microwave Attenuators
Attenuation (Scattering Parameter S_{ij}) on a Vector Automatic Network Analyzer

| Connector Type | Quantity | Range in dB | Frequency in Hz | Best Uncertainty (\pm) in dB ^{note 1} |
|----------------|------------|---------------|-----------------|--|
| 7 mm | $ S_{ij} $ | (0 to 70) | 300 k to 18.0 G | 0.02 to 0.35 |
| 7 mm | $ S_{ij} $ | (>70 to 100) | 250 M to 18.0 G | 0.25 to 1.5 |
| APC-3.5 | $ S_{ij} $ | (0 to 80) | 50 M to 26.5 G | 0.02 to 1.5 |

NVLAP Code: 20/R17
RF/Microwave Power Meters
CW Microwave Power Meter

| Connector Type | Quantity | Range in dBm | Frequency in Hz | Best Uncertainty (\pm) in dB ^{note 1} |
|----------------|----------|--------------|-----------------|--|
| N | Power | (-60 to -20) | 100 k to 18 G | (0.13 to 0.30) |
| N | Power | (-20 to +20) | 100 k to 18 G | (0.10 to 0.25) |

THERMODYNAMICS

NVLAP Code: 20/T02
Humidity

| Range | Best Uncertainty (\pm) ^{note 1} | Remarks |
|-----------------------|--|---------|
| -80 to 10 Frost Point | 2 °C Frost Point | |
| 5 to 98 % RH | 0.8 % RH | |

NVLAP Code: 20/T03
Laboratory Thermometers

| Range in °C | Best Uncertainty (\pm) in °C ^{note 1} | Remarks |
|-------------|--|---------|
| | | |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

-60 to 260 0.03
> 260 to 420 0.06

Comparison to SPRTs
Comparison to SPRTs

NVLAP Code: 20/T04
Leak Artifacts

| Range | Best Uncertainty (\pm) % of Reading ^{note 1} | Remarks |
|---|---|--------------------------|
| 10 x 10 ⁻⁷ cc/sec STP | 5 | Comparison Method Helium |
| 10 x 10 ⁻⁸ cc/sec STP | 5 | Comparison Method Helium |
| 10 x 10 ⁻⁹ cc/sec STP | 5 | Comparison Method Helium |
| 10 x 10 ⁻¹⁰ cc/sec STP | 6 | Comparison Method Helium |
| 10 x 10 ⁻⁶ cc/sec STP | 4 | Comparison Method Argon |
| 10 x 10 ⁻⁷ cc/sec STP | 4 | Comparison Method Argon |
| 10 x 10 ⁻² cc/sec STP to 10 x 10 ⁻⁸ cc/sec STP | 6.7 | Rate of Rise Method |

NVLAP Code: 20/T05
Pressure

| Range | Best Uncertainty (\pm) % of Reading ^{note 1} | Remarks |
|------------------------------|---|----------------|
| 0 psia to 500 psia | 0.015 | |
| >500 psia to 12 000 psia | 0.06 | |
| 0 psia to 500 psig | 0.015 | |
| >500 psig to 12 000 psig | 0.06 | |
| >12 000 psig to 40 000 psig | 0.05 | |
| >40 000 psig to 100 000 psig | 0.07 | |
| 0.002 psid to 500 psid | 0.08 | |

NVLAP Code: 20/T07
Resistance Thermometry - SPRT

| Range in °C | Best Uncertainty (\pm) in °C ^{note 1} | Remarks |
|--------------------|--|------------------------|
| -40 to 232 | 0.005 | Fixed Point Cell |
| -190 to 100 | 0.020 | SPRT Comparison Method |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

NVLAP Code: 20/T08
Thermocouples

| <i>Range in °C</i> | <i>Best Uncertainty (±) in ^{note 1}</i> | <i>Remarks</i> |
|--------------------|--|----------------|
| -75 to 485 | 0.4 | Type J |
| 0 to 500 | 0.75 | Type K |
| >500 to 1100 | 1.0 | Type K |
| >1100 to 1250 | 1.4 | Type K |
| -75 to 260 | 0.25 | Type T |

NVLAP Code: 20/T09
Thermocouple Calibrator/ Reader (Simulated or Measured)

| <i>Range in °C</i> | <i>Best Uncertainty (±) in °C ^{note 1}</i> | <i>Remarks</i> |
|--------------------|---|----------------|
| 100 to 1820 | 0.34 to 0.39 | Type B |
| -270 to 1000 | 0.36 to 0.63 | Type E |
| -200 to 1200 | 0.34 to 0.36 | Type J |
| -200 to 1350 | 0.34 to 0.38 | Type K |
| -50 to 1750 | 0.34 to 0.39 | Type R |
| -50 to 1750 | 0.34 to 0.40 | Type S |
| -200 to 400 | 0.34 | Type T |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200108-0

-
1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.
 2. L is in inches.
 3. L is in meters.
 4. Best uncertainty is for steel blocks.
 5. ASTM E74 compliant calibrations. Best uncertainty includes the uncertainty of a high quality artifact under test. Specific artifact uncertainty is determined by the statistics of the test and the artifact under test.
 6. Using redundant weighing designs and assuming density standard uncertainty of ± 0.005 g /cc for subject weights

2009-01-01 through 2009-12-31

Effective dates

A handwritten signature in cursive script that reads 'Sally S. Bruce'.

For the National Institute of Standards and Technology