



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Asanuma Giken Co., Ltd.
 4079-1 Koto-Cho, Hamamatsu-shi
 Shizuoka-ken 431-1103
 JAPAN
 Mr. Susumu Asanuma
 Phone: 81-53-486-1110 Fax: 81-53-486-1109
 E-mail: info@asanuma-g.jp
 URL: http://www.asanuma-g.jp/

CALIBRATION LABORATORIES

NVLAP LAB CODE 200536-0
 Scope Revised: 2008-11-24

DIMENSIONAL

NVLAP Code: 20/D16
 Coordinate Measuring Machines

<i>Range</i>	<i>Best Uncertainty (\pm)^{note 1}</i>	<i>Remarks</i>
0 mm to 150 mm	3.0 μ m	Artifact Calibration
0 in to 5.91 in	118.1 μ in	Artifact Calibration

1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.
2. Calibration, including two- and three-dimensional measurements of the Asanuma Giken manufactured Quality Master Gauge.

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology