



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Technical Calibration Labs, Inc.
 51 Peddler Hill Road
 Monroe, NY 10950
 Mr. John Tasca
 Phone: 845-783-4235 Fax: 845-783-3573
 E-mail: colorcal@optionline.net
 URL: www.technicalibration.com

CALIBRATION LABORATORIES

NVLAP LAB CODE 200797-0

OPTICAL RADIATION

NVLAP Code: 20/O02
 Color Temperature

<i>Range in K</i>	<i>Best Uncertainty (±) in K ^{note 1}</i>
2300	42
2856	54

1. Represents an expanded uncertainty using a coverage factor, k = 2, at an approximate level of confidence of 95 %.

2008-10-01 through 2009-09-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology