



National Voluntary Laboratory Accreditation Program



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

ORNL Metrology Laboratory

P.O. Box 2008, MS-6366

1 Bethel Valley Road

Oak Ridge, TN 37831

Mr. Mike Duncan

Phone: 865-574-7349 Fax: 865-241-4644

E-mail: duncanml@ornl.gov

URL: <http://www.ornl.gov/sci/metrology>

CALIBRATION LABORATORIES

NVLAP LAB CODE 200659-0

NVLAP Code: 20/A01

ANSI/NCSL Z540-1-1994; Part 1

Compliant

NVLAP Code: 20/D03

Gage Blocks

Range in Inches

Best Uncertainty (\pm) in μ inches^{note 1}

Remarks

0 to 4

4.1 + 1.1L

L in inches for Steel

0 to 4

4.2 + 0.6L

L in inches for Chrome Carbide

ELECTROMAGNETICS - DC/LOW FREQUENCY

NVLAP Code: 20/E02

AC Current Output

**Best Uncertainty (\pm) in A^{note 1}
Frequency in Hertz**

Range in A	10	20	40	1 k	5 k	10 k	Remarks
0.00022	1.16 E-06	1.16 E-06	5.78 E-07	3.05 E-07	4.19 E-07	4.19 E-07	Source / Measure
0.0022	1.16 E-05	1.16 E-05	5.78 E-06	3.03 E-06	4.19 E-06	4.19 E-06	Source / Measure
0.022	1.16 E-04	1.16 E-04	5.78 E-05	3.02 E-05	4.13 E-05	4.13 E-05	Source / Measure
0.22	1.16 E-03	1.16 E-03	6.01 E-04	4.63 E-04	9.37 E-04	9.37 E-04	Source / Measure

2008-10-01 through 2009-09-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200659-0

2.2	1.16 E-03	1.39 E-03	1.39 E-03	9.43 E-03	9.43 E-03	Source / Measure
11.0		2.91 E-02	2.95 E-02	3.59 E-02	3.67 E-02	Source / Measure

NVLAP Code: 20/E05

2 Wire Resistance

Range in Ω	Best Uncertainty (\pm) in Ω ^{note 1}	Remarks
0	2.89 E-01	Source / Measure
1	2.89 E-01	Source / Measure
1.9	2.89 E-01	Source / Measure
10	2.89 E-01	Source / Measure
19	2.90 E-01	Source / Measure
100	2.91 E-01	Source / Measure
190	2.92 E-01	Source / Measure
1 k	3.01 E-01	Source / Measure
1.9 k	3.19 E-01	Source / Measure
10 k	4.18 E-01	Source / Measure
19 k	5.89 E-01	Source / Measure
100 k	1.63 E + 00	Source / Measure
190 k	6.04 E + 00	Source / Measure
1 M	2.31 E + 01	Source / Measure
1.9 M	2.27 E + 02	Source / Measure
10 M	7.31 E + 02	Source / Measure
19 M	1.21 E + 04	Source / Measure
100 M	5.92 E + 04	Source / Measure

4 Wire Resistance

0	6.56 E-05	Source / Measure
1	9.59 E-05	Source / Measure
1.9	1.40 E-04	Source / Measure
10	2.69 E-04	Source / Measure
19	9.39 E-04	Source / Measure
100	2.11 E-03	Source / Measure

2008-10-01 through 2009-09-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200659-0

190	3.11 E-03	Source / Measure
1 k	1.31 E-02	Source / Measure
1.9 k	3.10 E-02	Source / Measure
10 k	1.31 E-01	Source / Measure
19 k	2.96 E-01	Source / Measure
100 k	1.37 E + 00	Source / Measure
190 k	5.76 E + 00	Source / Measure
1 M	2.28 E + 01	Source / Measure
1.9 M	2.27 E + 02	Source / Measure
10 M	7.31 E + 02	Source / Measure
19 M	1.21 E + 04	Source / Measure

NVLAP Code: 20/E06
DC Voltage

<i>Range in V</i>	<i>Best Uncertainty (\pm) in V^{note 1}</i>	<i>Remarks</i>
0	9.33 E-07	Source / Measure
0.22	3.27 E-06	Source / Measure
2.2	2.20 E-05	Source / Measure
11	1.52 E-04	Source / Measure
22	2.92 E-04	Source / Measure
220	2.73 E-03	Source / Measure
1100	1.26 E-02	Source / Measure

NVLAP Code: 20/E06
DC Voltage Fixed Points

<i>Range in V</i>	<i>Best Uncertainty (\pm) in V^{note 1}</i>	<i>Remarks</i>
0.1	1.22 E-06	Source / Measure
1	3.65 E-06	Source / Measure
10	2.46 E-05	Source / Measure
100	3.33 E-04	Source / Measure
1000	4.16 E-03	Source / Measure

2008-10-01 through 2009-09-30

Effective dates

Sally S. Bruce
For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200659-0

NVLAP Code: 20/E06

DC Current

Range in A	Best Uncertainty (\pm) in A^{note 1}	Remarks
0	3.47 E-09	Source / Measure
0.00022	1.38 E-08	Source / Measure
0.0022	1.14 E-07	Source / Measure
0.022	1.45 E-06	Source / Measure
0.22	3.75 E-05	Source / Measure
2.2	2.74 E-03	Source / Measure
11.0	4.89 E-03	Source / Measure

NVLAP Code: 20/E09

LF AC Voltage

Range	Best Uncertainty (\pm) in % + μV^{note 1}							
	Frequency in Hertz							
10 to 20	20 to 40	40 to 20 k	20 k to 50 k	50 k to 100 k	100 k to 300 k	300 k to 500 k	500 k to 1 M	
2.2 mV	0.209 + 4	0.209 + 4	0.18 + 4	0.808 + 4	1.697 + 5	1.680 + 10	1.680 + 10	1.683 + 20
22 mV	0.036 + 4	0.036 + 4	0.047 + 4	0.047 + 4	0.105 + 4	0.406 + 5	1.214 + 10	1.215 + 20
220 mV	0.034 + 12	0.032 + 7	0.032 + 7	0.047 + 7	0.105 + 07	0.405 + 17	1.214 + 20	1.215 + 25
2.2 V	0.034 + 40	0.032 + 15	0.032 + 15	0.046 + 8	0.104 + 10	0.404 + 30	1.213 + 80	1.214 + 200
22 V	0.048 + 400	0.047 + 150	0.047 + 150	0.052 + 50	0.150 + 100	0.150 + 200	1.790 + 600	1.791 + 2000

Range	Best Uncertainty (\pm) in % + mV^{note 1}						
220 V	0.071 + 4	0.070 + 1.5	0.070 + 1.5	0.150 + 0.6	0.358 + 1	0.358 + 2.5	1.749 + 16
1100 V	0.067 + 16	0.053 + 4	0.053 + 3.5				1.777 + 40

TIME AND FREQUENCY

2008-10-01 through 2009-09-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200659-0

NVLAP Code: 20/F01
Frequency Dissemination

Range in Hz	Best Uncertainty (\pm) (Fractional Frequency Error) ^{note 1}	Remarks
10 M	1.07×10^{-11}	GPS Master Oscillator

MECHANICAL

NVLAP Code: 20/M05
Flow Rate

Range	Best Uncertainty (\pm) ^{note 1}	Remarks
0.1 SLPM to 1.0 SLPM	0.6 % of Reading + 0.01 % FS	Laminar
3.0 SLPM to 30 SLPM	0.6 % of Reading + 0.01 % FS	Laminar
28 SLPM to 100 SLPM	0.5 % of Reading + 0.01 % FS	Sonic
67 SLPM to 250 SLPM	0.5 % of Reading + 0.01 % FS	Sonic
248 SLPM to 1000 SLPM	0.5 % of Reading + 0.01 % FS	Sonic
3 sccm to 40 000 sccm	0.56 % of Reading	Sierra

NVLAP Code: 20/M06
Torque Transducers

Range	Best Uncertainty (\pm) in % ^{note 1}	Remarks
4 to 50 in lb	0.24	
30 to 400 in lb	0.20	
80 to 1000 in lb	0.22	
20 to 250 ft lb	0.16	
60 to 600 ft lb	0.09	

Torque Wrenches

4 in lb to 600 ft lb	0.25
----------------------	------

2008-10-01 through 2009-09-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200659-0

NVLAP Code: 20/M08

Mass

Range	Best Uncertainty (\pm) ^{note 1}	Remarks
50 lb	13 mg	Echelon II
20 kg	7.8 mg	Echelon II
10	2.7 mg	Echelon II
5	2.3 mg	Echelon II
2	0.41 mg	Echelon II
1	0.12 mg	Echelon II
500 g	0.067 mg	Echelon II
200	0.042 mg	Echelon II
100	0.039 mg	Echelon II
50	0.024 mg	Echelon II
20	0.011 mg	Echelon II
10	8.3 μ g	Echelon II
5	4.5 μ g	Echelon II
2	2.1 μ g	Echelon II
1	1.8 μ g	Echelon II
500 mg	1.3 μ g	Echelon II
200	1.0 μ g	Echelon II
100	1.0 μ g	Echelon II
50	0.9 μ g	Echelon II
20	0.7 μ g	Echelon II
10	0.7 μ g	Echelon II
5	0.7 μ g	Echelon II
2	0.6 μ g	Echelon II
1	0.6 μ g	Echelon II

Balances ^{Note 2}

Readability in mg	Uncertainty(\pm) in mg	Remarks
0.001	0.0043	OIML Class E2 Weights
0.01	0.059	OIML Class E2 Weights
0.1	0.095	OIML Class E2 Weights
1	0.75	OIML Class E2 Weights
10	7.5	OIML Class E2 Weights
100	75	OIML Class E2 Weights

2008-10-01 through 2009-09-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200659-0

THERMODYNAMIC

NVLAP Code: 20/T02

Humidity

<i>Range in %</i>	<i>Best Uncertainty (\pm) in %^{Note 1}</i>	<i>Remarks</i>
10 to 95 RH	1.2	2500

NVLAP Code: 20/T03

Laboratory Thermometers

RTD's, Thermistors, and Direct Reading Thermometers

<i>Range in °C</i>	<i>Best Uncertainty (\pm) in °C^{Note 1}</i>	<i>Remarks</i>
0 to 100	0.0093	Water Bath
100 to 419	0.265	Fluidized Bath

Thermocouples

<i>Thermocouple Type</i>	<i>Best Uncertainty (\pm) in °C^{Note 1} (Range 0 to 100 °C in Water)</i>	<i>Best Uncertainty (\pm) in °C^{Note 1} (Range 100 to 419 °C in Fluidized Bath)</i>
B	n/a	1.359
C	0.197	0.368
E	0.177	0.357
J	0.178	0.358
K	0.179	0.358
R	0.282	0.390
S	0.279	0.391
T	0.179	0.358

2008-10-01 through 2009-09-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200659-0

NVLAP Code: 20/T05

Pressure

Absolute Mode

Range	Best Uncertainty (\pm) in ppm of Reading + mPa (psi) ^{note 1}	Remarks
1.0 Pa to 15 kPa	37 + 8	FPG 8601
0.000145 psia to 2.17 psia	37 + (1.2 E-6)	FPG 8601

Range	Best Uncertainty (\pm) in ppm of Reading + Pa ^{note 1}	Remarks
10 to 380 kPa (1.45 to 55 psia)	15.7 + 0.10	7601 (10 kPa/kg)
50 to 1900 kPa (7.3 to 275 psia)	18.0 + 0.12	7601 (50 kPa/kg)
200 to 7600 kPa (29 to 1100 psia)	19.9 + 0.25	7601 (200 kPa/kg)

Absolute Differential Mode

Range	Best Uncertainty (\pm) in ppm of Reading + mPa (psi) ^{note 1}	Remarks
0 Pa to 15 kPa	37 + 5	FPG 8601
0 psia to 2.17 psia	37 + (7.3 E-7)	FPG 8601

Gauge Mode	Best Uncertainty (\pm) in ppm of Reading + mPa (psi) ^{note 1}	Remarks
0 Pa to 15 kPa	37 + 5	FPG 8601
0 psi to 2.17 psi	37 + (7.3 E-7)	FPG 8601

Range	Best Uncertainty (\pm) in ppm ^{note 1}	Remarks
10 to 380 kPa (1.45 to 55 psia)	16.0 + 0.01 Pa	7601 (10 kPa/kg)
50 to 1900 kPa (7.3 to 275 psi)	18.3 + 0.06 Pa	7601 (50 kPa/kg)
200 to 7600 kPa (29 to 1100 psi)	20.2 + 0.23 Pa	7601 (200 kPa/kg)
1.379 MPa to 68.948 MPa	108	5202
200 psig to 10 000 psig	108	5202

2008-10-01 through 2009-09-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200659-0

NVLAP Code: 20/T07
Resistance Thermometry

Range in °C	Best Uncertainty (\pm) in mK	Remarks
0.01	1.0	TPW
29.7646	2.0	GaMP
231.928	4.0	Sn FP
419.527	8.0	Zn FP

-
1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.
 2. This parameter is accredited for field service, reported uncertainties may be higher due to environmental conditions.

2008-10-01 through 2009-09-30

Effective dates

For the National Institute of Standards and Technology