

## NC Job Risk Assessment

<b>Name(s) of Risk Team Members:</b> L. Davis, D. Elling, S. Hoey, W. Litzke, A. Piper	<b>Point Value Parameter</b> → ↓	1	2	3	4	5
<b>Job Title:</b> <b><i>Working with Vacuum Systems</i></b>	<b>Frequency (B)</b>	≤once/year	≤once/month	≤once/week	≤once/shift	>once/shift
<b>Job Number or Job Identifier:</b> NC-JRA-015						
<b>Job Description:</b> Work with systems subject to vacuum pumps and pressures less than atmospheric	<b>Severity (C)</b>	First Aid Only	Medical Treatment	Lost Time	Partial Disability	Death or Permanent Disability
Training Procedures List (Optional):	<b>Likelihood (D)</b>	Very Unlikely	Unlikely	Possible	Probable	Multiple
Approved by: S. Hoey Date: 12/29/07 Rev. #: 0						
<b>Stressors (if applicable, please list all)</b>		<b>Reason for Revision (if applicable):</b>			<b>Comments:</b>	

Activity	Hazard	Control(s)	Before Additional Controls					Control(s) Added to Reduce Risk	After Additional Controls					% Risk Reduction	
			Stressor	# of People A	Frequency B	Severity C	Likelihood D		Risk* AxBxCxD	Stressors	# of People A	Frequency B	Severity C		Likelihood D
1. Operation of Vacuum Systems	Electrical Shock	Proper grounding, feed-through protectors, strain relief, power cord condition, Tier 1 inspections, EEL inspections.	N	1	4	5	1	20							
	Release of toxic or potentially hazardous nanomaterials due to vacuum failure	Interim standard on nanomaterials, work permit, vacuum interlocks, Experimental safety reviews, HEPA filters	N	1	3	2	2	12							
	Inhalation of oil mists	Mist eliminators, venting to hoods or outdoors	N	1	4	2	2	16							
2. Assembly or disassembly of vacuum	Bending, twisting, overexertion, dead lifting, repetitive motion,	Proper lifting techniques, back safety training (if applicable) , Coordination wit co-workers, work planning , gloves	N	1	3	2	3	18							

## NC Job Risk Assessment

Activity	Hazard	Control(s)	Before Additional Controls						Control(s) Added to Reduce Risk	After Additional Controls					
			Stressor	# of People A	Frequency B	Severity C	Likelihood D	Risk* AxBxCxD		Stressors	# of People A	Frequency B	Severity C	Likelihood D	Risk* AxBxCxD
system components	working in areas with restricted access resulting in bumps and abrasions														
3. Leak checking vacuum systems	See NC-JRA-004 Work with compressed gas cylinders														
4. Maintenance of vacuum systems	See NC-JRA—11 Hand Tool Use														
	Bending, twisting, overexertion, dead lifting, repetitive motion	Proper lifting techniques, back safety training (if applicable), coordination with co workers, work planning, gloves	N	1	3	2	2	12							
	Potential exposure to toxic/hazardous materials (nanomaterials)		Y	1	2	3	2	12							
5. Bleeding up of vacuum systems to atmospheric	Release of vacuum potential overpressure from purge gas.	Follow manufacturers recommendations for maintenance, SOP's for maintenance, Regulate N2, See NC-JRA-004 Compressed gas.	N	1	2	3	2	12							
6. Heating of vacuum systems/bak eout	Contact with temperature extremes (burns)	Insulated Heating blankets/panels, warning signs, gloves, lab access controls.	Y	1	3	2	3	18							

<b>*Risk:</b>	<b>0 to 20</b>	<b>21 to 40</b>	<b>41-60</b>	<b>61 to 80</b>	<b>81 or greater</b>
	<b>Negligible</b>	<b>Acceptable</b>	<b>Moderate</b>	<b>Substantial</b>	<b>Intolerable</b>