



Quality Assurance Site Information for Azusa

This page updated February 10, 2005



AIRS Number	ARB Number	Site Start Date	Reporting Agency and Agency Code
060370002	70060	3/1/78	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
803 N. Loren Av, Azusa CA 91702	Los Angeles	South Coast	34° 8' 11"	117° 55' 26"	182



Site Survey Report

Siting Information

Site Name: Azusa	Audit Date: 2006-08-09	ARB Number: 70060	AIRS Number: 060370002
Address: 803 N. Loren Av Azusa, CA 91702	Latitude: 34° 8' 11"	Longitude: 117° 55' 26"	Elevation (m): 182
	Auditors: Long Liu Alvin Danque	Site Technician: Norman Broellos	Site Phone:
Operating Agency: South Coast AQMD		Site Report: Yes	Site Photos: Yes

General Siting Conditions

Station Temperature Controlled: Yes Recorded: Yes Inside Temp: 25 Degrees Celsius	Traffic Description: Urban Distance: 7 meters Count (Veh/Day): 1000	Topography Site: Level Region: Level	Predominant Wind Direction: West Arc Air Flow (Deg): 360 Degrees
		QA Manual Approved: Yes Agency: South Coast AQMD	Probe Clean: Yes Manifold Clean: Yes
			Cleaning Schedule: Semi Annually Autocalibrator Type: Environics 100
Meteorology Located With Instruments: Yes Shadowing: No Boom Orientation (Deg): 345 Temp(Motor/Natural): Natural	Non-vehicular Local Sources Description: Welding shop Distance: 10 meters Direction: WNW	Urbanization: Suburban	Site Survey Complete: Yes
		Ground Cover: Roof	Logbook Up To Date: Yes

Site Survey Report (Cont.)

Monitor Type	Carbon Monoxide	Nitrogen Dioxide	Ozone	PM10-SSI
Manufacturer/Model	Horiba APMA-360	API 200	API/Teledyne 400	Sierra Anderson 1200
Serial Number	0016511	e000221	534-S	07421
POC	1	2	1	2
Data For Record?	Yes	Yes	Yes	No
Purpose				
Objective				
Scale				
Height Above Ground	5.2	5.2	5.2	4.5
Height Above Platform	1.8	1.8	1.8	1.1
Sampler Spacing				
Current Manual Available?	Yes	Yes	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes	Yes	Yes
In-line Filter Change Date	2006-08-08	2006-08-08	2006-08-08	
Cal. Gas Cert. Date	2002-11-22	2002-11-22		
Calibration Current?	Yes	Yes	Yes	No
Calibration Date	2006-05-12	2006-05-12	2006-05-12	2005-12-23
Cal. Equipment Cert. Date	2006-02-10	2006-02-10	2006-02-10	2005-08-18
Obstacle Description	None	None	None	None
Distance to Obstacle	-	-	-	-
Height Above Inlet	-	-	-	-
Distance to Walls, etc.	-	-	-	-
Distance to Dripline	-	-	-	-
Dominant Influence	Industrial	Industrial	Industrial	
Residence Time (sec)	13.8	14.2	15.4	

Site Survey Report (Cont.)

Monitor Type	TSP	PM2.5	Xontech 920
Manufacturer/Model	GMW TSP	Sierra Anderson RAAS PM2.5	Xontech 924
Serial Number	1578	00317	20021471
POC	2	1	1
Data For Record?	Yes	Yes	Yes
Purpose			
Objective			
Scale			
Height Above Ground	4.5	5.5	4.5
Height Above Platform	1.1	1.9	1.5
Sampler Spacing		1.6	
Current Manual Available?	Yes	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes	Yes
In-line Filter Change Date			
Cal. Gas Cert. Date			
Calibration Current?	Yes	Yes	Yes
Calibration Date	2006-07-07	2006-10-24	2006-08-24
Cal. Equipment Cert. Date	2005-08-18	Not Available	Not Available
Obstacle Description	None	None	None
Distance to Obstacle	-	-	-
Height Above Inlet	-	-	-
Distance to Walls, etc.	-	-	-
Distance to Dripline	-	-	-
Dominant Influence	Vehicular	Vehicular	Industrial
Residence Time (sec)			