Quality Assurance
Site Information for Burbank



AIRS Number	ARB Number	Site Start Date	Reporting Agency and Agency Code
060371002	70069	3/1/80	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
228 W Palm Av, Burbank CA 91502	Los Angeles	South Coast	34° 10' 33"	118° 19' 1"	10


BURBANK
PO
McCambridge Park
BROADWAY
SAN FERNANDO
PACIFIC
11
Monterey Ave Sch
Burbank Junc
Methodist Sch
Burbank High Sch
Emerson Sch
Bellarmine Sch
Bellarmine-Jefferson H S
Hospital
Library
Community Hospital
Fire Station
City Hall
Main PO
BM 586
BM 568
595
605
BLVD
KEYSTONE
PARISH
REESE
BURBANK BOUNDARY
Edison Sch
Powerplant
WT
564
GRIFFITH
LOMITA
MAGNOLIA
VICTORY
MARIPOSA
BM 562
Central Sch
546
BM 535
Water Tank
ALAMEDA
WESTERN
PROVIDENCIA
LAKE
FLOWER
574
BUENA VISTA
LINCOLN
Burroughs High Sch
Olive Ave Recreation Center
Fire Station
McKinley Sch
CHANNEL
BURBANK GLENDALE
THOMPSON
Library
Lincoln Sch
Jordan Jr High Sch
St Finbar Sch
509
Drive-in Theater
500
Gaging Station
AVENUE
PARKSIDE AVE
Mountain View
Equestrian Area
GLENDALE
DRIVE
530
520
560
644
720
724

Site Survey Report

Siting Information

Site Name: Burbank	Audit Date: 2006-07-25	ARB Number: 70069	AIRS Number: 060371002
Address: 228 W Palm Av Burbank, CA 91502	Latitude: 34° 10' 33"	Longitude: 118° 19' 1"	Elevation (m): 10
	Auditors: Charles Pearson Fred Burriell	Site Technician: Norm Broellos	Site Phone:
Operating Agency: South Coast AQMD		Site Report: Yes	Site Photos: Yes

General Siting Conditions

Station Temperature	**Traffic**	**Topography**	Predominant Wind Direction: South
Controlled: Yes	Description: Arterial	Site: Level	Arc Air Flow (Deg): 360 Degrees
Recorded: Yes	Distance: 92 meters	Region: Valley	Probe Clean: Yes
Inside Temp: 26 Degrees Celsius	Count (Veh/Day): 50000	**QA Manual**	Manifold Clean: Yes
Meteorology	**Non-vehicular Local Sources**	Approved: Yes	Cleaning Schedule: Annually
Located With Instruments: Yes	Description: E.G. Plant	Agency: South Coast AQMD	Autocalibrator Type: Environics 100
Shadowing: No	Distance: 100 meters	Urbanization: Suburban	Site Survey Complete: Yes
Boom Orientation (Deg): West	Direction: E	Ground Cover: Roof	Logbook Up To Date: Yes
Temp(Motor/Natural): Natural			

Site Survey Report (Cont.)

Monitor Type	Carbon Monoxide	Sulfur Dioxide	Nitrogen Dioxide	Ozone
Manufacturer/Model	Horiba APMA-360	TECO 43A, 43B, 43C	API 200A	API/Teledyne 400
Serial Number	0016212	0016628	E000207	536-S
POC	1	2	2	1
Data For Record?	Yes	Yes	Yes	Yes
Purpose				
Objective				
Scale				
Height Above Ground	6.0	6.0	6.0	6.0
Height Above Platform	2.5	2.5	2.5	2.5
Sampler Spacing				
Current Manual Available?	Yes	Yes	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes	Yes	Yes
In-line Filter Change Date	2006-07-19	2006-07-19	2006-07-19	2006-07-19
Cal. Gas Cert. Date	Not Available	Not Available	Not Available	
Calibration Current?	Yes	Yes	Yes	Yes
Calibration Date	2006-01-27	2006-02-02	2006-07-11	2006-01-26
Cal. Equipment Cert. Date	Not Available	Not Available	Not Available	Not Available
Obstacle Description	None	None	None	None
Distance to Obstacle	-	-	-	-
Height Above Inlet	-	-	-	-
Distance to Walls, etc.	-	-	-	-
Distance to Dripline	-	-	-	-
Dominant Influence	Vehicular	Vehicular	Vehicular	Vehicular
Residence Time (sec)	9.2	13.1	9.9	9.6

Site Survey Report (Cont.)

Monitor Type	PM10-SSI	TEOM	BAM-PM2.5	PM2.5
Manufacturer/Model	Andersen 1200	R&P 1400a	MetOne BAM1020	Andersen RAAS2.5-3.0
Serial Number	4932	001011	132843	E000022
POC	2	2	2	1
Data For Record?	Yes	Yes	No	Yes
Purpose				
Objective				
Scale				
Height Above Ground	5.0	5.0	5.0	5.2
Height Above Platform	1.5	1.5	2.0	1.5
Sampler Spacing				
Current Manual Available?	Yes	Yes	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes	Yes	Yes
In-line Filter Change Date				
Cal. Gas Cert. Date				
Calibration Current?	Yes	Yes	Yes	Yes
Calibration Date	Not Available	2005-09-15	2006-06-28	2005-05-06
Cal. Equipment Cert. Date	Not Available	Not Available	Not Available	Not Available
Obstacle Description	None	None	None	None
Distance to Obstacle	-	-	-	-
Height Above Inlet	-	-	-	-
Distance to Walls, etc.	-	-	-	-
Distance to Dripline	-	-	-	-
Dominant Influence	Vehicular	Vehicular	Vehicular	Vehicular
Residence Time (sec)				

Site Survey Report (Cont.)

Monitor Type	PM2.5	Xontech 920	Xontech 920
Manufacturer/Model	Sierra Anderson RAAS PM2.5	Xontech 920	Xontech 920
Serial Number	00340	11179	20004221
POC	1	1	1
Data For Record?	Yes	Yes	Yes
Purpose			
Objective			
Scale			
Height Above Ground	5.2	4.9	4.9
Height Above Platform	1.5	1.2	1.2
Sampler Spacing			
Current Manual Available?	Yes	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes	Yes
In-line Filter Change Date			
Cal. Gas Cert. Date			
Calibration Current?	Yes	Yes	Yes
Calibration Date	2005-11-09	2006-06-27	2006-06-14
Cal. Equipment Cert. Date	Not Available	Not Available	Not Available
Obstacle Description	None	None	None
Distance to Obstacle	-	-	-
Height Above Inlet	-	-	-
Distance to Walls, etc.	-	-	-
Distance to Dripline	-	-	-
Dominant Influence	Vehicular	Vehicular	Vehicular
Residence Time (sec)			