



**Quality Assurance  
Site Information for Los Angeles-North Main Street**



AIRS Number	ARB Number	Site Start Date	Reporting Agency and Agency Code
060371103	70087	3/1/78	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
1630 North Main Street, Los Angeles CA 90012	Los Angeles	South Coast	34° 3' 59"	118° 13' 36"	87



# Site Survey Report

## Siting Information

Site Name: Los Angeles-North Main Street	Audit Date: 2006-07-26	ARB Number: 70087	AIRS Number: 060371103
Address: 1630 North Main Street Los Angeles, CA 90012	Latitude: 34° 3' 59"	Longitude: 118° 13' 36"	Elevation (m): 87
	Auditors: Charles Pearson Fred Burriell	Site Technician: Tom Mac	Site Phone:
Operating Agency: South Coast AQMD		Site Report: Yes	Site Photos: Yes

## General Siting Conditions

<b>Station Temperature</b>  Controlled: Yes  Recorded: Yes  Inside Temp: 26 Degrees Celsius	<b>Traffic</b>  Description: Commercial  Distance: 40 meters  Count (Veh/Day): 10000	<b>Topography</b>  Site: Level  Region: Level	Predominant Wind Direction: West
			Arc Air Flow (Deg): 360 Degrees
			Probe Clean: Yes
<b>Meteorology</b>  Located With Instruments: Yes  Shadowing: No  Boom Orientation (Deg): 345  Temp(Motor/Natural): Natural	<b>Non-vehicular Local Sources</b>  Description: None  Distance: N/A  Direction: N/A	<b>QA Manual</b>  Approved: Yes  Agency: South Coast AQMD	Manifold Clean: Yes
			Cleaning Schedule: Annually
		Urbanization: Suburban	Site Survey Complete: Yes
		Ground Cover: Roof	Logbook Up To Date: Yes

## Site Survey Report (Cont.)

Monitor Type	Carbon Monoxide	Sulfur Dioxide	Nitrogen Dioxide	Ozone
Manufacturer/Model	Horiba APMA-360	TECO 43A, 43B, 43C	API 200A	API/Teledyne 400
Serial Number	0016214	0016634	000210	520-S
POC	1	1	1	1
Data For Record?	Yes	Yes	Yes	Yes
Purpose				
Objective				
Scale				
Height Above Ground	12.1	12.1	12.1	12.1
Height Above Platform	2.0	2.0	2.0	2.0
Sampler Spacing				
Current Manual Available?	Yes	Yes	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes	Yes	Yes
In-line Filter Change Date	2006-07-25	2006-07-25	2006-07-25	2006-07-25
Cal. Gas Cert. Date	Not Available	Not Available	Not Available	
Calibration Current?	Yes	Yes	Yes	Yes
Calibration Date	2006-04-07	2006-04-07	2006-04-07	2006-05-24
Cal. Equipment Cert. Date	Not Available	Not Available	Not Available	Not Available
Obstacle Description	None	None	None	None
Distance to Obstacle	-	-	-	-
Height Above Inlet	-	-	-	-
Distance to Walls, etc.	-	-	-	-
Distance to Dripline	-	-	-	-
Dominant Influence	Vehicular	Vehicular	Vehicular	Vehicular
Residence Time (sec)	7.7	9.1	8.4	7.9

## Site Survey Report (Cont.)

Monitor Type	PM10-SSI	BAM-PM2.5	PM2.5	PM2.5
Manufacturer/Model	Andersen 1200	MetOne BAM1020	Andersen RAAS 2.5	Andersen RAAS 2.5
Serial Number	7363	20021149	E000005	E000006
POC	2	3	1	2
Data For Record?	Yes	Yes	Yes	Yes
Purpose				
Objective				
Scale				
Height Above Ground	11.8	12.1	12.0	12.0
Height Above Platform	1.7	2.0	1.9	1.9
Sampler Spacing			2.0	2.0
Current Manual Available?	Yes	Yes	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes	Yes	Yes
In-line Filter Change Date				
Cal. Gas Cert. Date				
Calibration Current?	Yes	Yes	Yes	Yes
Calibration Date	2006-06-26	2006-04-21	2006-06-29	2006-06-29
Cal. Equipment Cert. Date	Not Available	Not Available	Not Available	Not Available
Obstacle Description	None	None	None	None
Distance to Obstacle	-	-	-	-
Height Above Inlet	-	-	-	-
Distance to Walls, etc.	-	-	-	-
Distance to Dripline	-	-	-	-
Dominant Influence	Vehicular	Vehicular	Vehicular	Vehicular
Residence Time (sec)				

## Site Survey Report (Cont.)

Monitor Type	PM2.5	PM2.5	Xontech 920	Xontech 920
Manufacturer/Model	Sierra Andersen RAAS PM2.5	Sierra Anderson RAAS PM2.5	Xontech 920	Xontech 920
Serial Number	00347	00305	0015466	156
POC	2	1	1	1
Data For Record?	Yes	Yes	Yes	Yes
Purpose				
Objective				
Scale				
Height Above Ground	12.0	12.0	12.0	12.0
Height Above Platform	1.9	1.9	1.5	1.5
Sampler Spacing	2.0	2.0		
Current Manual Available?	Yes	Yes	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes	Yes	Yes
In-line Filter Change Date				
Cal. Gas Cert. Date				
Calibration Current?	Yes	Yes	Yes	Yes
Calibration Date	2006-10-19	2006-10-19	2006-06-28	2006-06-22
Cal. Equipment Cert. Date	Not Available	Not Available	Not Available	Not Available
Obstacle Description	None	None	None	None
Distance to Obstacle	-	-	-	-
Height Above Inlet	-	-	-	-
Distance to Walls, etc.	-	-	-	-
Distance to Dripline	-	-	-	-
Dominant Influence	Vehicular	Vehicular	Vehicular	Vehicular
Residence Time (sec)				

## Site Survey Report (Cont.)

Monitor Type	Xontech 920
Manufacturer/Model	Xontech 924
Serial Number	20021009
POC	1
Data For Record?	Yes
Purpose	
Objective	
Scale	
Height Above Ground	12.0
Height Above Platform	1.5
Sampler Spacing	
Current Manual Available?	Yes
Instrument Log Up-to-date?	Yes
In-line Filter Change Date	
Cal. Gas Cert. Date	
Calibration Current?	Yes
Calibration Date	Not Available
Cal. Equipment Cert. Date	Not Available
Obstacle Description	None
Distance to Obstacle	-
Height Above Inlet	-
Distance to Walls, etc.	-
Distance to Dripline	-
Dominant Influence	Vehicular
Residence Time (sec)	