



**Quality Assurance  
Site Information for Compton - MATES**

*This page updated February 10, 2005*



AIRS Number	ARB Number	Site Start Date	Reporting Agency and Agency Code
None	70112	1/1/04	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
700 North Bullis Rd, Compton CA 90221	Los Angeles	South Coast	33° 54' 01"	118° 16' 07"	25



# Site Survey Report

## Siting Information

Site Name: Compton - MATES	Audit Date: 2004-11-16	ARB Number: 70112	AIRS Number: None
Address: 700 North Bullis Rd Compton, CA 90221	Latitude: 33° 54' 01"	Longitude: 118° 16' 07"	Elevation (m): 25
	Auditors: Fred Burriell	Site Technician: Sumner Wilson	Site Phone:
Operating Agency: South Coast AQMD		Site Report: Yes	Site Photos: Yes

## General Siting Conditions

<b>Station Temperature</b>  Controlled: Yes  Recorded: Yes  Inside Temp: 25 Degrees Celsius	<b>Traffic</b>  Description: Residential  Distance: 15 meters  Count (Veh/Day): 2000	<b>Topography</b>  Site: level  Region: level	Predominant Wind Direction: West
		<b>QA Manual</b>  Approved: Yes  Agency: South Coast AQMD	Arc Air Flow (Deg): 360 Degrees
			Probe Clean: No
<b>Meteorology</b>  Located With Instruments: Yes  Shadowing: No  Boom Orientation (Deg): 345  Temp(Motor/Natural): Natural	<b>Non-vehicular Local Sources</b>  Description:  Distance: 0 meters  Direction:	Urbanization: Suburban	Manifold Clean: No
		Ground Cover: roof	Cleaning Schedule:
			Autocalibrator Type:
			Logbook Up To Date: Yes

## Action Items

## Comments

## Site Survey Report (Cont.)

Monitor Type	Xontech 920
Manufacturer/Model	Xontech 920
Serial Number	156
POC	
Data For Record?	No
Purpose	
Objective	
Scale	
Height Above Ground	4.6
Height Above Platform	1.6
Sampler Spacing	
Current Manual Available?	Yes
Instrument Log Up-to-date?	Yes
In-line Filter Change Date	
Cal. Gas Cert. Date	
Calibration Current?	Yes
Calibration Date	2004-11-16
Cal. Equipment Cert. Date	Not Available
Obstacle Description	None
Distance to Obstacle	-
Height Above Inlet	-
Distance to Walls, etc.	-
Distance to Dripline	-
Dominant Influence	Vehicular
Residence Time (sec)	