

Quality Assurance Site Information for Huntington Park - MATES

This page updated February 10, 2005



AIRS Number	ARB Number	Site Start Date	Reporting Agency and Agency Code
None	70114	1/1/04	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
Fire Station 164, 6301 S. Santa Fe Av, Huntington Park CA 90255	Los Angeles	South Coast	33° 59' 01"	118° 13' 49"	50



Site Survey Report

Siting Information

Site Name: Huntington Park - MATES	Audit Date: 2004-11-16	ARB Number: 70114	AIRS Number: None
Address: Fire Station 164, 6301 S. Santa Fe Av	Latitude: 33° 59' 01"	Longitude: 118° 13' 49"	Elevation (m): 50
Huntington Park, CA 90255	Auditors: Fred Burriell	Site Technician: Sumner Wilson	Site Phone:
Operating Agency: South Coast AQMD		Site Report: Yes	Site Photos: Yes

General Siting Conditions

Station Temperature Traffic		Topography	Predominant Wind Direction: West	
Controlled: No	Description: Arterial	Site: level	Arc Air Flow (Deg): 360 Degrees	
Recorded: No	Distance: 10 meters	Region: level	Probe Clean: No	
Inside Temp: 0 Degrees Celsius	Count (Veh/Day): 10000	QA Manual	Manifold Clean: No	
Meteorology	Non-vehicular Local Sources	Approved: Yes	Cleaning Schedule:	
Located With Instruments: Yes	Description:	Agency: South Coast AQMD	Autocalibrator Type:	
Shadowing: No	Distance: 0 meters	Urbanization: Suburban	Site Survey Complete: Yes	
Boom Orientation (Deg): 345			Logbook Up To Date: Yes	
Temp(Motor/Natural): Natural	Birodion.	Ground Cover: roof	Logsoon op 10 bate. 103	

Action Items

	Con	nments	

Site Survey Report (Cont.)

Monitor Type	Xontech 920
Manufacturer/Model	Xontech 920
Serial Number	0015479
POC	
Data For Record?	No
Purpose	
Objective	
Scale	
Height Above Ground	4.8
Height Above Platform	1.8
Sampler Spacing	
Current Manual Available?	Yes
Instrument Log Up-to-date?	Yes
In-line Filter Change Date	
Cal. Gas Cert. Date	
Calibration Current?	Yes
Calibration Date	2004-10-07
Cal. Equipment Cert. Date	Not Available
Obstacle Description	None
Distance to Obstacle	-
Height Above Inlet	-
Distance to Walls, etc.	-
Distance to Dripline	-
Dominant Influence	Vehicular
Residence Time (sec)	