


Quality Assurance Site Information for Indio-Jackson Street

This page updated February 10, 2005



AIRS Number	ARB Number	Site Start Date	Reporting Agency and Agency Code
060652002	33157	1/1/83	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
46990 Jackson St, Indio CA 92201	Riverside	Salton Sea	33° 42' 30"	116° 12' 57"	-4



Site Survey Report

Siting Information

Site Name: Indio-Jackson Street	Audit Date: 2006-08-02	ARB Number: 33157	AIRS Number: 060652002
Address: 46990 Jackson St Indio, CA 92201	Latitude: 33° 42' 30"	Longitude: 116° 12' 57"	Elevation (m): -4
	Auditors: Don Fitzell Alvin Danque	Site Technician: Rene Bermudez	Site Phone:
Operating Agency: South Coast AQMD		Site Report: Yes	Site Photos: Yes

General Siting Conditions

Station Temperature Controlled: Yes Recorded: Yes Inside Temp: 25 Degrees Celsius	Traffic Description: Arterial Distance: 100 meters Count (Veh/Day): 10000	Topography Site: Level	Predominant Wind Direction: Southeast
		Region: Level	Arc Air Flow (Deg): 360 Degrees
		QA Manual Approved: Yes Agency: South Coast AQMD Urbanization: Suburban Ground Cover: Asphalt	Probe Clean: Yes
Manifold Clean: Yes			
Cleaning Schedule: Annually			
Autocalibrator Type: Environics 100			
Meteorology Located With Instruments: Yes Shadowing: Yes Boom Orientation (Deg): 345 Temp(Motor/Natural): Natural	Non-vehicular Local Sources Description: None Distance: N/A Direction: N/A	Site Survey Complete: Yes	
		Logbook Up To Date: Yes	

Action Items

Comments

Site Survey Report (Cont.)

Monitor Type	Ozone	PM10-SSI	PM10-SSI	PM10-SSI
Manufacturer/Model	API/Teledyne 400	Tisch TE300-310	Tisch TE300-310	Tisch TE300-310
Serial Number	526-S	15397 A-1	1593-B	1559 A-2
POC	1	2	2	2
Data For Record?	Yes	No	No	No
Purpose				
Objective				
Scale				
Height Above Ground	6.0	2.0	2.0	2.0
Height Above Platform	3.0	1.5	1.5	1.5
Sampler Spacing		1.5	1.5	1.5
Current Manual Available?	Yes	Yes	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes	Yes	Yes
In-line Filter Change Date	2006-07-27			
Cal. Gas Cert. Date				
Calibration Current?	Yes	Yes	Yes	Yes
Calibration Date	2005-12-07	2006-06-14	2006-06-14	2006-06-14
Cal. Equipment Cert. Date	Not Available	2006-01-06	2006-01-06	2006-01-06
Obstacle Description	None	None	None	None
Distance to Obstacle	-	-	-	-
Height Above Inlet	-	-	-	-
Distance to Walls, etc.	-	-	-	-
Distance to Dripline	-	-	-	-
Dominant Influence	Vehicular			
Residence Time (sec)	9.4			

Site Survey Report (Cont.)

Monitor Type	PM2.5	PM2.5
Manufacturer/Model	Andersen 300	Andersen 300
Serial Number	0099	442
POC	1	2
Data For Record?	Yes	Yes
Purpose		
Objective		
Scale		
Height Above Ground	2.9	2.9
Height Above Platform	1.9	1.9
Sampler Spacing	3.0	3.0
Current Manual Available?	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes
In-line Filter Change Date		
Cal. Gas Cert. Date		
Calibration Current?	Yes	Yes
Calibration Date	2006-09-13	2006-09-13
Cal. Equipment Cert. Date	Not Available	Not Available
Obstacle Description	None	None
Distance to Obstacle	-	-
Height Above Inlet	-	-
Distance to Walls, etc.	-	-
Distance to Dripline	-	-
Dominant Influence	Vehicular	Vehicular
Residence Time (sec)		