


Quality Assurance Site Information for Los Angeles - LAX

This page updated February 10, 2005



AIRS Number	ARB Number	Site Start Date	Reporting Agency and Agency Code
060375005	70111	4/12/04	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
7201 W Westchester Parkway, Los Angeles CA 90045	Los Angeles	South Coast	33° 57' 15"	118° 25' 49"	10

Site Survey Report

Siting Information

Site Name: Los Angeles - LAX	Audit Date: 2006-07-27	ARB Number: 70111	AIRS Number: 060375005
Address: 7201 W Westchester Parkway Los Angeles, CA 90045	Latitude: 33° 57' 15"	Longitude: 118° 25' 49"	Elevation (m): 10
	Auditors: Fred Burriell Charles Pearson	Site Technician: Tom Mac	Site Phone:
Operating Agency: South Coast AQMD		Site Report: Yes	Site Photos: Yes

General Siting Conditions

Station Temperature Controlled: Yes Recorded: Yes Inside Temp: 25 Degrees Celsius	Traffic Description: Residential Distance: 1000 meters Count (Veh/Day): 1000	Topography Site: Level Region: Level	Predominant Wind Direction: Northeast
			Arc Air Flow (Deg): 360 Degrees
		Meteorology Located With Instruments: Yes Shadowing: No Boom Orientation (Deg): 345 Temp(Motor/Natural): Motor	Non-vehicular Local Sources Description: Distance: 0 meters Direction:
Manifold Clean: Yes			
Cleaning Schedule: As Needed			
Autocalibrator Type: Environics 100			
Site Survey Complete: Yes			
	Ground Cover: Asphalt	Logbook Up To Date: Yes	

Action Items

Comments

Site Survey Report (Cont.)

Monitor Type	Carbon Monoxide	Sulfur Dioxide	Nitrogen Dioxide	Ozone
Manufacturer/Model	Horiba APMA-360	TECO 43A, 43B, 43C	API 200A	API/Teledyne 400
Serial Number	0015557	0016627	E000209	524-S
POC	1	1	1	1
Data For Record?	Yes	Yes	Yes	Yes
Purpose				
Objective				
Scale				
Height Above Ground	4.5	4.5	4.5	4.5
Height Above Platform	1.5	1.5	1.5	1.5
Sampler Spacing				
Current Manual Available?	Yes	Yes	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes	Yes	Yes
In-line Filter Change Date	2006-07-25	2006-07-25	2006-07-25	2006-07-25
Cal. Gas Cert. Date	Not Available	Not Available	Not Available	
Calibration Current?	Yes	Yes	Yes	Yes
Calibration Date	2005-12-20	2006-01-03	2006-03-31	2006-02-21
Cal. Equipment Cert. Date	Not Available	Not Available	Not Available	Not Available
Obstacle Description	None	None	None	None
Distance to Obstacle	-	-	-	-
Height Above Inlet	-	-	-	-
Distance to Walls, etc.	-	-	-	-
Distance to Dripline	-	-	-	-
Dominant Influence	Vehicular	Vehicular	Vehicular	Vehicular
Residence Time (sec)	7.7	9.2	8.9	8.5

Site Survey Report (Cont.)

Monitor Type	PM10-SSI	TSP
Manufacturer/Model	Andersen 1200	GMW 1200
Serial Number	54394	54394
POC	1	1
Data For Record?	No	No
Purpose		
Objective		
Scale		
Height Above Ground	2.0	2.0
Height Above Platform	1.5	1.5
Sampler Spacing		
Current Manual Available?	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes
In-line Filter Change Date		
Cal. Gas Cert. Date		
Calibration Current?	Yes	Yes
Calibration Date	2006-02-22	2006-02-22
Cal. Equipment Cert. Date	Not Available	Not Available
Obstacle Description	None	None
Distance to Obstacle	-	-
Height Above Inlet	-	-
Distance to Walls, etc.	-	-
Distance to Dripline	-	-
Dominant Influence		Vehicular
Residence Time (sec)		