



Quality Assurance Site Information for North Long Beach (Long Beach)

This page updated February 10, 2005



AIRS Number	ARB Number	Site Start Date	Reporting Agency and Agency Code
060374002	70072	5/1/69	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
3648 N. Long Beach Bl, Long Beach CA 90807	Los Angeles	South Coast	33° 49' 25"	118° 11' 19"	6



Site Survey Report

Siting Information

Site Name: North Long Beach (Long Beach)	Audit Date: 2005-11-29	ARB Number: 70072	AIRS Number: 060374002
Address: 3648 N. Long Beach Bl Long Beach, CA 90807	Latitude: 33° 49' 25"	Longitude: 118° 11' 19"	Elevation (m): 6
	Auditors: Fred Burriell	Site Technician: Sumner Wilson	Site Phone:
Operating Agency: South Coast AQMD		Site Report: Yes	Site Photos: Yes

General Siting Conditions

Station Temperature Controlled: Yes Recorded: Yes Inside Temp: 26 Degrees Celsius	Traffic Description: Commercial Distance: 2 meters Count (Veh/Day): 10000	Topography Site: Level Region: Valley QA Manual Approved: Yes Agency: South Coast AQMD Urbanization: City Center Ground Cover: roof	Predominant Wind Direction: Southwest
			Arc Air Flow (Deg): 360 Degrees
			Probe Clean: Yes
Meteorology Located With Instruments: Yes Shadowing: Yes Boom Orientation (Deg): 345 Temp(Motor/Natural): Natural	Non-vehicular Local Sources Description: None Distance: N/A Direction: N/A		Manifold Clean: Yes
			Cleaning Schedule: Annually
			Autocalibrator Type: Environics 100
			Site Survey Complete: Yes
			Logbook Up To Date: Yes

Action Items

Comments

Site Survey Report (Cont.)

Monitor Type	PM2.5	Xontech 920	Xontech 920
Manufacturer/Model	Andersen RAAS2.5-300	Xontech 920	Xontech 920
Serial Number	E000003	0015477	08898
POC	1	1	7
Data For Record?	Yes	Yes	Yes
Purpose			
Objective			
Scale			
Height Above Ground	2.9	2.4	2.4
Height Above Platform	1.9	1.4	1.4
Sampler Spacing			
Current Manual Available?	Yes	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes	Yes
In-line Filter Change Date			
Cal. Gas Cert. Date			
Calibration Current?	Yes	Yes	Yes
Calibration Date	2005-11-08	Not Available	2005-11-02
Cal. Equipment Cert. Date	Not Available	Not Available	Not Available
Obstacle Description	None	None	None
Distance to Obstacle	-	-	-
Height Above Inlet	-	-	-
Distance to Walls, etc.	-	-	-
Distance to Dripline	-	-	-
Dominant Influence	Vehicular	Vehicular	Vehicular
Residence Time (sec)			