

AIRS Number	ARB Number	Site Start Date	Reporting Agency and Agency Code
060371301	70084	1/1/78	South Coast AQMD (061)

Josephine St

E El Segundo Blvd Copyright @ 1989-1999 Werosoft Corporation and/or its suppliers lattingits reserved, http://www.expediamaps.com

E Weber Ave

Compton

6 School

Codar Ave

Magoolia Ave

Agnes Ave

Welnut Ave

Contiand St

Emuood Ave

Virginia St


Louise St.

E 120th St

E 124th St

E 126th St

Site Address	County	Air Basin	Latitude	Longitude	Elevation
11220 Long Beach Bl, Lynwood CA 90262	Los Angeles	South Coast	33° 55' 44"	118° 12' 39"	40



Site Survey Report

Siting Information

Site Name: Lynwood	Audit Date: 2006-12-07	ARB Number: 70084	AIRS Number: 060371301
Address: 11220 Long Beach Bl Lynwood, CA 90262	Latitude: 33° 55' 44"	Longitude: 118° 12' 39"	Elevation (m): 40
	Auditors: Dustin Goto Long Liu	Site Technician: Rene Bermudez	Site Phone:
Operating Agency: South Coast AQMD		Site Report: Yes	Site Photos: No

General Siting Conditions

Station Temperature	Traffic	Topography	Predominant Wind Direction: South
Controlled: Yes	Description: Commercial	Site: Level	Arc Air Flow (Deg): 360 Degrees
Recorded: Yes	Distance: 15 meters	Region: Level	Probe Clean: N/A
Inside Temp: 0 Degrees Celsius	Count (Veh/Day): 40000	QA Manual	Manifold Clean: N/A
Meteorology	Non-vehicular Local Sources	Approved: Yes	Cleaning Schedule: N/A
Located With Instruments: Yes	Description: Flues	Agency: South Coast AQMD	Autocalibrator Type: Environics 100
Shadowing: No	Distance: 25 meters	Urbanization: City Center	Site Survey Complete: Yes
Boom Orientation (Deg): 345	Direction: South	Ground Cover: Roof	Logbook Up To Date: Yes
Temp(Motor/Natural): Natural			

Action Items

Comments

Site Survey Report (Cont.)

Monitor Type	PM2.5
Manufacturer/Model	Sierra Anderson RAAS PM2.5
Serial Number	00369
POC	1
Data For Record?	Yes
Purpose	
Objective	
Scale	
Height Above Ground	4.5
Height Above Platform	1.9
Sampler Spacing	
Current Manual Available?	Yes
Instrument Log Up-to-date?	Yes
In-line Filter Change Date	
Cal. Gas Cert. Date	
Calibration Current?	Yes
Calibration Date	2006-12-06
Cal. Equipment Cert. Date	Not Available
Obstacle Description	None
Distance to Obstacle	-
Height Above Inlet	-
Distance to Walls, etc.	-
Distance to Dripline	-
Dominant Influence	Vehicular
Residence Time (sec)	