



**Quality Assurance  
Site Information for Pico Rivera #2**

*This page updated February 10, 2005*



AIRS Number	ARB Number	Site Start Date	Reporting Agency and Agency Code
060371602	70185	9/12/05	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
4144 San Gabriel River Pkwy., Pico Rivera CA	Los Angeles	South Coast	34° 00' 37"	118° 4' 7"	75



# Site Survey Report

## Siting Information

Site Name: Pico Rivera #2	Audit Date: 2006-08-01	ARB Number: 70185	AIRS Number: 060371602
Address: 4144 San Gabriel River Pkwy. Pico Rivera, CA	Latitude: 34° 00' 37"	Longitude: 118° 4' 7"	Elevation (m): 75
	Auditors: Don Fitzell Alvin Danque	Site Technician: Tom Mac	Site Phone:
Operating Agency: South Coast AQMD		Site Report: Yes	Site Photos: Yes

## General Siting Conditions

<b>Station Temperature</b>  Controlled: Yes  Recorded: Yes  Inside Temp: 25 Degrees Celsius	<b>Traffic</b>  Description: Freeway  Distance: 1000 meters  Count (Veh/Day): 250000	<b>Topography</b>  Site: Level  Region: Valley	Predominant Wind Direction: Southeast
			Arc Air Flow (Deg): 360 Degrees
		<b>Meteorology</b>  Located With Instruments: Yes  Shadowing: Yes  Boom Orientation (Deg): Temp(Motor/Natural):	<b>Non-vehicular Local Sources</b>  Description: None  Distance: N/A  Direction: N/A
Urbanization: City Center  Ground Cover: Asphalt	Manifold Clean: Yes		
	Cleaning Schedule: as needed		
	Autocalibrator Type: Env. 100		
Site Survey Complete: Yes			
Logbook Up To Date: Yes			

## Action Items

--

## Comments

--

## Site Survey Report (Cont.)

Monitor Type	Carbon Monoxide	Nitrogen Dioxide	Ozone	PM10-SSI
Manufacturer/Model	API 300	API 200A	API/Teledyne 400	Tisch 1200
Serial Number	20017419	E000215	535-S	50461
POC	1	2	1	
Data For Record?	Yes	Yes	Yes	No
Purpose				
Objective				
Scale				
Height Above Ground	3.0	3.0	3.0	2.0
Height Above Platform	1.5	1.5	1.5	1.5
Sampler Spacing				1.0
Current Manual Available?	Yes	Yes	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes	Yes	Yes
In-line Filter Change Date	2006-07-25	2006-07-25	2006-07-25	
Cal. Gas Cert. Date	2002-11-22	2002-11-22		
Calibration Current?	Yes	Yes	Yes	Yes
Calibration Date	2006-06-22	2006-06-30	2006-06-21	2006-05-26
Cal. Equipment Cert. Date	Not Available	Not Available	Not Available	Not Available
Obstacle Description	None	None	None	None
Distance to Obstacle	-	-	-	-
Height Above Inlet	-	-	-	-
Distance to Walls, etc.	-	-	-	-
Distance to Dripline	-	-	-	-
Dominant Influence	Vehicular	Vehicular	Vehicular	Vehicular
Residence Time (sec)	8.9	10.6	9.2	

## Site Survey Report (Cont.)

Monitor Type	TSP	PM2.5	Barometric Pressure
Manufacturer/Model	Tisch TE 300-310	Sierra Andersen RAAS PM2.5	MET ONE 090C-26/31
Serial Number	60283	00511	unk.
POC	1	1	
Data For Record?	Yes	Yes	No
Purpose			
Objective			
Scale			
Height Above Ground	2.0	2.5	
Height Above Platform	1.5	1.5	
Sampler Spacing	1.0		
Current Manual Available?	Yes	Yes	No
Instrument Log Up-to-date?	Yes	Yes	No
In-line Filter Change Date			
Cal. Gas Cert. Date			
Calibration Current?	Yes	Yes	No
Calibration Date	2006-01-25	2006-11-29	Not Available
Cal. Equipment Cert. Date	Not Available	Not Available	Not Available
Obstacle Description	None	None	None
Distance to Obstacle	-	-	-
Height Above Inlet	-	-	-
Distance to Walls, etc.	-	-	
Distance to Dripline	-	-	
Dominant Influence	Vehicular	Vehicular	
Residence Time (sec)			