



Quality Assurance Site Information for Pomona

This page updated February 10, 2005



AIRS Number	ARB Number	Site Start Date	Reporting Agency and Agency Code
060371701	70075	1/1/78	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
924 N. Garey Av, Pomona CA 91767	Los Angeles	South Coast	34° 4' 0"	117° 45' 0"	270



Site Survey Report

Siting Information

Site Name: Pomona	Audit Date: 2005-09-01	ARB Number: 70075	AIRS Number: 060371701
Address: 924 N. Garey Av Pomona, CA 91767	Latitude: 34° 4' 0"	Longitude: 117° 45' 0"	Elevation (m): 270
	Auditors: Alvin Danque Long Liu	Site Technician: Eric Trinn	Site Phone:
Operating Agency: South Coast AQMD		Site Report: Yes	Site Photos: Yes

General Siting Conditions

Station Temperature Controlled: Yes Recorded: Yes Inside Temp: 25 Degrees Celsius	Traffic Description: Arterial Distance: 50 meters Count (Veh/Day): 10000	Topography Site: level Region: level	Predominant Wind Direction: Southwest
			Arc Air Flow (Deg): 360 Degrees
		Meteorology Located With Instruments: Yes Shadowing: No Boom Orientation (Deg): N/A Temp(Motor/Natural):	Non-vehicular Local Sources Description: NONE Distance: 0 meters Direction:
Approved: Yes	Manifold Clean: Yes		
Agency: South Coast AQMD	Cleaning Schedule: As needed		
Urbanization: Suburban	Autocalibrator Type: Environics 100		
Ground Cover: roof	Site Survey Complete: Yes		
	Logbook Up To Date: Yes		

Action Items

Comments

Site Survey Report (Cont.)

Monitor Type	Carbon Monoxide	Nitrogen Dioxide	Ozone
Manufacturer/Model	Horiba AQM-10,11,12	API 200	API/Teledyne 400
Serial Number	0015559	E000214	519-S
POC	1	2	1
Data For Record?	Yes	Yes	Yes
Purpose			
Objective			
Scale			
Height Above Ground	8.5	8.5	8.5
Height Above Platform	2.5	2.5	2.5
Sampler Spacing			
Current Manual Available?	Yes	Yes	Yes
Instrument Log Up-to-date?	Yes	Yes	Yes
In-line Filter Change Date	2005-08-17	2005-08-17	2005-08-17
Cal. Gas Cert. Date	2002-11-15	2002-11-15	
Calibration Current?	Yes	Yes	Yes
Calibration Date	2005-04-05	2005-06-16	2005-05-09
Cal. Equipment Cert. Date	2005-03-15	2005-03-15	2005-03-15
Obstacle Description	None	None	None
Distance to Obstacle	-	-	-
Height Above Inlet	-	-	-
Distance to Walls, etc.	-	-	-
Distance to Dripline	-	-	-
Dominant Influence	Vehicular	Vehicular	Commercial
Residence Time (sec)	9.2	10.5	9.9