



## Quality Assurance Site Information for Riverside-Magnolia

*This page updated February 10, 2005*



AIRS Number	ARB Number	Site Start Date	Reporting Agency and Agency Code
060651003	33146	6/1/76	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
7002 Magnolia Av, Riverside CA 92506	Riverside	South Coast	33° 56' 47"	117° 24' 4"	258



# Site Survey Report

## Siting Information

Site Name: Riverside-Magnolia	Audit Date: 2006-12-12	ARB Number: 33146	AIRS Number: 060651003
Address: 7002 Magnolia Av Riverside, CA 92506	Latitude: 33° 56' 47"	Longitude: 117° 24' 4"	Elevation (m): 258
	Auditors: Don Fitzell	Site Technician: Mike Hamdan	Site Phone:
Operating Agency: South Coast AQMD		Site Report: Yes	Site Photos: Yes

## General Siting Conditions

<b>Station Temperature</b>  Controlled: No  Recorded: No  Inside Temp: 0 Degrees Celsius	<b>Traffic</b>  Description: City  Distance: 10 meters  Count (Veh/Day): 40000	<b>Topography</b>  Site: Level  Region: Level  <b>QA Manual</b>  Approved: Yes  Agency: South Coast AQMD  Urbanization: City Center  Ground Cover: Asphalt	Predominant Wind Direction: West
			Arc Air Flow (Deg): 360 Degrees
			Probe Clean: N/A
<b>Meteorology</b>  Located With Instruments: No  Shadowing: No  Boom Orientation (Deg): Temp(Motor/Natural):	<b>Non-vehicular Local Sources</b>  Description: None  Distance: N/A  Direction: N/A		Manifold Clean: N/A
			Cleaning Schedule: N/A
			Autocalibrator Type: Environics 100
			Site Survey Complete: Yes
			Logbook Up To Date: Yes

## Action Items

## Comments

## Site Survey Report (Cont.)

Monitor Type	PM2.5
Manufacturer/Model	Andersen 300
Serial Number	00348
POC	1
Data For Record?	Yes
Purpose	
Objective	
Scale	
Height Above Ground	5.0
Height Above Platform	1.9
Sampler Spacing	
Current Manual Available?	Yes
Instrument Log Up-to-date?	Yes
In-line Filter Change Date	
Cal. Gas Cert. Date	
Calibration Current?	Yes
Calibration Date	2006-09-12
Cal. Equipment Cert. Date	Not Available
Obstacle Description	None
Distance to Obstacle	-
Height Above Inlet	-
Distance to Walls, etc.	-
Distance to Dripline	-
Dominant Influence	Vehicular
Residence Time (sec)	