



Quality Assurance Site Information for Upland

This page updated February 10, 2005



AIRS Number	ARB Number	Site Start Date	Reporting Agency and Agency Code
060711004	36175	5/1/92	South Coast AQMD (061)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
1350 San Bernardino Rd, Upland CA 91786	San Bernardino	South Coast	34° 6' 14"	117° 37' 45"	379



Site Survey Report

Siting Information

Site Name: Upland	Audit Date: 2003-07-17	ARB Number: 36175	AIRS Number: 060711004
Address: 1350 San Bernardino Rd Upland, CA 91786	Latitude: 34° 6' 14"	Longitude: 117° 37' 45"	Elevation (m): 379
	Auditors: Don Fitzell Josh Berghouse	Site Technician: Mike Hamdan	Site Phone:
Operating Agency: South Coast AQMD		Site Report: Yes	Site Photos: Yes

General Siting Conditions

Station Temperature Controlled: Yes Recorded: Yes Inside Temp: 25 Degrees Celsius	Traffic Description: Residential Distance: 10 meters Count (Veh/Day): 20	Topography Site: Level Region: valley	Predominant Wind Direction: North Arc Air Flow (Deg): 360 Degrees Probe Clean: Yes
		QA Manual Approved: Yes Agency: South Coast AQMD Urbanization: Suburban Ground Cover: gravel	Manifold Clean: Yes Cleaning Schedule: Annually Autocalibrator Type: Environics 100
			Site Survey Complete: Yes Logbook Up To Date: Yes
Meteorology Located With Instruments: Yes Shadowing: Yes Boom Orientation (Deg): Temp(Motor/Natural): Natural	Non-vehicular Local Sources Description: None Distance: N/A Direction: N/A		

Action Items

Comments

Site Survey Report (Cont.)

Monitor Type	Ozone	TSP	TEOM	Barometric Pressure
Manufacturer/Model	API/Teledyne 400	GMW	R&P 1400a	MET ONE 091
Serial Number	012667	5050	20005938	U1997
POC	2	1		
Data For Record?	Yes	Yes	No	No
Purpose				
Objective				
Scale				
Height Above Ground	4.8	2.0	4.9	2.0
Height Above Platform	1.4	1.0	1.5	
Sampler Spacing				
Current Manual Available?	Yes	Yes	Yes	No
Instrument Log Up-to-date?	Yes	Yes	Yes	No
In-line Filter Change Date	2003-06-24			
Cal. Gas Cert. Date				
Calibration Current?	Yes	Yes	Yes	Yes
Calibration Date	2003-01-31	2003-06-27	2003-03-21	2002-09-19
Cal. Equipment Cert. Date	Not Available	2002-12-13	Not Available	Not Available
Obstacle Description	None	None	None	None
Distance to Obstacle	-	-	-	-
Height Above Inlet	-	-	-	-
Distance to Walls, etc.	-	-	-	
Distance to Dripline	-	-	-	
Dominant Influence	Vehicular	Vehicular	Vehicular	
Residence Time (sec)	7.5			