

**DEPARTMENT OF HEALTH AND HUMAN SERVICES****Food and Drug Administration**

[Docket No. 2004N-0515]

**Agency Information Collection Activities; Proposed Collection; Comment Request; Medical Device Labeling Regulations****AGENCY:** Food and Drug Administration, HHS.**ACTION:** Notice.

**SUMMARY:** The Food and Drug Administration (FDA) is announcing an opportunity for public comment on the proposed collection of certain information by the agency. Under the Paperwork Reduction Act of 1995 (the PRA), Federal agencies are required to publish notice in the **Federal Register** concerning each proposed collection of information, including each proposed extension of an existing collection of information, and to allow 60 days for public comment in response to the notice. This notice solicits comments on medical device labeling regulations.

**DATES:** Submit written or electronic comments on the collection of information by February 14, 2005.

**ADDRESSES:** Submit electronic comments on the collection of information to: <http://www.fda.gov/dockets/ecomments>. Submit written comments on the collection of information to the Division of Dockets Management (HFA-305), Food and Drug Administration, 5630 Fishers Lane, rm. 1061, Rockville, MD 20852. All comments should be identified with the docket number found in brackets in the heading of this document.

**FOR FURTHER INFORMATION CONTACT:** Peggy Robbins, Office of Management Programs (HFA-250), Food and Drug Administration, 5600 Fishers Lane Rockville, MD 20857, 301-827-1223.

**SUPPLEMENTARY INFORMATION:** Under the PRA (44 U.S.C. 3501-3520), Federal agencies must obtain approval from the Office of Management and Budget (OMB) for each collection of information they conduct or sponsor. "Collection of information" is defined in 44 U.S.C. 3502(3) and 5 CFR 1320.3(c) and includes agency requests or requirements that members of the public submit reports, keep records, or provide information to a third party. Section 3506(c)(2)(A) of the PRA (44 U.S.C. 3506(c)(2)(A)) requires Federal agencies to provide a 60-day notice in the **Federal Register** concerning each proposed collection of information, including each proposed extension of an

existing collection of information, before submitting the collection to OMB for approval. To comply with this requirement, FDA is publishing notice of the proposed collection of information set forth in this document.

With respect to the following collection of information, FDA invites comments on these topics: (1) Whether the proposed collection of information is necessary for the proper performance of FDA's functions, including whether the information will have practical utility; (2) the accuracy of FDA's estimate of the burden of the proposed collection of information, including the validity of the methodology and assumptions used; (3) ways to enhance the quality, utility, and clarity of the information to be collected; and (4) ways to minimize the burden of the collection of information on respondents, including through the use of automated collection techniques, when appropriate, and other forms of information technology.

**Medical Device Labeling Regulations—21 CFR Parts 800, 801, and 809 (OMB Control No. 0910-0485)—Extension**

Section 502 of the Federal Food, Drug, and Cosmetic Act (the act) (21 U.S.C. 352), among other things, establishes requirements for the label or labeling of a medical device so that it is not misbranded and subject to regulatory action. Certain provisions of section 502 of the act require that manufacturers, importers, and distributors of medical devices disclose information about themselves or their devices on the labels or labeling of the devices. Section 502(b) of the act requires that, if the device is in a package, the label must contain the name and place of business of the manufacturer, packer, or distributor and an accurate statement of the quantity of the contents. Section 502(f) of the act provides that the labeling of a device must contain adequate directions for use. FDA may grant an exemption from the adequate directions for use requirement, if FDA determines that adequate directions for use are not necessary for the protection of the public health.

FDA regulations in parts 800, 801, and 809 (21 CFR parts 800, 801, and 809) require manufacturers, importers, and distributors of medical devices to disclose to health professionals and consumers specific information about themselves or their devices on the label or labeling of their devices. FDA issued these regulations under the authority of sections 201, 301, 502, and 701 of the act (21 U.S.C. 321, 331, 352, and 371). Most of the regulations in parts 800, 801, and 809 derive from the

requirements of section 502 of the act, which provides, in part, that a device shall be misbranded if, among other things, its label or labeling fails to bear certain required information concerning the device, is false or misleading in any particular way, or fails to contain adequate directions for use.

Section 800.12 requires that packages of contact lens cleaning solutions include a tamper-resistant feature to prevent malicious adulteration. Sections 800.10(a)(3) and 800.12(c) require that the label of contact lens cleaning solutions contain a prominent statement alerting consumers to the tamper-resistant feature.

Section 800.10(b)(2) requires that the labeling of liquid ophthalmic preparations packed in multiple-dose containers include information as to duration of use and necessary warnings to afford adequate protection from contamination during use.

Section 801.1 requires that the label of a device in package form contain the name and place of business of the manufacturer, packer, or distributor.

Section 801.5 requires that the labeling of devices include directions under which the layman can use a device safely and for the purposes for which it is intended. Section 801.4 defines "intended use". Where necessary, the labeling should include: (1) Statements of all conditions, purposes, or uses for which the device is intended, unless the device is a prescription device subject to the requirements of § 801.109; (2) quantity of dose; (3) frequency of administration or application; (4) duration of administration or application; (5) time of administration, e.g. in relation to meals, onset of symptoms, etc.; (6) route of method or application; and (7) preparation for use.

Section 801.61 requires that the principal display panel of an over-the-counter (OTC) device in package form must include a statement of the identity of the device. The statement of the identity of the device must include the common name of the device followed by an accurate statement of the principal intended actions of the device.

Section 801.62 requires that the label of an OTC device in package form must include a declaration of the net quantity of contents. The label must express the net quantity in terms of weight, measure, numerical count, or a combination of numerical count and weight, measure, or size.

Section 801.109 establishes labeling requirements for prescription devices. A prescription device is defined as a device which, because of its potential for harmful effect, the method of its use,

or the collateral measures necessary to its use, is not safe except under the supervision of a practitioner licensed by law to use the device and, therefore, for which adequate directions for use by a layperson cannot be developed.

The label of the device must include: (1) The statement "Caution: Federal law restricts this device to sale by or on the order of a \_\_\_\_\_". The blank is to be filled in by a term such as "physician," "dentist," or other appropriate term; and (2) the method of its application or use.

Labeling must include information for use, including indications, effects, routes, methods, frequency and duration of administration, and any relevant hazards, contraindications, side effects, and precautions under which practitioners licensed by law to administer the device can use the device safely and for the purpose for which it is intended, including all purposes for which it is advertised or represented. Information may be omitted from the dispensing package if, but only if, the article is a device for which directions, hazards, warnings, and other information are commonly known to practitioners licensed by law to use the device.

Section 801.110 establishes a labeling requirement for a prescription device delivered to the ultimate purchaser or user upon the prescription of a licensed practitioner. The device must be accompanied by labeling bearing the name and address of the licensed practitioner and the directions for use and cautionary statements, if any, contained in the order.

Section 801.405 establishes labeling requirements for articles intended for lay use in repairing and refitting dentures. The labeling must: (1) Limit directions for use for denture repair kits to emergency repair pending unavoidable delay in obtaining professional reconstruction of the denture; (2) limit directions for use for denture reliners, pads, and cushions to temporary refitting pending unavoidable delay in obtaining professional reconstruction of the denture; and (3) contain the word "emergency" preceding and modifying each indication-for-use statement for denture repair kits and the word "temporary" preceding and modifying each indication-for-use statement for reliners, pads, and cushions.

Section 801.410(f) requires that results of impact tests and description of the test method and apparatus be kept for a period of 3 years.

Section 801.410(f) is designed to protect the eyeglass wearer from potential eye injury resulting from

shattering of ordinary eyeglass lenses and requires that eyeglasses and sunglasses be fitted with impact-resistant lenses. Examination of data available on the frequency of eye injuries resulting from the shattering of ordinary crown glass lenses indicates that the use of such lenses constitutes an avoidable hazard to the eye of the wearer. According to the Vision Council of America, 60 percent of the population, or 161 million Americans, wear prescription eyewear; 81 percent have eyeglasses, 3 percent have contact lenses only and 16 percent have both eyeglasses and contact lenses.

Section 801.420(c) requires that the manufacturers or distributors of hearing aids develop a User Instructional Brochure, which accompanies the device and is provided to the prospective user by the dispenser of the hearing aid. The brochure must contain detailed information on the use and maintenance of the hearing aid.

Section 801.421(b) requires the hearing aid dispenser to provide the prospective user a copy of the User Instructional Brochure and an opportunity to review the comments with him/her orally or in the predominant method of communication used during the sale.

Section 801.421(c) requires the hearing aid dispenser to provide, upon request, to the prospective purchaser of any hearing aid (s)he dispenses, a copy of the User Instructional Brochure or the name and address of the manufacturer or distributor from whom the brochure may be obtained.

Section 801.421(d) requires the hearing aid dispenser to retain copies of all physician statements or any waivers of medical evaluation for 3 years from the time of dispensing.

Section 801.435 requires condom manufacturers to include an expiration date in the labeling of the condom. The manufacturer must support the expiration date by data from quality control tests demonstrating physical and mechanical integrity of three random lots of the same product which were stored under accelerated and real time conditions.

Section 809.10(a) provides that a label for an in vitro diagnostic product must contain the following information:

1. The proprietary and established name;
2. The intended use or uses of the product;
3. For a reagent, a declaration of the established name, if any, and the quantity, proportion, and concentration of each reactive ingredient;
4. A statement of warnings and precautions for users;

5. For a reagent, appropriate storage instructions;

6. For a reagent, a means by which the user may be assured that the product meets the appropriate standards of identity, strength, quality, and purity;

7. For a reagent, a declaration of the net quantity of contents;

8. Name and place of business of the manufacturer, packer, and distributor; and

9. A lot or control number.

Section 809.10(b) provides that the labeling (package insert) accompanying the device must contain the following:

1. Proprietary name and established name, if any;
2. The intended use or uses;
3. A summary and explanation of the test;
4. The chemical, physical, physiological, or biological principles of the procedure;
5. Information about the reagents;
6. Information about the instruments;
7. Information about the specimen collection and preparation for analysis;
8. Information about the procedure;
9. Information about the results;
10. Information about the limitations of the procedure;
11. Expected values;
12. Specific performance characteristics;
13. A bibliography of pertinent references; and
14. Date of issuance of the last revision of the labeling.

Section 809.10(d) provides that the labeling for general purpose laboratory reagents may be exempt from the labeling requirements in § 809.10(a) and (b), if the labeling contains the following:

1. The proprietary name and established name of the reagent;
2. The established name and the quantity, proportion, and concentration of the reagent ingredient;
3. A statement of the purity and quality of the reagent;
4. A statement of warnings and precautions for users;
5. Appropriate storage instructions;
6. A declaration of the net quantity of contents;
7. Name and place of business of the manufacturer, packer, or distributor; and
8. A lot or control number.

Section 809.10(e) requires manufacturers of analyte specific reagents to include the following in the labeling:

1. The proprietary name and established name, if any, of the reagent;
2. A declaration of established name, if any, and quantity, proportion or concentration of the reagent ingredient;

3. A statement of the purity and quality of the reagent;  
 4. A statement of warnings or precautions for users;  
 5. Appropriate storage instructions;  
 6. A declaration of the net quantity of contents;  
 7. Name and place of business of the manufacturer, packer, or distributor;  
 8. A lot or control number; and  
 9. The statement, "For analyte specific reagent use only. Analytical and performance characteristics are not established."

Section 809.10(f) requires that the labeling for OTC test sample collection systems for drugs of abuse testing bear the following information in a language appropriate for the intended users:

1. Adequate instructions for specimen collection and handling;  
 2. An identification system to ensure that specimens are not mixed up or otherwise misidentified at the laboratory;

3. The intended use or uses of the product;  
 4. A statement that confirmatory testing will be conducted on all samples that initially test positive;  
 5. A statement of warnings or precautions for users;  
 6. Adequate instructions on how to obtain test results from a person who can explain their meaning, including the probability of false positive and false negative results, as well as how to contact a trained health professional if additional information on interpretation of test results or followup counseling is desired; and

7. Name and place of business of the manufacturer, packer, or distributor.  
 Section 809.30(d) requires that manufacturers of analyte specific reagents (ASRs) assure that advertising and promotional materials for ASRs:  
 1. Include the identity and purity of the ASR and the identity of the analyte; and

2. Do not include any statement regarding analytical or clinical performance.

These estimates are based on FDA's registration and listing database for medical device establishments, agency communications with industry, and FDA's knowledge of and experience with device labeling. We have not estimated a burden for those requirements where the information to be disclosed is information that has been supplied by FDA. Also, we have not estimated a burden for that information that is disclosed to third parties as a usual and customary part of a medical device manufacturer, distributor, or importer's normal business activities. We do not include any burden for time that is spent designing labels to improve the format or presentation.

FDA estimates the burden of this collection of information as follows:

TABLE 1.—ESTIMATED ANNUAL REPORTING BURDEN<sup>1</sup>

21 CFR Section	No. of Respondents	Annual Frequency per Response	Total Annual Responses	Hours per Response	Total Hours
800.10(a)(3) and 800.12(c)	4	10	40	1	40
800.10(b)(2)	4	10	40	40	1,600
801.1	20,000	3.5	70,000	0.1	7,000
801.5	2,000	3.5	7,000	22.35	156,450
801.61	1,000	3.5	3,500	1	3,500
801.62	200	5	1,000	1	1,000
801.109	18,000	3.5	63,000	17.77	1,119,510
801.110	10,000	50	500,000	0.25	125,000
801.405(b)	40	1	40	4	160
801.420(c)	40	5	200	40	8,000
801.421(b)	10,000	160	1,600,000	0.30	480,000
801.421(c)	10,000	5	50,000	0.17	8,500
801.435	45	1	45	96	4,320
809.10(a) and (b)	1,700	6	10,200	80	816,000
809.10(d)	300	2	600	40	24,000
809.10(e)	300	25	7,500	1	7,500
809.10(f)	20	1	20	100	2,000
809.30(d)	300	25	7,500	1	7,500
Total Burden Hours					2,772,080

<sup>1</sup> There are no capital costs or operating and maintenance costs associated with this collection of information.

TABLE 2.—ESTIMATED ANNUAL RECORDKEEPING BURDEN<sup>1</sup>

21 CFR Section	No. of Recordkeepers	Annual Frequency per Recordkeeping	Total Annual Records	Hours per Record	Total Hours
801.410(f)	30	769,000	23,070,000	641	19,225
801.421(d)	9,900	162	1,600,000	0.25	400,000
Total Hours					419,225

<sup>1</sup> There are no capital costs or operating and maintenance costs associated with this collection of information.

Dated: December 6, 2004.

**Jeffrey Shuren,**

*Assistant Commissioner for Policy.*

[FR Doc. 04-27333 Filed 12-13-04; 8:45 am]

BILLING CODE 4160-01-S

## DEPARTMENT OF HOMELAND SECURITY

### Bureau of Customs and Border Protection

#### Agency Information Collection Activities: User Fees

**AGENCY:** Bureau of Customs and Border Protection, Department of Homeland Security.

**ACTION:** Proposed collection; comments requested.

**SUMMARY:** The Bureau of Customs and Border Protection (CBP) of the Department of Homeland Security has submitted the following information collection request to the Office of Management and Budget (OMB) for review and approval in accordance with the Paperwork Reduction Act of 1995: User Fees. This is a proposed extension of an information collection that was previously approved. CBP is proposing that this information collection be extended with a change to the burden hours. This document is published to obtain comments from the public and affected agencies. This proposed information collection was previously published in the **Federal Register** (69 FR 56448) on September 21, 2004, allowing for a 60-day comment period. This notice allows for an additional 30 days for public comments. This process is conducted in accordance with 5 CFR 1320.10.

**DATES:** Written comments should be received on or before January 13, 2005.

**ADDRESSES:** Written comments and/or suggestions regarding the items contained in this notice, especially the estimated public burden and associated response time should be directed to the Office of Management and Budget, Office of Information and Regulatory Affairs, Attention: Department of

Treasury Desk Officer, Washington, DC 20503. Additionally comments may be submitted to OMB via facsimile to (202) 395-6974.

**SUPPLEMENTARY INFORMATION:** The Bureau of Customs and Border Protection (CBP) encourages the general public and affected Federal agencies to submit written comments and suggestions on proposed and/or continuing information collection requests pursuant to the Paperwork Reduction Act of 1995 (Pub. L. 104-13). Your comments should address one of the following four points:

(1) Evaluate whether the proposed collection of information is necessary for the proper performance of the functions of the agency/component, including whether the information will have practical utility;

(2) Evaluate the accuracy of the agencies/components estimate of the burden of The proposed collection of information, including the validity of the methodology and assumptions used;

(3) Enhance the quality, utility, and clarity of the information to be collected; and

(4) Minimize the burden of the collections of information on those who are to respond, including the use of appropriate automated, electronic, mechanical, or other technological collection techniques or other forms of information technology, e.g., permitting electronic submission of responses.

#### Title: User Fees.

*OMB Number:* 1651-0052.

*Form Number:* Form CBP-339.

*Abstract:* The information collected on the User Fee Form CBP-339 is necessary in order for CBP to account for, and track user fees from private and commercial vessels, private aircraft, operators of commercial trucks, and passenger and freight railroad cars entering the United States and recipients of certain dutiable mail entries for certain official services.

*Current Actions:* This submission is being submitted to extend the expiration date with a change in the burden hours.

*Type of Review:* Extension (with change).

*Affected Public:* Businesses or other for-profit.

*Estimated Number of Respondents:* 45,000.

*Estimated Time Per Respondent:* 5-10 minutes.

*Estimated Total Annual Burden Hours:* 4,166.

*Estimated Total Annualized Cost on the Public:* \$62,490.

If additional information is required contact: Tracey Denning, Bureau of Customs and Border Protection, 1300 Pennsylvania Avenue NW., Room 3.2.C, Washington, DC 20229, at 202-344-1429.

Dated: December 7, 2004.

**Tracey Denning,**

*Agency Clearance Officer, Information Services Branch.*

[FR Doc. 04-27326 Filed 12-13-04; 8:45 am]

BILLING CODE 4820-02-P

## DEPARTMENT OF THE INTERIOR

### Bureau of Land Management

[ES-020-04-1320-EL]

#### Notice of the Solicitation of Public Comments on Fair Market Value and Maximum Economic Recovery; Coal Lease by Application KYES-51002

**AGENCY:** Bureau of Land Management, Interior.

**ACTION:** Notice.

**SUMMARY:** The United States Department of Agriculture Forest Service, Daniel Boone National Forest, has prepared an Environmental Impact Statement (EIS) to address coal lease application KYES-51002. The EIS has been prepared in cooperation with the Bureau of Land Management's (BLM) Eastern States, Jackson Field Office, the Office of Surface Mining and the Kentucky Department of Surface Mining Reclamation and Enforcement. A public hearing, requesting input to the EIS, was held on November 24, 2003, at the Leslie County Extension Services office in Hyden, Kentucky.

Public comments are requested on the fair market value (FMV) and maximum