

interested persons between 9 a.m. and 4 p.m., Monday through Friday.

1. Gastroenterology and Urology Devices Panel of the Medical Devices Advisory Committee transcript, August 7, 1997.

2. Lewis J.H. *et al.*, "A Way to Help Your Patients Who Use Vacuum Devices," *Contemporary Urology*, vol. 3, No. 12: 15–24, 1991.

3. Montague, D.K. *et al.*, "Clinical Guidelines Panel on Erectile Dysfunction: Summary Report on the Treatment of Erectile Dysfunction," *Journal of Urology*, 156: 2007–2011, 1996.

4. NIH Consensus Statement, "Impotence," *National Institutes of Health*, vol. 10, No. 4, 1992.

List of Subjects in 21 CFR Part 876

Medical devices.

Therefore, under the Federal Food, Drug, and Cosmetic Act and under authority delegated to the Commissioner of Food and Drugs, 21 CFR part 876 is amended as follows:

PART 876—GASTROENTEROLOGY-UROLOGY DEVICES

1. The authority citation for 21 CFR part 876 continues to read as follows:

Authority: 21 U.S.C. 351, 360, 360c, 360e, 360j, 360l, 371.

2. Section 876.1 is amended by adding paragraph (e) to read as follows:

§ 876.1 Scope.

* * * * *

(e) Guidance documents referenced in this part are available on the Internet at <http://www.fda.gov/cdrh/guidance.html>.

3. Section 876.5020 is added to subpart F to read as follows:

§ 876.5020 External penile rigidity devices.

(a) *Identification.* External penile rigidity devices are devices intended to create or maintain sufficient penile rigidity for sexual intercourse. External penile rigidity devices include vacuum pumps, constriction rings, and penile splints which are mechanical, powered, or pneumatic devices.

(b) *Classification.* Class II (special controls). The devices are exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in § 876.9. The special control for these devices is the FDA guidance document entitled "Class II Special Controls Guidance Document: External Penile Rigidity Devices." See § 876.1(e) for the availability of this guidance document.

Dated: December 15, 2004.

Linda S. Kahan,

Deputy Director, Center for Devices and Radiological Health.

[FR Doc. 04–28252 Filed 12–27–04; 8:45 am]

BILLING CODE 4160–01–S

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

21 CFR Part 884

[Docket No. 2004N–0530]

Medical Devices; Obstetrical and Gynecological Devices; Classification of the Assisted Reproduction Laser System

AGENCY: Food and Drug Administration, HHS.

ACTION: Final rule.

SUMMARY: The Food and Drug Administration (FDA) is classifying the assisted reproduction laser system into class II (special controls). The special control that will apply to the device is the guidance document entitled "Class II Special Controls Guidance Document: Assisted Reproduction Laser Systems." The agency is classifying this device into class II (special controls) in order to provide a reasonable assurance of safety and effectiveness of the device. Elsewhere in this issue of the **Federal Register**, FDA is publishing a notice of availability of the guidance document that is the special control for this device.

DATES: This rule is effective January 27, 2005. The classification was effective November 4, 2004.

FOR FURTHER INFORMATION CONTACT: Michael Bailey, Center for Devices and Radiological Health (HFZ–400), Food and Drug Administration, 9200 Corporate Blvd., Rockville, MD 20850, 301–594–1180, ext. 130.

SUPPLEMENTARY INFORMATION:

I. Background

In accordance with section 513(f)(1) of the Federal Food, Drug, and Cosmetic Act (the act) (21 U.S.C. 360c(f)(1)), devices that were not in commercial distribution before May 28, 1976, the date of enactment of the Medical Device Amendments of 1976 (the amendments), generally referred to as postamendments devices, are classified automatically by statute into class III without any FDA rulemaking process. These devices remain in class III and require premarket approval, unless and until the device is classified or reclassified into class I or II or FDA issues an order finding the device to be substantially equivalent, in accordance with section 513(i) of the act, to a predicate device that does not require premarket approval. The agency determines whether new devices are substantially equivalent to previously marketed

devices by means of premarket notification procedures in section 510(k) of the act (21 U.S.C. 360(k)) and part 807 (21 CFR part 807) of FDA's regulations.

Section 513(f)(2) of the act provides that any person who submits a premarket notification under section 510(k) of the act for a device that has not previously been classified may, within 30 days after receiving an order classifying the device in class III under section 513(f)(1) of the act, request FDA to classify the device under the criteria set forth in section 513(a)(1) of the act. FDA shall, within 60 days of receiving such a request, classify the device by written order. This classification shall be the initial classification of the device. Within 30 days after the issuance of an order classifying the device, FDA must publish a document in the **Federal Register** announcing such classification (section 513(f)(2) of the act).

In accordance with section 513(f)(1) of the act, FDA issued a document on August 10, 2004, classifying the Hamilton Thorne Zona Infrared Laser Optical System (ZILOS-tkr) into class III, because it was not substantially equivalent to a device that was introduced or delivered for introduction into interstate commerce for commercial distribution before May 28, 1976, or a device which was subsequently reclassified into class I or class II. On August 25, 2004, Hamilton Thorne Biosciences, Inc., submitted a petition requesting classification of this device under section 513(f)(2) of the act. The manufacturer recommended that the device be classified into class II (Ref. 1).

In accordance with section 513(f)(2) of the act, FDA reviewed the petition in order to classify the device under the criteria for classification set forth in 513(a)(1) of the act. Devices are to be classified into class II if general controls, by themselves, are insufficient to provide reasonable assurance of safety and effectiveness, but there is sufficient information to establish special controls to provide reasonable assurance of the safety and effectiveness of the device for its intended use. After review of the information submitted in the petition, FDA has determined that the device can be classified in class II with the establishment of special controls. FDA believes that class II special controls, in addition to general controls, will provide reasonable assurance of the safety and effectiveness of the device.

The device is assigned the generic name assisted reproduction laser system and it is identified as a device that images, targets, and controls the power and pulse duration of a laser beam used to ablate a small tangential hole in, or

to thin, the zona pellucida of an embryo for assisted hatching or other assisted reproduction procedures.

The potential risks to health associated with the device are: (1) Damage to the embryo, (2) ineffective treatment, (3) hazards associated with electrical equipment, and (4) electromagnetic interference and electrostatic discharge hazards. The special controls guidance document entitled "Class II Special Controls Document: Assisted Reproduction Laser Systems" aids in mitigating the risks by recommending performance characteristics, safety testing, and appropriate labeling.

Thus, in addition to the general controls of the act, an assisted reproduction laser system, is subject to the special controls guidance document. FDA believes that following the class II special controls guidance document generally addresses the risks to health identified in the previous paragraph. On November 4, 2004, FDA issued an order to the petitioner classifying the device as described previously into class II and is codifying this classification by adding 21 CFR 884.6200.

Following the effective date of this final classification rule, any firm submitting a 510(k) (premarket notification) will need to address the issues covered in the special controls guidance. However, the firm need only show that its device meets the recommendations of the guidance or in some other way provides equivalent assurances of safety and effectiveness.

Section 510(m) of the act provides that FDA may exempt a class II device from the premarket notification requirements under section 510(k) of the act, if FDA determines that premarket notification is not necessary to provide reasonable assurance of the safety and effectiveness of the device. For this type of device, FDA has determined that premarket notification is necessary to provide reasonable assurance of the safety and effectiveness; therefore, the device is not exempt from premarket notification requirements. Thus, persons who intend to market this type of device must submit to FDA a premarket notification, prior to marketing the device, which contains information about the device they intend to market.

II. Environmental Impact

The agency has determined under 21 CFR 25.34(b) that this action is of a type that does not individually or cumulatively have a significant effect on the human environment. Therefore, neither an environmental assessment nor an environmental impact statement is required.

III. Analysis of Impacts

FDA has examined the impacts of the final rule under Executive Order 12866, the Regulatory Flexibility Act (5 U.S.C. 601–612), and the Unfunded Mandates Reform Act of 1995 (Public Law 104–4). Executive Order 12866 directs agencies to assess all costs and benefits of available regulatory alternatives and, when regulation is necessary, to select regulatory approaches that maximize net benefits (including potential economic, environmental, public health and safety, and other advantages; distributive impacts; and equity). The agency believes that this final rule is not a significant regulatory action under the Executive Order.

The Regulatory Flexibility Act requires agencies to analyze regulatory options that would minimize any significant impact of a rule on small entities. Because classification of this device type into class II will relieve manufacturers of the device of the cost of complying with the premarket approval requirements of section 515 of the act (21 U.S.C. 360e), and may permit small potential competitors to enter the marketplace by lowering their costs, the agency certifies that the final rule will not have a significant impact on a substantial number of small entities.

Section 202(a) of the Unfunded Mandates Reform Act of 1995 requires that agencies prepare a written statement, which includes an assessment of anticipated costs and benefits, before proposing "any rule that includes any Federal mandate that may result in the expenditure by State, local, and tribal governments, in the aggregate, or by the private sector, of \$100,000,000 or more (adjusted annually for inflation) in any one year." The current threshold after adjustment for inflation is \$110 million. FDA does not expect this final rule to result in any 1-year expenditure that would meet or exceed this amount.

IV. Federalism

FDA has analyzed this final rule in accordance with the principles set forth in Executive Order 13132. FDA has determined that the rule does not contain policies that have substantial direct effects on the States, on the relationship between the National Government and the States, or on the distribution of power and responsibilities among the various levels of government. Accordingly, the agency has concluded that the rule does not contain policies that have federalism implications as defined in the Executive order and, consequently, a federalism summary impact statement is not required.

V. Paperwork Reduction Act of 1995

This final rule contains no collections of information. Therefore, clearance by the Office of Management and Budget under the Paperwork Reduction Act of 1995 is not required.

VI. Reference

The following reference has been placed on display in the Division of Dockets Management (HFA–305), Food and Drug Administration, 5630 Fishers Lane, rm. 1061, Rockville, MD 20852, and may be seen by interested persons between 9 a.m. and 4 p.m., Monday through Friday.

1. Petition from Hamilton Thorne Biosciences, Inc., dated August 25, 2004.

List of Subjects in 21 CFR Part 884

Medical devices.

■ Therefore, under the Federal Food, Drug, and Cosmetic Act and under authority delegated to the Commissioner of Food and Drugs, 21 CFR part 884 is amended as follows:

PART 884—OBSTETRICAL AND GYNECOLOGICAL DEVICES

■ 1. The authority citation for 21 CFR part 884 continues to read as follows:

Authority: 21 U.S.C. 351, 360, 360c, 360e, 360j, 371.

■ 2. Section 884.6200 is added to subpart G to read as follows:

§ 884.6200 Assisted reproduction laser system.

(a) *Identification.* The assisted reproduction laser system is a device that images, targets, and controls the power and pulse duration of a laser beam used to ablate a small tangential hole in, or to thin, the zona pellucida of an embryo for assisted hatching or other assisted reproduction procedures.

(b) *Classification.* Class II (special controls). The special control is FDA's guidance document entitled "Class II Special Controls Guidance Document: Assisted Reproduction Laser Systems." See § 884.1(e) for the availability of this guidance document.

Dated: December 15, 2004.

Linda S. Kahan,

Deputy Director, Center for Devices and Radiological Health.

[FR Doc. 04–28251 Filed 12–27–04; 8:45 am]

BILLING CODE 4160–01–S