



# National Voluntary Laboratory Accreditation Program



## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

### Sandia National Laboratories

Primary Standard Laboratory  
P.O. Box 5800, Mail Stop 0665  
Albuquerque, NM 87185-0665

Dr. Larry Azevedo

Phone: 505-844-8700 Fax: 505-844-4372

E-mail: ljazeve@sandia.gov

URL: <http://www.sandia.gov/psl/PSLHOME.HTML>

### CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

#### DIMENSIONAL

**NVLAP Code:** 20/D01

Angular

<i>Range</i>	<i>Best Uncertainty (±) <sup>note 1</sup></i>	<i>Remarks</i>
Angle Blocks	0.47 arc second	Standard Sizes, 1 arc second to 45°
Optical Squares	0.46 arc second	
True Squares	0.28 arc second	

Angular/Rotary Index Tables and Optical Polygons

<i>Range</i>	<i>Best Uncertainty (±) <sup>notes 1, 2</sup></i>	<i>Remarks</i>
30° increments	0.08 arc second	Stack method
30° increments	0.50 arc second	Comparison method

**NVLAP Code:** 20/D03

Gage Blocks

<i>Range</i>	<i>Best Uncertainty (±) <sup>notes 1, 6</sup></i>	<i>Remarks</i>
to 100 mm	(35 + 0.53 L) nm, L in mm	Interferometry, single wring
to 4 in	(1.38 + 0.53 L) μin, L in inches	Interferometry, single wring
< 1 mm	42 nm	Mechanical comparison to masters <sup>notes 2, 3, 4</sup>
< 0.04 in	1.65 μin	Mechanical comparison to masters <sup>notes 2, 3, 4</sup>

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

1 mm to 100 mm	(37 + 0.60 L) nm, L in mm	Mechanical comparison to masters <i>notes 2, 3, 4</i>
0.04 in to 4 in	(1.46 + 0.60 L) μin, L in inches	Mechanical comparison to masters <i>notes 2, 3, 4</i>
125 mm to 500 mm	(127 + 0.30 L) nm, L in mm	Mechanical comparison to masters <i>notes 2, 3, 4</i>
5 in to 20 in	(5.0 + 0.30 L) μin, L in inches	Mechanical comparison to masters <i>notes 2, 3, 4</i>

**NVLAP Code:** 20/D07  
Thread Measuring Wires

<i>Range</i>	<i>Best Uncertainty (±) notes 1, 2</i>	<i>Remarks</i>
All Standard 29° and 60°	160 nm (6.3 μin)	Comparison to NIST-calibrated masters

**NVLAP Code:** 20/D08  
Optical Reference Planes

<i>Range</i>	<i>Best Uncertainty (±) notes 1, 2</i>	<i>Remarks</i>
to 10 inch diameter	1.2 μin (30 nm)	Comparison to NIST-calibrated masters

**NVLAP Code:** 20/D09  
Roundness

<i>Range - Diameter</i>	<i>Best Uncertainty (±) notes 1, 2</i>	<i>Remarks</i>
to 100 mm	5.4 nm + 5.1% of value	Spindle error deconvolution at limited points.
to 4 in	0.2 μin + 5.1% of value	Spindle error deconvolution at limited points.
to 350 mm	10.6 nm + 6.8% of value	Spindle-compensated trace. Uncertainty will increase for large artifacts.
to 14 in	0.42 μin + 6.8% of value	Spindle-compensated trace. Uncertainty will increase for large artifacts.

**NVLAP Code:** 20/D11  
Spherical Diameter; Plug/Ring Gages

<i>Range</i>	<i>Best Uncertainty (±) notes 1, 2</i>	<i>Remarks</i>
1 mm to 25 mm	150 nm	Gaging Balls Calibration Spheres; Comparison to NIST-calibrated masters
0.03125 in to 1 in	5.91 μin	Gaging Balls Calibration Spheres; Comparison to NIST-calibrated masters
to 250 mm	(230 + 17 L) nm, L in mm	Plain plug/ring gages

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

to 10 in (9.0 + 1.7 L)  $\mu$ in, L in inches

Plain plug/ring gages

*NVLAP Code:* 20/D12  
Surface Texture

<i>Range</i>	<i>Best Uncertainty (<math>\pm</math>)</i> <sup>notes 1, 2</sup>	<i>Remarks</i>
0.5 $\mu$ in to 500 $\mu$ in (0.0125 $\mu$ m to 12.5 $\mu$ m)		
0.5 $\mu$ in	0.1 $\mu$ in	Step Height Standards
15 $\mu$ in	0.3 $\mu$ in	Step Height Standards
100 $\mu$ in	2.0 $\mu$ in	Step Height Standards
500 $\mu$ in	8.0 $\mu$ in	Step Height Standards
0.5 $\mu$ in to 125 $\mu$ in (0.0125 $\mu$ m to 3.1 $\mu$ m)		
0.5 $\mu$ in	0.1 $\mu$ in	Roughness Average (Ra)
15 $\mu$ in	0.6 $\mu$ in	Roughness Average (Ra)
100 $\mu$ in	2.7 $\mu$ in	Roughness Average (Ra)

## DC/LOW FREQUENCY

*NVLAP Code:* 20/E01  
Current/AC-DC Difference

<i>Range</i>	<i>Best Uncertainty (<math>\pm</math>) in ppm</i> <sup>note 1</sup>		
	<i>100</i>	<i>50 k</i>	<i>100 k</i>
10 mA	55	99	134
25 mA	54	99	134
50 mA	58	100	140
100 mA	59	107	149
250 mA	56	109	155
0.5 A	56	136	162
1.0 A	58	119	171
2.5 A	80	130	182

2009-01-01 through 2009-12-31

*Effective dates*

*For the National Institute of Standards and Technology*



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

5.0 A	68	159	212
10.0 A	77	176	248
16.0 A	100	216	279

**NVLAP Code:** 20/E03  
Capacitance Voltage Dividers - Pulsed High-Voltage Condition

Range	Best Uncertainty ( $\pm$ ) in percent <sup>note 1</sup>	Remarks
1 kV to 350 kV	2.0	1 $\mu$ s to 30 $\mu$ s Pulse

**NVLAP Code:** 20/E05  
DC Resistance

Range in ohms	Best Uncertainty ( $\pm$ ) in ppm <sup>note 1</sup>	Remarks
0.0001 to 0.001	5	6010 + 6011 System Time of Test
0.001 to 0.01	0.35	6010 + 6011 System Time of Test
0.01 to 0.1	0.29	6010 + 6011 System Time of Test
0.1 to 1	0.21	6010 + 6011 System Time of Test
1	0.055	6010 + 4220 + double substitution Time of Test
1 to 10	0.076	6010 + 4220 System Time of Test
10 to 100	0.09	6010 + 4220 System Time of Test
100 to 1000	0.11	6010 + 4220 System Time of Test
1000 to 10 k	0.15	6010 + 4220 System Time of Test
10 <sup>4</sup> to 10 <sup>5</sup>	0.20	6000 + 4220A System Time of Test
10 <sup>5</sup> to 10 <sup>6</sup>	0.25	6000 + 4220A System Time of Test
10 <sup>6</sup> to 10 <sup>7</sup>	0.29	6000 + 4220A System Time of Test
10 <sup>7</sup> to 10 <sup>8</sup>	0.58	6010 + 4220 System Time of Test
10 <sup>8</sup> to 10 <sup>9</sup>	5.0	6000 + 4220A System Time of Test
10 <sup>9</sup> to 10 <sup>10</sup>	470	Terraohm Meter
10 <sup>10</sup> to 10 <sup>11</sup>	670	Terraohm Meter
10 <sup>11</sup> to 10 <sup>12</sup>	1400	Terraohm Meter
10 <sup>12</sup> to 10 <sup>13</sup>	2000	Terraohm Meter
10 <sup>13</sup> to 10 <sup>14</sup>	3300	Terraohm Meter
10 <sup>14</sup> to 10 <sup>15</sup>	6700	Terraohm Meter
10 <sup>15</sup> to 10 <sup>16</sup>	7.0%	Terraohm Meter

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

### Shunts

<i>Range</i>	<i>Best Uncertainty (±) in ppm <sup>note 1</sup></i>	<i>Remarks</i>
100 mA to 2500 A	2.5	

<i>Range (nominal values) in ohms</i>	<i>Best Uncertainty (±) in ppm <sup>note 1</sup></i>	<i>Remarks</i>
1 and 1.9	97	Comparison method
10 and 19	23	Comparison method
100 and 190	10	Comparison method
1 k, 1.9 k, 10 k, 19 k	9	Comparison method
100 k and 190 k	11	Comparison method
1 M and 1.9 M	21	Comparison method
10 M and 19 M	51	Comparison method
100 M	102	Comparison method

### DC Current

<i>Range</i>	<i>Best Uncertainty (±) <sup>note 1</sup></i>	<i>Remarks</i>
220 µA	43 ppm + 6 nA	Comparison method
2.2 mA	37 ppm + 7 nA	Comparison method
22 mA	35 ppm + 40 nA	Comparison method
220 mA	48 ppm + 0.7 µA	Comparison method
2.2 A	89 ppm + 12 µA	Comparison method
11 A	367 ppm + 490 µA	Comparison method

NVLAP Code: 20/E06

### DC Voltage

<i>Range in Volts</i>	<i>Best Uncertainty (±) in ppm <sup>note 1</sup></i>	<i>Remarks</i>
1.0 to 1.018	0.14	Josephson Array Systems, Zeners, and DVMs
10.0	0.017	Josephson Array Systems, Zeners, and DVMs

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

Pressure coefficient of Voltage for Solid State Voltage Standards

<i>Pressure Coefficient of Voltage</i>	<i>Best Uncertainty (<math>\pm</math>)<sup>note 1</sup></i>	<i>Remarks</i>
> 5 nV/hPa	2 nV/hPa	- 2000 m to + 3000 m elevation (Difference Measurement)
0 nV/hPa to 5 nV/hPa	0.1 nV/hPa	- 2000 m to + 3000 m elevation (Josephson Array Measurement)

Voltage divider - Potentiometer combination

1.5 V to 1500 V	2.5 ppm	Volt Box-Potentiometer, k = 2
x1.0 range to 1.05 V	0.5 ppm of reading + 0.1 $\mu$ V	Potentiometer only, k = 3
x1.0 range above 1.05 V	1.0 ppm of reading + 0.1 $\mu$ V	Potentiometer only, k = 3
x0.1 range	1.5 ppm of reading + 0.01 $\mu$ V	Potentiometer only, k = 3
x0.01 range	2.5 ppm of reading + 0.005 $\mu$ V	Potentiometer only, k = 3

Automated Intermediate Voltage System

0 V to 10.0 V	0.5 ppm + 0.2 $\mu$ V	Automated Potentiometer System
10 V to 30 V	1.4 ppm	Automated Potentiometer System
300 V to 1200 V	4.0 ppm	Automated Potentiometer System

High Voltage - Electrostatic Voltmeters, etc.

<i>Range in kV</i>	<i>Best Uncertainty (<math>\pm</math>)<sup>note 1</sup></i>	<i>Remarks</i>
to 100	106 ppm	200 kV system
100 to 200	140 ppm	200 kV system
to 10	2000 ppm	10 kV system

Ratio/Bridges

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

1:1 to 1:100 000                      .05 (ratio)                      For ratio based on 20 step first dial (k = 3). For bridges, uncertainty combines ratio and resistance uncertainties

Range	Best Uncertainty ( $\pm$ ) in ppm + $\mu V$ <sup>note 1</sup>	Remarks
220 mV	12.0 + 0.4	Comparison method
2.2 V	5.8 + 0.7	Comparison method
11 V	4.3 + 2.5	Comparison method
22 V	4.3 + 4	Comparison method
220 V	6.1 + 40	Comparison method
1100 V	7.3 + 400	Comparison method

NVLAP Code: 20/E08  
Inductive Voltage Dividers

Range in Volts	Best Uncertainty ( $\pm$ ) in ppm <sup>note 1</sup>	Remarks
15, 35 and 100	55	@ 60 Hz, 1 kHz and 10 kHz

NVLAP Code: 20/E09  
Voltage/AC-DC Difference

Range in Volts	Best Uncertainty ( $\pm$ ) in ppm <sup>note 1</sup>											
	Frequency in Hertz											
	10	20	50	100	200	500	1 k	2 k	5 k	10 k	20 k	50 k
1	21	12	11	8	7	5	6	5	6	5	7	8
2	16	16	10	8	6	5	8	7	15	13	6	7
3	16	14	11	6	6	6	10	10	10	9	9	9
4	18	13	11	7	6	6	8	6	6	7	8	7
6	23	15	14	13	15	14	13	10	10	9	8	8
10	17	13	11	8	10	7	11	11	13	13	12	13
12	19	13	11	8	10	11	10	11	10	10	11	9
20	31	14	11	6	6	7	7	8	7	8	9	9
30	19	15	11	8	6	8	8	9	8	8	9	12
40	30	14	11	17	6	8	18	9	8	8	16	21
60	20	14	13	7	7	9	8	8	8	8	9	12
100	35	14	12	7	7	8	9	8	8	8	10	13
120	102	23	21	22	21	9	10	9	9	9	11	15

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

200	101	21	21	22	21	10	12	12	13	11	15	23
300	101	21	21	21	21	12	14	14	11	11	14	24
400	101	21	21	23	21	13	14	15	15	15	18	28
600	102	22	21	21	21	16	16	15	16	15	21	32
1200	102	21	22	22	21	17	18	17	22	19	22	26

### LF AC Voltage

*Best Uncertainty (±) in ppm<sup>note 1</sup>*  
*Frequency in Hertz*

<i>Range in Volts</i>	<i>100 k</i>	<i>200 k</i>	<i>500 k</i>	<i>1 M</i>
1	4	75	82	86
2	10	74	74	76
3	11	72	77	81
4	10	71	71	71
6	9	73	75	81
10	11	72	71	74
12	13	72	73	72
20	11	71	72	71
30	14	72	77	78
40	21	91	106	71
60	14	72	74	73
100	23	90	91	134
120	19			
200	35			
300	45			
400	51			
600	59			
1200	49			

### AC Voltage

*Best Uncertainty (±) in ppm + μV (below 2.2V)<sup>note 1</sup>*  
*Frequency in Hertz*

<i>Range</i>	<i>10 to 20</i>	<i>20 to 40</i>	<i>40 to 20 k</i>	<i>20 k to 50 k</i>	<i>50 k to 100 k</i>	<i>100 k to 300 k</i>	<i>300 k to 500 k</i>	<i>500 k to 1 M</i>
2.2 mV	1700 + 1.3	770 + 1.3	500 + 1.3	990 + 2.0	1500 + 2.5	2600 + 4.0	2700 + 8.0	3700 + 8.0

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology





# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

7 mV	880 + 1.3	430 + 1.3	350 + 1.3	700 + 2.0	1100 + 2.5	1800 + 4.0	1800 + 8.0	2600 + 8.0
22 mV	380 + 1.3	290 + 1.3	250 + 1.3	610 + 2.0	970 + 2.5	1500 + 4.0	1600 + 8.0	2100 + 8.0
70 mV	260 + 1.5	120 + 1.5	81 + 1.5	130 + 2.0	260 + 2.5	540 + 4.0	690 + 8.0	1100 + 8.0
220 mV	230 + 1.5	92 + 1.5	62 + 1.5	78 + 2.0	170 + 2.5	310 + 4.0	420 + 8.0	1010 + 8.0
700 mV	230 + 1.5	83 + 1.5	65 + 1.5	110 + 2.0	110 + 2.5	320 + 4.0	400 + 8.0	1000 + 8.0
2.2 V	220	69	45	110	110	310	370	940
7 V	210	70	50	90	140	200	440	1300
22 V	210	72	50	90	140	200	440	1300
70 V	230	100	80	120	200	820	900	1400
220 V	230	100	80	130	200	830	940	
700 V	210	120	100	170	520			
1000 V	210	120	100	190	520			

### AC Current

*Best Uncertainty (±) in ppm + nA <sup>note 1</sup>*  
*Frequency in Hertz*

Range	10 to 20	20 to 40	40 to 1 k	1 k to 5 k	5 k to 10 k
220 µA	250 + 16	160 + 10	130 + 8	280 + 12	1100 + 65
2.2 mA	250 + 40	160 + 35	120 + 35	200 + 110	1100 + 650
22 mA	250 + 400	160 + 350	130 + 350	210 + 550	1100 + 5000
220 mA	250 + 4000	160 + 3500	170 + 2500	230 + 3500	1200 + 10000

*Best Uncertainty (±) in ppm + µA <sup>note 1</sup>*  
*Frequency in Hertz*

Range	20 to 1 K	1 k to 5 k	5 k to 10 k
2.2 A	260 + 35	530 + 80	700 + 160

NVLAP Code: 20/E10

LF Capacitance

Range in pF

0.01 to 1000

*Best Uncertainty (±) in ppm <sup>note 1</sup>*

5

Remarks

@ 1 kHz

NVLAP Code: 20/E11

LF Inductance

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 105002-0**

<i>Range</i>	<i>Best Uncertainty (±) in % <sup>note 12</sup></i>		
	<i>Frequency in Hertz</i>		
	<i>100</i>	<i>1 k</i>	<i>10 k</i>
50 μH	0.40	0.31	0.31
100 μH	0.18	0.16	0.16
200 μH	0.094	0.076	0.075
500 μH	0.037	0.035	0.081
1 mH	0.032	0.032	0.080
2 mH	0.042	0.040	0.083
5 mH	0.034	0.034	0.080
10 mH	0.032	0.031	0.080
20 mH	0.032	0.033	0.089
50 mH	0.031	0.031	0.077
100 mH	0.032	0.034	0.090
200 mH	0.03	0.03	
500 mH	0.03	0.03	
1 H	0.03	0.081	
2 H	0.032	0.075	
5 H	0.031	0.157	
10 H	0.032	0.31	

**NVLAP Code:** 20/E18  
Resistive Voltage Dividers - Pulsed High-Voltage Condition

<i>Range in kV</i>	<i>Best Uncertainty (±) in % <sup>note 1</sup></i>	<i>Remarks</i>
1 to 350	1.0	1 μs to 30 μs Pulse

**NVLAP Code:** 20/E20  
Oscilloscopes  
Pulse Amplitude (digitized waveform)

<i>Range</i>	<i>Best Uncertainty (±) <sup>note 1</sup></i>	<i>Remarks</i>
0.1 V to 200 V	0.25 % of pulse amplitude	NIST VPMS

2009-01-01 through 2009-12-31

*Effective dates*

*Sally S. Bruce*

*For the National Institute of Standards and Technology*



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 105002-0**

**TIME AND FREQUENCY**

*NVLAP Code:* 20/F01  
Frequency Dissemination

<i>Range in MHz</i>	<i>Best Uncertainty (<math>\pm</math>)<sup>note 1</sup></i>	<i>Remarks</i>
0.1	$5 \times 10^{-13}$	NIST FMAS
1	$5 \times 10^{-13}$	NIST FMAS
5	$5 \times 10^{-13}$	NIST FMAS
10	$5 \times 10^{-13}$	NIST FMAS

*NVLAP Code:* 20/F04  
Oscillator Characterization

<i>Range</i>	<i>Best Uncertainty (<math>\pm</math>)<sup>note 1</sup></i>	<i>Remarks</i>
Phase -180° to 360°	$\pm 0.005^\circ$	

*NVLAP Code:* 20/F04  
Pulse Waveform  
Transition Duration

<i>Range</i>	<i>Best Uncertainty (<math>\pm</math>)<sup>note 1</sup></i>	<i>Remarks</i>
< 17.5 ps	5 ps	Fast Rise Pulse
Time Interval		
> 500 ps	500 ps	

*NVLAP Code:* 20/F05  
Stopwatches and Timers

<i>Range</i>	<i>Best Uncertainty (<math>\pm</math>)<sup>note 1</sup></i>	<i>Remarks</i>
--------------	---	----------------

2009-01-01 through 2009-12-31

*Effective dates*

*Sally S. Bruce*

*For the National Institute of Standards and Technology*



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

1 s to 24 hrs                                      0.05 s/day                                      Time base methods

## IONIZING RADIATION

*NVLAP Code:* 20/I04  
Radioactive Sources

<i>Range</i>	<i>Best Uncertainty (±) <sup>note 1</sup></i>	<i>Remarks</i>
1 Bq to 2 x 10 <sup>5</sup> Bq into 2 π	2.0%	Alpha Emission Rate
10 Bq to 10 <sup>4</sup> Bq into 2π	4.0%	Beta Emission Rate
3 MeV to 8 MeV	30 keV	Alpha Energy

## MECHANICAL

*NVLAP Code:* 20/M05  
Gas Flow

<i>Range</i>	<i>Best Uncertainty (±) <sup>note 1</sup></i>
50 cc/min to 1200 L/min	0.5%

*NVLAP Code:* 20/M06  
Force

<i>Range in lbf</i>	<i>Best Uncertainty (±) in % <sup>notes 1, 2, 5</sup></i>	<i>Remarks</i>
100 to 1000	0.0052	Primary Standard (Deadweight)
1000 to 100 000	0.025	Secondary Standards

*NVLAP Code:* 20/M08  
Mass – Metric

<i>Range</i>	<i>Best Uncertainty (±) in mg (k=2)</i>	<i>Remarks</i>
25 kg	18	Echelon II
20 kg	12	Echelon II
10 kg	4	Echelon II
5 kg	1.7	Echelon II
3 kg	0.7	Echelon II
2 kg	0.6	Echelon II

2009-01-01 through 2009-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 105002-0**

1 kg	0.11	Echelon II
500 g	0.061	Echelon II
300 g	0.069	Echelon II
200 g	0.037	Echelon II
100 g	0.033	Echelon II
50 g	0.021	Echelon II
30 g	0.020	Echelon II
20 g	0.013	Echelon II
10 g	0.0098	Echelon II
5 g	0.0061	Echelon II
3 g	0.0086	Echelon II
2 g	0.0050	Echelon II
1 g	0.0050	Echelon II
500 mg	0.0033	Echelon II
300 mg	0.0031	Echelon II
200 mg	0.0030	Echelon II
100 mg	0.0029	Echelon II
50 mg	0.0027	Echelon II
30 mg	0.0026	Echelon II
20 mg	0.0025	Echelon II
10 mg	0.0023	Echelon II
5 mg	0.0020	Echelon II
3 mg	0.0018	Echelon II
2 mg	0.0016	Echelon II
1 mg	0.0014	Echelon II

*NVLAP Code:* 20/M11  
Acceleration / Vibration

<i>Range in gn</i>	<i>Frequency in Hz</i>	<i>Best Uncertainty (±) in % <sup>note 1</sup></i>
1 to 50	10 to 10 000	2.5
Acceleration / Vibration – Adding Long Stroke Capability		
0.05	1	2.5
0.2	2	2.5

2009-01-01 through 2009-12-31

*Effective dates*

*For the National Institute of Standards and Technology*



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 105002-0**

1	5	2.5
Shock		
<i>Range in gn</i>	<i>Best Uncertainty (±) in % <sup>note 1</sup></i>	<i>Remarks</i>
500 to 10 000	4	
Static Acceleration - Centrifuge		
<i>Range in gn</i>	<i>Best Uncertainty (±) in % <sup>note 1</sup></i>	<i>Remarks</i>
25 to 1000	0.5	

**RF MICROWAVE**

*NVLAP Code:* 20/R05  
HF Capacitance

<i>Range in pF</i>	<i>Best Uncertainty (±) in percent <sup>note 1</sup></i>				
	<i>100</i>	<i>1 k</i>	<i>10 k</i>	<i>100 k</i>	<i>1 M</i>
0.01		0.20		1.3	
0.1		0.05		1.3	
1		0.02		0.04	
10		0.01		0.02	
100		0.01		0.01	
1000		0.01		0.03	
1		0.02		0.2	0.30
2		0.02		0.35	0.60
5		0.02		0.22	0.26
10		0.10		0.14	0.15
20		0.10		0.13	0.11
50				0.03	0.02
100				0.02	0.02
200				0.01	0.01

2009-01-01 through 2009-12-31

*Effective dates*

*Sally S. Bruce*

*For the National Institute of Standards and Technology*



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

500				0.02	0.01
1000				0.02	0.03
10		0.0001			
100		0.0001			
1	0.01	0.002	0.006	0.02	0.18
10	0.01	0.002	0.0025	0.0025	0.003
100	0.01	0.002	0.0044	0.0044	0.004
1000	0.01	0.002	0.0043	0.0048	0.003

NVLAP Code: 20/R06  
HF Inductance

*Best Uncertainty (±) in percent<sup>note 1</sup>*  
*Frequency in Hertz*

<i>Range</i>	<i>10 k</i>	<i>100 k</i>	<i>1 M</i>	<i>10 M</i>
0.1 µH		3.13	4.72	
0.2 µH		1.63	1.67	
0.5 µH		0.93	0.97	
1 µH		0.75	0.77	
2 µH		0.36	0.76	
5 µH		0.27	0.69	
10 µH		0.29	0.72	
25 µH		0.27	0.65	
50 µH		0.22	0.68	
100 µH		0.21	0.62	
250 µH		0.26	0.22	
500 µH		0.26	0.3	
1 mH		0.29		
2.5 mH		0.25		
5 mH		0.18		
10 mH		0.17		
25 mH		0.25		
0.25 µH	1.2	1.4	1.7	0.8

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

1 $\mu$ H	0.4	0.5	0.9	0.6
10 $\mu$ H	0.4	0.4	0.6	0.1
100 $\mu$ H	0.2	0.2	0.2	

**NVLAP Code:** 20/R10  
Q Standards

Range	Best Uncertainty ( $\pm$ ) in% <sup>note 1</sup>	Remarks
Selected values from 95 to 607	1.2 to 4.5 dependent on Q value and frequency	Frequency range 50 kHz to 45 MHz

**NVLAP Code:** 20/R11  
Voltage/RF-DC Difference  
High Frequency TVC

Range	Best Uncertainty ( $\pm$ ) in percent <sup>note 1</sup>										
	Frequency in Hertz										
	1M	10 M	20 M	30 M	40 M	5 M	60 M	70 M	80 M	90 M	100 M
0.5V	0.03	0.05	0.06	0.13	0.12	0.21	0.24	0.30	0.38	0.39	0.41
1V	0.03	0.05	0.06	0.08	0.12	0.19	0.23	0.30	0.38	0.39	0.75
2V	0.03	0.05	0.06	0.09	0.12	0.19	0.23	0.29	0.37	0.37	0.71
3V	0.03	0.05	0.06	0.09	0.12	0.17	0.23	0.29	0.37	0.37	0.55
5V	0.03	0.05	0.06	0.14	0.51	0.17	0.24	0.30	0.38	0.39	0.44
10 V	0.03	0.06	0.06	0.10	0.12	0.18	0.25	0.31	0.39	0.41	0.43
20 V	0.04	0.06	0.07	0.10	0.12	0.18	0.24	0.30	0.38	0.38	0.45
50V	0.04	0.06		1.15		0.34					0.93

**NVLAP Code:** 20/R12  
RF/Microwave Bolometer Units  
Expanded Uncertainties <sup>notes 1, 2, 3</sup> on Effective Efficiency & Calibration Factor of HP bolometric power sensors

Connector Type	Quantity	Quantity Range	Frequency (MHz)			
			10 - 2000	2000 - 8000	8000 - 12 000	12 000 - 18 000
N	Calibration Factor	0.9 to 1	0.003 - 0.008	0.003 - 0.006	0.004 - 0.007	0.005 - 0.010

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology





# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

APC-3.5	Calibration Factor	0.9 to 1	-----	0.007 - 0.009	0.009 - 0.010	0.010 - 0.011
N	Effective Efficiency	0.9 to 1	0.003 - 0.008	0.003 - 0.005	0.004 - 0.006	0.005 - 0.010
APC-3.5	Effective Efficiency	0.9 to 1	-----	0.007 - 0.008	0.008 - 0.009	0.009 - 0.010

**NVLAP Code:** 20/R13  
RF/Microwave Attenuators

Reflection Coefficient (or Scattering Parameter  $S_{ii}$ )

### A. HP8510 Vector Network Analyzer Uncertainties

1. Expanded Uncertainties <sup>notes 1, 2, 3</sup> on one-port or two-port devices

Connector Type	Quantity	Quantity Range	Frequency (MHz)			
			50-2000	2000-8000	8000-18 000	18 000- 50 000
N	Sii	0 to 1	0.001 - 0.003	0.001 - 0.009	0.004 - 0.021	-----
APC-7	Sii	0 to 1	0.001 - 0.007	0.001 - 0.003	0.001 - 0.007	-----
APC-3.5	Sii	0 to 1	0.001 - 0.007	0.004 - 0.020	0.004 - 0.020	0.004 - 0.020 (to 26.5 GHz)
2.4 mm	Sii	0 to 1	0.002 - 0.006	0.001 - 0.007	0.001 - 0.010	0.001 - 0.032
N	Arg(Sii)	- 180° to + 180°	0.05 - 180	0.36 - 180	1.34 - 180	-----
APC-7	Arg(Sii)	- 180° to + 180°	0.15 - 180	0.16 - 180	0.33 - 180	-----
APC-3.5	Arg(Sii)	- 180° to + 180°	0.53 - 180	0.33 - 180	0.33 - 180	0.33 - 180 (to 26.5 GHz)

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

2.4 mm	Arg(S <sub>ii</sub> )	- 180° to + 180°	0.13 - 180	0.24 - 180	0.14 - 180	0.58 - 180
--------	-----------------------	------------------------	------------	------------	------------	------------

2. HP8510 Expanded Uncertainties <sup>notes 1, 2, 3</sup> on three-port or four-port devices

Connector Type	Quantity	Quantity Range	Frequency (MHz)			
			50 - 2000	2000 - 8000	8000 - 18 000	18 000 - 50 000
N	S <sub>ii</sub>	0 to 0.3	0.001 - 0.006	0.001 - 0.006	0.002 - 0.014	-----
N	Arg(S <sub>ii</sub> )	- 180° to + 180°	0.13 - 180	0.23 - 180	0.57 - 180	-----
N	Γ <sub>ge</sub>	0 to 0.3	0.001 - 0.014	0.002 - 0.019	0.004 - 0.019	-----
N	Arg(Γ <sub>ge</sub> )	- 180° to + 180°	10 - 180	4.6 - 180	6.3 - 180	-----
N	SR wcΔρ <i>note 9</i>	0 to 0.2	0.002 - 0.02	0.002 - 0.5	0.004 - 0.05	-----
APC-7	S <sub>ii</sub>	0 to 0.3	0.001 - 0.005	0.001 - 0.005	0.001 - 0.01	-----
APC-7	Arg(S <sub>ii</sub> )	- 180° to + 180°	0.15 - 180	0.16 - 180	0.5 - 180	-----
APC-7	Γ <sub>ge</sub>	0 to 0.3	0.001 - 0.005	0.001 - 0.007	0.003 - 0.012	-----
APC-7	Arg(Γ <sub>ge</sub> )	- 180° to + 180°	3 - 180	3 - 180	3 - 180	-----
APC-7	SR wcΔρ <i>note 9</i>	0 to 0.2	0.002 - 0.02	0.002 - 0.05	0.004 - 0.05	-----
APC-3.5	S <sub>ii</sub>	0 to 0.3	0.002 - 0.007	0.003 - 0.007	0.003 - 0.01	-----
APC-3.5	Arg(S <sub>ii</sub> )	- 180° to + 180°	0.6 - 180	0.6 - 180	0.6 - 180	-----
APC-3.5	Γ <sub>ge</sub>	0 to 0.3	0.003 - 0.013	0.004 - 0.012	0.004 - 0.12	-----
APC-3.5	Arg(Γ <sub>ge</sub> )	- 180° to + 180°	3 - 180	3 - 180	3 - 180	-----
2.4 mm	S <sub>ii</sub>	0 to 0.3	-----	0.001 - 0.007	0.001 - 0.01	0.003 - 0.02

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

2.4 mm	Arg(S <sub>ii</sub> )	- 180° to + 180°	----	0.24 - 180	0.14 - 180	0.33 - 180
2.4 mm	Γ <sub>ge</sub>	0 to 0.3	----	0.002 - 0.01	0.003 - 0.12	0.003 - 0.03
2.4 mm	Arg(Γ <sub>ge</sub> )	- 180° to + 180°	----	3 - 180	3 - 180	3 - 180

### B. HP8753 Vector Network Analyzer Expanded Uncertainties *notes 1, 2, 3*

#### 1. One-port or two-port devices

Connector Type	Quantity	Quantity Range	Frequency (MHz)		
			0.3 - 25	25 - 1000	1000 - 3000
N	S <sub>ii</sub>	0 to 1	0.001 - 0.013	0.001 - 0.009	0.003 - 0.016
APC-7	S <sub>ii</sub>	0 to 1	0.002 - 0.007	0.002 - 0.04	0.002 - 0.004
APC-3.5	S <sub>ii</sub>	0 to 1	0.001 - 0.004	0.006 - 0.02	0.006 - 0.035
N	Arg(S <sub>ii</sub> )	- 180° to + 180°	0.06 - 180	0.2 - 70	1 - 180
APC-7	Arg(S <sub>ii</sub> )	- 180° to + 180°	0.27 - 180	0.3 - 180	0.2 - 180
APC-3.5	Arg(S <sub>ii</sub> )	- 180° to + 180°	0.27 - 180	1 - 180	1.6 - 180

#### 2. Three-port or four-port devices *notes 1, 2, 3*

Connector Type	Quantity	Quantity Range	Frequency (MHz)	
			0.3 - 25	25 - 3000 (MHz)
N, APC-7, APC-3.5	S <sub>ii</sub>	0 to 0.3	0.001 - 0.006	0.001 - 0.020
N, APC-7, APC-3.5	Arg(S <sub>ii</sub> )	- 180° to + 180°	0.10 - 180	0.09 - 180
N, APC-7, APC-3.5	Γ <sub>ge</sub>	0 to 0.3	0.003 - 0.005	0.002 - 0.03
N, APC-7, APC-3.5	Arg(Γ <sub>ge</sub> )	- 180° to + 180°	10 - 180	1.9 - 180
N, APC-7	SR wcΔp <i>note 9</i>	0 to 0.2	0.001 - 0.006	0.001 - 0.020

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 105002-0**

Attenuation (or Scattering Parameter  $S_{ij}$ )

A. HP8510 Vector Network Analyzer Uncertainties

1. Expanded Uncertainties <sup>notes 1, 2, 3</sup> on two-port devices

Connector Type	Quantity	Quantity Range in dB	Frequency (MHz)			
			50 - 2000	2000 - 8000	8000 - 18 000	18 000 - 50 000
N	$ S_{ij} $	0 to 60	0.01 - 0.12	0.02 - 0.17	0.03 - 0.48	-----
APC-7	$ S_{ij} $	0 to 60	0.01 - 0.08	0.01 - 0.13	0.01 - 0.18	-----
APC-3.5	$ S_{ij} $	0 to 60	0.01 - 0.12	0.02 - 0.22	0.04 - 0.49	-----
2.4 mm	$ S_{ij} $	0 to 60	0.02 - 0.15	0.02 - 0.22	0.03 - 0.34	0.05 - 2.7
N	$\text{Arg}(S_{ij})$	$0^\circ$ to $360^\circ$ $ S_{ij}  < 60$	0.22 - 1.19	0.32 - 1.27	0.36 - 3.46	-----
APC-7	$\text{Arg}(S_{ij})$	$0^\circ$ to $360^\circ$ $ S_{ij}  < 60$	0.22 - 0.73	0.25 - 1.21	0.41 - 2.85	-----
APC-3.5	$\text{Arg}(S_{ij})$	$0^\circ$ to $360^\circ$ $ S_{ij}  < 60$	0.45 - 0.80	0.35 - 1.39	0.41 - 3.17	-----
2.4 mm	$\text{Arg}(S_{ij})$	$0^\circ$ to $360^\circ$ $ S_{ij}  < 60$	-----	1.1 - 1.7	1.1 - 1.7	1.3 - 21.0

2. HP8510 Expanded Uncertainties <sup>notes 1, 2, 3</sup> on three-port or four-port devices

Connector Type	Quantity	Quantity Range in dB	Frequency (MHz)			
			50 - 2000	2000 - 8000	8000 - 18 000	18 000 - 50 000
N, APC-7, APC-3.5	Mainline/ Coupling $ S_{ij} $	0 to 40	0.01 - 0.12	0.02 - 0.12	0.01 - 0.43	-----
N, APC-7, APC-3.5	Isolation $ S_{ij} $	40 to 80	0.2 - 16	0.4 - 9	0.9 - 17	-----
2.4 mm	Mainline/ Coupling $ S_{ij} $	0 to 40	-----	0.03 - 0.15	0.04 - 0.5	0.05 - 0.6
2.4 mm	Isolation $ S_{ij} $	40 to 85	-----	0.4 - 10	1.0 - 12	1.5 - 15

B. HP8753 Vector Network Analyzer Expanded Uncertainties <sup>notes 1, 2, 3</sup>

1. Two-port devices

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

Connector Type	Quantity	Quantity Range	Frequency (MHz)		
			0.3 - 25	25 - 1000	1000 - 3000
N	S <sub>ij</sub>	0 dB to 60 dB	0.02 - 0.25	0.003 - 0.5	0.004 - 1.2
APC-7	S <sub>ij</sub>	0 dB to 60 dB	0.01 - 0.16	0.002 - 0.6	0.003 - 0.9
APC-3.5	S <sub>ij</sub>	0 dB to 60 dB	0.01 - 0.22	0.003 - 0.6	0.003 - 1.0
N, APC-7, APC-3.5	Arg(S <sub>ij</sub> )	- 180° to + 180°	1.4 - 13	0.4 - 10	0.4 - 10

2. Three-port or four-port devices

Connector Type	Quantity	Quantity Range in dB	Frequency (MHz)		
			0.3 to 25	25 to 3000	
N, APC-7, APC-3.5	Mainline/Coupling  S <sub>ij</sub>	0 to 40	0.008 to 0.2	0.020 to 0.3	
N, APC-7, APC-3.5	Isolation  S <sub>ij</sub>	40 to 87	0.1 to 13.6	0.1 to 13.8	

C. Power Ratio Attenuation Expanded Uncertainties <sup>notes 1, 2, 3</sup>

Connector Type	Quantity	Quantity Range in dB	Frequency (MHz)			
			10 - 2000	2000 - 8000	8000 - 12 000	12 000 - 18 000
Fixed Attenuators						
N, APC-7, APC-3.5	S <sub>ij</sub>	0 to 11	0.007 - 0.014	0.008 - 0.021	0.010 - 0.025	0.016 - 0.026
Isolated Step/Variable Attenuators						
N, APC-7, APC-3.5	S <sub>ij</sub>	0 to 11	0.007 - 0.014	0.007 - 0.016	0.007 - 0.015	0.007 - 0.018

NVLAP Code: 20/R16

Group Delay Expanded Uncertainties <sup>notes 1, 2, 3, 4</sup>

Connector Type	Typical Atten. (dB)	Delay (ns)	Frequency (MHz)	
			50 - 1000	1000 - 50 000

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

APC-7, N, APC-3.5, 2.4 mm                      0 - 10                      5 - 500                      0.2 - 0.5                      0.02 - 1.0

**NVLAP Code:** 20/R17  
 RF/Microwave Power Meters  
 Power Meter Certification Uncertainties *notes 2, 3, 4, 10*

A. Low to Medium Power CW Microwave Power Meters at Type N Connector

<i>k</i> = 3 uncertainties Quantity	Quantity Range in dBm	Frequency (MHz)			
		10 to 2000	2000 to 4000	4000 to 12 400	12 400 to 18 000
Power (dBm)	- 30 to - 10	0.09 to 0.41	0.13 to 0.41	0.14 to 0.34	0.16 to 0.46
Power (dBm)	- 10 to 10	0.06 to 0.27	0.10 to 0.25	0.11 to 0.30	-----
Power (dBm)	10 to 30	0.06 to 0.25	0.10 to 0.21	0.11 to 0.24	-----

B. Low Power, Wide Range, CW Microwave Power Meters at Type N Connector

<i>k</i> = 3 uncertainties Quantity	Quantity Range in dBm	Frequency (MHz)		
		10 to 4000	4000 to 8000	8000 to 18 000
Power (dBm)	- 60 to - 50	0.20 to 0.48	0.24 to 0.43	0.24 to 0.43
Power (dBm)	- 50 to - 40	0.17 to 0.33	0.20 to 0.35	0.20 to 0.42
Power (dBm)	- 40 to - 30	0.14 to 0.27	0.16 to 0.32	0.20 to 0.36
Power (dBm)	- 30 to - 20	0.14 to 0.26	0.14 to 0.27	0.18 to 0.35

C. Medium Power CW Microwave Power Meters at Type N Connector

<i>k</i> = 3 uncertainties Quantity	Quantity Range	Frequency (MHz)	
		10 to 2000	2000 to 2500
Power (mW)	1 to 10	1.76 % to 3.30 %	-----
Power (mW)	1 to 160	-----	1.95 % to 4.29 %

D. Medium Power CW Microwave Power Meters at APC-3.5 Connector

Frequency (MHz)

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

<i>k = 3 uncertainties</i>	<i>Quantity</i>	<i>2000 to 4000</i>	<i>4000 to 8000</i>	<i>8000 to 18 000</i>
<i>Quantity</i>	<i>Range</i>			
Power (mW)	0.1 to 8	2.81 % to 4.02 %	2.99 % to 4.92 %	3.97 % to 5.83 %

### E. High Power CW Microwave Power Meters at APC 3.5 Connector

<i>k = 3 uncertainties</i>	<i>Quantity Range</i>	<i>Frequency (MHz)</i>	
<i>Quantity</i>		<i>10 to 300</i>	<i>300 to 3000</i>
Power (Watts)	0.2 to 10	8.97 % to 9.1 %	3.26 % to 10.61 %
Power (Watts)	10 to 200	4.36 % to 10.13 %	9.60 % to 10.60 %

### F. Pulse Power Meters at Type N Connector

<i>k = 3 uncertainties</i>	<i>Quantity</i>	<i>Pulse Width Range</i>	<i>PRF Range</i>	<i>2000 to 18 000</i>
<i>Quantity</i>	<i>Range</i>			
Power (mW)	10 to 100	0.2 $\mu$ s to 5 $\mu$ s	1 Hz to 20 kHz	7.32 % to 8.16 %

## THERMODYNAMICS

NVLAP Code: 20/T02

### Humidity – High

<i>Range in % rH</i>	<i>Best Uncertainty (<math>\pm</math>) in % rH <sup>note 1</sup></i>	<i>Remarks</i>
10 to 50	0.5	
50 to 95	1.0	

### Dew Point

<i>Range</i>	<i>Best Uncertainty (<math>\pm</math>) <sup>note 1</sup></i>	<i>Remarks</i>
(-15 to +20)° C	0.1° C	

### Humidity – Low

<i>Range</i>	<i>Best Uncertainty (<math>\pm</math>) <sup>note 1</sup></i>	<i>Remarks</i>

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

(-80 to 0) ° C Frost Point

0.25 °C

**NVLAP Code:** 20/T04

Leak Artifacts

<i>Range</i>	<i>Best Uncertainty (±) in percent <sup>note 1</sup></i>	<i>Remarks</i>
Gas Leak - PΔV Technique		
1 x 10 <sup>-7</sup> moles/s	0.7	Total gas measurement
1 x 10 <sup>-8</sup> moles/s	0.9	Total gas measurement
1 x 10 <sup>-9</sup> moles/s	1.0	Total gas measurement
1 x 10 <sup>-10</sup> moles/s	1.0	Total gas measurement
Gas Leak - Accumulate - Dump Technique		
1 x 10 <sup>-10</sup> moles/s to 1 x 10 <sup>-14</sup> moles/s	1.0	1 to 200 Atomic Mass Units for any non-reactive, non-hazardous, non-radioactive gas
Gas Leak - Comparison Technique		
1 x 10 <sup>-10</sup> moles/s	2.5	Helium
1 x 10 <sup>-11</sup> moles/s	2.4	Helium
1 x 10 <sup>-12</sup> moles/s	2.3	Helium
1 x 10 <sup>-13</sup> moles/s	2.3	Helium
1 x 10 <sup>-14</sup> moles/s	7.0	Helium

2009-01-01 through 2009-12-31

*Effective dates*

*For the National Institute of Standards and Technology*





# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

*NVLAP Code:* 20/T05

Pressure

**Range** **Best Uncertainty ( $\pm$ ) in ppm <sup>note 1</sup>** **Remarks**

Pneumatic Deadweight Piston Gauges (absolute mode) - Direct Pressure Comparison

0.2 psia to 24 psia [ $\approx$ 1.4 kPa to 170 kPa]	31	Nitrogen
2.0 psia to 70 psia [ $\approx$ 14 kPa to 480 kPa]	28	Nitrogen
52 psia to 1000 psia [ $\approx$ 0.4 MPa to 7.0 MPa]	46	Nitrogen

Pneumatic Deadweight Piston Gauges (gauge mode) - Direct Pressure Comparison

0.2 psig to 24 psig [ $\approx$ 1.4 kPa to 170 kPa]	29	Nitrogen
2.0 psig to 70 psig [ $\approx$ 14 kPa to 480 kPa]	26	Nitrogen
52 psig to 1000 psig [ $\approx$ 0.4 MPa to 7.0 MPa]	44	Nitrogen

Hydraulic Deadweight Piston Gauges (gauge mode) - Direct Pressure Comparison

0.4 kpsig to 4.0 kpsig [ $\approx$ 2.8 MPa to 28 MPa]	44	Oil
2.0 kpsig to 20 kpsig [ $\approx$ 14 MPa to 140 MPa]	61	Oil
4.0 kpsig to 40 kpsig [ $\approx$ 28 MPa to 280 MPa]	59	Oil

Pneumatic Deadweight Piston Gauges - Cross Float (effective area)

0.2 psig to 24 psig [ $\approx$ 14 kPa to 170 kPa]	35	Nitrogen
2.0 psig to 70 psig [ $\approx$ 14 kPa to 480 kPa]	33	Nitrogen
52 psig to 1000 psig [ $\approx$ 0.4 MPa to 7.0 MPa]	46	Nitrogen

Hydraulic Deadweight Piston Gauges - Cross Float (effective area)

0.4 kpsig to 4.0 kpsig [ $\approx$ 2.8 MPa to 28 MPa]	46	Oil
2.0 kpsig to 20 kpsig [ $\approx$ 14 MPa to 140 MPa]	67	Oil
4.0 kpsig to 40 kpsig [ $\approx$ 28 MPa to 280 MPa]	61	Oil

Secondary Pressure - Low Range Absolute

**Range** **Best Uncertainty ( $\pm$ ) in psia <sup>note 1</sup>** **Remarks**

2009-01-01 through 2009-12-31

*Effective dates*

*Sally S. Bruce*

*For the National Institute of Standards and Technology*



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

0.2 psia [ ≈ 1.4 kPa]	0.0013	Nitrogen
1.0 psia [ ≈ 7.0 kPa]	0.0013	Nitrogen
6.0 psia [ ≈ 41 kPa]	0.0017	Nitrogen
10 psia [ ≈ 70 kPa]	0.0021	Nitrogen
15 psia [ ≈ 100 kPa]	0.0028	Nitrogen

### Secondary Pressure - Low Range Gauge or Absolute

<i>Range</i>	<i>Best Uncertainty (±) in psi <sup>note 1</sup></i>	<i>Remarks</i>
20 psi [ ≈ 140 kPa]	0.009	Nitrogen
40 psi [ ≈ 280 kPa]	0.010	Nitrogen
60 psi [ ≈ 410 kPa]	0.011	Nitrogen
80 psi [ ≈ 550 kPa]	0.013	Nitrogen
100 psi [ ≈ 690 kPa]	0.014	Nitrogen

### Secondary Pressure - Mid-Range Gauge or Absolute

200 psi [ ≈ 1.4 MPa]	0.137	Nitrogen
500 psi [ ≈ 3.4 MPa]	0.157	Nitrogen
1.0 kpsi [ ≈ 7.0 MPa]	0.201	Nitrogen
1.5 kpsi [ ≈ 10 MPa]	0.247	Nitrogen
2.0 kpsi [ ≈ 14 MPa]	0.280	Nitrogen
4.0 kpsi [ ≈ 28 MPa]	0.6	Nitrogen
6.0 kpsi [ ≈ 41 MPa]	0.8	Nitrogen
8.0 kpsi [ ≈ 55 MPa]	1.0	Nitrogen
10 kpsi [ ≈ 70 MPa]	1.0	Nitrogen

**NVLAP Code:** 20/T07  
Resistance Thermometry

<i>Range in °C</i>	<i>Best Uncertainty (±) in m °C <sup>note 1</sup></i>	<i>Material/Equilibrium State</i>
- 189.3442	0.53	Ar/Triple Point
- 38.8344	0.30	Hg/Triple Point

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

0.01	0.16	H <sub>2</sub> O/Triple Point
29.7646	0.12	Ga/Melting Point
156.5985	2.00	In/Freezing Point
231.928	0.92	Sn/Freezing Point
419.527	1.10	Zn/Freezing Point
660.323	5.0	Al/Freezing Point

### Standard Platinum Resistance Thermometer Calibrations

- 189.3442	1.6	Ar/Triple Point
- 38.8344	0.7	Hg/Triple Point
0.01	0.6	H <sub>2</sub> O/Triple Point
29.7646	1.0	Ga/Melting Point
156.5985	2.6	In/Freezing Point
231.928	1.8	Sn/Freezing Point
419.527	2.1	Zn/Freezing Point
660.323	5.2	Al/Freezing Point

### Comparison Calibrations

<i>Range in °C</i>	<i>Best Uncertainty (±) in °C<sup>note 1</sup></i>	<i>Type of Device</i>
- 80 to 0	0.10	Thermocouples
10 to 150	0.10	Thermocouples
150 to 660	0.22	Thermocouples
660 to 700	0.47	Thermocouples
700 to 1100	2.5	Thermocouples
1100 to 1300	2.8	Thermocouples
- 80 to 0	0.06	RTD/IPRT/PRT
10 to 150	0.09	RTD/IPRT/PRT
150 to 660	0.21	RTD/IPRT/PRT
- 80 to 0	0.05	Liquid in Glass
10 to 150	0.06	Liquid in Glass
- 80 to 0	0.06	Thermistors
10 to 150	0.09	Thermistors

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

Thermocouple Simulator/Readout Calibration Methods

Type	ITS-90 Temperature Range (°C)	Best Uncertainty (±) in °C <sup>notes 1,7</sup>	NIST Monograph 175 Reference Table <sup>note 8</sup>
K	- 200 to 1370	0.10 to 0.30	7.3.3
J	- 200 to 1200	0.08 to 0.22	6.3.3
E	- 240 to 1000	0.07 to 0.38	5.3.3
T	- 240 to 400	0.09 to 0.53	9.3.3
R	- 50 to 1750	0.38 to 1.09	3.3.3
S	- 50 to 1750	0.43 to 1.02	4.3.3
B	100 to 1750	0.43 to 4.45	2.3.3
C	0 to 2300	0.24 to 0.82	
N	-200 to 1300	0.5	8.3.3

NVLAP Code: 20/T10

Vacuum

Range	Best Uncertainty (±) in percent <sup>note 1</sup>	Remarks
Ionization Gage Reference for direct comparison		
$1.3 \times 10^{-6} \text{ Pa} < \text{reading} \leq 1.3 \times 10^{-5} \text{ Pa}$	4.8	N <sub>2</sub> ; 10 <sup>-8</sup> Torr
$1.3 \times 10^{-5} \text{ Pa} < \text{reading} \leq 1.3 \times 10^{-4} \text{ Pa}$	4.7	N <sub>2</sub> ; 10 <sup>-7</sup> Torr
$1.3 \times 10^{-4} \text{ Pa} < \text{reading} \leq 1.3 \times 10^{-3} \text{ Pa}$	4.7 to 2.5	N <sub>2</sub> ; 10 <sup>-6</sup> Torr

Spinning Rotor Gage Reference for direct comparison

$1.3 \times 10^{-4} \text{ Pa} < \text{reading} \leq 1.3 \times 10^{-3} \text{ Pa}$	4.3 to 2.1	N <sub>2</sub> ; 10 <sup>-6</sup> Torr
$1.3 \times 10^{-3} \text{ Pa} < \text{reading} \leq 1.3 \text{ Pa}$	2.1	N <sub>2</sub> ; 10 <sup>-5</sup> Torr to 10 <sup>-3</sup> Torr
$1.3 \text{ Pa} \leq \text{reading} \leq 13. \text{ Pa}$	2.2	N <sub>2</sub> ; 10 <sup>-3</sup> Torr

Capacitance Diaphragm Gage Reference for direct comparison

$1.3 \times 10^{-1} \text{ Pa} \leq \text{reading} \leq 13.3 \text{ Pa}$	2.1 to 0.7	N <sub>2</sub> ; 0.1 Torr
--	------------	---------------------------

Force Balance and Deadweight Piston Gauges - direct pressure comparison (Absolute, Absolute Differential and Gauges Modes)

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 105002-0

Range	Best Uncertainty ( $\pm$ ) <sup>note 1</sup>	Remarks
0.3 Pa to < 6.0 Pa	30 ppm + 0.025 Pa	N <sub>2</sub> ; 0.1 Torr range
> 6.0 Pa to 15.0 kPa	30 ppm + 0.008 Pa	N <sub>2</sub> ; 0.1, 1, 10 100 Torr range
15.0 kPa to 133.3 kPa	31 ppm	N <sub>2</sub> ; 1000 Torr range

Secondary Capacitance Diaphragm Gages Reference for direct comparison (Note: uncertainty in % of reading)

Range	Best Uncertainty ( $\pm$ )in percent <sup>note 1</sup>	Remarks
$1.3 \times 10^{-1}$ Pa $\leq$ 13.3 Pa	2.3 to 0.5	N <sub>2</sub> ; 0.1 Torr range
13.3 Pa $\leq$ reading $\leq$ 133.3	0.6	N <sub>2</sub> ; 1 Torr range
133.3 Pa $\leq$ reading $\leq$ 1.3 kPa	0.4	N <sub>2</sub> ; 10 Torr range
1.3 kPa $\leq$ reading $\leq$ 13.3 kPa	0.3	N <sub>2</sub> ; 100 Torr range
13.3 kPa $\leq$ reading $\leq$ 133.3 kPa	0.3 to 0.05	N <sub>2</sub> ; 1000 Torr range

1. Represents uncertainty with coverage factor of  $k = 2$ , unless otherwise specified.
2. Approximate value. Actual value determined by test results.
3. The uncertainty ranges are the lowest and highest uncertainty ( $\pm$ ) values within the specified frequency range and quantity range.
4. Certification uncertainty consists of an appropriate combination of the measurement uncertainty (which includes all significant sources of uncertainty associated with the calibration process) and uncertainties due to use, environment, handling or variation with time over the certification interval.
5. ASTM loading range classes (e.g., A, AA) are not used or reported.
6. Uncertainties listed are linearized forms ( $A' + B'L$ ) of uncertainties calculated as root sum squares of constant and length-dependent terms  $\{A^2 + (BL)^2\}^{1/2}$ .  $A'$  and  $B'$  are calculated by fitting a straight line through the RSS uncertainty values at the upper and lower limits of range.
7. Uncertainty is dependent on the specific temperature point tested.

2009-01-01 through 2009-12-31

Effective dates

*Sally S. Bruce*

For the National Institute of Standards and Technology



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 105002-0**

- 8. Referenced tables in NIST Monograph 175 (April 1993) provide values for emf E output/input of the thermocouple simulator/readout and the Seebeck coefficient S for the specific temperature points with the specified ranges. The best uncertainty (at  $k = 2$ ) of the emf E in  $\mu\text{V}$  is equal to the product of  $U * S$ , where U is the best uncertainty (at  $k = 2$ ) of the temperature point tested.
- 9. SR  $w\Delta\rho$  is the worst-case error in measurement of reflection coefficient,  $\rho$ , ( $0.5 \geq \rho \geq 0$ ) as made by a 4-port scalar reflectometer which has rf access to all 4 ports. Uncertainty shown in the table is the uncertainty on value of SR  $w\Delta\rho$ .
- 10. User mismatch uncertainty not included.
- 11. Using multiple double substitutions or redundant weighing designs.
- 12. Represents uncertainty with coverage factor of  $k = 3$ .

2009-01-01 through 2009-12-31

*Effective dates*

*For the National Institute of Standards and Technology*