



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Scantek, Inc. Calibration Laboratory
 7060 Oakland Mills Rd. #L
 Columbia, MD 21046
 Mr. Richard J. Peppin
 Phone: 410-290-7726 Fax: 410-290-9167
 E-mail: peppinr@scantekinc.com
 URL: http://www.scantekinc.com

CALIBRATION LABORATORIES

NVLAP LAB CODE 200625-0

NVLAP Code: 20/A01 ANSI/NCSL Z540-1-1994; Part 1 Compliant

NVLAP Code: 20/E05
 DC Current

<i>Signal Level Range</i>	<i>Best Uncertainty (±) (% of reading + % of range) ^{notes 1, 4}</i>	<i>Remarks</i>
1 mA to 10 mA	0.1 + 0.02	34401A
10 mA to 100 mA	0.12 + 0.005	34401A
100 mA to 1 mA	0.25 + 0.01	34401A
1 A to 3 A	0.2 + 0.02	34401A

DC Resistance (4-wires)

<i>Signal Level Range</i>	<i>Best Uncertainty (±) (% of reading + % of range) ^{notes 1, 4}</i>	<i>Remarks</i>
1 mΩ to 100 Ω	0.1 + 0.006	34401A
100 Ω to 1 MΩ	0.1 + 0.004	34401A

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200625-0

NVLAP Code: 20/E06

DC Voltage

Signal Level Range	Best Uncertainty (\pm) (% of reading + % of range) ^{notes 1, 4}	Remarks
1 mV to 100 mV	0.0055 + 0.0035	34401A
100 mV to 1 V	0.0046 + 0.0007	34401A
1 V to 10 V	0.0042 + 0.0005	34401A
10 V to 100 V	0.005 + 0.0006	34401A
100 V to 1000 V	0.005 + 0.0010	34401A

NVLAP Code: 20/E09

AC Voltage

Measuring equipment: generate

Signal Level Range	Frequency Domain	Frequency Best Uncertainty (\pm) ^{notes 1, 4}	Level Best Uncertainty (\pm) ^{notes 1, 4}	Remarks
10 μ V _{pp} to 40 V _{pp}	0.001 Hz to 200 kHz	54 ppm + 4 mHz	1.1	SR DS360 Generate: Sine Square, two-tone
10 μ V _{pp} to 40 V _{pp}	0.001 Hz to 200 kHz ON: 1/2, 1 to 65534 periods Rep. Rate 1 to 65535 Off atten to 10 kHz: >70 dB	54 ppm + 4 mHz	1.1	SR DS360 Generate: Sine or square bursts

AC Voltage: true RMS

Range	Frequency Domain	Best Uncertainty (\pm) ^{notes 1, 4}	Remarks
50 μ V to 1 mV	20 Hz to 20 kHz	0.2 dB	N-1504A System
	3 Hz to 100 kHz	0.3 dB	N-1504A System

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200625-0

Range	Frequency Domain	Best Uncertainty (\pm)(% of reading + % of range) ^{notes 1, 4}	Remarks
1 mV to 100 mV	3 Hz to 5 Hz	1.00 + 0.04	34401A
	5 Hz to 10 Hz	0.37 + 0.04	34401A
	10 Hz to 20 kHz	0.13 + 0.04	34401A
	20 kHz to 50 kHz	0.17 + 0.05	34401A
	50 kHz to 100 kHz	0.62 + 0.08	34401A
	100 kHz to 300 kHz	4.0 + 0.50	34401A
100 mV to 750 V	3 Hz to 5 Hz	1.00 + 0.03	34401A
	5 Hz to 10 Hz	0.37 + 0.03	34401A
	10 Hz to 20 kHz	0.13 + 0.03	34401A
	20 kHz to 50 kHz	0.16 + 0.05	34401A
	50 kHz to 100 kHz	0.62 + 0.08	34401A
	100 kHz to 300 kHz	4.0 + 0.50	34401A

Self Generated Noise	Frequency Domain	Best Uncertainty (\pm) in dB ^{notes 1, 4}	Remarks
1 μ V to 10 V	0.1 Hz to 20 kHz	0.85	840 RTA
>30 μ V	20 Hz to 80 kHz	2	HP 8903A

Parameter Range	Frequency Domain	Signal Level Range	Best Uncertainty (\pm) ^{notes 1, 4}	Remarks
Signal distortion >0.01%	20 kHz to 100 kHz	50 mV to 300 V	2 dB	HP 8903A
Signal to noise ratio <80 dB	50 Hz to 100 kHz	50 mV to 300 V	1 dB	HP 8903A

TIME & FREQUENCY

NVLAP Code: 20/F01

Frequency dissemination: Frequency and period

Signal Level Range	Frequency Domain	Best Uncertainty (\pm) in % of reading ^{notes 1, 4}	Remarks
100 mV to 750 V	3 Hz to 5 Hz	0.1	34401A
	5 Hz to 10 Hz	0.05	34401A
	10 Hz to 40 Hz	0.03	34401A
	40 Hz to 300 kHz	0.01	34401A

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200625-0

Parameter Range	Best Uncertainty (\pm) ^{notes 1, 4}	Remarks
Time intervals >8s	1 s	Chronometer

MECHANICAL

NVLAP Code: 20/M01

Acoustic

Sensitivity or open circuit sensitivity: direct and comparison methods

Range	Frequency Domain in Hz	Best Uncertainty (\pm) in dB ^{notes 1,5}	Remarks
-70 dB to <-50 dB re 1 V/Pa (0.3 mV/Pa to 3 mV/Pa)	250	0.12	Comparison and direct
-50 dB to 0 dB re 1 V/Pa (3 mV/Pa to 1V/Pa)	250	0.09	Comparison and direct
-70 dB to <-50 dB re 1 V/Pa (0.3 mV/Pa to 3 mV/Pa)	1 k	0.13	Comparison
-50 dB to 0 dB re 1 V/Pa (3 mV/Pa to 1 V/Pa)	1 k	0.11	Comparison

Frequency Response: electrostatic excitation (applicable for condenser microphones with removable grid)

Range	Frequency Domain in Hz	Best Uncertainty (\pm) in dB ^{notes 1, 4}		Remarks
		Actuator response	Free-field and diffuse field responses ^{note 2}	
-70 dB to 0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	31.5 to 100	0.20	0.20	
-70 dB to 0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	100 to 1.25 k	0.14	0.18	
-70 dB to 0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	1.25 k to 4 k	0.14	0.23	
-70 dB to 0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	4 k to 8 k	0.17	0.45	
-70 dB to 0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	8 k to 10 k	0.38	0.57	
-70 dB to 0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	10 k to 16 k	0.38	0.77	

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200625-0

-70 dB to 0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	16 k to 20 k	0.59	0.89
-70 dB to 0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	20 k to 50 k	0.8	2.1
-70 dB to 0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	50 k to 100 k	1.1	4.2

Frequency Response: acoustical method

Microphone Sensitivity Range	Frequency Domain in Hz	Best Uncertainty (\pm) in dB ^{notes 1, 4}		Remarks
		Pressure response	Free-field and diffuse field responses ^{note 2}	
-70 dB to <0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	31.5	0.3	0.3	4226 Calibrator
-70 dB to <0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	63	0.14	0.24	4226 Calibrator
-70 dB to <0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	125	0.14	0.24	4226 Calibrator
-70 dB to <0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	250	0.09	0.15	1253 Calibrator
-70 dB to <0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	500	0.14	0.16	4226 Calibrator
-70 dB to <0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	1k	0.11	0.13	1253 Calibrator
-70 dB to <0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	2 k	0.14	0.24	4226 Calibrator
-70 dB to <0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	4 k	0.14	0.32	4226 Calibrator
-70 dB to <0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	8 k	0.14	0.51	4226 Calibrator

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200625-0

-70 dB to <0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	12.5 k	0.14	1.00	4226 Calibrator
-70 dB to <0 dB re 1 V/Pa (0.3 mV/Pa to 1 V/Pa)	16 k	0.21	1.45	4226 Calibrator

Acoustical Calibrators and Pistonphones

Sound pressure level (SPL)

SPL Range	Frequency Domain	Best Uncertainty (\pm) in dB ^{note 1}	Remarks
90 dB to 170 dB re 20 μ Pa	31.5 Hz (\pm 2 Hz)	0.40	Comparison
	63 Hz to 800 Hz	0.10	Direct Method
	250 Hz (\pm 10 Hz)	0.095 ^{note 3}	Direct Method
	250 Hz (\pm 10 Hz)	0.092 ^{note 4}	Direct Method
	1 kHz (\pm 40 Hz)	0.11 ^{note 3}	Direct Method
	1 kHz (\pm 40 Hz)	0.10 ^{note 4}	Direct Method
	1250 Hz to 5 kHz	0.11	Direct Method
	6.3 kHz to 8 kHz	0.14	Direct Method
	10 kHz to 12.5 kHz	0.16	Direct Method
16 kHz	0.21	Direct Method	

Sound Frequency

Range	SPL Domain	Best Uncertainty (\pm)(% of reading) ^{notes 1,4}	Remarks
10 Hz to 40 Hz	70 dB to 170 dB re 20 μ Pa	0.03	34401A
40 Hz to 20 kHz	70 dB to 170 dB re 20 μ Pa	0.01	34401A

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200625-0

Frequency Stability

Stability Range	Frequency Domain	Best Uncertainty(±) in % ^{notes 1,4}	Remarks
>0.01%	10 Hz to 20 kHz	5% of the measured values	

Sound Distortion

Distortion Range	Frequency Domain	Sound Pressure Level Domain	Best Uncertainty (±) in % distortion	Remarks
>0.1%	10 Hz to 20 kHz	50 dB to 170 dB re 20 µPa	0.1	Analyzer Method

Sound Level Meters, Dosimeters, Real-time and FFT analyzers, Filter sets
Acoustical Tests
Sound Pressure Level: fixed points, tones

SPL Range in dB	Frequency Domain in Hz	Best Uncertainty (±) in dB ^{note 1}		Remarks
		Pressure Response	Free-field and diffuse field responses ^{note 2}	
94, 104, 114	31.5	0.3	0.3	4226 Calibrator
94, 104, 114	63	0.14	0.23	4226 Calibrator
94, 104, 114	125	0.14	0.23	4226 Calibrator
94, 104, 114	250	0.14	0.15	4226 Calibrator
124	250	0.09	0.11	1253 Calibrator
94, 104, 114	500	0.14	0.15	4226 Calibrator
94, 104, 114	1 k	0.13	0.13	4226 Calibrator
124	1 k	0.12	0.13	1253 Calibrator
94, 104, 114	2 k	0.14	0.23	4226 Calibrator
94, 104, 114	4 k	0.14	0.32	4226 Calibrator
94, 104, 114	8 k	0.14	0.50	4226 Calibrator
94, 104, 114	12.5 k	0.14	1.00	4226 Calibrator
94, 104, 114	16 k	0.3	1.45	4226 Calibrator

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200625-0

Tone Burst Sounds

SPL Range	Frequency Domain	Best Uncertainty (\pm) in $dB^{note 1}$	Remarks
80 dB to 100 dB re 20 μ Pa	2 kHz	0.15	4226 Calibrator

Timed sounds

Measuring Equipment (generate)

SPL Range in dB	Frequency Domain in kHz	Best Uncertainty (\pm) in % dose^{note 1}	Remarks
94	1	0.01	
104	1	0.03	
114	1	0.12	
124	1	0.26	

NVLAP Code: 20/M11

Vibration

Exciters and Vibration Calibrators

Acceleration: measure

Level Range	Frequency Domain	Best Uncertainty (\pm) in %^{note 1, 4}	Remarks
0.1 m/s ² to 200 m/s	40 Hz to 400 Hz	1.2	
0.1 m/s ² to 200 m/s	20 Hz to 5 kHz	1.8	
0.1 m/s ² to 200 m/s	>5 kHz to 10 kHz	3.0	
0.1 m/s ² to 200 m/s	10 Hz to 16 Hz	12	

Motion Frequency

Range in Hz	Acceleration Domain	Best Uncertainty (\pm) in % of reading^{notes 1, 4}	Remarks
3 to 5	0.01 m/s ² to 200 m/s ²	0.16	33401A
5 to 10	0.01 m/s ² to 200 m/s ²	0.18	33401A
10 to 40	0.01 m/s ² to 200 m/s ²	0.20	33401A
40 to 100	0.01 m/s ² to 200 m/s ²	0.06	33401A

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200625-0

100 to 300	0.01 m/s ² to 200 m/s ²	0.03	33401A
300 to 10 k	0.01 m/s ² to 200 m/s ²	0.01	33401A

Distortion of Motion

<i>Distortion Range</i>	<i>Frequency Domain</i>	<i>Vibration Level Domain</i>	<i>Best Uncertainty (±) in % notes 1,4</i>	<i>Remarks</i>
<0.1%	10 Hz to 10 kHz	0.1 m/s ² to 200 m/s ²	3% of measured value	Analyzer Method

Accelerometers

Accelerometer sensitivity: comparison method

<i>Range Charge</i>	<i>Frequency Domain</i>	<i>Best Uncertainty (±) in % notes 1,4</i>	<i>Remarks</i>
	10 Hz	12.0	
	12.5 Hz	7.2	
	16 Hz	3.7	
	20 Hz to 13.5 Hz	1.6	
	40 Hz to 400 Hz	0.9	
	>400 Hz to 5 kHz	1.3	
	>5 kHz to 10 kHz	2.9	
<i>Voltage</i>			
	10 Hz	12.0	
	12.5 Hz	7.3	
	16 Hz	3.7	
	20 Hz to 31.5 Hz	1.7	
	40 Hz to 400 Hz	1.1	
	400 Hz to 5 kHz	1.4	
	>5 kHz to 10 kHz	2.9	

2008-04-01 through 2009-03-31

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200625-0

-
1. Represents an expanded uncertainty using a coverage factor $k = 2$, at an approximate level of confidence of 95%.
 2. These characteristics are calculated using the measured actuator/pressure response and the correction coefficients provided by the manufacturer of the tested device.
 3. At reference conditions.
 4. At actual conditions.

2008-04-01 through 2009-03-31

Effective dates

A handwritten signature in cursive script that reads 'Sally S. Bruce'.

For the National Institute of Standards and Technology