



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Oak Ridge Metrology Center
P. O. Box 2009
Oak Ridge, TN 37831-8091
G. R. (Richard) Jones
Phone: 865-574-1549 Fax: 865-574-2802
E-mail: jonesgr@y12.doe.gov

CALIBRATION LABORATORIES

NVLAP LAB CODE 105000-0
Scope Revised: 2008-08-28

DIMENSIONAL

NVLAP Code: 20/D03
Gage Blocks, Steel and Chrome Only

<i>Range</i>	<i>Best Uncertainty (±) ^{note 1}</i>	<i>Remarks</i>
0.010 in to 0.090 in	3.6 µin	Mechanical Comparison
> 0.090 in to 1.000 in	3.0 µin	Mechanical Comparison
> 1.0 in to 4.0 in	(2.7 + 0.31L) µin ^{note 3}	Mechanical Comparison
0.30 mm to 2.5 mm	0.09 µm	Mechanical Comparison
> 2.50 mm to 25 mm	0.08 µm	Mechanical Comparison
> 25 mm to 100 mm	(0.07 + 0.31L) µm ^{note 4}	Mechanical Comparison

NVLAP Code: 20/D05
Length

<i>Range in m</i>	<i>Best Uncertainty (±) ^{notes 1, 4}</i>	<i>Remarks</i>
0 to 0.8	(0.32 + 0.37L) µm	Moore M32 CMM
0 to 1.2	(0.32 + 0.46L) µm	Moore M48 CMM

NVLAP Code: 20/D06
Line Standards

<i>Range in mm</i>	<i>Best Uncertainty (±) ^{notes 1, 4}</i>	<i>Remarks</i>
0 to 150	(0.32 + 0.20L) µm	CMM (optical)
>150 to 600	(0.16 + 1.6L) µm	CMM (optical)

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 105000-0

Scope Revised: 2008-08-28

NVLAP Code: 20/D08

Optical Grid Plates/Reference Planes

Range

Field of view ^{note 5}

Best Uncertainty (\pm) ^{notes 1, 4}

0.2 μm

Remarks

CMM (optical), Measurements taken within camera field of view

Range in mm

0 to 848

Best Uncertainty (\pm) ^{notes 1, 4}

(0.54 + 0.40L) μm

Remarks

CMM (optical), maximum length and width (600 x 600) mm

NVLAP Code: 20/D09

Roundness

Range

to 6 inches Diameter and 4 inches Height

Best Uncertainty (\pm) ^{note 1}

0.1 μm

Remarks

Roundness Instrument

NVLAP Code: 20/D12

Surface Texture

Range

41 μin to 120 μin (1.04 μm to 3.05 μm)
13 μin to 40 μin (0.33 μm to 1.02 μm)
12 μin (0.31 μm)

Best Uncertainty (\pm) ^{note 1}

5.03 μin (0.13 μm)
1.74 μin (0.044 μm)
0.85 μin (0.021 μm)

Remarks

Ra (Roughness Average)
Ra (Roughness Average)
Ra (Roughness Average)

NVLAP Code: 20/D15

Two Dimensional Gages

Range in m

0 to 0.9 Diagonal
0 to 1.3 Diagonal

Best Uncertainty (\pm) ^{notes 1, 4}

(0.45 + 0.52L) μm
(0.45 + 0.66L) μm

Remarks

Moore M32 CMM
Moore M48 CMM

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 105000-0

Scope Revised: 2008-08-28

NVLAP Code: 20/D16
Coordinate Measuring Machines

Range in m	Best Uncertainty (\pm) ^{notes 1, 4}	Remarks
0 to 1.0 Diagonal	(0.55 + 0.64L) μm	Spatial Moore M32 CMM
0 to 1.4 Diagonal	(0.55 + 0.80L) μm	Spatial Moore M48 CMM

NVLAP Code: 20/D18
Gears

Range	Best Uncertainty (\pm) ^{note 1}	Remarks
to 14 inches Diameter	0.9 μm	CMM for Involute Profile
to 6 inches Diameter and Infinite Lead	0.8 μm	CMM for Helix
to 6 inches Diameter and 99 inches Lead	0.9 μm	CMM for Helix
to 6 inches Diameter and 32 inches Lead	1.1 μm	CMM for Helix
to 6 inches Diameter and 16 inches Lead	1.2 μm	CMM for Helix
to 6 inches Diameter and 11 inches Lead	1.3 μm	CMM for Helix
to 6 inches Diameter (pin offset)	0.7 μm	CMM for Pin Master
to 6 inches Diameter (pin diameter)	0.5 μm	CMM for Pin Master
to 6 inches Diameter (pin roundness)	0.3 μm	Federal Formscan for Pin Master
to 24 inches Diameter	1.6 arcseconds	CMM for Index and Runout

TIME AND FREQUENCY

NVLAP Code: 20/F01
Frequency Dissemination

Range	Best Uncertainty (\pm) ^{note 1}	Remarks
1 MHz, 5 MHz, 10 MHz	8.8 x 10 ⁻¹²	NIST FMAS

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 105000-0

Scope Revised: 2008-08-28

MECHANICAL

NVLAP Code: 20/M08

Mass

Range	Best Uncertainty (\pm) ^{note 1}	Remarks
30 kg	22 mg	Echelon II
25 kg	21 mg	Echelon II
20 kg	14 mg	Echelon II
10 kg	14 mg	Echelon II
5 kg	2.6 mg	Echelon II
3 kg	1.9 mg	Echelon II
2 kg	1.9 mg	Echelon II
1 kg	0.13 mg	Echelon II

THERMODYNAMIC

NVLAP Code: 20/T05

Pressure

Range	Best Uncertainty (\pm) ppm ^{note 1}	Remarks
Pneumatic Deadweight Piston Gauge (absolute Mode) - Direct Pressure Comparison		
1.2 psia to 23.6 psia (8.3 kPa to 162.7 kPa)	500	Nitrogen
5.7 psia to 95.6 psia (39.3 kPa to 659.1 kPa)	101	Nitrogen
41.9 psia to 1001.6 psia (288.9 kPa to 6905.8 kPa)	45	Nitrogen
Pneumatic Deadweight Piston Gauge (Gauge Mode) - Direct Pressure Comparison		
1.2 psig to 23.6 psig (8.3 kPa to 162.7 kPa)	26	Nitrogen
5.7 psig to 95.6 psig (39.3 kPa to 659.1 kPa)	22	Nitrogen
41.9 psig to 1001.6 psig (288.9 kPa to 6905.8 kPa)	43	Nitrogen
Hydraulic Deadweight Piston Gauge (Gauge Mode) - Direct Comparison		
203 psig to 3771 psig (1.4 MPa to 26 MPa)	60	Oil
2031 psig to 19 870 psig (14 MPa to 137 MPa)	70	Oil

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 105000-0

Scope Revised: 2008-08-28

4061 psig to 39 595 psig (28 MPa to 273 MPa)

70

Oil

NVLAP Code: 20/T07

Resistance Temperature Devices

Range in °C

0.01 to 29.7646

Best Uncertainty (\pm)^{note 1}

0.0027 °C

Remarks

Comparison

-
1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.
 2. Realizable uncertainty depends on frequency being measured, customer requirements, and suitability of customer's equipment.
 3. L is length in inches.
 4. L is length in meters.
 5. Glass Reticles, Stage Micrometers, Glass Magnification Scales, Orthogonality Standards, and Calibration Charts.

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology