



National Voluntary Laboratory Accreditation Program



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

National Standards of Puerto Rico
 281 Jesus T. Pintero Avenue
 P.O. Box 11936
 Rio Piedras, PR 00922-1936
 Mr. Denio Hernandez Osorio
 Phone: 787-765-5564 Fax: 787-751-5920
 E-mail: denio@inscointernational.com

CALIBRATION LABORATORIES

NVLAP LAB CODE 200454-0
 Scope Revised: 2008-10-16

NVLAP Code: 20/A01 ANSI/NCSL Z540-1-1994; Part 1 Compliant

MECHANICAL

NVLAP Code: 20/M08
 Mass

<i>Nominal Value</i> ^{note 6}	<i>Best Uncertainty</i> (\pm) ^{note 1}	<i>Remarks</i>
20 kg	5.7 mg	Echelon I
10 kg	1.5 mg	Echelon I
5 kg	0.75 mg	Echelon I
3 kg	0.75 mg	Echelon I
2 kg	0.24 mg	Echelon I
1 kg	49 μ g	Echelon I
500 g	28 μ g	Echelon I
300 g	28 μ g	Echelon I
200 g	19 μ g	Echelon I
100 g	15 μ g	Echelon I
50 g	9.6 μ g	Echelon I
30 g	9.6 μ g	Echelon I
20 g	6.0 μ g	Echelon I
10 g	4.3 μ g	Echelon I
5 g	1.7 μ g	Echelon I
3 g	1.7 μ g	Echelon I

2008-07-01 through 2009-06-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200454-0

Scope Revised: 2008-10-16

2 g	1.2 µg	Echelon I
1 g	1.0 µg	Echelon I
500 mg	0.58 µg	Echelon I
300 mg	0.58 µg	Echelon I
200 mg	0.44 µg	Echelon I
100 mg	0.43 µg	Echelon I
50 mg	0.51 µg	Echelon I
30 mg	0.51 µg	Echelon I
20 mg	0.42 µg	Echelon I
10 mg	0.47 µg	Echelon I
5 mg	0.41 µg	Echelon I
3 mg	0.41 µg	Echelon I
2 mg	0.20 µg	Echelon I
1 mg	0.26 µg	Echelon I
20 kg	5.5 mg	Echelon II
10 kg	1.9 mg	Echelon II
5 kg	0.80 mg	Echelon II
3 kg	0.80 mg	Echelon II
2 kg	0.37 mg	Echelon II
1 kg	48 µg	Echelon II
500 g	30 µg	Echelon II
300 g	30 µg	Echelon II
200 g	22 µg	Echelon II
100 g	21 µg	Echelon II
50 g	10 µg	Echelon II
30 g	10 µg	Echelon II
20 g	7.1 µg	Echelon II
10 g	6.1 µg	Echelon II
5 g	2.7 µg	Echelon II
3 g	2.7 µg	Echelon II
2 g	2.6 µg	Echelon II
1 g	2.6 µg	Echelon II
500 mg	3.2 µg	Echelon II

2008-07-01 through 2009-06-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200454-0

Scope Revised: 2008-10-16

300 mg	3.2 µg	Echelon II
200 mg	1.8 µg	Echelon II
100 mg	2.7 µg	Echelon II
50 mg	2.0 µg	Echelon II
30 mg	2.0 µg	Echelon II
20 mg	2.1 µg	Echelon II
10 mg	1.9 µg	Echelon II
5 mg	1.7 µg	Echelon II
3 mg	1.7 µg	Echelon II
2 mg	1.5 µg	Echelon II
1 mg	1.5 µg	Echelon II
Single Substitution		
20 kg	5.7 mg	Echelon III
10 kg	1.7 mg	Echelon III
5 kg	0.79 mg	Echelon III
3 kg	0.79 mg	Echelon III
2 kg	0.27 mg	Echelon III
1 kg	51 µg	Echelon III
500 g	35 µg	Echelon III
300 g	35 µg	Echelon III
200 g	32 µg	Echelon III
100 g	22 µg	Echelon III
10 g	4.7 µg	Echelon III
5 g	3.1 µg	Echelon III
3 g	3.1 µg	Echelon III
2 g	1.3 µg	Echelon III
1 g	1.9 µg	Echelon III
Modified Substitution		
50 kg	14 mg	Echelon III
30 kg	14 mg	Echelon III
20 kg	14 mg	Echelon III
10 kg	1.8 mg	Echelon III

2008-07-01 through 2009-06-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200454-0

Scope Revised: 2008-10-16

5 kg	1.4 mg	Echelon III
3 kg	1.2 mg	Echelon III
2 kg	1.2 mg	Echelon III
1 kg	1.2 mg	Echelon III
500 g	0.12 mg	Echelon III
300 g	0.12 mg	Echelon III
200 g	0.12 mg	Echelon III
100 g	0.12 mg	Echelon III
50 g	0.12 mg	Echelon III
30 g	10 µg	Echelon III
20 g	7.3 µg	Echelon III
10 g	6.0 µg	Echelon III
5 g	5.3 µg	Echelon III
3 g	18 µg	Echelon III
2 g	18 µg	Echelon III
1 g	18 µg	Echelon III
500 mg	18 µg	Echelon III
300 mg	6.1 µg	Echelon III
200 mg	6.0 µg	Echelon III
100 mg	6.0 µg	Echelon III
50 mg	6.0 µg	Echelon III
30 mg	6.0 µg	Echelon III
20 mg	6.0 µg	Echelon III
10 mg	6.0 µg	Echelon III
5 mg	6.0 µg	Echelon III
3 mg	6.0 µg	Echelon III
2 mg	6.0 µg	Echelon III
1 mg	6.0 µg	Echelon III

2008-07-01 through 2009-06-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200454-0
Scope Revised: 2008-10-16

THERMODYNAMIC

NVLAP Code: 20/T03
Liquid in Glass Thermometers *note 2*

Range in °C	Best Uncertainty (±) in mk <i>note 1</i>
-80 to -40	52.2
-40 to -20	52.9
-20 to 0	54.9
0 (Ice Point)	52.2
0 to 50	55.1
50 to 150	53.4
150 to 200	53.4
200 to 250	54.9

NVLAP Code: 20/T04
Platinum Resistance Thermometers by Comparison *note 3*

Range in °C	Best Uncertainty (±) in mk <i>note 1</i>
-196 to -80	9.6
-80 to -40	5.7
-40 to -20	10.3
-20 to 0	17.9
0.01	2.0
0 to 50	18.8
50 to 100	18.8
100 to 200	13.0
200 to 250	18.4
250 to 420	29.7

2008-07-01 through 2009-06-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200454-0

Scope Revised: 2008-10-16

NVLAP Code: 20/T07

Resistance Thermometry – ITS 90 Fixed Point

<i>Value in °C</i>	<i>Best Uncertainty (±) in mk ^{note 1}</i>	<i>Remarks</i>
419.527	8.8	Zn FP
231.928	7.2	Sn FP
156.5985	4.8	In FP
0.01 (TPW)	2.0	H ₂ O TP
-38.8344	2.8	Hg TP
-189.3442	9.6	Liq. LN2

NVLAP Code: 20/T06

Thermocouples ^{note 4}

<i>Range in °C</i>	<i>Best Uncertainty (±) in mk ^{note 1}</i>
-196 to -80	29.9
-80 to -40	21.2
-40 to -20	21.4
-20 to 0	45.0
0 (Ice Point)	15.4
0 to 50	127.2
50 to 100	139.4
100 to 200	125.2
200 to 250	124.3
250 to 420	626.2
420 to 800	1645
800 to 1100	1575

2008-07-01 through 2009-06-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200454-0

Scope Revised: 2008-10-16

NVLAP Code: 20/T08
Digital Thermometer ^{note 5}

Range in °C	Best Uncertainty (±) in mk ^{note 1}
-196 to -80	42.8
-80 to -40	42.0
-40 to -20	42.9
-20 to 0	58.6
0 (Ice Point)	40.5
0 to 50	51.7
50 to 100	76.9
100 to 200	46.4
200 to 250	44.1
250 to 420	49.8
420 to 800	1522
800 to 1100	1446

1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.
2. Uncertainty components for DUT repeatability, resolution, and drift at ice point are calculated at time of test; for the purpose of this scope, typical values were used.
3. Uncertainty components for DUT repeatability, resolution, and hysteresis are calculated at time of test; for the purpose of this scope, typical values were used.
4. Uncertainty components for DUT repeatability, resolution, and inhomogeneity are calculated at time of test; for the purpose of this scope, typical values were used.
5. Uncertainty components for DUT repeatability, resolution, and hysteresis are calculated at time of test; for the purpose of this scope, typical values were used.
6. Calibration of mass values other than those listed are available at higher uncertainties.

2008-07-01 through 2009-06-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology