



National Voluntary Laboratory Accreditation Program



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Unified Industries Incorporated Standards Laboratory

7530 Fullerton Court
Springfield, VA 22153-2829
Mr. Richard A. Cozier, Jr.
Phone: 703-569-0670 Fax: 703-569-5760
E-mail: cozier@uui.com
URL: <http://www.uui.com>

CALIBRATION LABORATORIES

NVLAP LAB CODE 200597-0

NVLAP Code: 20/A01 ANSI/NCSL Z540-1-1994; Part 1 Compliant

ELECTROMAGNETICS - DC/LOW FREQUENCY

NVLAP Code: 20/E02
AC Current

Range	Best Uncertainty (\pm) in ppm ^{note 1}			Remarks
	400	1 k	5 k	
10 mA	80	80	80	Measure
20 mA	80	80	80	Measure
30 mA	90	90	90	Measure
100 mA	90	90	90	Measure
200 mA	90	90	90	Measure
300 mA	150	150	150	Measure
500 mA	150	150	150	Measure
1 A	150	150	150	Measure
2 A	40	40	40	Measure
3 A	40	40	40	Measure
5 A	45	45	50	Measure
10 A	50	55	50	Measure
20 A	80	80	85	Measure

2008-07-01 through 2009-06-30

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200597-0

10 mA	80	80	80	Generate
20 mA	80	80	80	Generate
30 mA	90	90	90	Generate
100 mA	90	90	90	Generate
200 mA	90	90	90	Generate
300 mA	150	150	150	Generate
500 mA	150	150	150	Generate
1 A	150	150	150	Generate
2 A	40	40	40	Generate
3 A	40	40	40	Generate
5 A	45	45	50	Generate
10 A	50	55	50	Generate
20 A	80	80	85	Generate

NVLAP Code: 20/E05
DC Current

<i>Range in A</i>	<i>Best Uncertainty in (±) ppm ^{note 1}</i>		<i>Remarks</i>
3		22	Measure
6		22	Measure
9		22	Measure
12		22	Measure
15		22	Measure
20		27	Measure
40		27	Measure
60		33	Measure
80		33	Measure
100		19	Measure
3		22	Generate
6		22	Generate
9		22	Generate
12		22	Generate
15		22	Generate

2008-07-01 through 2009-06-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200597-0

20	27	Generate
40	27	Generate
60	33	Generate
80	33	Generate
100	19	Generate

NVLAP Code: 20/E05

DC Resistance

Range	Best Uncertainty (\pm) in ppm ^{note 1}	Remarks
0.001 Ω	6.0	Measure
0.01 Ω	2.2	Measure
0.1 Ω	0.9	Measure
1 Ω	0.25	Measure
10 Ω	0.25	Measure
100 Ω	0.25	Measure
1 k Ω	0.25	Measure
10 k Ω	0.25	Measure
100 k Ω	0.5	Measure
1 M Ω	0.8	Measure
10 M Ω	8.4	Measure
0.001 Ω	6.0	Generate
0.01 Ω	2.2	Generate
0.1 Ω	0.9	Generate
1 Ω	0.25	Generate
10 Ω	0.25	Generate
100 Ω	0.25	Generate
1 k Ω	0.25	Generate
10 k Ω	0.25	Generate
100 k Ω	0.5	Generate
1 M Ω	0.8	Generate
10 M Ω	8.4	Generate

2008-07-01 through 2009-06-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200597-0

NVLAP Code: 20/E06

DC Voltage

Range	Best Uncertainty (\pm) in ppm ^{note 1}	Remarks
100 mV	0.7	Measure
1 V	1.1	Measure
1.018 V	2.1	Measure
10 V	0.7	Measure
100 V	0.7	Measure
1000 V	0.9	Measure
0 to 1000 V	0.7	Measure
0 to 1000 V	0.7	Generate

NVLAP Code: 20/E09

AC Voltage

Range	Best Uncertainty (\pm) in ppm ^{note 1}												Remarks	
	Frequency in Hz													
	10	20	40	100	1 k	10 k	20 k	50 k	100 k	300 k	500 k	800 k	1 M	
22 mV	100	90	80	110	85	85	80	190	280	360	580	480	610	Measure & Generate
220 mV	50	50	40	30	30	30	30	45	85	150	250	340	380	Measure & Generate
700 mV	30	40	20	14	14	14	14	35	35	120	150	190	280	Measure & Generate
2.2 V	35	35	25	16	14	14	14	25	25	100	110	85	85	Measure & Generate
7 V	40	30	20	12	12	12	12	25	25	110	110	85	90	Measure & Generate
22 V	40	35	25	12	12	12	14	20	25	110	120	95	110	Measure & Generate
70 V	30	35	25	14	14	16	14	45	25	85				Measure & Generate
220 V	45	35	25	25	18	18	20	45	55					Measure & Generate
1000 V	60	30	30	30	30	30	35	65	90					Measure & Generate

2008-07-01 through 2009-06-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200597-0

NVLAP Code: 20/E10

LF Capacitance

<i>Range in pF</i>	<i>Best Uncertainty (\pm) in ppm ^{note 1}</i>	<i>Remarks</i>
1000	8	Measure at 1 kHz
100	5	Measure at 1 kHz
10	17	Measure at 1 kHz
1000	8	Generate at 1 kHz
100	5	Generate at 1 kHz
10	17	Generate at 1 kHz

NVLAP Code: 20/E15

Phase Meters

<i>Range in degrees</i>	<i>Volts rms</i>	<i>Frequency kHz</i>	<i>Best Uncertainty (\pm) in mdegree ^{note 1}</i>	<i>Remarks</i>
0 to 360	10	0.06	6.3	Generate
0 to 360	10	0.4	6.3	Generate
0 to 360	10	1	6.3	Generate
0 to 360	10	2	6.3	Generate
0 to 360	10	5	6.3	Generate
0 to 360	10	10	10.7	Generate
0 to 360	10	20	10.7	Generate
0 to 360	10	50	10.7	Generate
0 to 360	50	0.06	6.3	Generate
0 to 360	50	0.04	6.3	Generate
0 to 360	100	0.06	6.3	Generate
0 to 360	100	0.4	6.3	Generate
0 to 360	10	0.06	30	Measure
0 to 360	10	0.4	30	Measure
0 to 360	10	1	30	Measure
0 to 360	10	2	30	Measure
0 to 360	10	5	30	Measure

2008-07-01 through 2009-06-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200597-0

0 to 360	10	10	30	Measure
0 to 360	10	20	30	Measure
0 to 360	10	50	30	Measure
0 to 360	50	0.06	30	Measure
0 to 360	50	0.04	30	Measure
0 to 360	100	0.06	30	Measure
0 to 360	100	0.4	30	Measure

NVLAP Code: 20/E19
Voltage Transformers

Range	Best Uncertainty (\pm) in ppm^{note 1}	Remarks
0 to 1 ^{note 2}	0.6	Ratio

TIME AND FREQUENCY

NVLAP Code: 20/F01
Frequency Dissemination

Range in MHz	Best Uncertainty (\pm)^{note 1}	Remarks
10	1×10^{-13} (7 day average)	Measure
10	1×10^{-13} (7 day average)	Generate

MECHANICAL

NVLAP Code: 20/M06
Force: Compression and Tension

Range in lbf	Best Uncertainty (\pm) in %^{note 1}	Remarks
10 to 500	0.02	500 lb Deadweight Tester
500 to 2 k	0.04	2 k Load Cell
2 k to 5 k	0.04	5 k Load Cell
5 k to 10 k	0.04	10 k Load Cell
10 k to 25 k	0.04	25 k Load Cell
25 k to 50 k	0.05	50 k Load Cell

2008-07-01 through 2009-06-30

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200597-0

70	0.0025	Measure
80	0.0029	Measure
90	0.0032	Measure
100	0.0035	Measure
0.5	0.0005	Generate
10	0.0007	Generate
20	0.0009	Generate
30	0.0012	Generate
40	0.0015	Generate
50	0.0019	Generate
60	0.0022	Generate
70	0.0025	Generate
80	0.0029	Generate
90	0.0032	Generate
100	0.0035	Generate

-
1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.
 2. Actual ratio could be ≤ 1.1 .

2008-07-01 through 2009-06-30

Effective dates

For the National Institute of Standards and Technology