



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Eli Lilly and Company Corporate Standards Lab

1402 South Dakota Street

Indianapolis, IN 46225

Mr. Ronald L. Holt

Phone: 317-655-0658 Fax: 317-651-0281

E-mail: holt_ronald_1@lilly.com

CALIBRATION LABORATORIES

NVLAP LAB CODE 200717-0

MECHANICAL

NVLAP Code: 20/M08

Mass

<i>Range</i>	<i>Best Uncertainty (±) ^{note 1}</i>	<i>Remarks</i>
10 kg	3.2 mg	Echelon II
5 kg	2.9 mg	Echelon II
3 kg	2.8 mg	Echelon II
2 kg	2.8 mg	Echelon II
1 kg	0.14 mg	Echelon II
500 g	0.080 mg	Echelon II
300 g	0.096 mg	Echelon II
200 g	0.073 mg	Echelon II
100 g	0.062 mg	Echelon II
50 g	0.031 mg	Echelon II
30 g	0.019 mg	Echelon II
20 g	0.013 mg	Echelon II
10 g	0.0083 mg	Echelon II
5 g	0.0043 mg	Echelon II
3 g	0.0028 mg	Echelon II
2 g	0.0021 mg	Echelon II
1 g	0.0018 mg	Echelon II

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200717-0

500 mg	0.0012 mg	Echelon II
300 mg	0.00098 mg	Echelon II
200 mg	0.00091 mg	Echelon II
100 mg	0.00094 mg	Echelon II
50 mg	0.00085 mg	Echelon II
30 mg	0.00080 mg	Echelon II
20 mg	0.00079 mg	Echelon II
10 mg	0.00083 mg	Echelon II
5 mg	0.00076 mg	Echelon II
3 mg	0.00073 mg	Echelon II
2 mg	0.00073 mg	Echelon II
1 mg	0.00075 mg	Echelon II

-
1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.

2008-04-01 through 2009-03-31

Effective dates

For the National Institute of Standards and Technology