



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

ELECTROMAGNETICS – DC/LOW FREQUENCY

NVLAP Code: 20/E02

AC Current

*Best Uncertainty (±) in % ^{note 1}
Frequency in Hz*

| <i>Range</i> | <i>10 to 20</i> | <i>20 to 40 k</i> | <i>40 to 10 k</i> | <i>10 k to 30 k</i> |
|--------------|-----------------|-------------------|-------------------|---------------------|
| 100 µA | 0.02 | 0.01 | 0.009 | 0.014 |
| 200 µA | 0.02 | 0.009 | 0.009 | 0.014 |
| 300 µA | 0.02 | 0.01 | 0.007 | 0.013 |
| 1 mA | 0.022 | 0.009 | 0.006 | 0.008 |
| 2 mA | 0.02 | 0.009 | 0.005 | 0.007 |
| 10 mA | 0.024 | 0.009 | 0.005 | 0.006 |
| 20 mA | 0.025 | 0.009 | 0.005 | 0.007 |
| 50 mA | 0.024 | 0.009 | 0.006 | 0.012 |
| 100 mA | 0.024 | 0.009 | 0.005 | 0.008 |
| 200 mA | 0.024 | 0.009 | 0.006 | 0.008 |
| 300 mA | 0.024 | 0.009 | 0.006 | 0.012 |
| 1 A | 0.024 | 0.009 | 0.007 | 0.012 |
| 2 A | 0.024 | 0.009 | 0.006 | 0.012 |
| 3 A | 0.027 | 0.01 | 0.008 | 0.016 |
| 5 A | 0.027 | 0.01 | 0.008 | 0.02 |
| 10 A | 0.049 | 0.042 | 0.042 | 0.044 |
| 20 A | 0.11 | 0.11 | 0.11 | 0.11 |

AC Current ^{note 2}

Measuring Equipment and Measure

*Best Uncertainty (±) in % + A ^{note 1}
Frequency in Hz*

| <i>Range</i> | <i>10 to 20</i> | <i>20 to 45</i> | <i>45 to 100</i> | <i>100 to 5 k</i> |
|------------------|-----------------|-----------------|------------------|-------------------|
| (0 to 100) µA | 0.46 + 30 n | 0.18 + 30 n | 0.08 + 30 n | 0.08 + 30 n |
| (0.1 to 1) mA | 0.46 + 200 n | 0.18 + 200 n | 0.07 + 200 n | 0.07 + 200 n |
| (1 to 10) mA | 0.46 + 2 µ | 0.17 + 2 µ | 0.07 + 2 µ | 0.04 + 2 µ |
| (10 to 100) mA | 0.46 + 20 µ | 0.18 + 20 µ | 0.69 + 20 µ | 0.35 + 20 µ |
| (100 to 1000) mA | 0.46 + 200 µ | 0.19 + 200 µ | 0.1 + 200 µ | 0.12 + 200 µ |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

NVLAP Code: 20/E05

DC Resistance

Measuring Equipment and Measure

| <i>Range</i> | <i>Best Uncertainty (±)^{note 1}</i> | <i>Remarks</i> |
|-----------------|--|---|
| 0 Ω to 0.1 mΩ | 1.4 nΩ | Standard Resistors w/Low Thermal Switch |
| 0.1 mΩ to 1 mΩ | 7.1 nΩ | Standard Resistors w/Low Thermal Switch |
| 1 mΩ to 10 mΩ | 5.4 ppm | Standard Resistors w/Low Thermal Switch |
| 10 mΩ to 100 mΩ | 8.2 ppm | Standard Resistors w/Low Thermal Switch |
| 0.1 Ω to 1 Ω | 0.36 ppm | Standard Resistors w/Low Thermal Switch |
| 1 Ω to 10 Ω | 0.5 ppm | Standard Resistors w/Guildline 9975 Bridge |
| 10 Ω to 100 Ω | 0.84 ppm | Standard Resistors w/Guildline 9975 Bridge |
| 100 Ω to 1 kΩ | 0.42 ppm | Standard Resistors w/Guildline 9975 Bridge |
| 1 kΩ to 10 kΩ | 0.31 ppm | Standard Resistors w/Guildline 9975 Bridge |
| 19 kΩ | 0.78 ppm | Standard Resistors w/Fluke 8508A in transfer mode |
| 100 kΩ | 2.9 ppm | Standard Resistors w/Fluke 8508A in transfer mode |
| 190 kΩ | 2.8 ppm | Standard Resistors w/Fluke 8508A in transfer mode |
| 1 MΩ | 3.8 ppm | Standard Resistors w/Fluke 8508A in transfer mode |
| 1.9 MΩ | 5.1 ppm | Standard Resistors w/Fluke 8508A in transfer mode |
| 10 MΩ Source | 4.4 ppm | Standard Resistors w/Fluke 8508A in transfer mode |
| 10 MΩ Measure | 5.0 ppm | Standard Resistors w/Fluke 8508A in transfer mode |
| 19 MΩ Measure | 8.3 ppm | Standard Resistors w/Fluke 8508A in transfer mode |
| 100 MΩ | 13 ppm | Standard Resistors w/Fluke 8508A in transfer mode |
| 1 GΩ Source | 63 ppm | Standard Resistors w/Fluke 8508A in transfer mode |
| 1 GΩ Measure | 200 ppm | Standard Resistors w/Fluke 8508A in transfer Mode |

Resistance Ratio

2009-01-01 through 2009-12-31

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

| | | |
|---|--------------------------|---------------------------|
| 1 Ω to 1 k Ω | 0.33 ppm | Guidline 9975 Bridge |
| Measuring Equipment and Measure ^{note 2} | | |
| 10 m Ω to 10 Ω | 19 ppm + 0.05 m Ω | HP3458A w/Decade Resistor |
| 10 Ω to 1 k Ω | 15 ppm + 0.5 m Ω | HP3458A w/Decade Resistor |
| 1 k Ω to 10 k Ω | 12 ppm + 5 m Ω | HP3458A w/Decade Resistor |
| 10 k Ω to 100 k Ω | 14 ppm + 50 m Ω | HP3458A w/Decade Resistor |
| 100 k Ω to 1 M Ω | 23 ppm + 2 Ω | HP3458A w/Decade Resistor |
| 1 M Ω to 10 M Ω | 73 ppm + 50 Ω | HP3458A w/Decade Resistor |
| 10 M Ω to 100 M Ω | 630 ppm + 1 k Ω | HP3458A w/Decade Resistor |
| 100 M Ω to 1 G Ω | 0.6 % + 10 k Ω | HP3458A w/Decade Resistor |

| | | |
|---------------------------------------|--------|-----------------|
| Measuring Equipment ^{note 2} | | |
| 10 G Ω to 100 G Ω | 1.16 % | Biddle Mega Dek |

DC Current
Measuring Equipment and Measure

| Range | Best Uncertainty (\pm) ^{note 1} | Remarks |
|---------------------|--|----------------------------------|
| 0 A to 100 μ A | 4 ppm + 230 pA | Standard Shunts w/current source |
| 100 μ A to 1 mA | 3.1 ppm + 1.2 nA | Standard Shunts w/current source |
| 1 mA to 10 mA | 3.2 ppm + 12 nA | Standard Shunts w/current source |
| 10 mA to 100 mA | 3.2 ppm + 115 nA | Standard Shunts w/current source |
| 100 mA to 1 A | 8.8 ppm + 1.4 μ A | Standard Shunts w/current source |
| 1 A to 10 A | 8.8 ppm + 12 μ A | Standard Shunts w/current source |
| 10 A to 100 A | 160 ppm + 7 mA | Standard Shunts w/current source |

| | | |
|---------------------------------|----------------------|-----------------------------|
| DC Current ^{note 2} | | |
| Measuring Equipment and Measure | | |
| 0 μ A to 100 μ A | 26 ppm + 0.8 nA | HP3458A w/current source |
| 100 μ A to 1 mA | 26 ppm + 5 nA | HP3458A w/current source |
| 1 mA to 10 mA | 26 ppm + 50 nA | HP3458A w/current source |
| 10 mA to 100 mA | 42 ppm + 0.5 μ A | HP3458A w/current source |
| 100 mA to 1 A | 0.013 % + 10 μ A | HP3458A w/current source |
| 1 A to 20 A | 0.12 % | Fluke 5520A w/current shunt |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

800 V to 1000 V

21 ppm + 100 μ V

3458A (002) w/5700A

NVLAP Code: 20/E09

AC Voltage

| Frequency in Hz | Best Uncertainty (\pm) in % ^{note 1} | | | Remarks |
|---------------------|---|--------|--------|------------|
| | 6 mV | 10 mV | 20 mV | |
| 10 | 0.19 | 0.03 | 0.03 | Fluke 792A |
| 20 | 0.14 | 0.03 | 0.02 | Fluke 792A |
| (40, 100) | 0.14 | 0.03 | 0.014 | Fluke 792A |
| (1, 10, 20) k | 0.14 | 0.03 | 0.014 | Fluke 792A |
| 50 k | 0.14 | 0.05 | 0.03 | Fluke 792A |
| 100 k | 0.17 | 0.07 | 0.02 | Fluke 792A |
| 300 k | 0.22 | 0.12 | 0.014 | Fluke 792A |
| 500 k | 0.40 | 0.15 | 0.014 | Fluke 792A |
| 1 M | 0.45 | 0.15 | 0.13 | Fluke 792A |
| 220 mV Range | | | | |
| | 20 mV | 60 mV | 200 mV | |
| 10 | 0.04 | 0.035 | 0.02 | Fluke 792A |
| 20 | 0.03 | 0.021 | 0.008 | Fluke 792A |
| (40, 100) | 0.02 | 0.017 | 0.004 | Fluke 792A |
| (1, 10, 20) k | 0.02 | 0.017 | 0.004 | Fluke 792A |
| 50 k | 0.03 | 0.021 | 0.01 | Fluke 792A |
| 100 k | 0.05 | 0.041 | 0.015 | Fluke 792A |
| 300 k | 0.07 | 0.068 | 0.045 | Fluke 792A |
| 500 k | 0.11 | 0.11 | 0.06 | Fluke 792A |
| 1 M | 0.2 | 0.17 | 0.06 | Fluke 792A |
| 700 mV Range | | | | |
| | 200 mV | 600 mV | | |
| 10 | 0.022 | 0.007 | | Fluke 792A |
| 20 | 0.008 | 0.007 | | Fluke 792A |
| (40, 100) | 0.005 | 0.003 | | Fluke 792A |
| (1, 10, 20) k | 0.005 | 0.003 | | Fluke 792A |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

| | | | |
|-------|------|-------|------------|
| 50 k | 0.01 | 0.005 | Fluke 792A |
| 100 k | 0.02 | 0.006 | Fluke 792A |
| 300 k | 0.05 | 0.014 | Fluke 792A |
| 500 k | 0.06 | 0.043 | Fluke 792A |
| 1 M | 0.06 | 0.06 | Fluke 792A |

2.2 V Range

| | 600 mV | 1 V | 2 V | |
|--------------------|--------|-------|--------|------------|
| 10 | 0.02 | 0.02 | 0.02 | Fluke 792A |
| 20 | 0.007 | 0.006 | 0.006 | Fluke 792A |
| 40 | 0.003 | 0.003 | 0.003 | Fluke 792A |
| (0.1, 1, 10, 20) k | 0.003 | 0.004 | 0.0011 | Fluke 792A |
| 50 k | 0.005 | 0.005 | 0.004 | Fluke 792A |
| 100 k | 0.006 | 0.012 | 0.005 | Fluke 792A |
| 300 k | 0.012 | 0.043 | 0.012 | Fluke 792A |
| 500 k | 0.043 | 0.045 | 0.043 | Fluke 792A |
| 1 M | 0.06 | 0.045 | 0.045 | Fluke 792A |

7 V Range

| | 2 V | 6 V | |
|---------------------|-------|--------|------------|
| 10 | 0.02 | 0.02 | Fluke 792A |
| 20 | 0.007 | 0.006 | Fluke 792A |
| 40 | 0.003 | 0.003 | Fluke 792A |
| (0.01, 1, 10, 20) k | 0.003 | 0.0011 | Fluke 792A |
| 50 k | 0.005 | 0.004 | Fluke 792A |
| 100 k | 0.006 | 0.005 | Fluke 792A |
| 300 k | 0.012 | 0.012 | Fluke 792A |
| 500 k | 0.044 | 0.043 | Fluke 792A |
| 1 M | 0.05 | 0.046 | Fluke 792A |

22 V Range

| | 6 V | 10 V | 20 V | |
|----|-------|-------|-------|------------|
| 10 | 0.02 | 0.02 | 0.02 | Fluke 792A |
| 20 | 0.007 | 0.006 | 0.006 | Fluke 792A |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

| | | | | |
|--------------------|-------|-------|-------|------------|
| 40 | 0.003 | 0.003 | 0.003 | Fluke 792A |
| (0.1, 1, 10, 20) k | 0.003 | 0.002 | 0.002 | Fluke 792A |
| 50 k | 0.005 | 0.004 | 0.004 | Fluke 792A |
| 100 k | 0.006 | 0.005 | 0.005 | Fluke 792A |
| 300 k | 0.012 | 0.012 | 0.012 | Fluke 792A |
| 500 k | 0.044 | 0.043 | 0.043 | Fluke 792A |
| 1 M | 0.05 | 0.05 | 0.05 | Fluke 792A |

70 V Range

| | 20 V | 60 V | |
|--------------------|-------|-------|------------|
| 10 | 0.02 | 0.02 | Fluke 792A |
| 20 | 0.007 | 0.006 | Fluke 792A |
| 40 | 0.003 | 0.003 | Fluke 792A |
| (0.1, 1, 10, 20) k | 0.003 | 0.002 | Fluke 792A |
| 50 k | 0.006 | 0.005 | Fluke 792A |
| 100 k | 0.007 | 0.007 | Fluke 792A |
| 300 k | 0.013 | 0.013 | Fluke 792A |

220 V Range

| | 60 V | 100 V | 200 V | |
|--------------------------|-------|-------|-------|------------|
| 10 | 0.02 | 0.02 | 0.02 | Fluke 792A |
| 20 | 0.007 | 0.007 | 0.006 | Fluke 792A |
| (0.04, 0.1, 1, 10, 20) k | 0.004 | 0.003 | 0.003 | Fluke 792A |
| 50 k | 0.007 | 0.007 | 0.006 | Fluke 792A |
| 100 k | 0.007 | 0.007 | 0.007 | Fluke 792A |
| 200 k | 0.01 | 0.01 | 0.01 | Fluke 792A |

1000 V Range

| | 200 V | 600 V | 1000 V | |
|--------------------------|-------|-------|--------|------------|
| 10 | 0.02 | | | Fluke 792A |
| 20 | 0.009 | | | Fluke 792A |
| (0.04, 0.1, 1, 10, 20) k | 0.004 | 0.004 | 0.003 | Fluke 792A |
| 50 k | 0.007 | 0.007 | 0.006 | Fluke 792A |
| 100 k | 0.007 | 0.008 | 0.007 | Fluke 792A |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

(5 to 80) kV @ 60 Hz

0.054

HV Divider
Comparison

AC Voltage – Measure ^{note 2}

| Range | Frequency in Hz | Best Uncertainty (\pm) ^{note 1} | Remarks |
|----------------|--------------------------------|--|----------------|
| (0 to 10) mV | 1 to 40 | 0.06 % + 3 μ V | 3458A |
| | 40 to 1 k | 0.05 % + 1 μ V | 3458A |
| | 1 k to 20 k | 0.05 % + 1 μ V | 3458A |
| | 20 k to 50 k | 0.12 % + 1 μ V | 3458A |
| | 50 k to 100 k | 0.6 % + 1 μ V | 3458A |
| | 100 k to 300 k | 4.6 % + 2 μ V | 3458A |
| (10 to 100) mV | 1 to 40 | 0.01 % + 4 μ V | 3458A |
| | 40 to 1 k | 0.01 % + 2 μ V | 3458A |
| | 1 k to 20 k | 0.018 % + 2 μ V | 3458A |
| | 20 k to 50 k | 0.036 % + 2 μ V | 3458A |
| | 50 k to 100 k | 0.1 % + 2 μ V | 3458A |
| | 100 k to 300 k 300 k to 1 M | 0.35 % + 10 μ V 1.2 % + 10 μ V | 3458A 3458A |
| (100 m to 1) V | 1 to 40 | 0.009 % + 40 μ V | 3458A |
| | 40 to 1 k | 0.009 % + 20 μ V | 3458A |
| | 1 k to 20 k | 0.017 % + 20 μ V | 3458A |
| | 20 k to 50 k | 0.035 % + 20 μ V | 3458A |
| | 50 k to 100 k | 0.09 % + 20 μ V | 3458A |
| | 100 k to 300 k 300 k to 1 M | 0.35 % + 100 μ V 1.2 % + 100 μ V | 3458A 3458A |
| (1 to 10) V | 1 to 40 | 0.009 % + 0.4 mV | 3458A |
| | 40 to 1 k | 0.009 % + 0.2 mV | 3458A |
| | 1 k to 20 k | 0.017 % + 0.2 mV | 3458A |
| | 20 k to 50 k | 0.035 % + 0.2 mV | 3458A |
| | 50 k to 100 k | 0.09 % + 0.2 mV | 3458A |
| | 100 k to 300 k 300 k to 1 M | 0.35 % + 1 mV 1.2 % + 1 mV | 3458A 3458A |
| (10 to 100) V | 1 to 40 | 0.025 % + 4 mV | 3458A |
| | 40 to 1 k | 0.025 % + 2 mV | 3458A |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

| | | | |
|----------------|----------------|----------------|-------|
| | 1 k to 20 k | 0.025 % + 2 mV | 3458A |
| | 20 k to 50 k | 0.04 % + 2 mV | 3458A |
| | 50 k to 100 k | 0.14 % + 2 mV | 3458A |
| | 100 k to 300 k | 0.5 % + 10 mV | 3458A |
| | 300 k to 1 M | 1.8 % + 10 mV | 3458A |
| (100 to 700) V | 1 to 40 | 0.05 % + 40 mV | 3458A |
| | 40 to 1 k | 0.05 % + 20 mV | 3458A |
| | 1 k to 20 k | 0.07 % + 20 mV | 3458A |
| | 20 k to 50 k | 0.14 % + 20 mV | 3458A |
| | 50 k to 100 k | 0.35 % + 20 mV | 3458A |

AC Voltage ^{note 2}
Measuring Equipment

| Range | Frequency in Hz | Best Uncertainty (\pm) ^{note 1} | Remarks |
|----------------|------------------------|--|----------------|
| (0 to 2.2) mV | 10 to 40 | 0.6 % + 5 μ V | 5700A / 5725A |
| | 40 to 50 k | 0.15 % + 5 μ V | 5700A / 5725A |
| | 50 k to 100 k | 0.1 % + 7 μ V | 5700A / 5725A |
| | 100 k to 300 k | 0.13 % + 13 μ V | 5700A / 5725A |
| | 300 k to 500 k | 0.2 % + 30 μ V | 5700A / 5725A |
| | 500 k to 1 M | 0.4 % + 30 μ V | 5700A / 5725A |
| (2.2 to 22) mV | 10 to 20 | 0.08 % + 5 μ V | 5700A / 5725A |
| | 20 to 20 k | 0.05 % + 5 μ V | 5700A / 5725A |
| | 20 k to 50 k | 0.06 % + 5 μ V | 5700A / 5725A |
| | 50 k to 100 k | 0.13 % + 7 μ V | 5700A / 5725A |
| | 100 k to 300 k | 0.15 % + 12 μ V | 5700A / 5725A |
| | 300 k to 500 k | 0.13 % + 25 μ V | 5700A / 5725A |
| | 500 k to 1 M | 0.4 % + 25 μ V | 5700A / 5725A |
| (22 to 220) mV | 10 to 20 | 0.07 % + 13 μ V | 5700A / 5725A |
| | 20 to 40 | 0.03 % + 10 μ V | 5700A / 5725A |
| | 40 to 20 k | 0.015 % + 10 μ V | 5700A / 5725A |
| | 20 k to 50 k | 0.04 % + 8 μ V | 5700A / 5725A |
| | 50 k to 100 k | 0.1 % + 25 μ V | 5700A / 5725A |
| | 100 k to 300 k | 0.14 % + 25 μ V | 5700A / 5725A |
| | 300 k to 500 k | 0.2 % + 35 μ V | 5700A / 5725A |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

| | | | |
|------------------|----------------|----------------------|---------------|
| | 500 k to 1 M | 0.34 % + 80 μ V | 5700A / 5725A |
| (220 m to 2.2) V | 10 to 20 | 0.06 % + 100 μ V | 5700A / 5725A |
| | 20 to 40 | 0.02 % + 25 μ V | 5700A / 5725A |
| | 40 to 20 k | 0.01 % + 6 μ V | 5700A / 5725A |
| | 20 k to 50 k | 0.014 % + 16 μ V | 5700A / 5725A |
| | 50 k to 100 k | 0.03 % + 70 μ V | 5700A / 5725A |
| | 100 k to 300 k | 0.06 % + 130 μ V | 5700A / 5725A |
| | 300 k to 500 k | 0.13 % + 350 μ V | 5700A / 5725A |
| | 500 k to 1 M | 0.26 % + 85 μ V | 5700A / 5725A |
| (2.2 to 22) V | 10 to 20 | 0.06 % + 0.8 mV | 5700A / 5725A |
| | 20 to 40 | 0.02 % + 0.3 mV | 5700A / 5725A |
| | 40 to 20 k | 0.009 % + 60 μ V | 5700A / 5725A |
| | 20 k to 50 k | 0.014 % + 0.2 mV | 5700A / 5725A |
| | 50 k to 100 k | 0.03 % + 0.4 mV | 5700A / 5725A |
| | 100 k to 300 k | 0.06 % + 1.5 mV | 5700A / 5725A |
| | 300 k to 500 k | 0.15 % + 5 mV | 5700A / 5725A |
| | 500 k to 1 M | 0.32 % + 9 mV | 5700A / 5725A |
| (22 to 220) V | 10 to 20 | 0.06 % + 8 mV | 5700A / 5725A |
| | 20 to 40 | 0.02 % + 3 mV | 5700A / 5725A |
| | 40 to 20 k | 0.01 % + 1 mV | 5700A / 5725A |
| | 20 k to 50 k | 0.03 % + 4 mV | 5700A / 5725A |
| | 50 k to 100 k | 0.06 % + 8 mV | 5700A / 5725A |
| | 100 k to 300 k | 0.18 % + 8 mV | 5700A / 5725A |
| (220 to 1100) V | 40 to 1 k | 0.01 % + 4 mV | 5700A / 5725A |
| | 1 k to 20 k | 0.02 % + 6 mV | 5700A / 5725A |
| | 20 k to 30 k | 0.07 % + 11 mV | 5700A / 5725A |
| (220 to 750) V | 30 k to 50 k | 0.07 % + 11 mV | 5700A / 5725A |
| | 50 k to 100 k | 0.3 % + 45 mV | 5700A / 5725A |

NVLAP Code: 20/E10

Capacitance – Source: (100 to 10 k) Hz

Range

Best Uncertainty (\pm)^{note 1}

Remarks

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

| | | |
|---------|--------|------------------------------------|
| 0.01 pF | 0.02 % | Fixed Capacitors w/GR1615-A Bridge |
| 0.1 pF | 0.02 % | Fixed Capacitors w/GR1615-A Bridge |
| 1 pF | 0.02 % | Fixed Capacitors w/GR1615-A Bridge |
| 10 pF | 0.02 % | Fixed Capacitors w/GR1615-A Bridge |
| 100 pF | 0.02 % | Fixed Capacitors w/GR1615-A Bridge |
| 500 pF | 0.02 % | Fixed Capacitors w/GR1615-A Bridge |
| 1 nF | 0.002% | Fixed Capacitors w/GR1615-A Bridge |
| 10 nF | 0.02 % | Fixed Capacitors w/GR1615-A Bridge |
| 200 nF | 0.02 % | Fixed Capacitors w/GR1615-A Bridge |

Capacitance – Measure: (50 to 1M) Hz

| | |
|-------------------|-----------------|
| (1 a to 1.1 μ) F | 0.012 % + 30 aF |
| (1.1 μ to 10 m) F | 0.06 % + 30 aF |

GR1615-A Bridge
Quadtech 7600LCR Bridge

Capacitance – Measure Equipment ^{note 2}

| Range | Frequency in Hz | Best Uncertainty (±) ^{note 1} | Remarks |
|----------------------|------------------------|---|----------------|
| (0.19 to 1.0999) nF | 10 to 10 k | 0.60 % + 0.01 nF | 5520A |
| (1.1 to 3.2999) nF | 10 to 3 k | 0.60 % + 0.01 nF | 5520A |
| (3.3 to 10.9999) nF | 10 to 1 k | 0.30 % + 0.01 nF | 5520A |
| (11 to 109.999) nF | 10 to 1 k | 0.30 % + 0.01 nF | 5520A |
| (110 to 329.999) nF | 10 to 1 k | 0.30 % + 0.3 nF | 5520A |
| (0.33 to 1.09999) μF | 10 to 600 | 0.30 % + 1 nF | 5520A |
| (1.1 to 3.2999) μF | 10 to 300 | 0.30 % + 3 nF | 5520A |
| (3.3 to 10.9999) μF | 10 to 150 | 0.30 % + 10 nF | 5520A |
| (11 to 32.9999) μF | 10 to 120 | 0.48 % + 30 nF | 5520A |
| (33 to 109.9999) μF | 10 to 80 | 0.54 % + 100 nF | 5520A |
| (110 to 329.999) μF | DC to 50 | 0.54 % + 300 nF | 5520A |
| (0.33 to 1.09999) mF | DC to 20 | 0.54 % + 1 μF | 5520A |
| (1.1 to 3.29999) mF | DC to 6 | 0.54 % + 3 μF | 5520A |
| (3.3 to 10.9999) mF | DC to 2 | 0.54 % + 10 μF | 5520A |
| (11 to 32.9999) mF | DC to 0.6 | 0.90 % + 30 μF | 5520A |
| (33 to 110) mF | DC to 0.2 | 1.31 % + 100 μF | 5520A |

NVLAP Code: 20/E11

LF Inductance ^{note 2}

Source Only

| Range | Best Uncertainty (±) in % ^{note 1} | Remarks |
|--------------|--|---------------------------------|
| 1 mH | 0.07 | Fixed Inductors w/Quadtech 7600 |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

| | | |
|--------|------|---------------------------------|
| 10 mH | 0.07 | Fixed Inductors w/Quadtech 7600 |
| 100 mH | 0.07 | Fixed Inductors w/Quadtech 7600 |
| 1 H | 0.07 | Fixed Inductors w/Quadtech 7600 |

| | | |
|------------------------------------|------|---------------|
| Measure @ 1 kHz (10 μ to 100) H | 0.07 | Quadtech 7600 |
|------------------------------------|------|---------------|

NVLAP Code: 20/E15

AC Phase ^{note 2}

Measure

| Range | Frequency in Hz | Best Uncertainty (±) ^{note 1} | Remarks |
|------------|-----------------|--|----------------------------|
| 0° to 360° | 1 to 50 k | 3 m° | Clark-Hess 5002 Bridge Set |
| | 50 k to 200 k | 11 m° | Clark-Hess 5002 Bridge Set |

AC Phase – Generate ^{note 2}

50mV to 100V

| Range | Frequency in Hz | Best Uncertainty (±) ^{note 1} | Remarks |
|------------|-----------------|--|----------------------------------|
| 0° to 360° | 1 to 1 k | 13 m° | Clark-Hess 5500-2 Phase Standard |
| | 1 k to 6.25 k | 17 m° | Clark-Hess 5500-2 Phase Standard |
| | 6.25 k to 50 k | 21 m° | Clark-Hess 5500-2 Phase Standard |
| | 50 k to 200 k | 50 m° | Clark-Hess 5500-2 Phase Standard |

100V to 120V

| Range | Frequency in Hz | Best Uncertainty (±) ^{note 1} | Remarks |
|------------|-----------------|--|----------------------------------|
| 0° to 360° | 1 to 1 k | 1.7 m° | Clark-Hess 5500-2 Phase Standard |
| | 1 k to 6.25 k | 26 m° | Clark-Hess 5500-2 Phase Standard |
| | 6.25 k to 50 k | 37 m° | Clark-Hess 5500-2 Phase Standard |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

50 k to 200 k

95 m°

Clark-Hess 5500-2 Phase Standard

TIME AND FREQUENCY

NVLAP Code: 20/F01

Frequency – Source and Measure

| <i>Range in Hz</i> ^{note 3} | <i>Best Uncertainty (±)</i> ^{note 1} | <i>Remarks</i> |
|--------------------------------------|---|-----------------------------|
| 10 M | 5.8 x 10 ⁻¹¹ | Rubidium Frequency Standard |

MECHANICAL

NVLAP Code: 20/M06

Torque ^{note 2}

Measure

| <i>Range</i> | <i>Best Uncertainty (±) in %</i> ^{note 1} | <i>Remarks</i> |
|-------------------------|--|----------------|
| 10 lbf-in to 600 lbf-ft | 2.0 % | CDI |

NVLAP Code: 20/M08

Mass ^{note 2}

| <i>Range</i> | <i>Best Uncertainty (±)</i> ^{note 1} | <i>Remarks</i> |
|--------------|---|----------------|
| 8 kg | 12 mg | Echelon III |
| 7 kg | 12 mg | Echelon III |
| 6 kg | 12 mg | Echelon III |
| 5 kg | 9.3 mg | Echelon III |
| 4 kg | 8.8 mg | Echelon III |
| 2 kg | 6.9 mg | Echelon III |
| 1 kg | 3.6 mg | Echelon III |
| 500 g | 2 mg | Echelon III |
| 200 g | 0.68 mg | Echelon III |
| 100 g | 0.34 mg | Echelon III |
| 50 g | 0.17 mg | Echelon III |
| 20 g | 0.10 mg | Echelon III |
| 10 g | 68 µg | Echelon III |
| 5 g | 54 µg | Echelon III |
| 2 g | 54 µg | Echelon III |

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

| | | |
|--------|-------|-------------|
| 1 g | 54 µg | Echelon III |
| 500 mg | 43 µg | Echelon III |
| 200 mg | 43 µg | Echelon III |
| 100 mg | 43 µg | Echelon III |
| 50 mg | 43 µg | Echelon III |
| 20 mg | 43 µg | Echelon III |
| 10 mg | 43 µg | Echelon III |
| 5 mg | 43 µg | Echelon III |
| 2 mg | 43 µg | Echelon III |
| 1 mg | 43 µg | Echelon III |

ELECTROMAGNETICS – RF/MICROWAVE

NVLAP Code: 20/R11

RF-DC Voltage/Current Converters *note 2*

Sinewave Flatness

| Range in Hz | Best Uncertainty (±) in % <i>note 1</i> | Remarks |
|--------------------|--|--------------------|
| 30 k to 1 M | 0.014 | Thermal Converters |
| 1 M to 10 M | 0.08 | Thermal Converters |
| 10 M to 30 M | 0.17 | Thermal Converters |
| 30 M to 80 M | 0.71 | Thermal Converters |
| 80 M to 100 M | 0.84 | Thermal Converters |

NVLAP Code: 20/R17

RF Power *note 2*

Absolute

| Range | Frequency in Hz | Best Uncertainty (±) <i>note 1</i> | Remarks |
|------------------|------------------------|---|----------------|
| (+30 to -20) dBm | 0.1 M to 1.3 G | 0.10 dBm + M | 8902A |

Harmonic Distortion (50 and 600 Ω)

| Range | Best Uncertainty (±) <i>note 1</i> | Remarks |
|--------------|---|----------------|
|--------------|---|----------------|

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

0 dB @ (10 to 26 G) Hz

0.3 dB

8903B
71209A
8902A

THERMODYNAMIC

NVLAP Code: 20/T05

Pressure

Absolute Pressure Source – Pneumatic

Range

(0.2 to 100) psia
(100 to 1000) psia

Best Uncertainty (\pm) ^{note 1}

11 ppm + 0.07 m psia
12 ppm

Remarks

Ruska 2465
Ruska 2465

Gage Pressure Source – Gage

Range

(0 to 1.2) psi
(1.2 to 100) psi
(100 to 1000) psi
(-20 to 20) in H₂O

Best Uncertainty (\pm) ^{note 1}

0.013 m psi
11 ppm
12 ppm
11 ppm + 240 μ in H₂O

Remarks

Ruska 2465
Ruska 2465
Ruska 2465
Differential

Gage Pressure Source – Hydraulic

Range

(75 to 3000) psi
(725 to 30 000) psi

Best Uncertainty (\pm) ^{note 1}

16 ppm
36 ppm

Remarks

DHI PG7000
DHI PG7000

Determination of Piston Area

Range

(0.2 to 100) psi
(100 to 1000) psi
(40 to 10 000) psi

Best Uncertainty (\pm) ^{note 1}

16 ppm
17 ppm
35 ppm

Remarks

Ruska 2465
Ruska 2465
DHI 5300

Gage Pressure Source ^{note 2}

Range

(0.5 to 500) psi

Best Uncertainty (\pm) ^{note 1}

73 ppm

Remarks

Pressurements T2300

Hydraulic ^{note 2}

Range

Best Uncertainty (\pm) ^{note 1}

Remarks

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

(500 to 15 000) psi

0.03 %

Ametek T-150

NVLAP Code: 20/T07

Thermodynamic

Range in °C

Best Uncertainty (±) in C[•] *note 1*

Remarks

-10 to 110

0.044

Liquid Bath w/PRT

100 to 300

0.12

Dry Block Calibrator

300 to 600

0.23

Dry Block Calibrator

Measure only *note 2*

-195 to 660

0.044

PRT & Super
Thermometer

NVLAP Code: 20/T08

Thermocouple

Isothermal Block Verification *note 2*

Range

Best Uncertainty (±) *note 1*

Remarks

Ambient (~23 °C)

0.04 °C

Thermocouple Half
Junction

1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95%.
2. Onsite calibrations available.
3. Uncertainty values of derivatives of 10 MHz will differ due to resolution, noise, and gating errors.

2009-01-01 through 2009-12-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology