



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Alden Research Laboratory, Inc.

30 Shrewsbury Street
Holden, MA 01520

Mr. Philip Stacy

Phone: 508-829-6000 x425 Fax: 508-829-5939

E-mail: pstacy@aldenlab.com

URL: www.aldenlab.com

CALIBRATION LABORATORIES

NVLAP LAB CODE 200798-0

NVLAP Code: 20/A01 ANSI/NCSL Z540-1-1994; Part 1 Compliant

MECHANICAL

NVLAP Code: 20/M05
Flow Rate

<i>Range in GPM</i>	<i>Best Uncertainty (\pm)in % ^{note 1}</i>	<i>Remarks</i>
1500 to 20 000	0.10	Allen 100,000 lb water Gravimetric System
50 to 1500	0.08	Allen 10,000 lb water Gravimetric System
4 to 90	0.08	Allen 1,000 lb water Gravimetric System

1. Represents an expanded uncertainty using a coverage factor, k=2, at an approximate level of confidence of 95%.

2008-11-05 through 2009-09-30

Effective dates

For the National Institute of Standards and Technology