



# National Institute of Standards & Technology

## Certificate

### Standard Reference Material 4926E Hydrogen-3 Radioactivity Standard

This Standard Reference Material (SRM) consists of radioactive hydrogen-3, as water, in 20 mL of distilled water. The solution is contained in a 20-mL borosilicate-glass serum vial. The SRM is intended for the calibration of beta-particle counting instruments and for the monitoring of radiochemical procedures.

#### **Radiological Hazard**

The SRM vial contains hydrogen-3 with a total activity of approximately 105 kBq. Hydrogen-3 decays by beta-particle emission. None of the beta particles escape from the SRM vial. During the decay process no photons are emitted. Approximate unshielded dose rates at several distances (as of the reference time) are given in note [a]\*. There is no detectable external radiation. The SRM should be used only by persons qualified to handle radioactive material.

#### **Chemical Hazard**

The SRM vial contains only distilled water. There is no chemical hazard.

#### **Storage and Handling**

The SRM should be stored and used at a temperature between 5 and 35 °C. The solution in an unopened vial should remain stable and homogeneous until at least September 2003.

The vial (or any subsequent container) should always be clearly marked as containing radioactive material. If the vial is transported it should be packed, marked, labeled, and shipped in accordance with the applicable national, international, and carrier regulations. The solution in the vial is a dangerous good (hazardous material) because of the radioactivity.

#### **Preparation**

This Standard Reference Material was prepared in the Physics Laboratory, Ionizing Radiation Division, Radioactivity Group, L.R. Karam, Group Leader. The overall technical direction and physical measurements leading to certification were provided by L.L. Lucas and M.P. Unterweger of the Radioactivity Group.

The support aspects involved in the preparation, certification, and issuance of this SRM were coordinated through the Standard Reference Materials Program by J.W.L. Thomas.

Bert M. Coursey, Chief  
Ionizing Radiation Division

Gaithersburg, Maryland 20899  
June 1999  
Half-life and text revised October 2000

Nancy M. Trahey, Chief  
Standard Reference Materials Program

**This page intentionally left blank.**

PROPERTIES OF SRM 4926E

Certified values

Solution density	(0.998 ± 0.002) g·mL <sup>-1</sup> at 20.0 °C [b]*
Radionuclide	Hydrogen-3
Reference time	1200 EST, 3 September 1998
Massic activity of the solution [c]	5.038 kBq·g <sup>-1</sup>
Relative expanded uncertainty (k=2)	0.72% [d] [e]

Uncertified values

Physical Properties:			
Source description	Liquid in a 20-mL borosilicate-glass serum vial with a butyl-rubber stopper (septum)		
Solution mass	Approximately 20 g		
Chemical Properties:			
Solution composition	Chemical Formula	Concentration (mol·L <sup>-1</sup> )	Mass Fraction (g·g <sup>-1</sup> )
	H <sub>2</sub> O <sup>3</sup> HHO	55 5 × 10 <sup>-9</sup>	1.00 1 × 10 <sup>-10</sup>
Radiological Properties:			
Radionuclidic impurities	None detected [f]		
Half lives used	Hydrogen-3: (4500 ± 8) d [g]		
Calibration method and measuring instrument(s)	4πβ gas counting of SRM 4927E using the NIST length-compensated internal gas proportional counters and intercomparison of SRMs 4927E/4926E using two 4πβ liquid-scintillation counting systems [h]		

EVALUATION OF THE UNCERTAINTY OF THE MASSIC ACTIVITY [d]\*

Input Quantity $x_i$ , the source of uncertainty  (and individual uncertainty components where appropriate)	Method Used To Evaluate $u(x_i)$ , the standard uncertainty of $x_i$ (A) denotes evaluation by statistical methods (B) denotes evaluation by other methods	Relative Uncertainty Of Input Quantity, $u(x_i)/x_i$ , (%) [i]	Relative Sensitivity Factor, $ \partial y/\partial x_i  \cdot$ $(x_i/y)$ [j]	Relative Uncertainty Of Output Quantity, $u_i(y)/y$ , (%) [k]
Massic count rate of SRM 4927E, corrected for background and decay [h]	Standard deviation of the mean for 23 sets of gas counting measurements (A)	0.18	1.0	0.18
Gram-mole measurements	Estimated (B)	0.20	1.0	0.20
Live-time [p]	Estimated (B)	0.10	1.0	0.10
Extrapolation of count-rate-versus-energy to zero energy	Estimated (B)	0.20	1.0	0.20
Half life of H-3	Standard uncertainty of the half life (A)	0.18 [m]	0.009 [n]	0.002
Liquid-scintillation intercomparison of SRM 4926E and SRM 4927E	Standard deviation of the mean for 7 sets of liquid-scintillation measurements (A)	0.06	1.0	0.06
Radionuclidic impurities	Limit of detection (B) [q]	100.	0.0005	0.05
Relative Combined Standard Uncertainty of the Output Quantity, $u_c(y)/y$ , (%)				0.36
Coverage Factor, $k$				<u>x 2</u>
Relative Expanded Uncertainty of the Output Quantity, $U/y$ , (%)				0.72

## NOTES

- [a] The Sievert is the SI unit for dose equivalent. See reference [1]. One  $\mu\text{Sv}$  is equal to 0.1 mrem.  
Distance from vial (cm):           1     30     100  
Approximate Dose Rate ( $\mu\text{Sv/h}$ ):   <0.1 (Not detectable)
- [b] The stated uncertainty is two times the standard uncertainty.
- [c] **Massic activity** is the preferred name for the quantity activity divided by the total mass of the sample. See reference [1].
- [d] The reported value,  $y$ , of massic activity (activity per unit mass) at the reference time was not measured directly but was derived from measurements and calculations of other quantities. This can be expressed as  $y = f(x_1, x_2, x_3, \dots, x_n)$ , where  $f$  is a mathematical function derived from the assumed model of the measurement process.
- The value,  $x_i$ , used for each input quantity  $i$  has a **standard uncertainty**,  $u(x_i)$ , that generates a corresponding uncertainty in  $y$ ,  $u_i(y) \equiv |\partial y / \partial x_i| \cdot u(x_i)$ , called a **component of combined standard uncertainty** of  $y$ .
- The **combined standard uncertainty** of  $y$ ,  $u_c(y)$ , is the positive square root of the sum of the squares of the components of combined standard uncertainty.
- The combined standard uncertainty is multiplied by a **coverage factor** of  $k = 2$  to obtain  $U$ , the **expanded uncertainty** of  $y$ .
- Since it can be assumed that the possible estimated values of the massic activity are approximately normally distributed with approximate standard deviation  $u_c(y)$ , the unknown value of the massic activity is believed to lie in the interval  $y \pm U$  with a level of confidence of approximately 95 percent.
- For further information on the expression of uncertainties, see references [2] and [3].
- [e] The value of each standard uncertainty component, and hence the value of the expanded uncertainty itself, is a best estimate based upon all available information, but is only approximately known. That is to say, the "uncertainty of the uncertainty" is large and not well known. This is true for uncertainties evaluated by statistical methods (e.g., the relative standard deviation of the standard deviation of the mean for the massic response is approximately 50%) and for uncertainties evaluated by other methods (which could easily be over estimated or under estimated by substantial amounts). The unknown value of the expanded uncertainty is believed to lie in the interval  $U/2$  to  $2U$  (i.e., within a factor of 2 of the estimated value).
- [f] The estimated limit of detection for radionuclidic impurities is  $3 \text{ Bq} \cdot \text{g}^{-1}$ .
- [g] The stated uncertainty is the standard uncertainty. See reference [5].
- [h] Extensive gas-counting measurements were made on the SRM 4927E solution during 1998 and 1999. The SRM 4926E solution was intercompared with the SRM 4927E solution using liquid-scintillation counting.
- [i] Relative standard uncertainty of the input quantity  $x_i$ .

- [j] The relative change in the output quantity  $y$  divided by the relative change in the input quantity  $x_i$ . If  $|\partial y/\partial x_i| \cdot (x_i/y) = 1.0$ , then a 1% change in  $x_i$  results in a 1% change in  $y$ . If  $|\partial y/\partial x_i| \cdot (x_i/y) = 0.05$ , then a 1% change in  $x_i$  results in a 0.05% change in  $y$ .
- [k] Relative component of combined standard uncertainty of output quantity  $y$ , rounded to two significant figures or less. The relative component of combined standard uncertainty of  $y$  is given by  $u_i(y)/y \equiv |\partial y/\partial x_i| \cdot u(x_i)/y = |\partial y/\partial x_i| \cdot (x_i/y) \cdot u(x_i)/x_i$ . The numerical values of  $u(x_i)/x_i$ ,  $|\partial y/\partial x_i| \cdot (x_i/y)$ , and  $u_i(y)/y$ , all dimensionless quantities, are listed in columns 3, 4, and 5, respectively. Thus, the value in column 5 is equal to the value in column 4 multiplied by the value in column 3. The input quantities are independent, or very nearly so. Hence the covariances are zero or negligible.
- [m] The relative standard uncertainty of  $\lambda \cdot t$  is determined by the relative standard uncertainty of  $\lambda$  (i.e., of the half life). The relative standard uncertainty of  $t$  is negligible.
- [n]  $|\partial y/\partial x_i| \cdot (x_i/y) = |\lambda \cdot t|$
- [p] The live time is determined by counting the pulses from a gated crystal-controlled oscillator.
- [q] The standard uncertainty for each undetected impurity that might reasonably be expected to be present is estimated to be equal to the estimated limit of detection for that impurity, i.e.  $u(x_i)/x_i = 100\%$ .  $|\partial y/\partial x_i| \cdot (x_i/y) = \{(\text{response per Bq of impurity})/(\text{response per Bq of H-3})\} \cdot \{(\text{Bq of impurity})/(\text{Bq of H-3})\}$ . Thus  $u_i(y)/y$  is the relative change in  $y$  if the impurity were present with a massic activity equal to the estimated limit of detection.

#### REFERENCES

- [1] International Organization for Standardization (ISO), *ISO Standards Handbook - Quantities and Units*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900.
- [2] International Organization for Standardization (ISO), *Guide to the Expression of Uncertainty in Measurement*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900. (Listed under ISO miscellaneous publications as "ISO Guide to the Expression 1993".)
- [3] B. N. Taylor and C. E. Kuyatt, *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*, NIST Technical Note 1297, 1994. Available from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20407, U.S.A.
- [4] National Council on Radiation Protection and Measurements Report No. 58, *A Handbook of Radioactivity Measurements Procedures*, Second Edition, 1985. Available from the National Council on Radiation Protection and Measurements, 7910 Woodmont Avenue, Bethesda, MD 20814 U.S.A.
- [5] L.L. Lucas and M.P. Unterweger, *Comprehensive Review and Critical Evaluation of the Half-Life of Tritium*, J. Res. Natl. Inst. Stand. Technol. **105**, 541-549 (2000).