

UNITED STATES DEPARTMENT OF COMMERCE
NATIONAL BUREAU OF STANDARDS
WASHINGTON 25, D. C.

DISCONTINUED

National Bureau of Standards
Certificate

BENZOIC ACID THERMOMETRIC STANDARD


943

Cell Number 77

The freezing temperature at its maximum in this cell was found to be 122.359°C on the International Practical Temperature Scale of 1948 when the procedures indicated by the accompanying instructions were followed.

For the Director

by



Delmo Enagonio, Chemist
Standard Reference Materials Section
Analytical & Inorganic Chemistry
Division

Washington, D. C.
Date: September 12, 1962