

5.1.4. GRAVIMETRIC STANDARDS

Calibration of Working Standards

In May 1999 the three-channel gas chromatograph system used to calibrate standards for the in situ monitoring program was replaced by a four-channel, custom-built GC. The four-channel GC is similar to the four-channel CATS instruments used for in situ monitoring at the observatories. The GC is operated under slightly different conditions compared to those at the observatories in order to optimize chromatography for cylinder calibration (Table 5.6).

The new four-channel instrument was compared to the old GC to ensure that the new GC would provide similar calibration results. For calibrations of ambient-level working standards, the new GC agreed to within 0.1% for N₂O, 0.4% for CFC-113, 0.5% for CH₃CCl₃, and 1% for CFC-11 and CCl₄. Initial tests revealed a 2-3% discrepancy for CFC-12. This experiment was repeated several months later, and the discrepancy could not be reproduced. Testing in December 1999 showed that the GCs agreed to within 0.2% for CFC-12. Following this CFC-12 experiment, the old three-channel GC instrument was retired. The four-channel GC is now used to calibrate working standards. Each standard is typically analyzed for 2 full days over a 2-3 week period (8-12 injections each day) (Table 5.7).

Preparation of Working Standards

Working standards for use with laboratory and in situ GCs continue to be filled at the C-1 site at Niwot Ridge, Colorado (40.04°N, 105.54°W, elevation 3013 m). Cylinders are filled using a three-stage, SA-6B compressor from Rix Industries (Oakland, California). Cylinders have occasionally been contaminated with trace amounts of unknown compounds that elute near CFC-11 and CFC-113 on the columns used in the CMDL in situ program. Contaminated standards have traditionally been associated with warm temperatures, while clean standards have been obtained during winter when ambient air temperatures are often below 5°C. The contaminants are thought to result from outgassing of material from the piston rings of one or more of the compressor stages (Ray Weiss, personal communication, 1999). Outgassing may be more severe at higher operating temperatures. To reduce the operating temperature of the first stage of the compressor, the compressor was modified in order to allow a small amount of distilled, degassed water to be injected into the air stream just upstream of

TABLE 5.6. Configuration of the Four-Channel GC for Calibration of Working Standards

Channel	Column	Detector	Compounds Resolved
1	Porapak Q	Valco ECD	N ₂ O, SF ₆
2	Unibeads 1s	Shimadzu ECD	CFC-12, CFC-11, halon-1211
3	OV-101	Valco ECD	CFC-12, CFC-11, halon-1211, CFC-113, CHCl ₃ , CH ₃ CCl ₃ , CCl ₄
4	Poraplot Q	Shimadzu ECD	HCFC-22, CH ₃ Cl, CH ₃ Br

TABLE 5.7. Analytical Precision Typically Obtained for a Complete Calibration of a Working Standard on the Four-Channel GC-ECD

Compound*	Precision (%)
N ₂ O	0.05
CFC-12	0.08
CFC-11	0.10
CFC-113	0.15
halon-1211†	0.2
CHCl ₃	1.0
CH ₃ CCl ₃	0.2
CCl ₄	0.1
SF ₆	0.8

*Calibration of HCFC-22, CH₃Cl and CH₃Br is only performed on the flask GC-MS instrument at this time.

†The calibration scale for halon-1211 is determined on the flask GC-ECD instrument. Halon-1211 calibration on the four-channel instrument is complicated by a co-eluting peak occurring on both channels 2 and 3.

the first stage. Injection of water in this manner reduces the operating temperature of the first stage cylinder head from >100°C to ~50°C. Contamination is substantially reduced by injecting water at a rate of 10-20 mL min⁻¹ (Figure 5.22).

Standards Preparation

A total of 66 natural-air standards were filled at Niwot Ridge during the 1998-1999 period. Fifty-eight of these were used by CMDL or related projects; eight standards were prepared for outside laboratories. A total of 10 gravimetric standards were prepared. Four were prepared for CMDL use, while 6 were prepared for outside laboratories.

Calibration Scales

A summary of the calibration scales used in conjunction with measurements performed by the HATS group is reported here. Table 5.8 lists the year that each calibration scale was established. Table 5.9 lists the gravimetric standards and mixing ratios used to define each calibration scale. Factors that




Fig. 5.22. Chromatograms of air samples from tanks pumped at Niwot Ridge, Colorado, with and without water injection (see text). The chromatogram (top) without water injection shows more contamination peaks.

TABLE 5.8. Reference Dates for HATS Calibration Scales

Compound	Year*	Reference‡
N ₂ O	1993	Swanson <i>et al.</i> [1993]
CFC-12	1997	Butler <i>et al.</i> [1998b]
CFC-11	1992	Elkins <i>et al.</i> [1993]
CFC-113	1993	Swanson <i>et al.</i> [1993]
CH ₃ CCl ₃	1996	Butler <i>et al.</i> [1998b]
CCl ₄	1996	Butler <i>et al.</i> [1998b]
HCFC-22	1992	Montzka <i>et al.</i> [1993]
HCFC-141b	1994	Montzka <i>et al.</i> [1994]
HCFC-142b	1994	Montzka <i>et al.</i> [1994]
HCFC-134a	1995	Montzka <i>et al.</i> [1996b]
CH ₃ Cl	1996	Butler <i>et al.</i> [1998b]
CH ₃ Br	1996	Butler <i>et al.</i> [1998b]
CH ₂ Cl ₂ †	1992	Spivakovskiy <i>et al.</i> [2000]
C ₂ Cl ₄ †	1992	Hurst <i>et al.</i> [1997]
CHCl ₃ †	1992	Hurst <i>et al.</i> [1997]
halon-2402†	1992	Butler <i>et al.</i> [1998b]
halon-1211	1996	Butler <i>et al.</i> [1998a]
halon-1301	1990	Butler <i>et al.</i> [1998a]
SF ₆	1994	Geller <i>et al.</i> [1997]
H ₂	1995	Novelli <i>et al.</i> [1999]
CO	1991	Novelli <i>et al.</i> [1991]

*Calibration scales are defined by a particular set of gravimetric standards (see Table 5.9). In some cases only a single standard is used. The year shown for each species refers to either the year during which the standards were prepared or the year in which a significant change in calibration scale occurred. Thus data reported prior to the year listed are associated with a different calibration scale than those reported in subsequent years.

†These species are measured by CMDL, but their calibration is defined only by an incomplete set of standards.

‡The reference listed for each species refers to a publication that describes a calibration change or a publication that includes recent results.

influence the number of gravimetric standards used to define a particular scale include the time-history of the measurements, the range of concentrations investigated, and the linearity of response (e.g., CFC-12 measurements were made over more extensive spatial and temporal scales than CH₃Br, thus a large number of CFC-12 standards were prepared over the years). All of the standards listed in Table 5.9 were prepared in 5.9-L or 29.5-L Aculife-treated aluminum cylinders (Scott Specialty Gases, Plumsteadville, Pennsylvania).

Examples of calibration response curves determined on the CATS GC instrument are shown in Figure 5.23. The uncertainty associated with the calibration curves (expressed as the standard error of the residuals) is typically 0.2-0.5%.




Fig. 5.23. Sample calibration curves for CFC-12 (a), CH₃CCl₃ (b), and CCl₄ (c). Each filled circle represents the response of a gravimetric standard relative to a working standard. Solid lines are second-order polynomial fits to response data.

The number of gravimetric standards used to define scales for N₂O and SF₆ will be expanded in the near future. A number of standards with N₂O and SF₆ mixing ratios near ambient levels will be prepared. The addition of numerous standards at near-ambient concentrations will enable high-precision maintenance of the scales over time.

TABLE 5.9. Gravimetric Standards Used to Define Calibration Scales for the HATS Group

Compound	Cylinder Number	Prepared Gravimetric Concentration	Compound	Cylinder Number	Prepared Gravimetric Concentration
N ₂ O	*ALM-38408	96.2 ppb	SF ₆	*CLM-7506	3.1 ppt
	*ALM-26738	172.0 ppb		*CLM-7519	23.2 ppt
	*ALM-26743	331.0 ppb		*CLM-7494	60.6 ppt
	*ALM-26737	359.9 ppb		*CLM-7503	107.6 ppt
	*CLM-30135	801.1 ppb		†CLM-7490	5.3 ppt
CFC-11	*ALM-38408	25.5 ppt	CH ₃ Br	†ALM-62626	5.3 ppt
	*CLM-2431	49.5 ppt		†ALM-39771	25.6 ppt
	*ALM-38417	80.9 ppt	CH ₃ Cl	†ALM-39971	1024 ppt
	*CLM-2482	69.2 ppt		†ALM-26738	105 ppt
	*ALM-26735	75.16 ppt	HCFC-22	†ALM-26743	145 ppt
	*CLM-2413	136.3 ppt		†ALM-26735	105 ppt
	*ALM-26743	288.2 ppt		†ALM-38417	111.6 ppt
CFC-12	*ALM-26737	313.6 ppt	HCFC-141b	†ALM-26737	158 ppt
	*CLM-2426	77.8 ppt		†ALM-39758	5.0 ppt
	*CLM-2431	163.5 ppt		†ALM-39744	25.5 ppt
	*CLM-9015	189.2 ppt	HCFC-142b	†ALM-39749	51.3 ppt
	*CLM-2482	228.8 ppt		†ALM-39758	5.0 ppt
	*CLM-9041	274.4 ppt		†ALM-39744	25.4 ppt
	*CLM-9018	370.3 ppt	HCFC-134a	†ALM-39749	51.1 ppt
	*CLM-2413	450.7 ppt		†CLM-8952	5.44 ppt
	*CLM-9038	646.9 ppt		†CLM-9036	10.22 ppt
	*CLM-2482	19.1 ppt		*†CLM-9039	3.02 ppt
CFC-113	*CLM-2413	37.5 ppt	halon-1211	*†CLM-9024	5.41 ppt
	*†ALM-26735	48.0 ppt	halon-1301	*†CLM-2416	95.4 ppt
	†ALM-26738	48.0 ppt		*†FF-30071	6.08 ppb
	*†ALM-26743	82.8 ppt	halon-2402	†ALM-26748	47.5 ppt
	*†ALM-26737	90.0 ppt	CH ₂ Cl ₂	†ALM-26748	122.0 ppt
	*†ALM-52792	30.9 ppt	C ₂ Cl ₄	†ALM-26748	157.0 ppt
CH ₃ CCl ₃	*†ALM-52788	70.8 ppt	CHCl ₃	†ALM-26748	141.0 ppt
	*†ALM-59967	111.0 ppt			
	*†ALM-52811	151.4 ppt			
	*†ALM-52792	22.1 ppt			
	*†ALM-52788	50.6 ppt			
CCl ₄	*†ALM-52788	50.6 ppt			
	*†ALM-59967	79.4 ppt			
	*†ALM-52811	108.3 ppt			

*These standards are used to define the calibration scales associated with GC-ECD instruments.

†These standards are used to define the calibration scales associated with GC-MS instruments.