

§ 53.21

40 CFR Ch. I (7-1-08 Edition)

method determination under this paragraph (b).

(c) For each performance specification (except Range), the test procedure shall be initially repeated seven (7) times to yield 7 test results. Each result shall be compared with the corresponding specification in table B-1; a value higher than or outside that specified constitutes a failure. These 7 results for each parameter shall be interpreted as follows:

(1) Zero (0) failures: Candidate method passes the performance parameter.

(2) Three (3) or more failures: Candidate method fails the performance parameter.

(3) One (1) or two (2) failures: Repeat the test procedures for the parameter eight (8) additional times yielding a total of fifteen (15) test results. The combined total of 15 test results shall then be interpreted as follows:

(i) One (1) or two (2) failures: Candidate method passes the performance parameter.

(ii) Three (3) or more failures: Candidate method fails the performance parameter.

TABLE B-1—PERFORMANCE SPECIFICATIONS FOR AUTOMATED METHODS

Performance parameter	Units ¹	Sulfur dioxide	Photochemical oxidants	Carbon monoxide	Nitrogen dioxide	Definitions and test procedures
1. Range	Parts per million ...	0-0.5	0-0.5	0-50	0-0.5	Sec. 53.23(a).
2. Noisedo005	.005	.50	.005	Sec. 53.23(b).
3. Lower detectable limit	Parts per million01	.01	1.0	.01	Sec. 53.23(c).
4. Interference equivalentdododododo	Sec. 53.23(d).
Each interferant	Parts per million ...	±.02	±.02	±1.0	±0.02	
Total interferantdo06	.06	1.5	.04	
5. Zero drift, 12 and 24 hourdo	±.02	±.02	±1.0	±.02	Sec. 52.23(e).
6. Span drift, 24 hourdododododo	Do.
20 percent of upper range limit	Percent	±20.0	±20.0	±10.0	±20.0	
80 percent of upper range limitdo	±5.0	±5.0	±2.5	±5.0	
7. Lag time	Minutes	20	20	10	20	Do.
8. Rise timedo	15	15	5	15	Do.
9. Fall timedo	15	15	5	15	Do.
10. Precisiondododododo	Do.
20 percent of upper range limit	Parts per million01	.01	.5	.02	
80 percent of upper range limitdo015	.01	.5	.03	

¹ To convert from parts per million to µg/m³ at 25 °C and 760 mm Hg, multiply by M/0.02447, where M is the molecular weight of the gas.

(d) The tests for *zero drift*, *span drift*, *lag time*, *rise time*, *fall time*, and *precision* shall be combined into a single sequential procedure to be conducted at various line voltages and ambient temperatures specified in § 53.23(e). The tests for *noise*, *lower detectable limit*, and *interference equivalents* shall be made at any temperature between 20 °C. and 30 °C. and at any normal line voltage between 105 and 125 volts, and shall be conducted such that not more than three (3) test results for each parameter are obtained per 24 hours.

(e) All response readings to be recorded shall first be converted to concentration units according to the calibration curve constructed in accordance with § 53.21(b).

(f) All recorder chart tracings, records, test data and other documentation obtained from or pertinent

to these tests shall be identified, dated, signed by the analyst performing the test, and submitted.

NOTE: Suggested formats for reporting the test results and calculations are provided in Figures B-2, B-3, B-4, B-5, and B-6 in appendix A. Symbols and abbreviations used in this subpart are listed in table B-5, appendix A.

[40 FR 7049, Feb. 18, 1975, as amended at 40 FR 18168, Apr. 25, 1975; 41 FR 52694, Dec. 1, 1976]

§ 53.21 Test conditions.

(a) *Set-up and start-up* of the test analyzer shall be in strict accordance with the operating instructions specified in the manual referred to in § 53.4(b)(3). Allow adequate warm-up or stabilization time as indicated in the operating instructions before beginning the tests.

If the candidate method does not include an integral strip chart recorder, connect the output signal of the test analyzer to a suitable strip chart recorder of the servo, null-balance type. This recorder shall have a chart width of at least 25 centimeters, chart speeds up to 10 cm per hour, a response time of 1 second or less, a deadband of not more than 0.25 percent of full scale, and capability either of reading measurements at least 5 percent below zero or of offsetting the zero by at least 5 percent.

NOTE: Other data acquisition components may be used along with the chart recorder during conduct of these tests. Use of the chart recorder is intended only to facilitate evaluation of data submitted.

(b) *Calibration* of the test analyzer shall be as indicated in the manual referred to in § 53.4(b)(3) and as follows: If the chart recorder does not have below zero capability, adjust either the controls of the test analyzer or the chart recorder to obtain a +5% offset zero reading on the recorder chart to facilitate observing negative response or drift. If the candidate method is not capable of negative response, the test analyzer (not recorder) shall be operated with an offset zero. Construct and submit a calibration curve showing a plot of recorder scale readings (ordinate) against pollutant concentrations (abscissa). A plot of output units (volts, millivolts, milliamps, etc.) against pollutant concentrations shall also be shown for methods not including an integral chart recorder. All such plots shall consist of at least seven (7) approximately equally spaced, identifiable points, including 0 and 90±5 percent of full scale.

(c) Once the test analyzer has been set up and calibrated and the tests started, manual adjustment or normal periodic maintenance is permitted only every 3 days. Automatic adjustments which the test analyzer performs by itself are permitted at any time. The submitted records shall show clearly when any manual adjustment or periodic maintenance was made and describe the operations performed.

(d) If the test analyzer should malfunction during any of the performance tests, the tests for that parameter shall be repeated. A detailed explanation of the malfunction, remedial action taken, and whether recalibration was necessary (along with all pertinent records and charts) shall be submitted. If more than one malfunction occurs, all performance test procedures for all parameters shall be repeated.

(e) Tests for all performance parameters shall be completed on the same test analyzer, except that use of multiple test analyzers to accelerate testing will be permitted when alternate ranges of a multi-range candidate method are being tested.

§ 53.22 Generation of test atmospheres.

(a) Table B-2 specifies preferred methods for generating test atmospheres and suggested methods of verifying the concentrations. Only one means of establishing the concentration of a test atmosphere is normally required. If the method of generation can produce reproducible concentrations, verification is optional. If the method of generation is not reproducible, then establishment of the concentration by some verification method is required. However, when a method of generation other than that given in table B-2 is used, the test concentration shall be verified.

(b) The test atmosphere delivery system shall be designed and constructed so as not to significantly alter the test atmosphere composition or concentration during the period of the test. The delivery system shall be fabricated from borosilicate glass or FEP Teflon.

(c) The output of the test atmosphere generation system shall be sufficiently stable to obtain stable response during the required tests. If a permeation device is used for generation of a test atmosphere, the device, as well as the air passing over it, shall be controlled to ±0.1 °C.

(d) All diluent air shall be zero air free of contaminants likely to cause a detectable response on the test analyzer.