

## § 640.4

## 21 CFR Ch. I (4–1–01 Edition)

health, as indicated in part in paragraph (b) of this section.

[38 FR 32089, Nov. 20, 1973, as amended at 49 FR 23834, June 8, 1984; 50 FR 4138, Jan. 29, 1985; 51 FR 15611, Apr. 25, 1986; 55 FR 11013, Mar. 26, 1990; 64 FR 45371, Aug. 19, 1999; 66 FR 1836, Jan. 10, 2001]

### § 640.4 Collection of the blood.

(a) *Supervision.* Blood shall be drawn from the donor by a qualified physician or under his supervision by assistants trained in the procedure. A physician shall be present on the premises when blood is being collected, except that blood may be collected when a physician is not present on the premises, provided the establishment (1) maintains on the premises, and files with the Center for Biologics Evaluation and Research, a manual of standard procedures and methods, approved by the Director of the Center for Biologics Evaluation and Research, that shall be followed by employees who collect blood, and (2) maintains records indicating the name and qualifications of the person immediately in charge of the employees who collect blood when a physician is not present on the premises.

(b) *The donor center.* The pertinent requirements of §§ 600.10 and 600.11 of this chapter shall apply at both the blood establishment and at any other place where the bleeding is performed.

(c) *Blood containers.* Blood containers and donor sets shall be pyrogen-free, sterile and identified by lot number. The amount of anticoagulant required for the quantity of blood to be collected shall be in the blood container when it is sterilized. In addition, all container and donor set surfaces that come in contact with blood used in the processing of Heparin Whole Blood shall be water repellent.

(d) *The anticoagulant solution.* The anticoagulant solution shall be sterile and pyrogen-free. Anticoagulant solutions shall be compounded and used according to a formula approved by the Director, Center for Biologics Evaluation and Research.

(e) *Donor identification.* Each unit of blood shall be so marked or identified by number or other symbol as to relate it to the individual donor whose iden-

tity shall be established to the extent necessary for compliance with § 640.3.

(f) *Prevention of contamination of the blood.* The skin of the donor at the site of phlebotomy shall be prepared thoroughly and carefully by a method that gives maximum assurance of a sterile container of blood. The blood shall be collected by aseptic methods in a sterile system which may be closed or may be vented if the vent protects the blood against contamination.

(g) *Pilot samples for laboratory tests.* Pilot samples for laboratory tests shall meet the following standards:

(1) One or more pilot samples shall be provided with each unit of blood when issued or reissued except as provided in § 640.2(c)(2) and all pilot samples shall be from the donor who is the source of the unit of blood.

(2) All samples for laboratory tests performed by the manufacturer and all pilot samples accompanying a unit of blood shall be collected at the time of filling the final container by the person who collects the unit of blood.

(3) All containers for all samples shall bear the donor's identification before collecting the samples.

(4) All containers for pilot samples accompanying a unit of blood shall be attached to the whole blood container before blood collection in a tamperproof manner that will conspicuously indicate removal and reattachment.

(5) When CPDA-1 is used, pilot samples for compatibility testing shall contain blood mixed with CPDA-1.

(h) *Storage.* Immediately after collection, unless the blood is to be used as a source for Platelets, it shall be placed in storage at a temperature between 1 and 6 °C unless it must be transported from the donor clinic to the processing laboratory. In the latter case, the blood shall be placed in temporary storage having sufficient refrigeration capacity to cool the blood continuously toward a range between 1 and 6 °C until it arrives at the processing laboratory, where it shall be stored at a temperature between 1 and 6 °C. Blood from which Platelets is to be prepared shall be held in an environment maintained at a temperature range 20 to 24 °C until the platelets are separated. The red blood cells shall be placed in storage at

## Food and Drug Administration, HHS

## § 640.6

a temperature between 1 and 6 °C immediately after the platelets are separated.

[38 FR 32089, Nov. 20, 1973, as amended at 42 FR 59878, Nov. 22, 1977; 43 FR 34460, Aug. 4, 1978; 49 FR 23834, June 8, 1984; 50 FR 4138, Jan. 29, 1985; 55 FR 11013, Mar. 26, 1990; 64 FR 45372, Aug. 19, 1999; 66 FR 1836, Jan. 10, 2001]

### § 640.5 Testing the blood.

All laboratory tests shall be made on a pilot sample specimen of blood taken from the donor at the time of collecting the unit of blood, and these tests shall include the following:

(a) *Serological test for syphilis.* Whole Blood shall be negative to a serological test for syphilis.

(b) *Determination of blood group.* Each container of Whole Blood shall be classified as to ABO blood group. At least two blood group tests shall be made and the unit shall not be issued until grouping tests by different methods or with different lots of antisera are in agreement. Only those Anti-A and Anti-B Blood Grouping Reagents licensed under, or that otherwise meet the requirements of, the regulations of this subchapter shall be used, and the technique used shall be that for which the serum is specifically designed to be effective.

(c) *Determination of the Rh factors.* Each container of Whole Blood shall be classified as to Rh type on the basis of tests done on the pilot sample. The label shall indicate the extent of typing and the results of all tests performed. If the test, using Anti-D Blood Grouping Reagent, is positive, the container may be labeled "Rh Positive". If this test is negative, the results shall be confirmed by further testing which may include tests for the Rh<sub>0</sub> variant (D<sup>u</sup>) and for other Rh-Hr factors. Blood may be labeled "Rh Negative" if negative to tests for the Rh<sub>0</sub> (D) and Rh<sub>0</sub> variant (D<sup>u</sup>) factors. If the test using Anti-D Blood Grouping Reagent is negative, but not tested for the Rh<sub>0</sub> variant (D<sup>u</sup>), the label must indicate that this test was not done. Only Anti-Rh Blood Grouping Reagents licensed under, or that otherwise meet the requirements of, the regulations of this

subchapter shall be used, and the technique used shall be that for which the serum is specifically designed to be effective.

(d) *Sterility test.* Whole Blood intended for transfusion shall not be tested for sterility by a method that entails entering the final container before the blood is used for transfusion.

(e) *Inspection.* Whole Blood shall be inspected visually during storage and immediately prior to issue. If the color or physical appearance is abnormal or there is any indication or suspicion of microbial contamination the unit of Whole Blood shall not be issued for transfusion.

(f) *Test for antibody to HIV.* Whole Blood shall be tested for antibody to HIV as prescribed in §610.45 of this chapter.

[38 FR 32089, Nov. 20, 1973, as amended at 50 FR 4138, Jan. 29, 1985; 53 FR 117, Jan. 5, 1988; 53 FR 12764, Apr. 19, 1988; 64 FR 45372, Aug. 19, 1999; 66 FR 1836, Jan. 10, 2001]

### § 640.6 Modifications of Whole Blood.

Upon approval by the Director, Center for Biologics Evaluation and Research, of a supplement to the biologics license application for Whole Blood a manufacturer may prepare Whole Blood from which the antihemophilic factor has been removed, provided the Whole Blood meets the applicable requirements of this subchapter and the following conditions are met:

(a) The antihemophilic factor shall be removed in accordance with paragraphs (a), (b), and (c) of § 640.52.

(b) Although the closed system between the red blood cells and plasma shall be maintained, the red blood cells shall be maintained between 1 and 6° C at all times, including that time when the plasma is being frozen for removal of the antihemophilic factor.

[38 FR 32089, Nov. 20, 1973, as amended at 49 FR 23834, June 8, 1984; 50 FR 4138, Jan. 29, 1985; 55 FR 11013, Mar. 26, 1990; 59 FR 49351, Sept. 28, 1994; 64 FR 45372, Aug. 19, 1999; 64 FR 56453, Oct. 20, 1999]