

Laboratory Procedure Manual

Analyte: Urinary Albumin

Matrix: Urine

*Method: Fluorometer, Sequoia-Turner model 450,
digital (Sequoia-Turner Corp., Mountain View, CA)*

Microalbumin in Urine - NHANES 2001-2002

0. Public Release Data Set Information

This document details the Lab Protocol for NHANES 2001-2002 data.

A list of the released analytes follows:

| Lab | Analyte | SAS Label | Description |
|------------|----------------|--------------------------|--------------------|
| I16_b | URXUMA | Albumin, urine (ug/mL) | Albumin, urine |
| I16_b | URXUMASI | Albumin, urine (mg/L) SI | |

1. SUMMARY OF TEST PRINCIPLE AND CLINICAL RELEVANCE

A fluorescent immunoassay for the measurement of human urinary albumin is described by Chavers et al.¹ The methodology involves solid-phase, non-competitive, double-antibody reaction. Urine specimen albumin antigen reacts with albumin antibody that is covalently attached to polyacrylamide beads. This resulting solid-phase antibody complex is then reacted with fluorescein-labeled antibody. Unattached fluorescent antibody and other proteins are removed by centrifugation. The fluorescence of the stable solid-phase double-antibody complex is measured with a fluorometer and is directly proportional to the amount of urine albumin present. The standard line calibration material is human serum albumin with a range of 0.5 to 20 µg/mL.

Increased microalbuminuria is a sign of renal disease and may be predictive of nephropathy risk in patients with type 1 and type 2 diabetes. It is associated with hypertension and cardiac disease. The fluorescent immunoassay (FIA) has been in service for several years and has proven to be reliable, accurate, precise, and sensitive for the detection of urinary albumin excretion. It is especially useful for quantitative analysis of low levels of urinary albumin. The FIA assay resembles the radio-immunoassay (RIA) in technique and sensitivity without the potential health hazards associated with the handling of isotopes in the laboratory.

2. SPECIAL SAFETY PRECAUTIONS

- a. Follow the Laboratory Safety and General Laboratory Practice regulations from the College of American Pathologists (CAP), the Clinical Laboratory Improvement Act (CLIA), and Occupational Safety and Health Administration (OSHA). Observe Universal Blood and Body Substance Technique (UBBST) and the Centers for Disease Control (MMWR 36;2S;1987) precautions for prevention of HIV transmission in the health care setting.
- b. Wear laboratory coats and disposable gloves when handling urine specimens. Cover the work surface with disposable, absorbent toweling. Place contaminated tubes, pipette tips, gloves, toweling, etc., and residual urine specimens into an autoclave bag. Autoclave the contaminants prior to discard. Clean the work surfaces with 0.5% bleach.
- c. Recommend to laboratory personnel performing the assay that they receive the HBV vaccine. Maintain records of vaccination or signed declination forms in the laboratory.
- d. Label all reagents indicating the preparation date, expiration date, formula, lot number if applicable, hazards of the reagent, antidote of contact with hazard, and the initials of the technician.
- e. Note the location of the Material Data Safety Sheets for the pertinent chemicals used in this assay: Section Q of the notebook "Chemical Hygiene Plan, University of Minnesota, Department of Pediatrics".

3. COMPUTERIZATION; DATA SYSTEM MANAGEMENT

COMPUTERIZATION

Linear regression analysis for the calculation of test results is performed on a Dell IBM PC computer with Microsoft Excel Software. Format an Excel electronic spreadsheet to be used as a template document for the assay result calculations. Program the formulas for linear regression analysis in order to perform a fill-down of calculations for the unknown specimen values. On this document include sample identification information, (urine collection time and volume, if applicable), the specimen Albustix value, the dilution factor, the duplicate percent relative fluorescence (% RF) values, the specimen values converted to µg/mL, and the calculated total µg/mL albumin concentration. Prepare a spreadsheet from the template for each assay.

DATA SYSTEM MANAGEMENT

The integrity of the specimen data is established by routine verification of the transcribed information against the identification on the specimen tube and the hard copy that accompanies all specimens. Urine albumin results with pertinent specimen information are reviewed by the laboratory supervisor and are returned by hard copy or electronic transmission as per study protocols. (For the NHANES IV, results are returned by e-mail onto formatted worksheets that were transmitted to us by the NHANES contract laboratory, Westat). The data is stored on the computer hard drive and backed-up monthly onto zip disks. The hard copies are organized in notebooks. Worksheets and results are archived indefinitely and located in the laboratory.

4. SPECIMEN COLLECTION, STORAGE, AND HANDLING PROCEDURES; CRITERIA FOR SPECIMEN REJECTION

SPECIMEN COLLECTION PROCEDURE

- a. Timed or random urine collections are obtained from study subjects as per study protocols. (For the NHANES IV, random urine specimens are collected at the survey mobile examination centers.)
- b. No special instructions such as fasting or special diets are requested.

Microalbumin in Urine - NHANES 2001-2002

- c. The optimum specimen tube is a 3- to 5-mL screw-top cryogenic vial. Tubes are selected per study protocols.
- d. The optimum sample volume is 3 mL, and the minimum acceptable volume is 1 mL.
- e. Design study protocols to ensure consistent specimen collection procedures for the variables that affect urine albumin excretion (such as exercise, time of day, water loading, and posture).

SPECIMEN STORAGE PROCEDURE

- a. Urine specimens arrive frozen (on dry ice) or fresh (2-8 °C) as per study protocols. (For the NHANES IV, specimens are shipped on dry ice by overnight courier and stored at -70 °C upon arrival.) Frozen specimens are thawed for analysis at refrigeration temperature (2-8 °C) and remain at refrigeration temperature up to 3 weeks until analyses are complete.
- b. Analyzed specimens are returned to frozen storage (-20 °C or -70 °C) or discarded as per study protocols. Specimen stability at -20 °C for at least 1 year has been documented. (For the NHANES IV, specimens are returned to -70 °C freezer after analysis, then shipped on dry ice by overnight courier to the storage facility, McKesson Bioservices, Rockville, MD).

SPECIMEN HANDLING PROCEDURE

- a. Handle all urine specimens as if they are capable of transmitting any infectious agent.
- b. Return specimens to specified storage as soon as possible to avoid prolonged time at room temperature.

CRITERIA FOR SPECIMEN REJECTION

- a. Corrupted specimen integrity; cracked or leaking tube, unreadable or missing label.
- b. Visibly hematuric specimens.
- c. Note: Fluorescein dye used in retinal angiograms does not interfere with the assay.

5. PROCEDURES FOR MICROSCOPIC EXAMINATIONS; CRITERIA FOR REJECTION OF INADEQUATELY PREPARED SLIDES

Not applicable for this procedure.

6. EQUIPMENT AND INSTRUMENTATION, MATERIALS, REAGENT PREPARATION, CALIBRATORS (STANDARDS), AND CONTROLS

EQUIPMENT AND INSTRUMENTATION

1. Milli-Q deionized water system (Millipore Corporation, San Antonio, TX).
2. Fluorometer, Sequoia-Turner model 450, digital (cat. no. 450-000). Excitation energy is provided by a quartz halogen lamp (cat. no. 450-201). Wavelength selection is accomplished with a filter application set for fluorescein, SC 515 and NB 490 Sharp cut emission [cat no. 450-151] transmits >515 nm and narrow band excitation 490 nm peak, respectively). Adjust the filter at 450-151 for excitation = 485 nm, emission = 525 nm. (Sequoia-Turner Corp., Mountain View, CA).
3. Automatic diluter, Micromedic Systems model 25004, with 50-µL sampling and 200-µL dispensing pumps (Micromedic Systems, Division of Rohm and Haas, Horsham, PA).
4. Automatic dispenser, continuous with adjustable volume 0.5- to 20-mL, cat. no. 8885-047004 (Oxford, Division of Sherwood Medical, St. Louis, MO).
5. Water bath, model 185 (Precision Scientific, Inc., Chicago, IL).
6. Vortex-Genie, model K-550-G (Scientific Industries, Inc., Bohemia, NY).
7. Multi-tube Vortexer, model 2600 (Scientific Manufacturing Ind., Emeryville, CA).
8. Balance, top-loader, model H15, 1000g/0.1g (Mettler Instrument Corp., Hightstown, NJ).
9. Balance, top-loader, model XE-100A, 100g/0.1 mg (Fisher Scientific, Chicago, IL).
10. pH meter, Accumet model 15 (Fisher) and pH meter probe part no. 13-620-285 (Fisher).
11. Magnetic stir plate, Magnestir cat. no. 214-924 (Curtin Matheson Scientific Inc., Eden Prairie, MN) and stir bars (U-Stores, Univ. of Minn.).
12. Micro-centrifuge, model 235C (Fisher).

Microalbumin in Urine - NHANES 2001-2002

13. Centrifuge, IEC model UV, with swinging buckets to hold 12- X 75- mm tubes (International Equipment Co., Boston, MA).
14. Centrifuge, Beckman Avanti J-25, with JLA-16.250 fixed-angle rotor and six 250-mL centrifuge bottles, clear polycarbonate, part no. 356-013 (Beckman Instruments, Palo Alto, CA).
15. Autoclave, isothermal, Eagle series 2022 (AMSCO Scientific, Apex, NC).
16. Micro-pipette, 0.5- to 10- μ L, Oxford P-7000 series (Oxford) and Fisherbrand Oxford-style tips, cat no. 21-197-2F (Fisher).
17. Pipette, 20- μ L, Eppendorf (Brinkman Instruments Inc., Westbury, NY) and CLP-2000 universal tip, 200 μ L, (Continental Lab Products, San Diego, CA).
18. Pipetman, Gilson, P-100 and P-200 adjustable (Rainin Instrument Co., Inc., Woburn, MA) and CLP-2000 universal tip, 200 μ L (Continental Lab Products).
19. Pipetman, Gilson, P-1000 adjustable (Rainin) and tips 100- to 1000- μ L (Rainin).
20. Pipetman, Gilson, P-5000 adjustable (Rainin) and tips 500- to 5000- μ L (Rainin).
21. Repeater pipettor, Eppendorf, cat. no. P-3880-1 (Fisher) and Combitips sizes 500- μ l and 2.5-mL (Brinkman).
22. Calculator, model TI-30Xa Solar, and operator's manual (Texas Instruments, Dallas, TX).
23. Dell IBM PC compatible platform computer, Dell Dimension XPS, R400 Pentium II (Dell Computer Corporation, Round Rock, TX) and manufacturer's operating manuals (EPS Technologies, Jefferson, SD). Computer printer, HP DeskJet 722C (Hewlett-Packard Company, Palo Alto, CA).

MATERIALS

1. Type 1 water, deionized and distilled, Milli-Q (Millipore).
2. Type 2 water, distilled, 6 gallons (Kandiyohi Bottled Water Co., Willmar, MN).
3. Hydrochloric acid, HCl, F.W. 36.46, A.C.S. grade (EM Science, Gibbstown, NJ). Store in acid cabinet at 20-25 °C up to 2 years. HAZARD: Caustic and poisonous.
4. Potassium hydroxide, KOH, F.W. 56.11, A.C.S. grade (EM Science). Store at 20-25 °C up to 5 years. HAZARD: Caustic and poisonous.
5. Potassium phosphate monobasic, KH_2PO_4 , F.W. 136.09, A.C.S. grade (Mallinckrodt, Paris, KY). Store at 20-25 °C up to 5 years.
6. Sodium chloride, NaCl, F.W. 58.44, A.C.S. grade (Fisher). Store at 20-25 °C up to 5 years.
7. Sodium azide, NaN_3 , F.W. 65.02, Laboratory grade (Sigma, St. Louis, MO). Store at 20-25 °C up to 5 years. CAUTION: Toxic. Wear safety goggles, mask, and gloves when handling the dry powder.
8. Ovalbumin, chicken egg albumin, average F.W. 42,699, grade V, cat. no. A-5503 (Sigma). Store at 2-8 °C, with desiccant according to the package expiration date. CAUTION: Dust hazard. Wear mask.
9. Activated immunobead matrix ("beads"), 200-mg vial lyophilized polyacrylamide beads with phosphate buffer salts, cat. no. 15391 (Irvine Scientific, Santa Ana, CA). Store at 2-8 °C up to 5 years.
10. Goat anti-human albumin ("GAHA"), IgG fraction, 5-mL vial of 40-50 mg lyophilized antibody protein per vial, cat. no. 55028 (ICN Biomedical, Inc., Costa Mesa, CA). Store at -70 °C up to 5 years.

Microalbumin in Urine - NHANES 2001-2002

11. 1-ethyl-3-(3-dimethyl-aminopropyl) carbodiimide HCl ("EDAC"), M.W. 191.7, cat. no. E-6383, protein sequencing grade (Sigma). Allow for complete return to 20-25 °C before opening to avoid condensation. Store at -70 °C with desiccant up to 5 years. CAUTION: Toxic. Wear safety goggles, mask and gloves when handling the dry powder.
12. Fluorescein-conjugated goat anti-human albumin ("FI-GAHA"), IgG fraction, cat. no. 55162 (ICN). Store at -70 °C protected from light up to 5 years.
13. Dialysis tubing, Spectra-Por molecularporous membrane, reorder no. 132678, 2.0 mL/cm, 12-14,000 MWCO. (Spectrum Medical Industries, Los Angeles, CA). Store at 2-8 °C.
14. Urine specimen tubes: 3.5-mL cryogenic storage vials, cat. no. 223-9835, for most urine specimen protocols and for storage of aliquots of FI-GAHA (Bio-Rad Laboratories, Richmond, CA) (For the NHANES IV: 4.5-mL cryogenic storage vials, cat. no. MS-4505, (Vanguard International, Inc., Neptune, NJ).)
15. Micro-centrifuge tube, 0.65-mL, siliconized, colors, cat. no. C3259 (Intermountain Scientific, Bountiful, UT).
16. Micro-centrifuge tube, 1.5-mL, cat. no. 72-690 (Sarstedt, Inc., Princeton, NJ).
17. Glass test tubes, borosilicate, sizes 12- X 75-mm and 16- X 100-mm, (Kimble Co, Toledo, OH).
18. Micro-filters, external filter tips, 45-µm polyethylene, cat. no. OFT 2002-45M (Labsciences, Reno, NV).
19. Filterware, Nalgene, 0.2-µm, sterile, 115-mL, disposable (Nalge Company, Rochester, NY).
20. Albutix test strips, cat. no. 2191 (Bayer Corporation, Diagnostics Division, Elkhart, IN). Follow the package insert for explicit directions for use. Store at 20-25 °C according to package instructions and expiration date.
21. General laboratory supplies: gloves, lab coats, safety glasses, 5% bleach, 99% isopropyl alcohol, autoclave bags, disposable absorbent bench-top toweling, transfer pipettes, weighing boats, parafilm, pH meter solutions, four 100-mL wide-mouth Nalgene bottles for immunobead storage, eight 4-L wide-mouth Nalgene bottles for solution storage, graduated cylinders, two 500-mL beakers, one 50-mL volumetric flask (U-Stores).
22. Human serum albumin ("HSA"), cat. no. A-8763 (Sigma), approximately 99% pure by agarose electrophoresis, essentially globulin-free, derived from Sigma cat. no. A-1653 Cohn Fraction V. Store with desiccant at 2-8 °C according to the product expiration date. CAUTION: Handle as if capable of transmitting infectious agents.
23. CAP Urine Protein Reference Material, human albumin Cohn Fraction V (College of American Pathologists, Skokie, IL). Store according to product directions. CAUTION: Handle as if capable of transmitting infectious agents.

REAGENT PREPARATION

1. Sterile distilled water
Filter the Type 1 or the Type 2 (depending on procedure) distilled water through sterile, 115-mL, 0.2-µm, disposable Nalgene Filterware under vacuum. Discard unused water.
2. 1 mol/L HCl (pH adjusting solution)
Dilute 8.3 mL in 100 mL Type 2 water and mix well. Store at 20-25 °C up to 1 year.
CAUTION: Prepare in safety hood. Wear safety glasses and gloves. Add acid to water.
3. 5 mol/L KOH (pH adjusting solution)
Dissolve 70.1 g KOH in 250 mL Type 2 water and mix well. Store at 20-25 °C up to 1 year.
CAUTION: Prepare in safety hood.

Microalbumin in Urine - NHANES 2001-2002

4. 0.3 mol/L KH₂PO₄ stock solution
Dissolve 40.8 g of KH₂PO₄ in 1 L of Type 2 water and mix well. Store at 20-25 °C up to 1 year.
5. 0.003 mol/L KH₂PO₄ (dialysis buffer for immunobead coupling)
Dilute 20 mL of 0.3 mol/L KH₂PO₄ to 2 L with Type 2 water and mix well. Adjust the pH to 6.3 with 5 mol/L KOH. Prepare the day before use and store at 2-8 °C overnight. Use cold. Discard after use.
6. Phosphate buffered saline stock solution concentrated (20X PBS)
Dissolve 175.4 g NaCl and 2 g NaN₃ in 666 mL of 0.3 mol/L KH₂PO₄. Dilute to 1 L with Type 2 water and mix well. Store at 20-25 °C up to 1 year.

[WARNING: NaN₃ is toxic and may react explosively with lead and copper plumbing.]
7. Physiological phosphate buffered saline (1X PBS)
Dilute 200 mL of 20X PBS to 4 L with Type 2 water and mix well. Adjust pH to 7.2 with 5 mol/L KOH. The final concentration is 0.01 mol/L KH₂PO₄, 0.15 mol/L NaCl, 0.015 mmol/L NaN₃. Prepare 4L for assay performance and prepare 12 L for the immunobead coupling procedure. Use cold. Store at 2-8 °C up to 3 weeks.
8. 1.4 mol/L NaCl in PBS (NaCl-PBS)
Dissolve 245.5 g NaCl in 3 L PBS for immunobead coupling preparation. Mix well. Prepare the day before use and store at 2-8 °C overnight. Use cold and discard unused solution.
9. Ovalbumin in physiological phosphate buffered saline (oval/PBS)
Dissolve 0.5 g ovalbumin in 100 mL PBS and filter through 0.2-µm sterile Nalgene Filterware. Store in this filter unit at 2-8 °C up to 3 weeks.

10. Immunobeads: beads coupled to goat anti-human albumin, a two-day procedure

Day 1:

Prepare 400 mL. Do not prepare more than a 3-month supply. For an estimated 4,000 specimens per year prepare four grams (400 mL) of immunobeads on a quarterly basis. The final concentration of the immunobeads is 10 mg/mL in oval/PBS with 0.15 mmol/L NaN₃. (This procedure is adapted from the Bio-Rad package insert⁶.)

Reconstitute each of four vials of GAHA with 5 mL of sterile Type 2 water and incubate at 2-8 °C for 1 hour. Combine the vials into approximately 25 cm of dialysis tubing and use an additional 2 mL of sterile Type 2 water to rinse the vials. Dialyze against 2 L of 0.003 mol/L KH₂PO₄, pH 6.3, for 4 hours at 2-8 °C.

Reconstitute 4 grams (twenty 200-mg vials) of lyophilized beads in a 500-mL beaker with 400 mL sterile Type 2 water. Adjust the pH to 6.3 with 1 mol/L HCl as needed. Cover and incubate at 2-8 °C for 1 hour.

Add dialyzed GAHA to the beads, adjust the pH to 6.3 with 1 mol/L HCl or 5 mol/L KOH as needed and incubate at 2-8 °C for 1 hour. The final concentration is 40- to 50-mg of antibody protein per 1 gram of beads.

Perform the bead coupling procedure: Set the pH meter probe in the solution, stirring. Have ready the pH solutions 1 mol/L HCl and 5 mol/L KOH. Quickly add 800 mg of EDAC while stirring. The pH will oscillate rapidly. Use the pH solutions as needed to maintain the pH at 6.3 for 30 min during this equilibration process. Initially, 1 mol/L HCl will be required to maintain pH 6.3. After 30 minutes the pH will stabilize.

Divide the immunobeads equally into each of six 250-mL centrifuge bottles. Incubate at 2-8 °C overnight.

Microalbumin in Urine - NHANES 2001-2002

Prepare the Beckman Avanti J-25 centrifuge. Insert the JLA-16.250 rotor, set the temperature at 4 °C, and allow the centrifuge to cool overnight. Prepare the immunobead washing buffers (12 L of PBS and 3 L of 1.4 mol/L NaCl-PBS) and store at 2-8 °C overnight. Continue the procedure the following day.

Day 2:

Perform six washing procedures with cold buffers and centrifuge set at 4 °C. For each wash, fill the bottle with buffer, centrifuge at 2900 rpm (2000 x g) for 8 minutes with the brake set on high, decant, blot, and resuspend the immunobeads with a small volume of the next wash solution with a gentle rolling motion. Perform the wash procedure 2 times with PBS and then wash 2 times with 1.4 mol/L NaCl-PBS. Wash 2 times more with PBS. Resuspend the contents of each bottle with 100 mL of PBS and incubate at 2-8 °C for 3 hours. Fill the bottle with PBS and do a final centrifugation. Decant.

Resuspend the contents of each bottle with a small volume of oval/PBS. Combine and adjust the total volume to 400 mL in a 500-mL beaker. Add 0.4 g of NaN₃. Divide into four 100-mL portions in the 100 mL Nalgene bottles. Store at 2-8 °C for 3 months, never allow beads to freeze.

11. FI-GAHA stock solution

Reconstitute each vial of lyophilized product with 2 mL of sterile Type 2 water. Combine all vials together and place them at 2-8 °C, protected from light, for 1 hour. Aliquot 0.4 mL into 3.5-mL cryogenic storage vials (Bio-Rad). Use the repeater pipette 2 times set on #4 with the 2.5-mL Combipip to deliver 0.4 mL. Store the vials at -20 °C protected from light up to 2 years.

12. FI-GAHA working solution

- Prepare a 1:10 dilution. Add 3.6 mL of oval/PBS to the cryogenic vial containing the 0.4-mL aliquot of FI-GAHA stock solution (or other proportions as determined by the limiting curve data, see below). Before use on each assay day, use the IEC centrifuge at 2100 rpm (1000 x g) for 5 min to remove cryo-precipitates. Transfer the supernatant to a 16- X 100-mm tube and leave the supernatant in a light-protected container at room temperature until use. Return the excess to a light-protected container. Store at 2-8 °C up to 1 month.
- Test new lots of FI-GAHA to determine the optimum concentration to use when the fluorescein/protein ratio (F/P) varies. Plot a limiting curve: assay the "20 µg/mL HSA" from the standard line preparation against serial dilutions of FI-GAHA and graph the data. Select a dilution that is 2 times the concentration at the point (or "limit") where the graph detours from a straight line and begins to plateau.

CALIBRATORS (STANDARDS)

Standard line calibration material: human serum albumin (HSA)

Prepare frozen aliquots of a 1mg/mL solution of HSA in order to thaw and use one for every assay. Weigh 50 mg of lyophilized HSA on a balance of 0.1 mg sensitivity and dissolve in 50 mL of sterile Type 1 (Milli-Q) water using a 50-mL volumetric flask. Pipette 0.4 mL into 0.65-mL micro-centrifuge tubes. Use the repeater pipette 2 times set on #4 with the 2.5-mL Combipip to deliver 0.4 mL. Label each aliquot "HSA." Attempt to prepare adequate numbers of aliquots in order that one set of materials could last the duration of a study. When these materials change, perform a comparability study of at least 100 specimens to demonstrate the relationship of the values before and after the change. Store at -70 °C. CAUTION: Handle as if capable of transmitting infectious agents.

Primary standard reference material

The primary calibration standard for the fluorescent immunoassay is "CAP Urine Protein Reference Material, human albumin Cohn Fraction V". Reconstitute the material as described in the package insert and devise a dilution schematic that correlates with the concentrations of the standard line calibration material for comparison.

CONTROLS

Quality control material: control pool urine, three levels

Prepare frozen aliquots of three control pools in order to thaw and use one of each for every assay. The three levels of control represent low, middle, and high positions on the standard line. Pool and screen urine from a number of healthy volunteers to obtain albumin concentrations of approximately 1, 7, and 15 µg/mL. Filter each urine pool through sterile 0.2-µm Nalgene Filterware under vacuum. Mix well. Pipette 0.4 mL into 0.65-mL micro-centrifuge tubes using a different tube color for each concentration level. Use the repeater pipette 2 times set on #4 with the 2.5-mL Combipip to deliver 0.4 mL. Label each aliquot. Attempt to prepare

Microalbumin in Urine - NHANES 2001-2002

adequate numbers of aliquots in order that one set of materials could last for the duration of a study. When these materials change, perform a comparability study of at least 100 specimens to demonstrate the relationship of the values before and after the change in the controls. Store at -70 °C. CAUTION: Handle as if capable of transmitting infectious agents.

7. CALIBRATION AND CALIBRATION VERIFICATION PROCEDURES

CALIBRATION PROCEDURE

- a. For each assay prepare a calibration curve from the results of the dilutions of standard line material, 0 µg/mL to 20 µg/mL HSA, as described in Table 1.
- b. For each assay calibrate the fluorometer instrument. Adjust the fluorometer to read 0% relative fluorescence (RF) and 100% RF for the HSA standard line calibration limits 0 µg/mL and 20 µg/mL, respectively. This will establish a known relationship between the instrument measurement and the true value of the albumin analyte.

CALIBRATION VERIFICATION PROCEDURE

- a. Verify that the instrument remains in calibration during the fluorometry procedure by periodically re-checking that the 0 µg/mL and the 20 µg/mL HSA samples remain in adjustment at 0% RF and 100% RF, respectively.
- b. Verify that the standard line calibration curve is linear with a y-intercept near zero. For each assay, the calibration curve correlation coefficient (R^2) must be > 0.9900.
- c. Verify that at least two of the three quality control urine specimens are within the expected 95% limits as described by the Westgard Rules⁷. Check the other rules regarding "out of control" criteria as listed in the QC section.
- d. Verify the concentration of the "1 mg/mL HSA" standard line calibration material by checking it against the primary standard, "CAP Urine Protein Reference Material, human albumin Cohn Fraction V", a reference material of known value.

8. PROCEDURE OPERATING INSTRUCTIONS; CALCULATIONS; INTERPRETATION OF RESULTS

PROCEDURE OPERATING INSTRUCTIONS

a. Preliminaries prior to the assay day

1. Attach the micro-filters to the pipette tips.
2. Number single 1.5-mL micro-centrifuge tubes to correspond with each urine specimen.
3. Number duplicate 12- X 75-mm tubes (use a waterproof marker) for each of the three quality control urine samples, the standard line samples (labeled from 0 to 20 µg/mL), and for each urine specimen. Perform in triplicate any specimen that requires immediate reporting.
4. Prepare a worksheet to compile the information corresponding with the information on the Excel spreadsheet described in section 3. On the worksheet place one of the 3 quality control urine samples in the 0 position before specimen number 1, one near the middle, and the third after the specimens. Rotate these positions in subsequent assays.
5. Adjust the water bath for optimum volume and maintain at 37 °C temperature.

b. Sample preparation on the assay day

1. Standard line calibration material preparation for the assay day

Thaw and vortex one aliquot of the "1 mg/mL HSA" on the day of the assay. Transfer to a 12- X 75-mm tube. Filter and transfer to a second tube with the P-200 pipette and pipette tip with the attached micro-filter. Follow the schematics of Table 1 to prepare the dilutions of HSA in oval/PBS buffer. Use the Oxford micro-pipette and 12- X 75-mm tubes. Vortex. Assay with the same conditions as the unknown specimens. The dilution factor is 1. Use the Micromedic sampling syringe set at 40%.

Table 1
Preparation of Standards

| HSA Standard (µg/mL) (mL) | Volume of HSA (µL) | Volume of oval/PBS |
|------------------------------|--------------------|--------------------|
| 20 | 20 | 1 |
| 15 | 15 | 1 |
| 10 | 10 | 1 |
| 5 | 5 | 1 |
| 2 | 4 | 2 |
| 1 | 2 | 2 |

Microalbumin in Urine - NHANES 2001-2002

| | | |
|----|---|---|
| .5 | 1 | 2 |
| 0 | 0 | 1 |

2. Control pool urine preparation for the assay day

Thaw and vortex one aliquot of each of the three control pools on the day of an assay. Transfer each to a 12- X 75-mm tube. Filter and transfer each to a 1.5-mL micro-centrifuge tube with the P-200 pipette and pipette tip with the attached micro-filter. Centrifuge the tubes for 2 min in the micro-centrifuge. Assay with the same conditions as the unknown specimens. The dilution factor is 1. Use the Micromedic sampling syringe set at 40%.

3. Urine specimen preparation for the assay day

- a. Thaw and mix the urine specimens by inverting the tubes a minimum of eight times. Number the specimen tubes to correspond with the specimen identification numbers and the worksheet. Filter and transfer 200 µL of urine specimen to the corresponding 1.5-mL micro-centrifuge tube with the P-200 pipette and pipette tip with the attached micro-filter. Centrifuge the tubes for 2 min in the micro-centrifuge. Assay with the same conditions as the standard line material and controls. The dilution factor and the Micromedic sampling syringe setting is determined by the Albustix value as described below.
- b. Measure the Albustix value of each specimen as directed on the package insert. Knowledge of the approximate concentration of the specimen enables the technician to prepare the specimen so that it will fall within the standard curve, 0.5 µg/mL to 20 µg/mL. Record Albustix values and the corresponding dilution factors on the worksheet as described in Table 2. The Albustix value determines the dilution factor. Assay each specimen at the specified dilution and corresponding Micromedic sampling syringe setting as determined by the Albustix value:

Urine specimens with very low albustix values

Specimens with “negative” or “non-detect” (negative and colorless) Albustix values may require analysis at 2X the concentration (i.e. dilution factor of “2”). This is adjusted with the Micromedic sampling syringe set at 80%.

Note: Specimens that fail to fall on the standard curve at this setting are considered < LOD.

Urine specimens with low to moderate albustix values

Specimens with Albustix values of “negative”, or “negative-trace” have dilution factors “1”, and “.5”, respectively. They are adjusted directly on the Micromedic. The Micromedic sampling syringe is set at 40% and 20%, respectively.

Urine specimens with high albustix values

Specimens with Albustix values of “trace”, “+”, “++”, or “+++” must be diluted. Dilute with oval/PBS as directed in Table 2. The dilution factors will be “.1”, “.02”, “.01”, “.001”, respectively. The Micromedic sampling syringe is set at 40%.

Table 2
Preparation of Specimen Dilutions

| Albustix value | Dilution factor | Hand Dilution | | Micromedic Automatic Diluter | | |
|----------------|-----------------|--------------------|----------------------|---|--------------------|---|
| | | Sample volume (µL) | oval/PBS buffer (µL) | Micromedic sampling syringe setting (%) | Sample volume (µL) | Final volume with addition of 75µL immunobeads (µL) |
| Non-detect | 2 | – | – | 80 | 40 | 115 |
| Negative | 1 | – | – | 40 | 20 | 95 |
| Neg/Trace | .5 | – | – | 20 | 10 | 85 |
| Trace | .1 | 50 | 450 | 40 | 20 | 95 |
| + | .02 | 20 | 980 | 40 | 20 | 95 |
| ++ | .01 | 10 | 990 | 40 | 20 | 95 |
| +++ | .001 | 2 | 2000 | 40 | 20 | 95 |

c. Instrument Setup for the assay day

Waterbath

Preset and maintain at 37 °C. Adjust the volume with tap deionized water. Check and record the temperature on the assay day prior to the incubation. Verify temperature at the end of the assay incubation.

IEC Centrifuge

Set at 2100 rpm (1000 x g) and 10 min for two washes. Use a swinging bucket with carriers that fit 12- X 75-mm tubes. Perform centrifugation at 20-25 °C with the brake set at maximum.

Micromedic

Microalbumin in Urine - NHANES 2001-2002

1. Immerse the tubing end into constantly stirring immunobeads and prime the Micromedic by washing through the bead solution with the dials set for maximum volume dispensing, just enough to fill the lines without the presence of bubbles. Prevent the beads from settling in the tubing while the instrument is not in operation.
2. Set the instrument for automatic sampling and dispensing. The number dialed on each syringe pump equals the percentage of the syringe volume. The maximum volume of the left (sampling) syringe is 50 μ L; the maximum volume of the right (dispensing) syringe is 200 μ L.

To Prime -- push both ("set" and "run") buttons down.

To Set -- push the "set" button down and the "run" button up to dial in percentages. Set the left (sampling) syringe at 20%, 40% or 80% according to the dilution factor schematic in Table 2. Keep the right (dispensing) syringe at 37.5% to deliver 75 μ L of immunobeads to all samples.

To Run -- push the "run" button down and "set" button up. Insert the delivery tip into the specimen. Press the foot pedal to siphon the sample. Avoid disturbing the pellet and remove the drop adhering to the delivery tip. Press the foot pedal to eject the sample combined with the beads into the 12- X 75-mm tubes. The final volume of sample combined with 75 μ L of immunobeads is listed in Table 2.

3. Clean the Micromedic when finished by running several cycles of isopropyl alcohol followed by tap deionized water (approximately 5 mL of each). Store the Micromedic with water-filled tubing.

Oxford continuous automatic dispenser

Prime with PBS to fill the lines. Adjust settings for dispensing the 4-mL volume or the 2-mL volume as required in the procedure. When finished, flush the PBS out and leave the lines dry.

Fluorometer

Turn the instrument on and allow it to warm up for a minimum of 15 minutes prior to use. Set the GAIN dial at "5". The raw data is read as percent relative fluorescence (% RF). Calibrate the fluorometer with the ZERO and SCAN dials as described following. The 12- X 75-mm tubes function as disposable cuvettes.

d. Operation of the assay procedure

1. Place the bottle of immunobeads in a beaker of crushed ice. Stir gently on a magnetic stir plate.
2. Use the Micromedic automatic diluter to combine the samples with immunobeads as directed in Table 2 and the instrument set-up. Dispense into duplicate 12- x 75-mm tubes according to the dilutions as listed on the worksheet for the specimens, standard line tubes, and the three quality control urines.
3. Mix by gently shaking the rack of tubes X3, cover, and incubate for 1.5 hours in the 37 °C water bath.
4. Prepare sufficient quantity of the FI-GAHA working solution. Add 50- μ L of FI-GAHA working solution to each tube. Use the repeater pipette set on #5 with the 500- μ L Combitip. The pipette will deliver eight 50- μ L aliquots for each fill.
5. Mix by gently shaking the rack of tubes X3, cover, and incubate for 2 hours in the 37 °C water bath.
6. Perform two washing procedures to remove the unbound FI-GAHA from the immunobead-antigen complex. Use the automatic dispenser to add cold PBS and the IEC centrifuge as directed. For each wash, add 4 mL of PBS, centrifuge, decant the supernatant and blot, and vortex to resuspend the pellet in the residual buffer. Add final volume of 2 mL PBS after the washing procedure.
7. Perform fluorometry:
 - (a) Vortex and wipe the outside of each tube until clean and dry. Insert each tube into the fluorometer and read and record the displayed % RF immediately while the beads are in suspension. Read the samples as soon as possible because of possible deterioration of the fluorescein photosensitivity.
 - (b) Calibrate the 0 μ g/mL (blank) tube to 0% RF: insert the 0 μ g/mL tube and set the ZERO dial to "0". Calibrate the 20 μ g/mL tube to 100% RF: insert the 20 μ g/mL tube and set the SCAN dial to "100".
 - (c) Record the % RF values onto the worksheet for the remaining standard line tubes, controls, and the unknown specimens.
 - (d) Substantiate the continued accuracy of the test method by verification that the fluorometer dials remain in adjustment: periodically re-check that the 0 μ g/mL and the 20 μ g/mL samples are at 0% RF and 100% RF, respectively.
 - (e) Turn off fluorometer, decant samples into sink, discard tubes into glass waste.
8. Perform Micromedic daily maintenance and clean the laboratory work area according to safety guidelines.

e. Replacement and periodic maintenance of key components

1. Fluorometer

Clean the filter windows with glass cleaner and lens paper periodically. Replace the lamp as needed.

2. Micromedic and Pipettors

Microalbumin in Urine - NHANES 2001-2002

Calibrate and clean all pipettors every 6 months. Replace parts as needed. (Professional service is used.)

3. Balances and Centrifuges

Calibrate and perform parts and maintenance service every 6 months. (Professional service is used.)

4. Computer

Verify the linear regression calculations on the Excel software assay template annually.

CALCULATIONS AND INTERPRETATION OF RESULTS

1. Enter the % RF values of the standard line, controls, and specimens onto the formatted Excel spreadsheet.
2. Apply the equation for a straight line, $y = mx + b$ and perform linear regression analysis on the standard line results to obtain slope (m), y-intercept (b) and R^2 and fitted line. The % RF value is directly proportional to the $\mu\text{g/mL}$ of the standard line.
3. Convert the % RF values of the controls and specimens into $\mu\text{g/mL}$ by performing a "fill-down" of the regression analysis equations. The dilution factor is multiplied by the converted $\mu\text{g/mL}$ to obtain the total $\mu\text{g/mL}$ albumin concentration.
4. Print the Excel document and verify that every value (i. e., specimen identification number, dilution factor, and % RF) matches those on the worksheet.
5. Record the values for the three quality control urine samples on the Levy-Jennings charts to determine satisfactory analytic performance.

9. REPORTABLE RANGE OF RESULTS

- a. The range of test values for the fluorometer instrument is the 0 to 100% RF which corresponds with the range of linearity for the standard line, 0 to 20 $\mu\text{g/mL}$ HSA.
- b. Specimens with fluorometry values that are below the % RF of the standard line value 0.5 $\mu\text{g/mL}$ are not accepted. Specimens with unacceptable low % RF values at dilution factor "1" are re-analyzed at 2X the concentration (i.e. the dilution factor "2" setting on the Micromedic) in a subsequent assay. The lowest calculated concentration with dilution factor "2" is 0.25 $\mu\text{g/mL}$ which "rounds" to 0.3 $\mu\text{g/mL}$. Specimens with unacceptable values at dilution factor "2" are designated < LOD.
- c. Specimens with fluorometry values that are above the % RF of the standard line value 20 $\mu\text{g/mL}$ are not accepted and are re-analyzed after dilution in a subsequent assay.
- d. The lower limit of detection for the assay is 0.3 $\mu\text{g/mL}$; there is no upper limit of detection. Therefore, the "reportable range of results" is "0.3 $\mu\text{g/mL}$ and up" or " $\geq 0.3 \mu\text{g/mL}$ ".

10. QUALITY CONTROL (QC) PROCEDURES

a. Quality Control System

Internal QC procedures monitor analytical performance relative to medical goals and alert analysts to unsatisfactory analytical performance. Estimates of imprecision and permanent confidence limits are generated from 100-pair data (100 assay days) of three control urine pool levels. Calculated from the 100-pair data are the % CV (percent correlation of variance), the SDo (standard deviation overall), and the duplicate range (average of the differences between duplicates, average R). Tolerance limits for controls are established by using the Westgard Rules⁷ as guidelines. The tolerance limits are defined as the mean ± 2 SDo, or 95% warning limit, and the mean ± 3 SDo, or 99% action limit. When new control urine pools are initiated into the assay, the new mean is established from 20-pair data (20 assay days) and analyzed with the older controls during this time period. Calculations from these 20-pair (% CV, SDo, and duplicate range) are used only to monitor the 100-pair permanent confidence limits. Low, medium, and high range bench quality control urine pool samples are analyzed in duplicate in each analytical run (a set of assays performed on a given day) so that judgements may be made on the day of analysis. The three control urine samples are the same in all assays.

Prepare a Levy-Jennings Means Chart for each of the three urine pool levels. Chart the means and tolerance limits on the graph paper. Plot the duplicate values for each control as a pair of dots for each assay. On this chart highlight unacceptable control values and action taken, document reagent changes and other pertinent information. Prepare a cumulative table of duplicate values and the difference between duplicates for the three control urine samples. Highlight unacceptable duplicates according to the within batch duplicate range (average R) tolerance limits and make notation on the Levy-Jennings charts. If the assay is determined to be out of range, perform function checks on the reagents and instrumentation.

Provide assay results only after problems are corrected. The Levy-Jennings charts are reviewed and signed monthly by the medical director and quarterly by the clinical chemistry laboratory supervisor. (For NHANES IV, copies are mailed for review quarterly.)

The system is declared "out-of-control" if any of the following events occur on the Means Chart:

- A single run mean for one or more pools falls outside the upper or lower 99% limit.
- The run means for two or more pools fall either both above or both below the 95% limit.
- Two successive run means for a single pool fall either both above or both below the 95% limit.
- Eight successive run means for a single pool fall either all above or all below the centerline, establishing a trend.

The system is declared "out-of-control" if any of the following events occur on the Range Table:

- A single within-run range falls above the upper 99% limit.
- The within-run ranges for two or more pools fall above the upper 95% limit.
- Two successive within-run ranges for a single pool fall above the upper 95% limit.
- Eight successive within-run ranges for a single pool fall above the center line.

b. Quality Assurance System

A "Continuous Quality Improvement Plan and Documentation" program is in effect for the laboratory. General QA systems are in operation to detect errors, monitor tolerance limits on temperatures, ensure proper reagent labeling and equipment maintenance. The laboratory ensures that samples are collected, handled, shipped, preserved and stored correctly, and rejected if the criteria are not met.

c. External Proficiency Testing

Assay performance evaluation is conducted through the College of American Pathologists, CAP. For most study protocols a custom-designed external proficiency test is included to evaluate the assay performance. (For the NHANES IV, the inter-assay variation is monitored by repeating 2% of the specimens from the previous assay.)

d. Accreditation

The laboratory is certified by the Clinical Laboratory Improvement Act, CLIA.

11. REMEDIAL ACTION IF CALIBRATION OR QC SYSTEMS FAIL TO MEET ACCEPTABLE CRITERIA

- a. Retest specimens that are out of the range of the acceptable test values for the fluorometer instrument, the % RF of 0.5 µg/mL to 20 µg/mL HSA, at the appropriate dilution.
- b. Retest specimens with discordant duplicate % RF values. The acceptable duplicate ranges are as follows: 4 units for 0 – 20% RF, 5 units for 21 – 50% RF, 8 units for 51 – 100% RF.
- c. Retest the entire assay if the R² value is < 0.9900 or if the quality control urine samples are outside of the limits as described by the Westgard Rules⁷.
- d. Replace reagents or recalibrate instrumentation as necessary for corrective action. Introduce reagents with new lot numbers into the assay one at a time to facilitate identifying problems.
- e. Report results on affected specimens only after corrective action and documentation of statistical control.

12. LIMITATIONS OF METHOD; INTERFERING SUBSTANCES AND CONDITIONS

LIMITATIONS OF METHOD

The lower limit of detection for the assay is 0.3 µg/mL and is also the lowest reportable value. Specimens that are below 0.3 µg/mL are designated "< LOD". There is no upper limit of detection for the assay because specimens can be diluted for analysis.

INTERFERING SUBSTANCES

Female participants who are menstruating may have contaminated urine specimens resulting in falsely elevated albumin levels. Unusually high albumin values, > 1000 µg/mL, may reflect hematuria and are routinely tested by dipstick for hemoglobin. There is no interference from the fluorescein excreted into the urine following retinal angiography.

INTERFERING CONDITIONS

The interpretation of urine albumin measurements may be confounded by conditions such as urinary tract infection, marked hyperglycemia, severe hypertension, or congestive heart failure.

13. REFERENCE RANGES (NORMAL VALUES)

Table 3
Urinary Albumin in Healthy Subjects

| Number of Subjects | Author | Method | µg/mL | µg/min | mg/24 h |
|--------------------|------------------------|--------|-----------|----------|----------|
| 100 adults | Chavers ⁽²⁾ | FIA | 0.2-17.3* | 0.8-14.6 | 1.1-21.0 |

Microalbumin in Urine - NHANES 2001-2002

| | | | | | |
|-------------|-------------------------|-------|-------|------------|----------|
| 20 adults | Mogenson ⁽³⁾ | RIA | | 5-20 | 3.6-23 |
| 11 adults | Howey ⁽⁴⁾ | RIA | 8.13+ | 7.1**,+ | 10.2+ |
| 20 children | Fielding ⁽⁵⁾ | ELISA | | 1.2-15.9** | 1.7-22.9 |

* obtained from subject data in the published manuscript

** calculated from published units

+ mean values

14. CRITICAL CALL RESULTS (" PANIC VALUES")

There are no "panic values" which indicate immediate medical intervention for unusual urine albumin values. Diabetic nephropathy can be defined by increased urine albumin values when measured by urine albumin excretion rate or urine albumin/creatinine ratio.

15. SPECIMEN STORAGE AND HANDLING DURING TESTING

Specimens are stored at 2-8 °C during preparation of testing. Specimens may reach room temperature during the following procedures: pipeting, micro-centrifugation, and during the Micromedic procedure. A critical temperature during the testing procedure is the 37 °C incubation. When incubation ends, cold buffer is introduced to quench the reaction and to begin the centrifugation/washing procedure. During fluorometry, the samples are relatively cool and approach room temperature.

16. ALTERNATIVE METHODS FOR PERFORMING TEST OR STORING SPECIMENS IF TEST SYSTEM FAILS

- Manual pipettes can be substituted for the automatic dispensers (Micromedic and Oxford). If the fluorometer fails, no alternative instrumentation is available until it is returned to service.
- Return specimens to -20 °C or -70 °C if the test system is out of operation for > 3 weeks.

17. TEST RESULT REPORTING SYSTEM; PROTOCOL FOR REPORTING CRITICAL CALLS (IF APPLICABLE)

TEST RESULT REPORTING SYSTEM

- Test results are reported by hard copy or by electronic transmission as designated by individual study protocols.
- The albumin test result is reported as µg/mL with one decimal point.
- Other test result reporting systems for albumin are designated by individual study protocols. Examples: albumin excretion rate is calculated from timed (overnight or 24-hour) urine collections and reported as µg/min or mg/24 h and albumin-to-creatinine ratio is calculated from random urine spot collections and is reported as µg/mg or mg/mmol.
- Unusual albumin values that are designated "< LOD" or are > 1000 µg/mL are flagged for attention.

PROTOCOL FOR REPORTING CRITICAL CALLS

Not applicable for this procedure.

18. TRANSFER OR REFERRAL OF SPECIMENS; PROCEDURES FOR SPECIMEN ACCOUNTABILITY AND TRACKING

Specimens are logged into data books by the laboratory upon receipt. They are received by the laboratory in batches or as single specimens, frozen or fresh, arriving by overnight courier or hand-carried by study personnel from co-ordinating centers. (For the NHANES IV batches of frozen specimens are received by Fed-Ex. Completed specimens are re-frozen in the original tubes at -70°C and returned on dry ice to the contract storage facility McKesson Bioservices.) All communications, notebooks, discs, and files containing raw data, final data, QC information, etc. are saved in the laboratory. These serve as documentation for specimen accountability and tracking.


Microalbumin in Urine - NHANES 2001-2002

19. Summary Statistics and QC Graphs

Summary Statistics for Urinary Albumin by Lot

| Lot | N | Start Date | End Date | Mean | Deviation Standard | of Variation Coefficient |
|-----|-----|------------|------------|--------|--------------------|--------------------------|
| f | 186 | 1/3/2001 | 12/31/2002 | 1.546 | 0.221 | 14.29 |
| k | 143 | 1/3/2001 | 6/7/2002 | 6.598 | 0.427 | 6.47 |
| v | 143 | 1/3/2001 | 6/7/2002 | 13.784 | 1.192 | 8.65 |
| n | 43 | 6/12/2002 | 12/31/2002 | 6.200 | 0.345 | 5.57 |
| t | 43 | 6/12/2002 | 12/31/2002 | 14.902 | 0.898 | 6.03 |

2001-2002 Urinary Albumin Quality Control



REFERENCES

1. Chavers BM, Simonson J, Michael AF. A solid-phase fluorescent immunoassay for the measurement of human urinary albumin. *Kidney Int.* 1984; 25: 576-8.
2. Chavers BM, Mauer SM, Ramsay RC, Steffes MW. Relationship between retinal and glomerular lesions in IDDM patients. *Diabetes* 1994; 43: 441-6.
3. Mogenson CE. Renal function changes in diabetes. *Diabetes* 1976; 25: 872-9.
4. Howey JEA, Browning MCK, Fraser CG. Selecting the optimum specimen for assessing slight albuminuria, and a strategy for clinical investigation: novel uses of data on biological variation. *Clin Chem* 1987; 33: 2032-8.

Microalbumin in Urine - NHANES 2001-2002

5. Fielding BA, Price DA, Houlton CA. Enzyme immunoassay for urinary albumin. Clin Chem 1983; 29: 355-7.
6. Instructions for immunobead reagent coupling kits. Richmond (CA): Bio-Rad Laboratories, 1979: 2-4.
7. Westgard JO, Carey RN, Wold S. Criteria for judging precision and accuracy in method development and evaluation. Clin Chem 1974; 20: 825-33.