

Analysis of the September 1995 Performance Evaluation Testing Results for T-Lymphocyte Immunophenotyping Reported to the Centers for Disease Control and Prevention by Participating Laboratories

This report is an analysis of results furnished to the Centers for Disease Control and Prevention (CDC) by laboratories participating in the Model Performance Evaluation Program (MPEP) after they tested the T-lymphocyte immunophenotyping (TLI) performance evaluation specimens sent them in September 1995. Of those laboratories receiving specimen panels, 314 of 342 (91.8%) reported testing results. Additionally, two laboratories returned information regarding the procedures used in their laboratory for performing TLI, but were unable to provide testing results for the specimens they received.

Each laboratory received a total of six specimens: three HIV-1 antibody-positive and two HIV-1 antibody-negative whole blood specimens, and an instrument performance control sample consisting of fluorescencated beads. One of the HIV-1 antibody-positive whole blood specimens was sent to the participant laboratories in duplicate. Not all laboratories received the same panel of specimens. The first three pages immediately following the title page contain the specimen numbers and donor information for each performance evaluation specimen. No laboratories received specimens derived from two previously designated donors (donors 37 and 41) due to the inability of obtaining whole blood from these donors on the day of shipment.

The result form used for the September 1995 specimen shipment was designed to be consistent with the CDC guidelines for CD4⁺ cell testing (MMWR, vol. 43, no. RR-3, March 4, 1994). Laboratories have been encouraged by the MPEP to utilize these guidelines in performing TLI on specimens from HIV-infected patients.

For the first time, due to revisions in the MPEP result reporting booklet, laboratories were able to report results using 3-color analyses of the specimens. As can be seen in the graph on the following page, the ability to report 3-color data resulted in a decreased number of participant laboratories, 170 (54.1%) of 314, that used the 2-color monoclonal antibody panel recommended by the CDC for CD4⁺ cell testing. Twenty-eight laboratories performed 2-color analyses for some cell marker determinations and 3-color for other cell marker determinations on the same specimen. Three laboratories performed 2-color analyses all markers for some of the specimens and 3-color for all the markers on other specimens. Twenty-two laboratories performed only 3-color analyses for all the markers on all the specimens they tested.

Percentage of Participating Laboratories Using CDC Recommended Monoclonal Antibody Panel




Figure 1 of the report shows the methods used by the laboratories to prepare specimens for TLI. The majority of laboratories, 307 (97.2%) of 316, reported using a method of whole blood lysis to prepare specimens for TLI.

Figure 2 shows the methods used by the laboratories to fix their TLI specimens before flow cytometric analysis. Of laboratories reporting testing results, 23 (7.3%) of 316, reported that they did not fix their TLI specimens before analyzing them even though the panel sent to the laboratories contained known HIV antibody-positive specimens.

The types of flow cytometers used by the laboratories for TLI are shown in Figure 3. Those reported as used most often were: FACScan, 125 (39.6%); EPICS XL, 83 (26.3%); EPICS Profile II, 48 (15.2%); Ortho CytoronAbsolute, 23 (7.3%); EPICS Profile I, 13 (4.1%) and EPICS Elite, 13 (4.1%). Other types of flow cytometers were used, each with a frequency of less than 2%.

Results reported by those laboratories which used the 2-color monoclonal antibody panel recommended in the CDC CD4⁺ testing guidelines were used to calculate 90% reference ranges for each donor and cell marker involved in two-color analysis using the Statistical Analysis

System (SAS) procedure PROC UNIVARIATE. Before calculation, data were analyzed for possible outliers. There were 73 (1.23%) of 5926 results returned by laboratories using the monoclonal antibody panel recommended in the CDC guidelines that were considered to be outliers. Of the 73 outliers detected, 20 (27.4%) were results reported by two laboratories, and all data reported by these two laboratories were removed for calculation of reference ranges. No data from any laboratory, however, were removed from the table comparing values against the 90% reference ranges.

Results reported by all laboratories were grouped by donor number and cell marker, and compared to the 90% reference ranges as determined for each donor and marker involved in two-color analysis. Cell markers for each donor for which there were insufficient data for determination of 90% reference ranges are not included in the tables.

The overall frequency of 2-color laboratory results, by cell marker within, above, or below the 90% reference ranges are shown in Table 1. Single-color and 3-color results and results using reagents not included in the CDC CD4⁺ cell testing guidelines (e.g., CD2 and CD57) are not presented. The percentage of participating laboratory results within the 90% reference ranges for the cell markers specified in Table 1 are: CD3 average, 92.07%; CD4, 92.79%; CD8, 91.37%; CD14, 96.97%; CD19, 95.52%; CD45, 94.00%; and CD56/16, 94.98%.

Due to insufficient data, 90% reference ranges could not be calculated for CD16 alone and CD56 alone. Table 2 shows the entire range of laboratory results (maximum and minimum) reported for these two cell markers.

The percentage of laboratory results, by monoclonal antibody manufacturer, that were within the 90% reference ranges are shown in Table 3. Other manufacturers included Caltag, Dako, GenTrak, Immunotech, In-house, Ortho, Pharmingen, or Sigma.

The percentage of laboratory results, by flow cytometry instrument manufacturer, that were within the 90% reference ranges are shown in Table 4.

Laboratories indicated they sometimes use one manufacturer's flow cytometer and another manufacturer's monoclonal reagents; e.g., a laboratory using a flow cytometer manufactured by Coulter may be using monoclonal antibody reagents manufactured by Becton Dickinson, or vice versa. Laboratories also indicated that they do not use monoclonal antibodies from one manufacturer exclusively in the battery of tests used to analyze these specimens; e.g., the manufacturer of the CD3/CD4 reagent may be different from the manufacturer of the CD3/CD56+CD16 reagent. Laboratories also indicated that they may use antibodies from different manufacturers within a single tube, e.g., for a CD3/CD4 tube the laboratory might use the CD3 reagent from Coulter and the CD4 reagent from Becton-Dickinson. Analysis of reported cell marker percentages by flow cytometer and by monoclonal antibody manufacturer (Tables 1, 2, 3, and 4) indicated the cell marker percent positive values differed depending on the flow cytometer or monoclonal antibody manufacturer used by the laboratories. These associations were made by comparing the results with each monoclonal antibody manufacturer

to the respective 90% reference ranges as shown in Table 3, and by comparing the results with instrument manufacturer to the respective 90% reference ranges as shown in Table 4. It is unclear whether these differences are related to either the flow cytometer, the monoclonal antibody manufacturer, the combination of the flow cytometer and monoclonal antibody manufacturers, or whether other factors may be involved, e.g., the method used to prepare the specimens for analysis.

The percentages of laboratory results within, above, or below the 90% reference ranges for the fluorescencated bead instrument performance control sample are shown in Table 5. The results from the same laboratories that were used to generate the 90% reference ranges for Table 1 were used to generate the 90% reference ranges for Table 5. Most laboratory results, 1091 (91.4%) of 1194, were within the 90% reference ranges.

The overall frequency of laboratory results using 3-color analyses, by cell marker within, above, or below the 90% reference ranges are shown in Table 6. The ranges used for comparison are the same 2-color derived ranges shown in Table 1. Results for 3-color analyses, for nearly comparable cell markers, were compared against the same 2-color ranges, e.g., CD45+/CD3+/CD4+ and CD3+/CD4+/CD8- were compared against the 2-color range for CD3+/CD4+, CD45+/CD3+CD8+ and CD3+/CD4-/CD8+ were compared against the 2-color range for CD3+CD8+, CD45+/CD3-/CD19+ was compared against the 2-color range for CD3-/CD19+, and CD45+/CD3-/CD56&16+ was compared against the 2-color range for CD3-/CD56&16+. The percentage of participating laboratory results within the 90% reference ranges for the cell markers specified in Table 6 are: CD45+/CD3+/CD4+, 83.95%; CD45+/CD3+/CD8+, 85.07%; CD3+/CD4+/CD8-, 83.55%; CD3+/CD8+/CD4-, 84.87%; CD45+/CD3-/CD19+, 83.61%; and CD45+/CD3-/CD56/16+, 80.30%.

In summary, most laboratories performed well on the donor specimens in the September 1995 shipment. Not all laboratories used the 2-color monoclonal antibody combinations recommended in the CDC MMWR CD4⁺ cell testing guidelines. Of the 2-color results listed in Table 1, 6358 (93.9%) of 6772 were within the 90% reference range. Of the 3-color results listed in Table 6, 485 (83.8%) of 579 were within the 2-color derived 90% reference range. Differences in laboratory performance of cell marker analysis may be related to the use of the CDC CD4⁺ cell testing guidelines, the use of different flow cytometer and reagent manufacturer combinations, or to other factors associated with specimen preparation.

Figure 1. Methods used to prepare specimens for T-lymphocyte immunophenotyping, reported by participant laboratories to CDC for the September 1995 shipment.




Figure 2. Methods used to fix specimens for T-lymphocyte immunophenotyping, reported by participant laboratories to CDC for the September 1995 shipment.




Figure 3. Types of flow cytometers used for T-lymphocyte immunophenotyping, reported by participant laboratories to CDC for the September 1995 shipment.

