

**Analysis of the November 2000 Performance Evaluation Testing Results for
Human T-lymphotropic Virus Types I and II Antibody
Reported to the Centers for Disease Control and Prevention (CDC)
by Participant Laboratories in the Model Performance Evaluation Program**

This report is an analysis of results provided to the Centers for Disease Control and Prevention (CDC) by participant laboratories in the Model Performance Evaluation Program (MPEP) after they tested the human T-lymphotropic virus types I and II (HTLV-I/II) performance evaluation samples shipped to them in November 2000. Testing results for this analysis were provided by 189 (85.9%) of 220 laboratories sent sample panels. The testing results reported by the participant laboratories reflect their testing performance using manufactured kits to test performance evaluation samples and do not necessarily reflect an evaluation of these manufactured kits.

Samples in this shipment consisted of plasma from donors who were HTLV-I/II antibody-negative (donor numbers 1 and 4) and donors antibody-positive for either HTLV-I (donor number 3) or HTLV-II (donor number 2). All laboratories participating in this survey received identical samples. Before shipment, each donor sample was tested with two HTLV lysate-based enzyme immunoassay (EIA) kits licensed by the Food and Drug Administration (FDA) and with two HTLV Western blot (WB) kits. Additionally, each HTLV-I/II antibody-positive donor sample was tested by radioimmunoprecipitation assay (RIPA) and with an indirect immunofluorescent (IIF) antibody assay that can differentiate antibodies specific for HTLV-I or HTLV-II. Donor sample reactivity was determined by the CDC based on composite EIA, WB, and RIPA testing. The CDC MPEP interpretation of WB reactivity for each donor sample was consistent with the kit manufacturers' criteria for interpretation of WB results.

The cumulative frequency of test result interpretations reported by MPEP participant laboratories, arranged according to sample reactivity, for EIA, WB, and IIF methods are shown in Figure 1. There were three false-positive and three false-negative interpretations among the 1,113 EIA interpretations reported. Three indeterminate interpretations were among the 124 WB results reported. There were no false-positive or false-negative interpretations among the 18 Indirect Immunofluorescence test results reported. There were two false-positive and two false-negative interpretations among the 51 test results reported using HTLV-I/II testing methods classified as Other.

The types of laboratories that reported HTLV antibody testing results to CDC are shown in Figure 2. Each laboratory type is noted, by decreasing frequency, for each of the test methods. The "Other" category includes, for example, research laboratories, organ procurement laboratories, drug screening/toxicology laboratories, and sexually transmitted diseases clinics.

The combinations of EIA, WB and IIF test methods used by laboratories and frequency of use are shown in Figure 3. Most laboratories performed only EIA (82.0%), while some laboratories performed both EIA and supplemental tests (13.2%). Four laboratories (2.1%) performed WB alone or WB in combination with testing methods classified as "Other."

The types of test kits used, by manufacturer, for the EIA, WB, and IIF methods are shown, by decreasing frequency, in Figure 4. The Abbott HTLV-I/ HTLV-II EIA kit was used by 72.3% of laboratories reporting EIA results, the Organon Teknika HTLV-I/II kit was used by 18.5% of laboratories, and a variety of other test kits were used. The Genelabs Diagnostic WB kit was used by 70.4% of laboratories reporting WB results, and the remainder of the participant laboratories (29.6%) used the BioMerieux/Cambridge Biotech kit.

The results reported for the EIA, WB, and IIF methods, listed by kit manufacturer, for the CDC HTLV-I/II survey samples are shown in Figures 5, 6, and 7.

EIA Results

Laboratories using the Abbott HTLV-I/II EIA kit reported three false-positive and three false-negative EIA interpretations.

WB Results

Two indeterminate WB interpretations were reported for one of the HTLV-I/II antibody-negative samples (Donor number 1) by laboratories using WB test kits manufactured by Genelabs (one interpretation) and Cambridge Biotech (one interpretation). The CDC did not detect any bands in this HTLV-I/II antibody-negative sample using test kits manufactured by Genelabs and Cambridge Biotech. One indeterminate interpretation was reported for the HTLV-II antibody-positive sample (Donor number 2) by a laboratory using a WB test kit manufactured by Genelabs.

Of the 27 participant laboratories reporting WB results, 26 (96.3%) provided information regarding the criteria used for WB interpretations. Fourteen of these (53.8%) used interpretive criteria contained in the insert of the manufactured WB kit they used for testing. Other laboratories used the interpretive criteria published by the World Health Organization, four (15.4%); the Association of Public Health Laboratories (APHL), four (15.4%); or “Other” criteria, three (11.5%). One laboratory indicated using the Public Health Service (PHS) Working Group criteria. The WB interpretive criteria of these organizations and the WB test kit manufacturers are described in the table on the following page.

CRITERIA FOR INTERPRETATION OF HTLV WESTERN BLOT TESTS

Source of Interpretative Criteria	Criteria for Positive Test
Public Health Service (PHS) Working Group	p24 and gp46 or gp61/68
Association of Public Health Laboratories (APHL)	p19 or p24 and one env* band
Consortium for Retrovirus Serology Standardization (CRSS)	p19 or p24 and gp46 or gp61/68
World Health Organization (WHO)	One gag** and one env band
Cambridge Biotech	p24 and gp46 or rp21e
Genelabs Diagnostics	<p>HTLV-I p19 or p24 and gp46 or rgp46-I and rgp21</p> <p>HTLV-II p24 and rgp46-II and rgp21</p>

* env bands = gp21, gp46, gp61/68

** gag bands = p15, p19, and p24

Of the four laboratories using the WHO WB interpretative criteria, three used the Genelabs WB test kit, and one used the Cambridge Biotech WB test kit. Of the four laboratories using the APHL guidelines, one used the Cambridge Biotech WB test kit, and three used the Genelabs WB test kit. All three laboratories using the WB interpretative criteria described as “Other” used the Genelabs WB test kit. The one laboratory using the PHS criteria used a Cambridge Biotech WB test kit. These twelve laboratories are not using the WB interpretative criteria contained in the insert of the manufacturer’s kit they used to test the performance evaluation samples.

IIF Results

All results were correctly reported for the HTLV-I/II antibody-negative and antibody-positive samples.

Western Blot Band Patterns

The percentage and frequency of WB protein bands reported are shown in Figure 8. The frequency of a reported band is shown above or within the column, and the number of reports is listed in the far right column. For the HTLV-I/II antibody-positive donor samples (donor numbers 2-3), the participating laboratories detected antibodies to most of the native viral-specific proteins (e.g., p19, p24, p32/33, and gp46). The presence of recombinant gp46-type I (r46I) was correctly reported for the HTLV-I antibody-positive sample, and recombinant gp46-type II (r46II) was correctly reported for the

HTLV-II antibody-positive sample by the laboratories using commercially available WB strips designed to detect these bands. One laboratory using a test kit manufactured by Cambridge Biotech reported a p19 band for the HTLV-I/II antibody-negative donor number 1. One laboratory using a test kit manufactured by Genelabs reported p19 and p21/22 bands for donor number 1.

IIF Fluorescence Intensity

The fluorescence intensity patterns of HTLV-infected cells, as reported by participant laboratories, are shown in Figure 9. The number of reports received for each donor sample is listed in the far right column. Two laboratories reported IIF results. Generally, laboratories reported 1+ or greater fluorescence intensity in HTLV-infected cells for the HTLV-II antibody-positive samples (Donor 2) and 2+ or greater for the HTLV-I antibody positive samples (Donor 3). No IIF reactivity was reported for any of the HTLV-I/II antibody-negative samples (Donors 1 and 4) tested.

Results derived from testing methods classified as Other

The results reported by laboratories using testing methods classified as Other are shown in Figure 10. Two laboratories using a particle agglutination test kit, Serodia-HTLV-I manufactured by Fujirebio, Inc., correctly identified all HTLV-I/II antibody-positive and HTLV-I/II antibody-negative samples. A laboratory using an in-house HTLV-I specific RIPA method reported positive reactivity for the HTLV-I antibody-positive samples (Donor 3) and the HTLV-II antibody-positive sample (Donor 2). This laboratory used an in-house HTLV-II specific RIPA and correctly reported positive results for the HTLV-II antibody-positive sample (Donor 2). One laboratory using a chemiluminescence assay, manufactured by Abbott, correctly identified the five HTLV-I/II antibody-negative and antibody-positive samples it tested. A single laboratory using the line immunoassay (Inno-LIA HTLV-I/II, manufactured by Innogenetics) reported two false-positive and two false-negative interpretations.

Quality Control Testing

Although information was requested on the use of quality control (QC) materials not included with the manufacturer's kits, some laboratories continue to describe the kit controls as their only QC material. Positive and negative samples included in manufactured kits are internal kit control materials used to verify each lot's performance, calculate EIA test run cutoff values, and provide visual guidelines for determining band intensity for reading WB test results. They are often of limited value in assessing test performance over multiple lots of reagents. To verify the performance specifications of a test method and confirm that the accuracy and precision of a procedure are adequate, laboratories would benefit from testing external positive and negative QC samples, that is, samples which closely mimic patient specimens and are independent of the manufacturer's kit controls. An analysis of external control values over time allows a more accurate detection of shifts and trends in an analytic testing process

resulting from testing problems such as faulty pipettors, inadequate incubation conditions, or kit lot sensitivity.

Of the 184 laboratories reporting EIA test results, 180 laboratories responded to the question whether they used external EIA QC samples. Of these 180 laboratories, 135 (75.0%) indicated they used external EIA QC samples. Of the 135 affirmative external EIA QC responses, 109 (80.7%) indicated they obtained QC samples for EIA testing from commercial sources. Fifty-seven (42.2%) of the 135 responses indicated the use of a single serum/plasma and 79 (58.5%) indicated the use of multiple sera/plasma. Seventy-seven (57.0%) of the 135 responses indicated the use of a weakly positive external control. Eighty (59.3%) of the 135 responses indicated external EIA QC was used with each set/run of EIA plates and 47 (34.8%) of the 135 responses indicated external EIA QC was used with each plate.

Of the 27 laboratories responding to the question regarding the use of external QC samples in WB testing, 12 (44.4%) reported the use of external QC samples. Seven laboratories (58.3%) using external WB QC samples indicated they only used HTLV WB QC samples obtained in house. Four (33.3%) laboratories indicated they only used a commercial source for WB QC samples. One laboratory indicated it used WB QC samples obtained both commercially and in house. Seven (58.3%) of the 12 laboratories reported using external QC material with each set or run of WB strips.

One (50.0%) of two laboratories reporting IIF results reported using external QC samples.

Conclusion

Most of the laboratories participating in this survey correctly identified the HTLV-I/II antibody-positive and antibody-negative samples. The analytic sensitivity is 99.5%, the analytic specificity is 99.5% and the analytic performance is 99.5% for the EIA test. If indeterminate interpretations are considered correct for antibody-positive samples, the analytic sensitivity is 100%, the analytic specificity is 95.6%, and the analytic performance is 98.4% for the WB test. Since there were no incorrect interpretations for the IIF test, the analytic sensitivity, analytic specificity and analytic performance for the IIF test are 100%.