



HIGH THROUGHPUT TESTING FOR WEST NILE AND OTHER ARBOVIRUSES: A Case Study

The Arbovirus Laboratory
Wadsworth Center
New York State Dept of Health


Flow chart for virologic testing of vertebrate and mosquito specimens



RT-PCR assays on avian tissue, 2001



RT-PCR assays on mosquito pools, 2001



TRITURATION: Qiagen Mixer Mill MM 300



Disrupts
2 x 96 samples (1.2 ml)
or
2 x 24 samples (2.0 ml)
in 2-4 min

High Throughput Testing

- Automated Nucleic Acid Workstation
 - Automates sample and reaction preparation for nucleic acid analysis
 - increase in productivity
 - cost efficient
 - high quality of product
 - decreased cross-contamination
 - consistency and reproducibility

ABI Prism 6700

- Class II biosafety cabinet enclosure (**HEPA filtered**)
- Sets up **dilutions and replicate samples** in up to four **96-well output trays** for **TaqMan analysis**
- **Automatically seals output trays with a full cover optical blanket** and holds them at **4C**
- Completely **compatible** with **TaqMan sequence detection system** without additional manipulation
 - Software synergy
- One-step process from RNA purification to assay plate
- Disadvantage: Chemistry works on cells only







File Edit View Setup Instrument Help

Protocol | Deckspace

Instrument |

Instrument Status

Turn Off Peltiers

- Archive and Lysis Plates: 4 °C
 - ◆ Assay Plates: 11 °C
 - Master Mix, Standards, Controls, Dilution Plates: 4 °C
- Current Task: Peltier Cooling On
Status: Idle
Time Remaining: 0 Hr 0 Min 0 Sec

▶ Start

■ Pause

Run Log

Protocol Detail

Sample List

Consumable Details

2002-03-04, 01:07:43 PM - Assay Station temp: 11°C. Dilution Station temp: 3 °C. Archive Station temp: 4°C.
2002-03-04, 01:12:50 PM - Assay Station temp: 11°C. Dilution Station temp: 4 °C. Archive Station temp: 4°C.
2002-03-04, 01:25:22 PM - Begin sealing Output Plates
2002-03-04, 01:25:22 PM - Waiting for Sealer to come to correct temperature.
2002-03-04, 01:25:30 PM - Plate sealer is heated.
2002-03-04, 01:27:09 PM - Waiting for Sealer to come to correct temperature.
2002-03-04, 01:27:26 PM - Plate sealer is heated.
2002-03-04, 01:29:06 PM - Completed Sealing Output Plates.
2002-03-04, 01:29:06 PM - Completed placing Samples, Standards and Controls.
2002-03-04, 01:29:06 PM - Assay Protocol Complete
2002-03-04, 01:29:13 PM - Starting Peltiers cooling.
2002-03-04, 01:45:22 PM - ConductTransaction() failed in ThermalUpdateThread.run(): java.lang.Exception: Time out error. Robot not responding in ReceiveOnly()
2002-03-04, 01:45:28 PM - Peltiers failed to turn off while terminating Peltier thread: Time out error. Robot not responding in ReceiveOnly()

Print Tab

DESCRIPTION		ASSAY	PERSON HOURS
		COST per sample*	(Hands on labor)**
Tissue sorting, excision, homogenization	Birds/Mammals	\$0.75	2 hr per 96-well plate
	Mosquitoes	\$0.50	3 hr per 96-well plate
Isolation of RNA	RNeasy Method		
	Birds/Mammals	\$3.25	5 h per 96-well plate
ABI 6700 Robot	Mosquitoes	\$3.25	5 h per 96-well plate
	Birds/Mammals	\$2.00	1 h per 96-well plate
	Mosquitoes	\$2.75	1 h per 96-well plate
Real-time RT- PCR (TaqMan)	Manual setup	\$3.25	1 h per 96-well plate
	ABI 6700 Robot	\$3.85	30 m per 96-well plate

* Specific supplies only (no equipment, personnel, general supplies)

**Robot run time: 85 m RNA extraction; 45 m TaqMan set-up

Comparison of TaqMan Ct values on RNA samples from naturally infected bird kidneys

Crow kidney	ABI Prism 6700		RNeasy		Ratio (6700:RNeasy)
	Mean Ct	SD	Mean Ct	SD	
A	14.97	0.8	16.56	0.7	0.90
B	17.85	0.6	18.45	0.6	0.97
C	14.59	0.6	16.93	0.5	0.86
D	14.71	0.7	16.78	0.6	0.88
E	17.57	0.4	21.95	1.3	0.80
F	30.13	1.5	29.54	5.5	1.02
G	40.00	0.0	40.00	0.0	1.00

Reproducibility of robotic assays

(INTRA-assay variability)


Crow kidney	Ct Values					
	WNV NS1			WNV env		
	Mean	SD	CV	Mean	SD	CV
227	16.7	0.51	3.0%	15.8	0.39	2.5%
228	16.2	0.35	2.2%	15.4	0.22	1.5%
229	15.9	0.19	1.2%	15.3	0.11	0.7%
230	15.8	0.22	1.4%	15.2	0.13	0.8%
231	16.9	0.31	1.8%	16.0	0.32	2.0%
232	17.1	0.17	1.0%	16.1	0.03	0.2%
233	16.5	0.90	5.5%	15.7	0.62	3.9%
234	15.8	0.25	1.6%	15.8	0.25	1.6%
235	16.4	0.11	0.7%	16.4	0.11	0.7%
236	16.1	0.05	0.3%	16.1	0.05	0.3%
302	16.4	0.20	1.2%	16.4	0.20	1.2%
303	16.9	0.42	2.5%	16.9	0.42	2.5%
304	16.6	0.53	3.2%	16.6	0.53	3.2%

Reproducibility of robotic assays

(**INTER**-assay variability)

Bird ^a	Tissue Homogenate ^b	<i>mean Ct^c</i>	<i>SD^d</i>	<i>%CV^e</i>
2445	P126	16.1	0.94	5.9%
	P127	15.1	1.06	7.0%
	P128	16.6	0.77	4.6%
	P129	15.8	0.59	3.7%
	P130	15.4	1.10	7.1%
473	P152	18.1	0.39	2.2%
	P153	18.4	0.51	2.8%
	P154	17.8	0.38	2.1%
	P155	18.6	0.44	2.4%
	P156	18.3	1.51	8.2%
3200	P217	15.3	0.29	1.9%
	P218	15.7	0.30	1.9%
	P219	15.9	1.88	11.8%
	P220	15.6	1.27	8.1%

Comparison of linearity of assays on avian tissue



Comparative assays on infected mosquito parts*

Infected Mosquito Part	Ct Value	
	Robot	RNeasy
1 Leg	40.0	32.4
	31.8	31.8
2 Legs	29.9	30.6
	30.3	32.4
Abdomen	23.5	25.2
Head	26.7	24.7
	25.9	26.5
Thorax	22.5	23.0

* Added to pool of 50 uninfected mosquitoes

High Throughput Testing

Qiagen 9604 nucleic acid workstation



QIAGEN 9604 (Lanciotti data)

	Manual Ext.	Robot	Robot
PFU in sample	Nov 8 2000	nov 14 2000	nov 13 2000
1000	24.01	23.46	23.04
100	27.06	26.4	25.51
10	30.21	31.47	29.65
1	34.28	33.66	32.13
0.1	36.18	37.69	38.35
0.1	36.63	37.1	36.1
0.01	45	39.34	40.84
0.01	45	45	45

Summary of High Throughput Techniques

- Submission of sample data to laboratory on Excel spreadsheets
- High capacity mixer mill
- Robotic workstation for RNA extraction and real time RT-PCR setup
- Real time RT-PCR

Arbovirus Laboratory staff



2002