


Identifier: SOP-5171	Revision: 0	
Effective Date: 6/30/08	Next Review Date: 6/30/13	

Waste & Environmental Services

Standard Operating Procedure

for **ROUTINE VALIDATION OF TOTAL PETROLEUM HYDROCARBONS GASOLINE RANGE ORGANICS/DIESEL RANGE ORGANICS ANALYTICAL DATA (METHOD 8015B)**

APPROVAL SIGNATURES:

Subject Matter Expert:	Organization	Signature	Date
Nita P. Patel	WES-EDA	Signature on file	4/30/08
Quality Assurance Specialist:	Organization	Signature	Date
Laura Ortega	QA-IQ	Signature on file	5/7/08
Responsible Line Manager:	Organization	Signature	Date
Craig Eberhart	WES-EDA	Signature on file	5/6/08

CONTROLLED DOCUMENT

Users are responsible for ensuring they work to the latest approved revision.
Printed or electronically transmitted copies are uncontrolled.

Title: Routine Validation of Total Petroleum Hydrocarbons Gasoline Range Organics/Diesel Range Organics Analytical Data (Method 8015B)	No.: SOP-5171	Page 2 of 10
	Revision: 0	

1.0 PURPOSE AND SCOPE

This procedure represents the minimum standards for evaluating routine total petroleum hydrocarbons gasoline range and diesel range organics analytical data. This procedure is a mandatory document and shall be implemented by all Los Alamos National Laboratory (LANL or Laboratory) personnel and contractors who evaluate routine total petroleum hydrocarbon analytical data for the specific LANL projects.

2.0 BACKGROUND AND PRECAUTIONS

2.1 Background

This procedure conforms to the requirements of Environmental Protection Agency (EPA) Methodologies and the EPA document, "U.S. EPA Contract Laboratory Program National Functional Guidelines for Organic Data Review." LANL data validation is performed according to procedures based upon the NNSA Model Data Validation Procedure. Data qualifiers and reason codes are assigned according to the specifications in this method specific procedure.

2.2 Precautions

Nothing in this procedure precludes the data validator from going beyond the minimum requirements specified within this procedure. If additional directions are required, the data validator shall reference NNSA Model Data Validation Procedure, EPA method specific guidelines and/or National Functional Guidelines for Organic Data Review. Implementation of this procedure may be followed by a more focused and data use-specific evaluation of the data by the project chemist, especially if the implementation of this procedure indicates the data may contain technical deficiencies.

3.0 EQUIPMENT AND TOOLS

None.

4.0 STEP-BY-STEP PROCESS DESCRIPTION

4.1 Qualifications for Data Validators

Data Validator	1.	Possess a minimum of a bachelor's degree in chemistry, or one of the physical sciences
		AND
		either two (2) years of experience in generating analytical data in an environmental analytical laboratory
		AND
		two (2) years of data validation experience.
	2.	Complete Attachment 1, Data Validation Cover Sheet, and Attachment 2, TPH GRO and DRO Analytical Data Validation Checklist, during data validation.
	3.	Use Attachment 3, Guidance for the Qualifier and Reason Code Application, for additional guidance.

CONTROLLED DOCUMENT

Users are responsible for ensuring they work to the latest approved revision.
Printed or electronically transmitted copies are uncontrolled.

4.2 Records

- Data Validator 1. Submit the following records generated by this procedure to the Records Processing Facility:
- Completed Data Validation Cover Sheets; and
 - Completed TPH GRO and DRO Analytical Data Validation Checklists.

5.0 PROCESS FLOW CHART

For specific validation criteria follow the NNSA Model for Data Validation.

6.0 ATTACHMENTS

- Attachment 1 5171-1 Data Validation Cover Sheet (1 page)
Attachment 2 5171-2 TPH GRO and DRO Analytical Data Validation Checklist (3 pages)
Attachment 3 5171-3 Guidance for the Qualifier and Reason Code Application (3 pages)

7.0 REVISION HISTORY

Author: Nita P. Patel

Revision No. <i>[Enter current revision number, beginning with Rev.0]</i>	Effective Date <i>[DCC inserts effective date for revision]</i>	Description of Changes <i>[List specific changes made since the previous revision]</i>	Type of Change <i>[Technical (T) or Editorial (E)]</i>
0	6/30/08	New Document	T

[Using a CRYPTOCard, click here to record "self-study" training to this procedure.](#)

If you do not possess a CRYPTOCard or encounter problems, contact the EP training specialist.

CONTROLLED DOCUMENT

Users are responsible for ensuring they work to the latest approved revision.
Printed or electronically transmitted copies are uncontrolled.

ATTACHMENT 1: EXAMPLE OF A DATA VALIDATION COVER SHEET

5171-1

Example of a Data Validation Cover Sheet

Records Use only



Section I.

REQUEST NUMBER: _____ VALIDATION DATE: _____ LAB CODE: _____

CONTRACT LABORATORY NAME: _____

VALIDATOR: _____ ORGANIZATION: _____

ANALYTICAL SUITE (CHECK ALL THAT APPLY):

- | | | | |
|--|--|--|---|
| <input type="checkbox"/> TPH-GRO | <input type="checkbox"/> HIGH EXPLOSIVES | <input type="checkbox"/> DIOXIN FURANS | <input type="checkbox"/> LCMSMS PERCHLORATES |
| <input type="checkbox"/> TPH-DRO | <input type="checkbox"/> METALS | <input type="checkbox"/> PCB CONGENERS | <input type="checkbox"/> ORGANOCHLORINE
PESTICIDES/POLYCHLORINATED |
| <input type="checkbox"/> GENERAL CHEMISTRY | <input type="checkbox"/> RADIOCHEMISTRY | <input type="checkbox"/> LCMSMS HIGH
EXPLOSIVES | <input type="checkbox"/> BIPHENYLS |

OTHER (DESCRIBE): _____

Section II. Completeness Check

YES	NO	N/A	(CHECK ONE)	YES	NO	N/A	(CHECK ONE)
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	1. CHAIN-OF-CUSTODY FORM(S)	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	6. RAW/BSS DATA
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	2. CASE NARRATIVE	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	7. QUALITY CONTROL FORMS
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	3. SAMPLE RESULT FORMS	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	8. QUANTITATION REPORTS
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	4. SAMPLE CHROMATOGRAMS	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	9. TICS FORMS
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	5. STANDARD CHROMATOGRAMS	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	10. TICS MASS SPECTRA

Comments/problems noted (include information about requests for further information submitted to the contract laboratory and agreed-upon date of resolution and contract laboratory point of contact):

VALIDATOR'S SIGNATURE: _____ DATE: _____

SOP-5171, Revision 0.0

LOS ALAMOS
Environmental Restoration Project

CONTROLLED DOCUMENT

Users are responsible for ensuring they work to the latest approved revision.
Printed or electronically transmitted copies are uncontrolled.

ATTACHMENT 2: TPH GRO AND DRO ANALYTICAL DATA VALIDATION CHECKLIST

5171-2

TPH GRO and DRO Analytical Data Validation Checklist

Records Use only



Yes No N/A (Check One)				Assign Qualifier Listed Below If Criterion = Yes	
				Non-detected Analyte	Detected Analyte
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	1. The retention time criteria were not met.	R, DR0 or GR0	R, DR0 or GR0
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	2. Required retention time documentation is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.	R, DR0b or GR0b	R, DR0b or GR0b
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	3. The surrogate is less than 10%R. Follow external laboratory limits.	R, DR3 or GR3	J-, DR3 or GR3
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	4. The surrogate is less than the Lower Acceptance Limit, but greater than or equal to 10%R. Follow external laboratory limits.	UJ, DR3a or GR3a	J-, DR3a or GR3a
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	5. The surrogate %R value is greater than the Upper Acceptance Limit. Follow external laboratory limits.	N/A	J+, DR3b or GR3b
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	6. Required surrogate information is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.	R, DR3d or GR3d	R, DR3d or GR3d
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	7. The sample result is less than or equal to 5 times the concentration of the related analyte in the method blank.	N/A	U, DR4 or GR4
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	8. The affected analytes are considered estimated and biased high because this analyte was identified in the method blank, but was greater than 5x.	N/A	J+, DR4a or GR4a
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	9. The sample result is less than or equal to 5 times the concentration of the related analyte in the trip blank, rinsate blank or equipment blank.	U, DR4d or GR4d	N/A

CONTROLLED DOCUMENT

Users are responsible for ensuring they work to the latest approved revision.
Printed or electronically transmitted copies are uncontrolled.

Yes No N/A (Check One)				Assign Qualifier Listed Below If Criterion = Yes	
				Non-detected Analyte	Detected Analyte
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	10. Required method blank information is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.	R, DR4e or GR4e	R, DR4e or GR4e
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	11. The affected results were not analyzed with a valid 5-point calibration curve and/or a standard at the reporting limit.	UJ or R, DR7 or GR7	J, DR7 or GR7
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	12. The affected analytes were analyzed with an initial calibration curve that exceeded the %RSD criteria and/or the associated multipoint calibration correlation coefficient is less than 0.995.	UJ, DR7a or GR7a	J, DR7a or GR7a
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	13. The ICV and/or CCV were recovered outside the method specific limits.	UJ, DR7c or GR7c	J, DR7c or GR7c
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	14. The ICV and/or CCV were not analyzed at the appropriate method frequency.	UJ, DR7d or GR7d	J, DR7d or GR7d
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	15. Required calibration information is missing or samples were analyzed on an expired calibration. Contact the SMO or external laboratory for information.	R, DR7f or GR7f	R, DR7f or GR7f
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	16. The extraction/analytical holding time is greater than 1x and less than or equal to 2 times the applicable holding time requirement.	UJ, DR9 or GR9	J-, DR9 or GR9
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	17. The holding time was greater than 2 times the applicable holding time requirement.	R, DR9a or GR9a	J-, DR9a or GR9a
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	18. The LCS percent recovery was less than 10%. Follow the external laboratory limits.	R, DR12 or GR12	J-, DR12 or GR12
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	19. The LCS percent recovery was less than the Lower Acceptance Limit, but greater than or equal to 10%. Follow the external laboratory limits.	UJ, DR12a or GR12a	J-, DR12a or GR12a
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	20. The LCS percent recovery was greater than the Upper Acceptance Limit. Follow the external laboratory limits.	N/A	J+, DR12b or GR12b
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	21. The LCS documentation is missing. Data may not be acceptable for use. Contact the SMO or the external laboratory for information.	R, DR12c or GR12c	R, DR12c or GR12c

CONTROLLED DOCUMENT

Users are responsible for ensuring they work to the latest approved revision.
Printed or electronically transmitted copies are uncontrolled.

Yes No N/A (Check One)				Assign Qualifier Listed Below If Criterion = Yes	
				Non-detected Analyte	Detected Analyte
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	22. The MS/MSD percent recovery was less than 10%.	R, DR12d or GR12d	R, DR12d or GR12d
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	23. The MS/MSD percent recovery was greater than or equal to 10%, but less than 70%.	UJ, DR12e or GR12e	J, DR12e or GR12e
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	24. The MS/MSD percent recovery was greater than 130%.	N/A	J+, DR12f or GR12f
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	25. The MS/MSD relative percent difference was greater than 30%.	UJ, DR12g or GR12g	J, DR12g or GR12g
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	26. The affected analytes have elevated detection limits and may not meet project DQOs because the sample was diluted without any target analytes identified due to matrix interference. (Qualify as Reject if the analytical laboratory cannot provide proof for matrix interference.)	UJ, R, DR15 or GR15	R, DR15 or GR15
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	27. The LANL project chemist identified quality deficiencies in the reported data that requires further qualification. This code can only be used under advisement by the LANL project chemist.	UJ, R, DR19 or GR19	J, R, DR19 or GR19
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	28. Duplicate, dilution, or reanalysis.	UJ, DR88 or GR88	J, DR88 or GR88

CONTROLLED DOCUMENT

Users are responsible for ensuring they work to the latest approved revision.
Printed or electronically transmitted copies are uncontrolled.

ATTACHMENT 3: GUIDELINES FOR THE QUALIFIER AND REASON CODE APPLICATION

5171-3

Guidelines for the Qualifier and Reason Code Application

Records Use only



No.	Valid Flag Code Nondetect	Valid Flag Code Detect	Valid Reason Code	Valid Reason Description
1	R	R	DRO or GRO	The retention time criteria were not met.
2	R	R	DROb or GROb	Required retention time documentation is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.
3	R	J-	DR3 or GR3	The surrogate is less than 10%R, which indicates the potential for a severely low bias in the results. Follow the external laboratory limits.
4	UJ	J-	DR3a or GR3a	The surrogate is less than the Lower Acceptance Limit, but greater than or equal to 10%R, which indicates the potential for a low bias in the results. Follow the external laboratory limits.
5	N/A	J+	DR3b or GR3b	The surrogate %R value is greater than the Upper Acceptance Limit, which indicates a potential for a high bias in the results and a potential for false positive results. Follow the external laboratory limits.
6	R	R	DR3d or GR3d	Required surrogate information is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.
7	N/A	U	DR4 or GR4	The sample result is less than or equal to 5 times the concentration of the related analyte in the method blank, which indicates the reported detection is considered indistinguishable from contamination in the blank.

CONTROLLED DOCUMENT

Users are responsible for ensuring they work to the latest approved revision.
Printed or electronically transmitted copies are uncontrolled.

No.	Valid Flag Code Nondetect	Valid Flag Code Detect	Valid Reason Code	Valid Reason Description
8	N/A	J+	DR4a or GR4a	The affected analytes are considered estimated and biased high because this analyte was identified in the method blank but was greater than 5x.
9	U	N/A	DR4d or GR4d	The sample result is less than or equal to 5 times the concentration of the related analyte in the trip blank, rinsate blank, or equipment blank, which indicates the reported detection is considered indistinguishable from contamination in the blank.
10	R	R	DR4e or GR4e	Required method blank information is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.
11	UJ, R	J	DR7 or GR7	The affected results were not analyzed with a valid 5-point calibration curve and/or a standard at the reporting limit.
12	UJ	J	DR7a or GR7a	The affected analytes were analyzed with an initial calibration curve that exceeded the %RSD criteria and/or the associated multipoint calibration correlation coefficient is less than 0.995.
13	UJ	J	DR7c or GR7c	The ICV and/or CCV were recovered outside the method specific limits.
14	UJ	J	DR7d or GR7d	The ICV and/or CCV were not analyzed at the appropriate method frequency.
15	R	R	DR7f or GR7f	Required calibration information is missing or samples were analyzed on an expired calibration. Contact the SMO or external laboratory for information.
16	UJ	J-	DR9 or GR9	The extraction/analytical holding time is greater than 1x and less than or equal to 2 times the applicable holding time requirement.
17	R	J-	DR9a or GR9a	The extraction/analytical holding times were exceeded by more than 2x the published method for holding times.

No.	Valid Flag Code Nondetect	Valid Flag Code Detect	Valid Reason Code	Valid Reason Description
18	R	J-	DR12 or GR12	The LCS percent recovery was less than 10%. Follow the external laboratory limits.
19	UJ	J-	DR12a or GR12a	The LCS percent recovery was less than the Lower Acceptance Limit but greater than or equal to 10%. Follow the external laboratory limits.
20	N/A	J+	DR12b or GR12b	The LCS percent recovery was greater than the Upper Acceptance Limit. Follow the external laboratory limits.
21	R	R	DR12c or GR12c	The LCS documentation is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.
22	R	R	DR12d or GR12d	The MS/MSD percent recovery was less than 10%
23	UJ	J	DR12e or GR12e	The MS/MSD percent recovery was greater than or equal to 10% but less than 70%.
24	N/A	J+	DR12f or GR12f	The MS/MSD percent recovery was greater than 130%.
25	UJ	J	DR12g or GR12g	The MS/MSD relative percent difference was greater than 30%.
26	UJ, R	R	DR15 or GR15	The affected analytes have elevated detection limits and may not meet project DQOs because the sample was diluted without any target analytes identified due to matrix interference. (Qualify as Reject if the analytical laboratory cannot provide proof for matrix interference.)
27	UJ, R	J, R	DR19 or GR19	The project chemist identified quality deficiencies in the reported data that requires further qualification. This code can only be used under advisement by the project chemist.
28	UJ	J	DR88 or GR88	Duplicate, dilution, or reanalysis.

CONTROLLED DOCUMENT

Users are responsible for ensuring they work to the latest approved revision.
Printed or electronically transmitted copies are uncontrolled.