DEPARTMENT OF COLLERGE BUREAU OF STANDARDS WASHINGTON

Provisional Certificate

of Surface Area Analysis of

STANDARD PORTLAND CEMENT

no. 114

Surface area - - - - - - - - - 1900 sq cm per gram

Residue on no. 325 sieve, wet

method - - - - - - - 12.3 percent

(Signed) LYMAN J. BRIGGS, Director.

Washington, D.C. April 11, 1934.