



National Institute of Standards & Technology

Certificate

Standard Reference Material 2534

Ellipsometric Parameters Δ and ψ and Derived Thickness and Refractive Index for a Silicon Dioxide Layer on Silicon

Serial No.

This Standard Reference Material (SRM) consists of a 76-mm (3-in) diameter silicon wafer on which a uniform silicon dioxide layer was grown. It is certified for the ellipsometric parameters delta, Δ , and psi, ψ , at the vacuum wavelength $\lambda = 633.0$ nm using the High-Accuracy Ellipsometer built at NIST [1]. The SRM is issued primarily to evaluate the accuracy of ellipsometers and may also be used as an aid in the calibration of various other optical thickness-monitoring instruments. Each SRM unit is individually measured; thus the certified values provided on page 3 apply only to the unit identified by the above serial number.

This SRM is also certified for derived values of the thicknesses and indices of refraction of both layers of a two-layer optical model of an oxide film on a single-crystal silicon substrate. The model consists of a homogeneous, isotropic, top layer separated from a homogeneous, isotropic substrate by a homogeneous isotropic interlayer. The substrate is characterized by a complex index of refraction, n_s ; the interlayer is characterized by a thickness, t_i , and a real index of refraction, n_i ; and the top layer is characterized by a thickness, t_t , and a real index of refraction, n_t . The bottom layer is interpreted as single-crystal silicon and the top layer as amorphous silicon dioxide. The physical interpretation of the interlayer is less straightforward, being a research topic, but the fit to the ellipsometric data is greatly improved by inclusion of the interlayer. Some applications may not be sensitive to the interlayer in the optical model in the same way as is ellipsometry. For this reason, the uncertainties stated on this certificate apply only when the SRMs are used in ellipsometric measurements.

The certified values along with noncertified supplemental information are given on page 3 of this certificate. The supplemental information includes values of Δ and ψ at 50°, 55°, 60°, 65°, 70°, and 75° angles of incidence that were calculated using the two-layer model. The thickness for a one-layer model of the oxide is also given.

It is not possible to determine n_s , t_i , n_i , t_t , and n_t from the measurements of Δ and ψ on a single wafer. Therefore, the derivation of the values reported in this certificate was carried out on a batch of wafers as described in Ref. [2] with the following exceptions: wafers with nominal 14-nm or 25-nm oxides may have been included in the data reduction. A more accurate value, 0.0156, was used for the imaginary part of n_s [3].

The overall direction and coordination of the technical aspects of this SRM were performed by P. Roitman, B.J. Belzer, D. Chandler-Horowitz, and J.F. Marchiando of the NIST Semiconductor Electronics Division.

The SRM was fabricated in the Semiconductor Processing Research Laboratory by M.L. Miller and D.B. Novotny of the NIST Semiconductor Electronics Division.

The ellipsometric measurements leading to certification were performed by B.J. Belzer of the NIST Semiconductor Electronics Division.

Guidance on the statistical analysis was performed by M.C. Croarkin of the NIST Statistical Engineering Division.

The support aspects involved in the certification and issuance of this Standard Reference Material were coordinated through the Standard Reference Materials Program by N.M. Trahey.

Gaithersburg, MD 20899
July 15, 1992

William P. Reed, Chief
Standard Reference Materials Program

(over)

Cleaning and Handling:

The SRM should be stored in the clean wafer container supplied, and should be handled only with proper wafer tweezers. Even when the SRM is kept in its clean container, a film forms on the surface of the oxide. This unwanted film forms rapidly, and can be removed by rinsing the SRM with reagent grade ethanol or isopropanol and then with deionized water. Both solvents, the alcohol and the water, can be blown off the wafer surface with ultra-clean nitrogen. The sample must then be permitted to stabilize in the atmosphere 20 to 30 minutes prior to continuing with the measurements. When the surface is prepared in this manner, the measurements are reproducible for several hours or more. If the SRM becomes excessively dirty from mishandling, improper storage, or any other reason, the wafer should then be cleaned using a semiconductor processing grade detergent and rinsed thoroughly with deionized water. No other cleaning method is recommended.

REFERENCES

1. Candela, G.A., and Chandler-Horowitz, D., An Ellipsometry System for High Accuracy Metrology of Thin Films; SPIE Vol. 480 Integrated Circuit Metrology II, p 2 (1984).
2. Candela, G.A., Chandler-Horowitz, D., Marchiando, J.F., Novotny, D.B., Belzer, B.J., and Croarkin, M.C., Preparation and Certification of SRM-2530, Ellipsometric Parameters Δ and ψ and Derived Thickness and Refractive Index of a Silicon Dioxide Layer on Silicon; NIST Special Publication 260-109 (1988).
3. Geist, J., Schaefer, A.R., Song, J-F., Wang Y.H., and Zalewski, E.F., An Accurate Value for the Absorption Coefficient of Silicon at 633 nm; J. Res. NIST 95, 549-558 (1990).

In addition, work comparing measurements made with the NIST High-Accuracy Ellipsometer and a NIST profilometer was carried out and reported in Ref. [1]. Examination of the data presented in this reference shows agreement between the two methods to within 3 nm for specimen SRMs having oxides nominally 50-nm thick (the "on-chrome" profilometry data agree with the ellipsometrically derived data to within 1 nm). The NIST Precision Engineering Division has provided unpublished data that demonstrates the estimated uncertainty for these profilometer measurements to be ± 3.8 nm, expressed in the same manner as the estimated uncertainty stated in the previous paragraph.

REFERENCE

1. Candela, G.A., Chandler-Horowitz, D., Novotny, D.B., Vorbürger, T.V., and Giauque, C.H.W., Film Thickness and Refractive Index Standard Reference Material Calibrated by Ellipsometry and Profilometry, Proc. SPIE **661**, 402-407 (1986).