U. S. Department of Commerce Malcolar Baldrige Secretary National Bureau of Standards Ernest Ambler, Director

# National Bureau of Standards Certificate

## Standard Reference Material 4202D

## Radioactivity Standard

Radionuclide Cadmium-109

Source identification SRM 4202-D-

Point source on polyester tape(1)\* Source description

 $_{\gamma} s^{-1} (2)$ 88-keV gamma-ray-emission rate

> Reference time 1200 EST October 1, 1986

1.4 percent(3) Overall uncertainty

 $22_{\text{Na}}$ ,  $60_{\text{Co}}$ ,  $65_{\text{Zn}}(5)$ Photon-emitting impurities

> Half life  $462.6 \pm 0.4 \, days^{(6)}$

Measuring instrument " $4\pi$ " NaI(T1) crystal<sup>(7)</sup>

This Standard Reference Material was prepared in the NBS Center for Radiation Research, Ionizing Radiation Division, Radioactivity Group, D.D. Hoppes, Group Leader, with the assistance of Guest Scientist, C. Ballaux, of the Nuclear Chemistry Department, Studie Centrum voor Kernenergie, Mol, Belgium.

Gaitherburg, MD. 20899 October, 1986

Stanley D. Rasberry, Chief Office of Standard Reference Materials

\*Notes on back

### **NOTES**

- (1) The source consists of a dried deposit of cadmium as the chloride on a disk of 0.55 mg/cm<sup>2</sup> MYLAR supported by 8.1 mg/cm<sup>2</sup> polyester tape and covered by a second layer of the thicker tape only. The tape is retained on an aluminum annulus of 5.4-cm outer diameter and 3.8-cm inner diameter.
- (2) The 88-keV gamma-ray emission rate (N $\gamma$ ) is the most useful parameter for this SRM, but measurements were also made of the source activity. The NBS activity for this source, N<sub>0</sub>, in units of Bq, may be obtained by multiplying the emission rate by 27.115, which was the ratio of the activity to the emission rate results. The overall uncertainty in the activity is 1.5 percent<sup>(4)</sup>.
- (3) The overall uncertainty in the gamma-ray-emission rate,  $N\gamma$ , was formed by taking three times the quadratic combination of the standard deviations of the mean, or approximations thereof, for the following:

a)	counting statistics	0.16	percent
<b>b</b> )	gravimetric	0.10	percent
c)	background	0.01	percent
d)	adsorption	0.01	percent
e)	impurities	0.01	percent
f)	live-time correction	0.05	percent
g)	extrapolation to 39.0 keV	0.07	percent
h)	corrections for peak overlap	0.41	percent
i)	detector dimensions	0.04	percent
j)	source positioning	0.04	percent
k)	energy calibration	0.07	percent
	combined uncertainty	0.47	percent

(4) The overall uncertainty in the activity,  $N_0$ , was formed by taking three times the quadratic combination of the standard deviations of the mean, or approximations thereof, for the following:

a)	9 liquid-scintillation measurements	0.15	percent
ъ)	source preparation	0.07	percent
c)	dead time	0.05	percent
d)	corrections for peak overlap	0.4	percent
e )	scintillator stability	0.05	percent
f)	gamma-ray efficiency correction	0.23	percent
g)	gamma-ray probability per decay	0.04	percent
	combined uncertainty	0.50	percent

#### NOTES, CONTINUED

- (5) The material from which this standard was prepared was examined with germanium detectors, and traces of  $^{22}\mathrm{Na}$ ,  $^{60}\mathrm{Co}$ , and  $^{65}\mathrm{Zn}$  were observed at levels of less than 0.1 parts per million.
- (6) NCRP Report 58, 2nd Edition, p.418 (1985).
- (7) "109Pd and 109Cd Activity Standardization and Decay Data", C. Ballaux, B.M. Coursey and D.D. Hoppes, submitted for publication (1986) in International Journal of Radiation Applications and Instrumentation, Part A.

For further information please call B.M. Coursey or J.M. Calhoun at (301) 975-5539.

4202D-